

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: YANN01
 Lab Code: ACE SDG NO.: Q3238
 Lab File ID: VW032288.D Date Analyzed: 09/29/2025
 Instrument ID: MSVOA_W Time Analyzed: 09:18
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	268675	7.96	467141	8.85	421624	11.63
UPPER LIMIT	537350	8.459	934282	9.349	843248	12.129
LOWER LIMIT	134338	7.459	233571	8.349	210812	11.129
EPA SAMPLE NO.						
RCA	166736	7.96	358094	8.85	380188	11.63
VW0929SBL01	170986	7.95	366797	8.85	375398	11.63
VW0929SBS01	225036	7.96	419951	8.85	396829	11.63
VW0929SBSD01	235838	7.97	436378	8.86	387451	11.64

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.