

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: YANN01
 Lab Code: ACE SDG NO.: Q3238
 Lab File ID: VW032298.D Date Analyzed: 09/30/2025
 Instrument ID: MSVOA_W Time Analyzed: 12:54
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	244399	7.96	443869	8.85	407884	11.63
UPPER LIMIT	488798	8.459	887738	9.349	815768	12.129
LOWER LIMIT	122200	7.459	221935	8.349	203942	11.129
EPA SAMPLE NO.						
RCARE	175613	7.96	403251	8.85	393328	11.63
VW0930SBL01	181583	7.96	391867	8.85	387829	11.63
VW0930SBS01	217107	7.97	424498	8.85	400367	11.63

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.