

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>YANN01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3238</u>
Lab File ID:	<u>VW032288.D</u>	Date Analyzed:	<u>09/29/2025</u>
Instrument ID:	<u>MSVOA_W</u>	Time Analyzed:	<u>09:18</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>Y</u>

	IS4 AREA #	RT #				
12 HOUR STD	216076	13.556				
UPPER LIMIT	432152	14.056				
LOWER LIMIT	108038	13.056				
EPA SAMPLE NO.						
RCA	175337	13.56				
VW0929SBL01	185238	13.56				
VW0929SBS01	199844	13.56				
VW0929SBSD01	191625	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.