

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3246</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>VX047958.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1002WBSD01	Dichlorodifluoromethane	20	20.0	ug/L	100	3		32	152	20
	Chloromethane	20	16.5	ug/L	83	5		50	139	20
	Vinyl chloride	20	19.0	ug/L	95	8		58	137	20
	Bromomethane	20	19.4	ug/L	97	2		53	141	20
	Chloroethane	20	18.9	ug/L	95	8		60	138	20
	Trichlorofluoromethane	20	19.8	ug/L	99	4		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	20.2	ug/L	101	2		70	136	20
	1,1-Dichloroethene	20	18.4	ug/L	92	6		71	131	20
	Acetone	100	73.6	ug/L	74	4		39	160	20
	Carbon disulfide	20	17.2	ug/L	86	6		64	133	20
	Methyl tert-butyl Ether	20	17.2	ug/L	86	8		71	124	20
	Methyl Acetate	20	17.1	ug/L	86	10		56	136	20
	Methylene Chloride	20	18.0	ug/L	90	7		74	124	20
	trans-1,2-Dichloroethene	20	18.3	ug/L	92	7		75	124	20
	1,1-Dichloroethane	20	18.8	ug/L	94	8		77	125	20
	Cyclohexane	20	17.8	ug/L	89	1		71	130	20
	2-Butanone	100	78.6	ug/L	79	14		56	143	20
	Carbon Tetrachloride	20	20.3	ug/L	102	8		72	136	20
	cis-1,2-Dichloroethene	20	18.8	ug/L	94	10		78	123	20
	Bromochloromethane	20	21.9	ug/L	110	12		78	123	20
	Chloroform	20	18.9	ug/L	95	4		79	124	20
	1,1,1-Trichloroethane	20	19.3	ug/L	97	6		74	131	20
	Methylcyclohexane	20	18.7	ug/L	94	3		72	132	20
	Benzene	20	19.4	ug/L	97	5		79	120	20
	1,2-Dichloroethane	20	18.9	ug/L	95	10		73	128	20
	Trichloroethene	20	19.9	ug/L	100	8		79	123	20
	1,2-Dichloropropane	20	20.1	ug/L	101	12		78	122	20
	Bromodichloromethane	20	19.8	ug/L	99	7		79	125	20
	4-Methyl-2-Pentanone	100	89.4	ug/L	89	14		67	130	20
	Toluene	20	19.7	ug/L	99	7		80	121	20
	t-1,3-Dichloropropene	20	18.7	ug/L	94	9		73	127	20
	cis-1,3-Dichloropropene	20	19.2	ug/L	96	6		75	124	20
	1,1,2-Trichloroethane	20	20.3	ug/L	102	13		80	119	20
	2-Hexanone	100	85.0	ug/L	85	14		57	139	20
	Dibromochloromethane	20	20.3	ug/L	102	10		74	126	20
	1,2-Dibromoethane	20	19.5	ug/L	98	10		77	121	20
	Tetrachloroethene	20	20.1	ug/L	101	7		74	129	20
	Chlorobenzene	20	19.8	ug/L	99	7		82	118	20
	Ethyl Benzene	20	19.5	ug/L	98	7		79	121	20
	m/p-Xylenes	40	39.4	ug/L	99	6		80	121	20
	o-Xylene	20	19.5	ug/L	98	7		78	122	20
	Styrene	20	19.5	ug/L	98	9		78	123	20
	Bromoform	20	19.3	ug/L	97	9		66	130	20
	Isopropylbenzene	20	18.7	ug/L	94	7		72	131	20
	1,1,2,2-Tetrachloroethane	20	18.2	ug/L	91	12		71	121	20
	1,3-Dichlorobenzene	20	18.5	ug/L	93	6		80	119	20
	1,4-Dichlorobenzene	20	18.4	ug/L	92	4		79	118	20
	1,2-Dichlorobenzene	20	18.7	ug/L	94	10		80	119	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3246</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>VX047958.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1002WBSD01	1,2-Dibromo-3-Chloropropane	20	15.2	ug/L	76	13		62	128	20
	1,2,4-Trichlorobenzene	20	16.5	ug/L	83	6		69	130	20
	1,2,3-Trichlorobenzene	20	16.9	ug/L	85	7		69	129	20