

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3246</u>
Lab File ID:	<u>VX047953.D</u>	Date Analyzed:	<u>10/02/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>08:50</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	107762	12				
UPPER LIMIT	215524	12.5				
LOWER LIMIT	53881	11.5				
EPA SAMPLE NO.						
BP-VPB-184-TB-20250925	177276	12.01				
BP-VPB-184-GW-460-462	122681	12.01				
BP-VPB-184-GW-480-482	168922	12.01				
BP-VPB-184-GW-500-502	156960	12.01				
BP-VPB-184-DUP-20250925	150587	12.01				
BP-VPB-184-GW-520-522	167184	12.01				
BP-VPB-184-GW-540-542	143261	12.01				
VX1002WBL01	160196	12.01				
VX1002WBS01	117908	12.01				
VX1002WBSD01	118503	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.