

LAB CHRONICLE

OrderID:	Q3247	OrderDate:	9/30/2025 9:46:00 AM
Client:	Alliance Technical Group, LLC - Newark	Project:	NJ AIR PT
Contact:	Mohammad Ahmed	Location:	VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3247-01	RR-VOAS-AIR	Air	VOCMS Group3	TO-15	09/29/25		10/15/25	09/30/25
Q3247-01DL	RR-VOAS-AIRDL	Air	VOCMS Group3	TO-15	09/29/25		10/15/25	09/30/25

Hit Summary Sheet
SW-846

SDG No.: Q3247
Client: Alliance Technical Group, LLC - Newark

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID: Q3247-01	RR-VOAS-AIR RR-VOAS-AIR	Air	m/p-Xylene	220	E	1.78	4.34	ug/m3
			Total Voc :	220				
			Total Concentration:	220				
Client ID: Q3247-01DL	RR-VOAS-AIRDL RR-VOAS-AIRDL	Air	m/p-Xylene	245	D	3.56	8.69	ug/m3
			Total Voc :	245				
			Total Concentration:	245				



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Project : NJ AIR PT

Sampling Date : 09/29/25

Field Id Number : RR-VOAS-AIR

Analysis Date : 10/15/25

Laboratory Id Number : Q3247-01

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Chloroethane	75-00-3	64.52	0.15	U	0.4		
Heptane	142-82-5	100.2	0.17	U	0.7		
Bromodichloromethane	75-27-4	163.8	0.06	U	0.4		
m/p-Xylene	179601-23-1	106.2	50.7	E	220		

Project : NJ AIR PT

Sampling Date : 09/29/25

Field Id Number : RR-VOAS-AIRDL

Analysis Date : 10/15/25

Laboratory Id Number : Q3247-01DL

Target Analyts : Air Results

Chemical	Cas Number	Molecular Weight	Insert Results in PPBV	Qualifier	Generate Results in ug/m3	QAS Decision	Foot Notes
Chloroethane	75-00-3	64.52	0.3	UD	0.79		
Heptane	142-82-5	100.2	0.34	UD	1.39		
Bromodichloromethane	75-27-4	163.8	0.13	UD	0.87		
m/p-Xylene	179601-23-1	106.2	56.3	D	244		

Method	Matrix	Parameter	Instrument ID	Column ID	MDL (ppbv)	RL (ppbv)	MB Data	MB Datafile	StdevVal
TO-15	Air	1,1,1,2-Tetrachloroethane	MSVOA_L	RTX-1	0.089	0.50	0		0.028115408
TO-15	Air	1,1,1-Trichloroethane	MSVOA_L	RTX-1	0.016	0.030	0		0.005089672
TO-15	Air	1,1,2,2-Tetrachloroethane	MSVOA_L	RTX-1	0.018	0.030	0		0.005472877
TO-15	Air	1,1,2-Trichloroethane	MSVOA_L	RTX-1	0.081	0.50	0		0.02572751
TO-15	Air	1,1,2-Trichlorotrifluoroethane	MSVOA_L	RTX-1	0.14	0.50	0		0.042983939
TO-15	Air	1,1-Dichloroethane	MSVOA_L	RTX-1	0.13	0.50	0		0.040178175
TO-15	Air	1,1-Dichloroethene	MSVOA_L	RTX-1	0.15	0.50	0		0.047056197
TO-15	Air	1,2,4-Trichlorobenzene	MSVOA_L	RTX-1	0.21	0.50	0		0.064291005
TO-15	Air	1,2,4-Trimethylbenzene	MSVOA_L	RTX-1	0.18	0.50	0		0.054467115
TO-15	Air	1,2-Dibromoethane	MSVOA_L	RTX-1	0.085	0.10	0		0.026726124
TO-15	Air	1,2-Dichlorobenzene	MSVOA_L	RTX-1	0.13	0.50	0		0.040999419
TO-15	Air	1,2-Dichloroethane	MSVOA_L	RTX-1	0.093	0.50	0		0.029277002
TO-15	Air	1,2-Dichloropropane	MSVOA_L	RTX-1	0.13	0.50	0		0.039880775
TO-15	Air	1,3,5-Trimethylbenzene	MSVOA_L	RTX-1	0.18	0.50	0		0.054598099
TO-15	Air	1,3-Butadiene	MSVOA_L	RTX-1	0.051	0.50	0		0.016035675
TO-15	Air	1,3-Dichlorobenzene	MSVOA_L	RTX-1	0.054	0.50	0		0.017043362
TO-15	Air	1,4-Dichlorobenzene	MSVOA_L	RTX-1	0.13	0.50	0		0.039460649
TO-15	Air	1,4-Dioxane	MSVOA_L	RTX-1	0.22	0.50	0		0.067330033
TO-15	Air	2,2,4-Trimethylpentane	MSVOA_L	RTX-1	0.14	0.50	0		0.044131837
TO-15	Air	2-Butanone	MSVOA_L	RTX-1	0.056	0.50	0		0.017728105
TO-15	Air	2-Chlorotoluene	MSVOA_L	RTX-1	0.17	0.50	0		0.052599113
TO-15	Air	2-Hexanone	MSVOA_L	RTX-1	0.11	0.50	0		0.033523268
TO-15	Air	4-Ethyltoluene	MSVOA_L	RTX-1	0.21	0.50	0		0.066332496
TO-15	Air	4-Methyl-2-Pentanone	MSVOA_L	RTX-1	0.097	0.50	0		0.030783422
TO-15	Air	Acetone	MSVOA_L	RTX-1	0.18	0.50	0		0.055592052
TO-15	Air	Allyl Chloride	MSVOA_L	RTX-1	0.11	0.50	0		0.033380918
TO-15	Air	Benzene	MSVOA_L	RTX-1	0.079	0.50	0		0.025071327
TO-15	Air	Benzyl Chloride	MSVOA_L	RTX-1	0.089	0.50	0		0.028284271
TO-15	Air	Bromodichloromethane	MSVOA_L	RTX-1	0.063	0.50	0		0.019880596
TO-15	Air	Bromoethene	MSVOA_L	RTX-1	0.2	0.50	0		0.062678317
TO-15	Air	Bromoform	MSVOA_L	RTX-1	0.048	0.50	0		0.015118579
TO-15	Air	Bromomethane	MSVOA_L	RTX-1	0.077	0.50	0		0.024494897
TO-15	Air	Carbon Disulfide	MSVOA_L	RTX-1	0.08	0.50	0		0.02544836
TO-15	Air	Carbon Tetrachloride	MSVOA_L	RTX-1	0.025	0.030	0		0.007690439
TO-15	Air	Chlorobenzene	MSVOA_L	RTX-1	0.077	0.50	0		0.024397502
TO-15	Air	Chlorodifluoromethane	MSVOA_L	RTX-1	0.12	0.50	0		0.037859389
TO-15	Air	Chloroethane	MSVOA_L	RTX-1	0.15	0.50	0		0.045250625
TO-15	Air	Chloroform	MSVOA_L	RTX-1	0.096	0.50	0		0.030394235
TO-15	Air	Chloromethane	MSVOA_L	RTX-1	0.098	0.50	0		0.03101459
TO-15	Air	cis-1,2-Dichloroethene	MSVOA_L	RTX-1	0.099	0.50	0		0.031471832
TO-15	Air	cis-1,3-Dichloropropene	MSVOA_L	RTX-1	0.11	0.50	0		0.032071349
TO-15	Air	Cyclohexane	MSVOA_L	RTX-1	0.22	0.50	0		0.06972736
TO-15	Air	Dibromochloromethane	MSVOA_L	RTX-1	0.085	0.50	0		0.026726124
TO-15	Air	Dichlorodifluoromethane	MSVOA_L	RTX-1	0.11	0.50	0		0.033380918
TO-15	Air	Dichlorotetrafluoroethane	MSVOA_L	RTX-1	0.15	0.50	0		0.045460606
TO-15	Air	Ethanol	MSVOA_L	RTX-1	0.32	0.50	0		0.098848128
TO-15	Air	Ethyl Acetate	MSVOA_L	RTX-1	0.12	0.50	0		0.036384193
TO-15	Air	Ethyl Benzene	MSVOA_L	RTX-1	0.19	0.50	0		0.058554004
TO-15	Air	Heptane	MSVOA_L	RTX-1	0.17	0.50	0		0.051130086
TO-15	Air	Hexachloro-1,3-Butadiene	MSVOA_L	RTX-1	0.13	0.50	0		0.038297084
TO-15	Air	Hexane	MSVOA_L	RTX-1	0.16	0.50	0		0.04894117
TO-15	Air	Isopropyl Alcohol	MSVOA_L	RTX-1	0.23	0.50	0		0.070676998
TO-15	Air	Isopropylbenzene	MSVOA_L	RTX-1	0.16	0.50	0		0.048205908
TO-15	Air	m/p-Xylene	MSVOA_L	RTX-1	0.41	1.00	0		0.130091543
TO-15	Air	Methyl Methacrylate	MSVOA_L	RTX-1	0.14	0.50	0		0.044131837
TO-15	Air	Methyl tert-Butyl Ether	MSVOA_L	RTX-1	0.23	0.50	0		0.07296444
TO-15	Air	Methylene Chloride	MSVOA_L	RTX-1	0.23	0.50	0		0.07204496
TO-15	Air	Naphthalene	MSVOA_L	RTX-1	0.013	0.10	0		0.004076647
TO-15	Air	Naphthalene,2-methyl-	MSVOA_L	RTX-1	0.22	0.50	0		0.069213266
TO-15	Air	n-Butylbenzene	MSVOA_L	RTX-1	0.18	0.50	0		0.055805786
TO-15	Air	n-propylbenzene	MSVOA_L	RTX-1	0.2	0.50	0		0.061644414
TO-15	Air	o-Xylene	MSVOA_L	RTX-1	0.21	0.50	0		0.06491753
TO-15	Air	p-Isopropyltoluene	MSVOA_L	RTX-1	0.19	0.50	0		0.059561893
TO-15	Air	Propene	MSVOA_L	RTX-1	0.077	0.50	0		0.024397502
TO-15	Air	sec-Butylbenzene	MSVOA_L	RTX-1	0.2	0.50	0		0.061913919
TO-15	Air	Styrene	MSVOA_L	RTX-1	0.16	0.50	0		0.05
TO-15	Air	t-1,3-Dichloropropene	MSVOA_L	RTX-1	0.15	0.50	0		0.046496288
TO-15	Air	tert-Butyl alcohol	MSVOA_L	RTX-1	0.16	0.50	0		0.048892496
TO-15	Air	tert-Butylbenzene	MSVOA_L	RTX-1	0.22	0.50	0		0.067928534
TO-15	Air	Tetrachloroethene	MSVOA_L	RTX-1	0.015	0.030	0		0.004725816
TO-15	Air	Tetrahydrofuran	MSVOA_L	RTX-1	0.08	0.50	0		0.025166115
TO-15	Air	Toluene	MSVOA_L	RTX-1	0.16	0.50	0		0.050142654
TO-15	Air	trans-1,2-Dichloroethene	MSVOA_L	RTX-1	0.12	0.50	0		0.036968455
TO-15	Air	Trichloroethene	MSVOA_L	RTX-1	0.024	0.030	0		0.007587584
TO-15	Air	Trichlorofluoromethane	MSVOA_L	RTX-1	0.17	0.50	0		0.053452248
TO-15	Air	Vinyl Acetate	MSVOA_L	RTX-1	0.26	0.50	0		0.079671946
TO-15	Air	Vinyl Chloride	MSVOA_L	RTX-1	0.025	0.030	0		0.007955232

Students	Spike_Amt_A	Datafile1	Concentration1	Adate1	Datafile2	Concentration2	Adate2
Value	dded						
3.143	0.4	VL041877.D	0.48	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.03	VL041883.D	0.037	01102025T00:17:00	VL041884.D	0.04	01102025T00:47:00
3.143	0.03	VL041883.D	0.045	01102025T00:17:00	VL041884.D	0.04	01102025T00:47:00
3.143	0.4	VL041877.D	0.46	01092025T21:16:00	VL041878.D	0.52	01092025T21:46:00
3.143	0.4	VL041877.D	0.53	01092025T21:16:00	VL041878.D	0.54	01092025T21:46:00
3.143	0.4	VL041877.D	0.53	01092025T21:16:00	VL041878.D	0.51	01092025T21:46:00
3.143	0.4	VL041877.D	0.5	01092025T21:16:00	VL041878.D	0.53	01092025T21:46:00
3.143	0.4	VL041877.D	0.45	01092025T21:16:00	VL041878.D	0.47	01092025T21:46:00
3.143	0.4	VL041877.D	0.46	01092025T21:16:00	VL041878.D	0.43	01092025T21:46:00
3.143	0.4	VL041877.D	0.48	01092025T21:16:00	VL041878.D	0.45	01092025T21:46:00
3.143	0.4	VL041877.D	0.52	01092025T21:16:00	VL041878.D	0.51	01092025T21:46:00
3.143	0.4	VL041877.D	0.52	01092025T21:16:00	VL041878.D	0.54	01092025T21:46:00
3.143	0.4	VL041877.D	0.53	01092025T21:16:00	VL041878.D	0.49	01092025T21:46:00
3.143	0.4	VL041877.D	0.44	01092025T21:16:00	VL041878.D	0.44	01092025T21:46:00
3.143	0.4	VL041877.D	0.52	01092025T21:16:00	VL041878.D	0.54	01092025T21:46:00
3.143	0.4	VL041877.D	0.49	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.4	VL041877.D	0.48	01092025T21:16:00	VL041878.D	0.47	01092025T21:46:00
3.143	0.4	VL041877.D	0.49	01092025T21:16:00	VL041878.D	0.45	01092025T21:46:00
3.143	0.4	VL041877.D	0.47	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.4	VL041877.D	0.47	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.4	VL041877.D	0.44	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.4	VL041877.D	0.41	01092025T21:16:00	VL041878.D	0.38	01092025T21:46:00
3.143	0.4	VL041877.D	0.44	01092025T21:16:00	VL041878.D	0.44	01092025T21:46:00
3.143	0.4	VL041877.D	0.44	01092025T21:16:00	VL041878.D	0.43	01092025T21:46:00
3.143	0.4	VL041877.D	0.6	01092025T21:16:00	VL041878.D	0.62	01092025T21:46:00
3.143	0.4	VL041877.D	0.46	01092025T21:16:00	VL041878.D	0.51	01092025T21:46:00
3.143	0.4	VL041877.D	0.47	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.4	VL041877.D	0.31	01092025T21:16:00	VL041878.D	0.29	01092025T21:46:00
3.143	0.4	VL041877.D	0.45	01092025T21:16:00	VL041878.D	0.44	01092025T21:46:00
3.143	0.4	VL041877.D	0.49	01092025T21:16:00	VL041878.D	0.53	01092025T21:46:00
3.143	0.4	VL041877.D	0.36	01092025T21:16:00	VL041878.D	0.35	01092025T21:46:00
3.143	0.4	VL041877.D	0.49	01092025T21:16:00	VL041878.D	0.52	01092025T21:46:00
3.143	0.4	VL041877.D	0.41	01092025T21:16:00	VL041878.D	0.45	01092025T21:46:00
3.143	0.03	VL041883.D	0.036	01102025T00:17:00	VL041884.D	0.038	01102025T00:47:00
3.143	0.4	VL041877.D	0.5	01092025T21:16:00	VL041878.D	0.48	01092025T21:46:00
3.143	0.4	VL041877.D	0.47	01092025T21:16:00	VL041878.D	0.47	01092025T21:46:00
3.143	0.4	VL041877.D	0.57	01092025T21:16:00	VL041878.D	0.59	01092025T21:46:00
3.143	0.4	VL041877.D	0.51	01092025T21:16:00	VL041878.D	0.5	01092025T21:46:00
3.143	0.4	VL041877.D	0.54	01092025T21:16:00	VL041878.D	0.51	01092025T21:46:00
3.143	0.4	VL041877.D	0.48	01092025T21:16:00	VL041878.D	0.45	01092025T21:46:00
3.143	0.4	VL041877.D	0.41	01092025T21:16:00	VL041878.D	0.4	01092025T21:46:00
3.143	0.4	VL041877.D	0.46	01092025T21:16:00	VL041878.D	0.47	01092025T21:46:00
3.143	0.4	VL041877.D	0.39	01092025T21:16:00	VL041878.D	0.4	01092025T21:46:00
3.143	0.4	VL041877.D	0.47	01092025T21:16:00	VL041878.D	0.51	01092025T21:46:00
3.143	0.4	VL041877.D	0.53	01092025T21:16:00	VL041878.D	0.52	01092025T21:46:00
3.143	0.4	VL042501.D	0.885	05062025T17:36:00	VL042502.D	0.76	05062025T18:07:00
3.143	0.4	VL041877.D	0.46	01092025T21:16:00	VL041878.D	0.45	01092025T21:46:00
3.143	0.4	VL041877.D	0.43	01092025T21:16:00	VL041878.D	0.44	01092025T21:46:00
3.143	0.4	VL041877.D	0.43	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.4	VL041877.D	0.49	01092025T21:16:00	VL041878.D	0.5	01092025T21:46:00
3.143	0.4	VL041877.D	0.46	01092025T21:16:00	VL041878.D	0.47	01092025T21:46:00
3.143	0.4	VL041877.D	0.59	01092025T21:16:00	VL041878.D	0.6	01092025T21:46:00
3.143	0.4	VL041877.D	0.45	01092025T21:16:00	VL041878.D	0.45	01092025T21:46:00
3.143	0.4	VL041877.D	0.91	01092025T21:16:00	VL041878.D	0.91	01092025T21:46:00
3.143	0.4	VL041877.D	0.44	01092025T21:16:00	VL041878.D	0.41	01092025T21:46:00
3.143	0.4	VL041877.D	0.4	01092025T21:16:00	VL041878.D	0.47	01092025T21:46:00
3.143	0.4	VL041877.D	0.68	01092025T21:16:00	VL041878.D	0.61	01092025T21:46:00
3.143	0.04	VL042504.D	0.035	05062025T19:10:00	VL042505.D	0.04	05062025T19:42:00
3.143	0.4	VL041877.D	0.34	01092025T21:16:00	VL041878.D	0.34	01092025T21:46:00
3.143	0.4	VL041877.D	0.44	01092025T21:16:00	VL041878.D	0.43	01092025T21:46:00
3.143	0.4	VL041877.D	0.44	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.4	VL041877.D	0.45	01092025T21:16:00	VL041878.D	0.45	01092025T21:46:00
3.143	0.4	VL041877.D	0.45	01092025T21:16:00	VL041878.D	0.44	01092025T21:46:00
3.143	0.4	VL041877.D	0.53	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.4	VL041877.D	0.48	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.4	VL041877.D	0.42	01092025T21:16:00	VL041878.D	0.4	01092025T21:46:00
3.143	0.4	VL041877.D	0.41	01092025T21:16:00	VL041878.D	0.36	01092025T21:46:00
3.143	0.4	VL041877.D	0.5	01092025T21:16:00	VL041878.D	0.45	01092025T21:46:00
3.143	0.4	VL041877.D	0.46	01092025T21:16:00	VL041878.D	0.46	01092025T21:46:00
3.143	0.03	VL041883.D	0.037	01102025T00:17:00	VL041884.D	0.04	01102025T00:47:00
3.143	0.4	VL041877.D	0.47	01092025T21:16:00	VL041878.D	0.43	01092025T21:46:00
3.143	0.4	VL041877.D	0.45	01092025T21:16:00	VL041878.D	0.43	01092025T21:46:00
3.143	0.4	VL041877.D	0.45	01092025T21:16:00	VL041878.D	0.52	01092025T21:46:00
3.143	0.03	VL041883.D	0.047	01102025T00:17:00	VL041884.D	0.047	01102025T00:47:00
3.143	0.4	VL041877.D	0.53	01092025T21:16:00	VL041878.D	0.52	01092025T21:46:00
3.143	0.4	VL041877.D	0.43	01092025T21:16:00	VL041878.D	0.47	01092025T21:46:00
3.143	0.03	VL042498.D	0.018	05062025T16:01:00	VL042500.D	0.032	05062025T17:04:00

Datafile3	Concentration3	Adate3	Datafile4	Concentration4	Adate4	Datafile5	Concentration5
VL041879.D	0.43	01092025T22:17:00	VL041909.D	0.42	01152025T00:52:00	VL041910.D	0.4
VL041885.D	0.038	01102025T01:18:00	VL041913.D	0.033	01152025T02:53:00	VL041914.D	0.037
VL041885.D	0.044	01102025T01:18:00	VL041913.D	0.033	01152025T02:53:00	VL041914.D	0.032
VL041879.D	0.48	01092025T22:17:00	VL041909.D	0.44	01152025T00:52:00	VL041910.D	0.47
VL041879.D	0.55	01092025T22:17:00	VL041909.D	0.5	01152025T00:52:00	VL041910.D	0.47
VL041879.D	0.57	01092025T22:17:00	VL041909.D	0.5	01152025T00:52:00	VL041910.D	0.48
VL041879.D	0.58	01092025T22:17:00	VL041909.D	0.49	01152025T00:52:00	VL041910.D	0.49
VL041879.D	0.48	01092025T22:17:00	VL041909.D	0.37	01152025T00:52:00	VL041910.D	0.36
VL041879.D	0.47	01092025T22:17:00	VL041909.D	0.37	01152025T00:52:00	VL041910.D	0.39
VL041879.D	0.47	01092025T22:17:00	VL041909.D	0.46	01152025T00:52:00	VL041910.D	0.43
VL041879.D	0.53	01092025T22:17:00	VL041909.D	0.44	01152025T00:52:00	VL041910.D	0.43
VL041879.D	0.5	01092025T22:17:00	VL041909.D	0.47	01152025T00:52:00	VL041910.D	0.46
VL041879.D	0.5	01092025T22:17:00	VL041909.D	0.43	01152025T00:52:00	VL041910.D	0.42
VL041879.D	0.43	01092025T22:17:00	VL041909.D	0.34	01152025T00:52:00	VL041910.D	0.34
VL041879.D	0.51	01092025T22:17:00	VL041909.D	0.5	01152025T00:52:00	VL041910.D	0.51
VL041879.D	0.48	01092025T22:17:00	VL041909.D	0.46	01152025T00:52:00	VL041910.D	0.44
VL041879.D	0.51	01092025T22:17:00	VL041909.D	0.43	01152025T00:52:00	VL041910.D	0.41
VL041879.D	0.5	01092025T22:17:00	VL041909.D	0.45	01152025T00:52:00	VL041910.D	0.47
VL041879.D	0.46	01092025T22:17:00	VL041909.D	0.37	01152025T00:52:00	VL041910.D	0.39
VL041879.D	0.49	01092025T22:17:00	VL041909.D	0.44	01152025T00:52:00	VL041910.D	0.44
VL041879.D	0.46	01092025T22:17:00	VL041909.D	0.37	01152025T00:52:00	VL041910.D	0.38
VL041879.D	0.39	01092025T22:17:00	VL041909.D	0.34	01152025T00:52:00	VL041910.D	0.36
VL041879.D	0.44	01092025T22:17:00	VL041909.D	0.33	01152025T00:52:00	VL041910.D	0.33
VL041879.D	0.43	01092025T22:17:00	VL041909.D	0.39	01152025T00:52:00	VL041910.D	0.37
VL041879.D	0.68	01092025T22:17:00	VL041909.D	0.53	01152025T00:52:00	VL041910.D	0.53
VL041879.D	0.5	01092025T22:17:00	VL041909.D	0.47	01152025T00:52:00	VL041910.D	0.5
VL041879.D	0.46	01092025T22:17:00	VL041909.D	0.42	01152025T00:52:00	VL041910.D	0.41
VL041879.D	0.28	01092025T22:17:00	VL041909.D	0.26	01152025T00:52:00	VL041910.D	0.34
VL041879.D	0.44	01092025T22:17:00	VL041909.D	0.4	01152025T00:52:00	VL041910.D	0.4
VL041879.D	0.55	01092025T22:17:00	VL041909.D	0.44	01152025T00:52:00	VL041910.D	0.52
VL041879.D	0.37	01092025T22:17:00	VL041909.D	0.33	01152025T00:52:00	VL041910.D	0.34
VL041879.D	0.53	01092025T22:17:00	VL041909.D	0.47	01152025T00:52:00	VL041910.D	0.5
VL041879.D	0.43	01092025T22:17:00	VL041909.D	0.41	01152025T00:52:00	VL041910.D	0.4
VL041885.D	0.048	01102025T01:18:00	VL041913.D	0.026	01152025T02:53:00	VL041914.D	0.033
VL041879.D	0.52	01092025T22:17:00	VL041909.D	0.45	01152025T00:52:00	VL041910.D	0.46
VL041879.D	0.51	01092025T22:17:00	VL041909.D	0.42	01152025T00:52:00	VL041910.D	0.42
VL041879.D	0.52	01092025T22:17:00	VL041909.D	0.46	01152025T00:52:00	VL041910.D	0.5
VL041879.D	0.53	01092025T22:17:00	VL041909.D	0.45	01152025T00:52:00	VL041910.D	0.46
VL041879.D	0.53	01092025T22:17:00	VL041909.D	0.52	01152025T00:52:00	VL041910.D	0.45
VL041879.D	0.49	01092025T22:17:00	VL041909.D	0.43	01152025T00:52:00	VL041910.D	0.42
VL041879.D	0.38	01092025T22:17:00	VL041909.D	0.34	01152025T00:52:00	VL041910.D	0.34
VL041879.D	0.5	01092025T22:17:00	VL041909.D	0.35	01152025T00:52:00	VL041910.D	0.35
VL041879.D	0.41	01092025T22:17:00	VL041909.D	0.33	01152025T00:52:00	VL041910.D	0.38
VL041879.D	0.56	01092025T22:17:00	VL041909.D	0.46	01152025T00:52:00	VL041910.D	0.5
VL041879.D	0.56	01092025T22:17:00	VL041909.D	0.46	01152025T00:52:00	VL041910.D	0.46
VL042503.D	0.611	05062025T18:39:00	VL042518.D	0.831	05072025T15:02:00	VL042519.D	0.842
VL041879.D	0.47	01092025T22:17:00	VL041909.D	0.41	01152025T00:52:00	VL041910.D	0.38
VL041879.D	0.44	01092025T22:17:00	VL041909.D	0.34	01152025T00:52:00	VL041910.D	0.35
VL041879.D	0.46	01092025T22:17:00	VL041909.D	0.37	01152025T00:52:00	VL041910.D	0.37
VL041879.D	0.51	01092025T22:17:00	VL041909.D	0.42	01152025T00:52:00	VL041910.D	0.43
VL041879.D	0.49	01092025T22:17:00	VL041909.D	0.42	01152025T00:52:00	VL041910.D	0.38
VL041879.D	0.63	01092025T22:17:00	VL041909.D	0.5	01152025T00:52:00	VL041910.D	0.5
VL041879.D	0.46	01092025T22:17:00	VL041909.D	0.37	01152025T00:52:00	VL041910.D	0.38
VL041879.D	0.92	01092025T22:17:00	VL041909.D	0.72	01152025T00:52:00	VL041910.D	0.69
VL041879.D	0.43	01092025T22:17:00	VL041909.D	0.36	01152025T00:52:00	VL041910.D	0.35
VL041879.D	0.42	01092025T22:17:00	VL041909.D	0.33	01152025T00:52:00	VL041910.D	0.52
VL041879.D	0.63	01092025T22:17:00	VL041909.D	0.54	01152025T00:52:00	VL041910.D	0.57
VL042506.D	0.04	05062025T20:13:00	VL042521.D	0.031	05072025T16:49:00	VL042522.D	0.032
VL041879.D	0.33	01092025T22:17:00	VL041909.D	0.28	01152025T00:52:00	VL041910.D	0.25
VL041879.D	0.42	01092025T22:17:00	VL041909.D	0.33	01152025T00:52:00	VL041910.D	0.34
VL041879.D	0.43	01092025T22:17:00	VL041909.D	0.36	01152025T00:52:00	VL041910.D	0.33
VL041879.D	0.47	01092025T22:17:00	VL041909.D	0.35	01152025T00:52:00	VL041910.D	0.35
VL041879.D	0.44	01092025T22:17:00	VL041909.D	0.34	01152025T00:52:00	VL041910.D	0.36
VL041879.D	0.52	01092025T22:17:00	VL041909.D	0.5	01152025T00:52:00	VL041910.D	0.48
VL041879.D	0.45	01092025T22:17:00	VL041909.D	0.38	01152025T00:52:00	VL041910.D	0.36
VL041879.D	0.38	01092025T22:17:00	VL041909.D	0.34	01152025T00:52:00	VL041910.D	0.31
VL041879.D	0.38	01092025T22:17:00	VL041909.D	0.36	01152025T00:52:00	VL041910.D	0.33
VL041879.D	0.52	01092025T22:17:00	VL041909.D	0.47	01152025T00:52:00	VL041910.D	0.41
VL041879.D	0.46	01092025T22:17:00	VL041909.D	0.36	01152025T00:52:00	VL041910.D	0.37
VL041885.D	0.034	01102025T01:18:00	VL041913.D	0.037	01152025T02:53:00	VL041914.D	0.034
VL041879.D	0.41	01092025T22:17:00	VL041909.D	0.41	01152025T00:52:00	VL041910.D	0.42
VL041879.D	0.44	01092025T22:17:00	VL041909.D	0.37	01152025T00:52:00	VL041910.D	0.36
VL041879.D	0.54	01092025T22:17:00	VL041909.D	0.47	01152025T00:52:00	VL041910.D	0.48
VL041885.D	0.035	01102025T01:18:00	VL041913.D	0.029	01152025T02:53:00	VL041914.D	0.033
VL041879.D	0.54	01092025T22:17:00	VL041909.D	0.45	01152025T00:52:00	VL041910.D	0.47
VL041879.D	0.56	01092025T22:17:00	VL041909.D	0.35	01152025T00:52:00	VL041910.D	0.44
VL042515.D	0.021	05072025T13:17:00	VL042516.D	0.021	05072025T13:50:00	VL042517.D	0.014

Adate5	Datafile6	Concentration6	Adate6	Datafile7	Concentration7	Adate7	Department
01152025T01:23:00	VL041932.D	0.41	01202025T22:30:00	VL041933.D	0.43	01202025T23:00:00	VOC
01152025T03:23:00	VL041936.D	0.028	01212025T00:31:00	VL041940.D	0.027	01212025T09:44:00	VOC
01152025T03:23:00	VL041936.D	0.034	01212025T00:31:00	VL041940.D	0.034	01212025T09:44:00	VOC
01152025T01:23:00	VL041932.D	0.46	01202025T22:30:00	VL041933.D	0.49	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.46	01202025T22:30:00	VL041933.D	0.44	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.45	01202025T22:30:00	VL041933.D	0.47	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.47	01202025T22:30:00	VL041933.D	0.43	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.34	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.35	01202025T22:30:00	VL041933.D	0.33	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.45	01202025T22:30:00	VL041933.D	0.4	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.45	01202025T22:30:00	VL041933.D	0.49	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.49	01202025T22:30:00	VL041933.D	0.47	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.5	01202025T22:30:00	VL041933.D	0.47	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.33	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.49	01202025T22:30:00	VL041933.D	0.52	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.46	01202025T22:30:00	VL041933.D	0.45	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.42	01202025T22:30:00	VL041933.D	0.41	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.42	01202025T22:30:00	VL041933.D	0.3	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.41	01202025T22:30:00	VL041933.D	0.37	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.47	01202025T22:30:00	VL041933.D	0.46	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.36	01202025T22:30:00	VL041933.D	0.33	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.35	01202025T22:30:00	VL041933.D	0.31	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.31	01202025T22:30:00	VL041933.D	0.3	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.38	01202025T22:30:00	VL041933.D	0.37	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.58	01202025T22:30:00	VL041933.D	0.54	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.43	01202025T22:30:00	VL041933.D	0.43	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.47	01202025T22:30:00	VL041933.D	0.43	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.29	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.42	01202025T22:30:00	VL041933.D	0.42	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.39	01202025T22:30:00	VL041933.D	0.41	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.34	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.47	01202025T22:30:00	VL041933.D	0.52	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.38	01202025T22:30:00	VL041933.D	0.38	01202025T23:00:00	VOC
01152025T03:23:00	VL041936.D	0.029	01212025T00:31:00	VL041940.D	0.027	01212025T09:44:00	VOC
01152025T01:23:00	VL041932.D	0.5	01202025T22:30:00	VL041933.D	0.48	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.51	01202025T22:30:00	VL041933.D	0.49	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.52	01202025T22:30:00	VL041933.D	0.49	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.46	01202025T22:30:00	VL041933.D	0.47	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.5	01202025T22:30:00	VL041933.D	0.48	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.42	01202025T22:30:00	VL041933.D	0.41	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.35	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.37	01202025T22:30:00	VL041933.D	0.33	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.38	01202025T22:30:00	VL041933.D	0.36	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.48	01202025T22:30:00	VL041933.D	0.51	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.45	01202025T22:30:00	VL041933.D	0.45	01202025T23:00:00	VOC
05072025T15:37:00	VL042520.D	0.766	05072025T16:13:00	VL042569.D	0.664	05212025T16:53:00	VOC
01152025T01:23:00	VL041932.D	0.39	01202025T22:30:00	VL041933.D	0.4	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.32	01202025T22:30:00	VL041933.D	0.31	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.37	01202025T22:30:00	VL041933.D	0.33	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.44	01202025T22:30:00	VL041933.D	0.43	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.4	01202025T22:30:00	VL041933.D	0.36	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.46	01202025T22:30:00	VL041933.D	0.46	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.36	01202025T22:30:00	VL041933.D	0.35	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.64	01202025T22:30:00	VL041933.D	0.65	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.35	01202025T22:30:00	VL041933.D	0.33	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.35	01202025T22:30:00	VL041933.D	0.33	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.5	01202025T22:30:00	VL041933.D	0.48	01202025T23:00:00	VOC
05072025T17:24:00	VL042523.D	0.03	05072025T17:58:00	VL042567.D	0.034	05212025T15:49:00	VOC
01152025T01:23:00	VL041932.D	0.2	01202025T22:30:00	VL041933.D	0.17	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.31	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.31	01202025T22:30:00	VL041933.D	0.33	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.32	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.32	01202025T22:30:00	VL041933.D	0.32	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.5	01202025T22:30:00	VL041933.D	0.48	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.34	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.3	01202025T22:30:00	VL041933.D	0.3	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.27	01202025T22:30:00	VL041933.D	0.31	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.44	01202025T22:30:00	VL041933.D	0.38	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.3	01202025T23:00:00	VOC
01152025T03:23:00	VL041936.D	0.026	01212025T00:31:00	VL041940.D	0.03	01212025T09:44:00	VOC
01152025T01:23:00	VL041932.D	0.39	01202025T22:30:00	VL041933.D	0.41	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.34	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.46	01202025T22:30:00	VL041933.D	0.44	01202025T23:00:00	VOC
01152025T03:23:00	VL041936.D	0.033	01212025T00:31:00	VL041940.D	0.03	01212025T09:44:00	VOC
01152025T01:23:00	VL041932.D	0.42	01202025T22:30:00	VL041933.D	0.41	01202025T23:00:00	VOC
01152025T01:23:00	VL041932.D	0.33	01202025T22:30:00	VL041933.D	0.37	01202025T23:00:00	VOC
05072025T14:29:00	VL042575.D	0.034	05222025T11:27:00	VL042576.D	0.032	05222025T11:59:00	VOC

AVG Concentration	Recovery	Effective Date
0.43	108	5/22/2025
0.03	114	5/22/2025
0.04	125	5/22/2025
0.47	119	5/22/2025
0.50	125	5/22/2025
0.50	125	5/22/2025
0.50	125	5/22/2025
0.40	100	5/22/2025
0.40	100	5/22/2025
0.45	112	5/22/2025
0.48	120	5/22/2025
0.49	123	5/22/2025
0.48	119	5/22/2025
0.38	95	5/22/2025
0.51	128	5/22/2025
0.46	116	5/22/2025
0.45	112	5/22/2025
0.44	110	5/22/2025
0.42	105	5/22/2025
0.46	115	5/22/2025
0.40	100	5/22/2025
0.36	91	5/22/2025
0.37	93	5/22/2025
0.40	100	5/22/2025
0.58	146	5/22/2025
0.47	118	5/22/2025
0.45	111	5/22/2025
0.30	75	5/22/2025
0.42	106	5/22/2025
0.48	119	5/22/2025
0.35	86	5/22/2025
0.50	125	5/22/2025
0.41	102	5/22/2025
0.03	113	5/22/2025
0.48	121	5/22/2025
0.47	118	5/22/2025
0.52	130	5/22/2025
0.48	121	5/22/2025
0.50	126	5/22/2025
0.44	111	5/22/2025
0.36	91	5/22/2025
0.40	101	5/22/2025
0.38	95	5/22/2025
0.50	125	5/22/2025
0.49	123	5/22/2025
0.77	191	5/22/2025
0.42	106	5/22/2025
0.38	94	5/22/2025
0.40	100	5/22/2025
0.46	115	5/22/2025
0.43	106	5/22/2025
0.53	134	5/22/2025
0.40	101	5/22/2025
0.78	194	5/22/2025
0.38	95	5/22/2025
0.40	101	5/22/2025
0.57	143	5/22/2025
0.03	86	5/22/2025
0.27	68	5/22/2025
0.37	93	5/22/2025
0.38	95	5/22/2025
0.39	97	5/22/2025
0.38	95	5/22/2025
0.50	124	5/22/2025
0.40	100	5/22/2025
0.35	88	5/22/2025
0.35	86	5/22/2025
0.45	113	5/22/2025
0.39	98	5/22/2025
0.03	113	5/22/2025
0.42	105	5/22/2025
0.39	97	5/22/2025
0.48	120	5/22/2025
0.04	121	5/22/2025
0.48	119	5/22/2025
0.42	105	5/22/2025
0.02	82	5/22/2025



SAMPLE DATA

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	09/29/25
Project:	NJ AIR PT	Date Received:	09/30/25
Client Sample ID:	RR-VOAS-AIR	SDG No.:	Q3247
Lab Sample ID:	Q3247-01	Matrix:	Air
Analytical Method:	TO-15	Test:	VOCMS Group3
Sample Wt/Vol:	400 mL		

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
TARGETS									
75-00-3	Chloroethane	0.15	0.40	U	1	0.40	1.32	10/15/25 10:50	VL101525
142-82-5	Heptane	0.17	0.70	U	1	0.70	2.05	10/15/25 10:50	VL101525
75-27-4	Bromodichloromethane	0.060	0.40	U	1	0.40	3.35	10/15/25 10:50	VL101525
179601-23-1	m/p-Xylene	50.7	220	E	1	1.78	4.34	10/15/25 10:50	VL101525
SURROGATES									
460-00-4	1-Bromo-4-Fluorobenzene	10.4				65 - 135	104% SPK: 10		
INTERNAL STANDARDS									
74-97-5	Bromochloromethane	Area Count							
540-36-3	1,4-Difluorobenzene	93100							
3114-55-4	Chlorobenzene-d5	301000							
		266000							

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	09/29/25
Project:	NJ AIR PT	Date Received:	09/30/25
Client Sample ID:	RR-VOAS-AIRDL	SDG No.:	Q3247
Lab Sample ID:	Q3247-01DL	Matrix:	Air
Analytical Method:	TO-15	Test:	VOCMS Group3
Sample Wt/Vol:	400 mL		

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
TARGETS									
75-00-3	Chloroethane	0.30	0.79	UD	2	0.79	2.64	10/15/25 12:12	VL101525
142-82-5	Heptane	0.34	1.39	UD	2	1.39	4.10	10/15/25 12:12	VL101525
75-27-4	Bromodichloromethane	0.13	0.87	UD	2	0.87	6.70	10/15/25 12:12	VL101525
179601-23-1	m/p-Xylene	56.3	245	D	2	3.56	8.69	10/15/25 12:12	VL101525
SURROGATES									
460-00-4	1-Bromo-4-Fluorobenzene	10.6				65 - 135	106% SPK: 10		
INTERNAL STANDARDS									
		Area Count							
74-97-5	Bromochloromethane	89800							
540-36-3	1,4-Difluorobenzene	290000							
3114-55-4	Chlorobenzene-d5	255000							

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements



QC SUMMARY

Surrogate Summary

SDG No.: Q3247

Client: Alliance Technical Group, LLC - Newark

Analytical Method: SWTO-15

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3247-01	RR-VOAS-AIR	1-Bromo-4-Fluorobenzene	10	10.4	104		65	135
Q3247-01DL	RR-VOAS-AIRDL	1-Bromo-4-Fluorobenzene	10	10.6	106		65	135
VL1015ABL01	VL1015ABL01	1-Bromo-4-Fluorobenzene	10	10.1	101		65	135
VL1015ABS01	VL1015ABS01	1-Bromo-4-Fluorobenzene	10	10.3	103		65	135
VL1015ABSD01	VL1015ABSD01	1-Bromo-4-Fluorobenzene	10	10.5	105		65	135

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3247 **Analytical Method:** SWTO-15
Client: Alliance Technical Group, LLC - New **Datafile :** VL043116.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL1015ABS01	Chloroethane	10	9.10	ppbv	91			70	130	
	Heptane	10	8.20	ppbv	82			70	130	
	Bromodichloromethane	10	9.90	ppbv	99			70	130	
	m/p-Xylene	20	18.3	ppbv	92			70	130	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3247 **Analytical Method:** SWTO-15
Client: Alliance Technical Group, LLC - New **Datafile :** VL043122.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL1015ABSD01	Chloroethane	10	9.70	ppbv	97	6		70	130	25
	Heptane	10	8.20	ppbv	82	0		70	130	25
	Bromodichloromethane	10	10.6	ppbv	106	7		70	130	25
	m/p-Xylene	20	19.6	ppbv	98	6		70	130	25

VOLATILE METHOD BLANK SUMMARY

Client ID

VL1015ABL01

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3247

Lab File ID: VL043115.D

Lab Sample ID: VL1015ABL01

Date Analyzed: 10/15/2025

Time Analyzed: 09:34

GC Column: RTX-1 ID: 0.32 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_L

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VL1015ABS01	VL1015ABS01	VL043116.D	10/15/2025
RR-VOAS-AIR	Q3247-01	VL043117.D	10/15/2025
RR-VOAS-AIRDL	Q3247-01DL	VL043119.D	10/15/2025
VL1015ABSD01	VL1015ABSD01	VL043122.D	10/15/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3247

Lab File ID: VL043103.D

BFB Injection Date: 10/14/2025

Instrument ID: MSVOA_L

BFB Injection Time: 08:42

GC Column: RTX-1 ID: 0.32 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	8.0 - 40.0% of mass 95	21.9
75	30.0 - 66.0% of mass 95	53.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.0 (0.0) 1
174	50.0 - 120.0% of mass 95	61.7
175	4.0 - 9.0% of mass 174	4.8 (7.8) 1
176	93.0 - 101.0% of mass 174	59.6 (96.5) 1
177	5.0 - 9.0% of mass 176	3.9 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICCC010	VSTDICCC010	VL043105.D	10/14/2025	10:31
VSTDICCC002	VSTDICCC002	VL043106.D	10/14/2025	11:27
VSTDICCC001	VSTDICCC001	VL043107.D	10/14/2025	12:01
VSTDICCC0.5	VSTDICCC0.5	VL043108.D	10/14/2025	12:34
VSTDICCC0.1	VSTDICCC0.1	VL043109.D	10/14/2025	13:08
VSTDICCC0.03	VSTDICCC0.03	VL043110.D	10/14/2025	13:42
VSTDICCC015	VSTDICCC015	VL043111.D	10/14/2025	14:18

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3247

Lab File ID: VL043113.D

BFB Injection Date: 10/15/2025

Instrument ID: MSVOA_L

BFB Injection Time: 07:45

GC Column: RTX-1 ID: 0.32 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	8.0 - 40.0% of mass 95	22.9
75	30.0 - 66.0% of mass 95	57.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	0.7 (1.2) 1
174	50.0 - 120.0% of mass 95	58.4
175	4.0 - 9.0% of mass 174	4.3 (7.4) 1
176	93.0 - 101.0% of mass 174	56.7 (97.2) 1
177	5.0 - 9.0% of mass 176	3.9 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC010	VSTDCCC010	VL043114.D	10/15/2025	08:43
VL1015ABL01	VL1015ABL01	VL043115.D	10/15/2025	09:34
VL1015ABS01	VL1015ABS01	VL043116.D	10/15/2025	10:15
RR-VOAS-AIR	Q3247-01	VL043117.D	10/15/2025	10:50
RR-VOAS-AIRDL	Q3247-01DL	VL043119.D	10/15/2025	12:12
VL1015ABSD01	VL1015ABSD01	VL043122.D	10/15/2025	15:57

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ALLI03
 Lab Code: ACE SDG NO.: Q3247
 Lab File ID: VL043114.D Date Analyzed: 10/15/2025
 Instrument ID: MSVOA_L Time Analyzed: 08:43
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	98213	2.79	316115	3.96	288868	8.89
UPPER LIMIT	137498	3.12	442561	4.29	404415	9.22
LOWER LIMIT	58927.8	2.46	189669	3.63	173321	8.56
EPA SAMPLE NO.						
RR-VOAS-AIR	93140	2.79	301002	3.97	265788	8.89
RR-VOAS-AIRDL	89788	2.79	289768	3.97	254747	8.89
VL1015ABL01	94939	2.79	308682	3.96	273385	8.88
VL1015ABS01	93836	2.79	302461	3.96	271209	8.89
VL1015ABSD01	91615	2.79	294311	3.97	260330	8.89

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
 AREA LOWER LIMIT = -40% of internal standard area
 RT UPPER LIMIT = +0.33 minutes of internal standard RT
 RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client: Alliance Technical Group, LLC - Newark
Project: NJ AIR PT
Client Sample ID: VL1015ABL01
Lab Sample ID: VL1015ABL01
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected:
Date Received:
SDG No.: Q3247
Matrix: Air
Test: VOCMS Group3

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
TARGETS									
75-00-3	Chloroethane	0.15	0.40	U	1	0.40	1.32	10/15/25 09:34	VL101525
142-82-5	Heptane	0.17	0.70	U	1	0.70	2.05	10/15/25 09:34	VL101525
75-27-4	Bromodichloromethane	0.060	0.40	U	1	0.40	3.35	10/15/25 09:34	VL101525
179601-23-1	m/p-Xylene	0.41	1.78	U	1	1.78	4.34	10/15/25 09:34	VL101525
SURROGATES									
460-00-4	1-Bromo-4-Fluorobenzene	10.1				65 - 135	101% SPK: 10		
INTERNAL STANDARDS									
		Area Count							
74-97-5	Bromochloromethane	94900							
540-36-3	1,4-Difluorobenzene	309000							
3114-55-4	Chlorobenzene-d5	273000							

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client: Alliance Technical Group, LLC - Newark
Project: NJ AIR PT
Client Sample ID: VL1015ABS01
Lab Sample ID: VL1015ABS01
Analytical Method: TO-15
Sample Wt/Vol: 400 mL

Date Collected:
Date Received:
SDG No.: Q3247
Matrix: Air
Test: VOCMS Group3

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
TARGETS									
75-00-3	Chloroethane	9.10	24.0		1	0.40	1.32	10/15/25 10:15	VL101525
142-82-5	Heptane	8.20	33.6		1	0.70	2.05	10/15/25 10:15	VL101525
75-27-4	Bromodichloromethane	9.90	66.3		1	0.40	3.35	10/15/25 10:15	VL101525
179601-23-1	m/p-Xylene	18.3	79.5		1	1.78	4.34	10/15/25 10:15	VL101525
SURROGATES									
460-00-4	1-Bromo-4-Fluorobenzene	10.3				65 - 135	103% SPK: 10		
INTERNAL STANDARDS									
		Area Count							
74-97-5	Bromochloromethane	93800							
540-36-3	1,4-Difluorobenzene	302000							
3114-55-4	Chlorobenzene-d5	271000							

U = Not Detected
RL = Reporting Limit
MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	Alliance Technical Group, LLC - Newark	Date Collected:	
Project:	NJ AIR PT	Date Received:	
Client Sample ID:	VL1015ABSD01	SDG No.:	Q3247
Lab Sample ID:	VL1015ABSD01	Matrix:	Air
Analytical Method:	TO-15	Test:	VOCMS Group3
Sample Wt/Vol:	400 mL		

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qua.	DF	MDL ug/m3	LOQ / CRQL ug/m3	Date Ana.	BatchID
TARGETS									
75-00-3	Chloroethane	9.70	25.6		1	0.40	1.32	10/15/25 15:57	VL101525
142-82-5	Heptane	8.20	33.6		1	0.70	2.05	10/15/25 15:57	VL101525
75-27-4	Bromodichloromethane	10.6	71.0		1	0.40	3.35	10/15/25 15:57	VL101525
179601-23-1	m/p-Xylene	19.6	85.1		1	1.78	4.34	10/15/25 15:57	VL101525
SURROGATES									
460-00-4	1-Bromo-4-Fluorobenzene	10.5				65 - 135	105% SPK: 10		
INTERNAL STANDARDS									
		Area Count							
74-97-5	Bromochloromethane	91600							
540-36-3	1,4-Difluorobenzene	294000							
3114-55-4	Chlorobenzene-d5	260000							

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements



CALIBRATION SUMMARY

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\

Method File : VL101425AIR.M

Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022

Last Update : Wed Oct 15 02:12:58 2025

Response Via : Initial Calibration

Calibration Files

0.03=VL043110.D 0.1 =VL043109.D 0.5 =VL043108.D 1 =VL043107.D 2 =VL043106.D 10 =VL043105.D 15 =VL043111.D

Compound	0.03	0.1	0.5	1	2	10	15	Avg	%RSD

1) I Bromochloromethane	-----ISTD-----								
2) T Dichlorodifluo...			1.545	1.527	1.481	1.278	1.278	1.422	9.38
3) Chlorodifluoro...			1.672	1.632	1.577	1.451	1.359	1.538	8.46
4) Chloromethane			0.640	0.621	0.579	0.500	0.477	0.563	12.89
5) T Vinyl Chloride	0.723	0.554	0.562	0.557	0.528	0.481	0.462	0.552	15.33
6) T Bromomethane			0.336	0.338	0.317	0.260	0.263	0.303	12.78
7) Chloroethane			0.261	0.256	0.238	0.192	0.183	0.226	16.03
8) T Dichlorotetra...			1.175	1.265	1.230	1.040	1.036	1.149	9.26
9) T Propene			0.738	0.713	0.721	0.655	0.585	0.682	9.20
10) T Heptane			2.184	2.151	2.032	1.803	1.650	1.964	11.73
11) T Trichlorofluor...			1.510	1.459	1.477	1.242	1.265	1.391	9.11
12) T 1,1,2-Trichlor...			1.215	1.220	1.191	0.976	0.996	1.120	10.97
13) Ethanol			0.289	0.307	0.302	0.193	0.171	0.252	25.86
14) T Bromoethene			0.481	0.434	0.431	0.371	0.362	0.416	11.84
15) T Acetone			1.561	1.522	1.508	1.091	1.058	1.348	18.60
16) T 1,3-Butadiene			0.780	0.700	0.649	0.552	0.531	0.642	16.07
17) tert-Butyl alc...			2.010	2.047	1.934	1.689	1.571	1.850	11.31
18) T 1,1-Dichloroet...			0.519	0.539	0.512	0.437	0.434	0.488	10.07
19) T Isopropyl Alcohol			0.971	0.958	0.870	0.725	0.674	0.839	16.03
20) T Methylene Chlo...			0.738	0.612	0.508	0.406	0.400	0.533	26.98
21) T Allyl Chloride			1.097	1.049	1.040	0.910	0.869	0.993	9.86
22) T trans-1,2-Dich...			0.609	0.696	0.594	0.511	0.508	0.583	13.38
23) T Vinyl Acetate			2.148	1.904	2.292	1.623	1.905	1.974	13.03
24) T 1,1-Dichloroet...			1.169	1.237	1.190	1.031	1.009	1.127	8.99
25) T Ethyl Acetate			3.526	3.626	3.558	3.132	2.871	3.342	9.77
26) T Hexane			1.753	1.677	1.548	1.406	1.318	1.540	11.77
27) T Carbon Disulfide			1.423	1.540	1.505	1.315	1.272	1.411	8.26
28) T Methyl tert-Bu...			0.742	0.767	0.769	0.587	0.638	0.701	11.86
29) T Chloroform			2.437	2.397	2.296	2.030	1.997	2.232	9.22
30) T Cyclohexane			1.493	1.448	1.390	1.214	1.176	1.344	10.53
31) T cis-1,2-Dichlo...			1.491	1.594	1.516	1.405	1.348	1.471	6.55
32) T 1,1,1-Trichlor...	2.165	2.605	2.407	2.439	2.424	2.136	2.092	2.324	8.31

33) I 1,4-Difluorobenzene	-----ISTD-----								
34) T 2-Butanone			0.753	0.708	0.707	0.606	0.587	0.672	10.68
35) T Carbon Tetrach...	0.752	0.761	0.710	0.743	0.745	0.649	0.674	0.719	5.96
36) T Benzene			0.974	0.980	1.023	0.871	0.862	0.942	7.60
37) T 1,2-Dichloroet...			0.589	0.571	0.565	0.485	0.502	0.543	8.49
38) T Trichloroethene	0.460	0.421	0.401	0.402	0.398	0.342	0.330	0.393	11.38
39) T 1,2-Dichloropr...			0.359	0.370	0.371	0.317	0.305	0.344	8.99

Method Path : Z:\voasrv\HPCHEM1\MSVOA_L\methods\ Method File : VL101425AIR.M										
40) T	1,4-Dioxane			0.166	0.166	0.162	0.141	0.135	0.154	9.56
41) T	Tetrahydrofuran			0.427	0.394	0.384	0.343	0.332	0.376	10.32
42) T	Bromodichlorom...			0.779	0.775	0.791	0.697	0.713	0.751	5.68
43) T	Methyl Methacr...			0.500	0.511	0.493	0.450	0.437	0.478	6.85
44) T	2,2,4-Trimethy...			1.695	1.706	1.683	1.460	1.401	1.589	9.22
45) T	t-1,3-Dichloro...			0.478	0.502	0.506	0.462	0.453	0.480	4.94
46) T	cis-1,3-Dichlo...			0.593	0.590	0.603	0.546	0.535	0.573	5.34
47) T	1,1,2-Trichlor...			0.416	0.383	0.384	0.338	0.338	0.372	9.07
48) T	Dibromochlorom...			0.619	0.650	0.667	0.589	0.582	0.622	5.96
49) T	Bromoform			0.503	0.515	0.544	0.487	0.478	0.505	5.14
50) T	4-Methyl-2-Pen...			0.969	0.989	0.974	0.873	0.833	0.928	7.55
51) T	2-Hexanone			0.683	0.708	0.738	0.692	0.658	0.696	4.31
52) T	Tetrachloroethene	0.354	0.345	0.349	0.366	0.353	0.312	0.298	0.340	7.27
53) T	Toluene			1.212	1.162	1.176	1.036	1.003	1.118	8.28
54) T	1,2-Dibromoethane		0.662	0.570	0.586	0.617	0.533	0.526	0.582	8.84
-----ISTD-----										
55) I	Chlorobenzene-d5									
56) T	1,1,1,2-Tetrac...			0.662	0.664	0.665	0.590	0.570	0.630	7.32
57) T	Chlorobenzene			1.010	1.009	1.000	0.849	0.810	0.936	10.47
58) T	Ethyl Benzene			1.801	1.815	1.824	1.599	1.543	1.716	7.84
59) T	m/p-Xylene			1.463	1.457	1.476	1.238	1.210	1.369	9.71
60) T	o-Xylene			1.433	1.471	1.454	1.224	1.177	1.352	10.34
61) T	Styrene			0.597	0.649	0.664	0.578	0.564	0.610	7.17
62) T	Isopropylbenzene			2.564	2.671	2.616	2.241	2.123	2.443	10.02
63) T	1,1,2,2-Tetrac...	1.029	1.019	0.900	0.924	0.933	0.777	0.755	0.905	11.78
64) T	n-propylbenzene			0.688	0.710	0.696	0.589	0.551	0.647	11.10
65) T	tert-Butylbenzene			2.342	2.383	2.364	1.901	1.796	2.157	13.20
66) T	Benzyl Chloride			0.229	0.259	0.250	0.241	0.221	0.240	6.35
67) T	sec-Butylbenzene			3.313	3.378	3.357	2.697	2.565	3.062	12.96
68) S	1-Bromo-4-Fluo...	0.765	0.780	0.768	0.770	0.779	0.781	0.798	0.777	1.44
69) T	p-Isopropyltol...			2.726	2.896	2.871	2.330	2.186	2.602	12.47
70) T	n-Butylbenzene			2.611	2.820	2.849	2.321	2.218	2.564	11.16
71) T	2-Chlorotoluene			1.959	1.997	2.018	1.698	1.645	1.863	9.53
72) T	4-Ethyltoluene			1.741	1.761	1.805	1.530	1.475	1.662	8.97
73) T	1,3,5-Trimethy...			1.434	1.466	1.477	1.261	1.217	1.371	8.95
74) T	1,2,4-Trimethy...			1.691	1.706	1.664	1.331	1.278	1.534	13.75
75) T	1,3-Dichlorobe...			0.903	0.956	1.000	0.818	0.773	0.890	10.57
76) T	1,4-Dichlorobe...			0.957	0.957	0.971	0.823	0.779	0.897	9.96
77) T	1,2-Dichlorobe...			0.944	0.954	0.914	0.766	0.727	0.861	12.40
78) T	Hexachloro-1,3...			0.751	0.764	0.746	0.555	0.516	0.667	18.08
79) T	Naphthalene		1.265	1.532	1.766	1.971	1.580	1.507	1.604	15.04
80) T	Naphthalene,2-...			0.255	0.355	0.515	0.478	0.485	0.418	26.20
81) T	1,2,4-Trichlor...			0.553	0.726	0.803	0.638	0.608	0.666	14.90

(#) = Out of Range

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ALLI03</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3247</u>
Instrument ID:	<u>MSVOA_L</u>	Calibration Date/Time:	<u>10/15/2025 08:43</u>
Lab File ID:	<u>VL043114.D</u>	Init. Calib. Date(s):	<u>10/14/2025 10/14/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>10:31 14:18</u>
GC Column:	<u>RTX-1</u>	ID:	<u>0.32</u> (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Chloroethane	0.226	0.195		-13.72	30
Heptane	1.964	1.654		-15.78	30
Bromodichloromethane	0.751	0.731		-2.66	30
m/p-Xylene	1.369	1.221		-10.81	30
1-Bromo-4-Fluorobenzene	0.777	0.786		1.16	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.