

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3247

Analytical Method: SWTO-15

Client: Alliance Technical Group, LLC - New

Datafile : VL043116.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VL1015ABS01	Chloroethane	10	9.10	ppbv	91			70	130	
	Heptane	10	8.20	ppbv	82			70	130	
	Bromodichloromethane	10	9.90	ppbv	99			70	130	
	m/p-Xylene	20	18.3	ppbv	92			70	130	