

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q3247

Lab File ID: VL043103.D

BFB Injection Date: 10/14/2025

Instrument ID: MSVOA_L

BFB Injection Time: 08:42

GC Column: RTX-1 ID: 0.32 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	8.0 - 40.0% of mass 95	21.9
75	30.0 - 66.0% of mass 95	53.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.0 (0.0) 1
174	50.0 - 120.0% of mass 95	61.7
175	4.0 - 9.0% of mass 174	4.8 (7.8) 1
176	93.0 - 101.0% of mass 174	59.6 (96.5) 1
177	5.0 - 9.0% of mass 176	3.9 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICCC010	VSTDICCC010	VL043105.D	10/14/2025	10:31
VSTDICCC002	VSTDICCC002	VL043106.D	10/14/2025	11:27
VSTDICCC001	VSTDICCC001	VL043107.D	10/14/2025	12:01
VSTDICCC0.5	VSTDICCC0.5	VL043108.D	10/14/2025	12:34
VSTDICCC0.1	VSTDICCC0.1	VL043109.D	10/14/2025	13:08
VSTDICCC0.03	VSTDICCC0.03	VL043110.D	10/14/2025	13:42
VSTDICCC015	VSTDICCC015	VL043111.D	10/14/2025	14:18