

ANALYTICAL RESULTS SUMMARY

SEMI-VOLATILE ORGANICS
VOLATILE ORGANICS

PROJECT NAME : NWIRP BETHPAGE 112G08005-WE13

TETRA TECH NUS, INC.

661 Andersen Drive

Suite 200

Pittsburgh, PA - 15220-2745

Phone No: 412-921-7090

ORDER ID : Q3248

ATTENTION : Ernie Wu



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	4
2.1) VOC-TCLVOA-10- Case Narrative	4
2.2) SVOC-SIMGroup1- Case Narrative	6
3) Qualifier Page	8
4) QA Checklist	9
5) VOC-TCLVOA-10 Data	10
6) SVOC-SIMGroup1 Data	70
7) Shipping Document	99
7.1) CHAIN OF CUSTODY	100
7.2) Lab Certificate	101
7.3) Internal COC	102

1
2
3
4
5
6
7

Cover Page

Order ID : Q3248

Project ID : NWIRP Bethpage 112G08005-WE13

Client : Tetra Tech NUS, Inc.

Lab Sample Number

Q3248-01
Q3248-02
Q3248-03
Q3248-04
Q3248-05
Q3248-06
Q3248-07
Q3248-08
Q3248-09
Q3248-10

Client Sample Number

BP-VPB-184-TB-20250930
BP-VPB-184-GW-560-562
BP-VPB-184-GW-580-582
BP-VPB-184-GW-600-602
BP-VPB-184-GW-620-622
BP-VPB-184-GW-640-642
BP-VPB-184-GW-660-662
BP-VPB-184-GW-685-687
BP-VPB-184-GW-685-687MS
BP-VPB-184-GW-685-687MSD

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 10/15/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: NWIRP Bethpage 112G08005-WE13

Project Manager# Ernie Wu

Order ID # Q3248

Test Name: VOC-TCLVOA-10

A. Number of Samples and Date of Receipt:

10 Water samples were received on 10/02/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: VOC-TCLVOA-10. This data package contains results for VOC-TCLVOA-10.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI. The analysis of VOC-TCLVOA-10 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Internal Standards Areas were met for all analysis.

The Retention Times were met for all analysis.

The MS {Q3248-09MS} with File ID: VX048005.D recoveries met the requirements for all compounds except for 1,2,4-Trichlorobenzene[68%], 1,2-Dichlorobenzene[75%], 1,3-Dichlorobenzene[73%] and 1,4-Dichlorobenzene[73%] due to matrix interference.

The MSD recoveries met the requirements for all compounds.

The sample # BP-VPB-184-GW-685-687MS is failing for 1,2,4-Trichlorobenzene, 1,2-Dichlorobenzene, 1,3-Dichlorobenzene and 1,4-Dichlorobenzene therefore the original sample(BP-VPB-184-GW-685-687) is reported with M flag for this compounds.

The RPD for {Q3248-10MSD} with File ID: VX048006.D met criteria except for 1,1,2,2-Tetrachloroethane[29%], 1,2,3-Trichlorobenzene[32%], 1,2,4-Trichlorobenzene[31%], 1,2-Dibromo-3-Chloropropane[31%], 1,2-Dibromoethane[23%], 1,2-Dichlorobenzene[31%], 1,3-Dichlorobenzene[32%], 1,4-Dichlorobenzene[30%], Bromodichloromethane[21%], Bromoform[28%], Chlorobenzene[26%], cis-1,3-Dichloropropene[21%], Ethyl Benzene[26%], Isopropylbenzene[30%], m/p-Xylenes[30%], o-Xylene[27%], Styrene[27%], t-1,3-Dichloropropene[22%], Tetrachloroethene[24%] and Toluene[21%] due to difference in results of MS and MSD.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

The Initial Calibration met the requirements.
The Continuous Calibration met the requirements.
The Tuning criteria met requirements.

E. Additional Comments:

The not QT review data is reported in the Miscellaneous.

The laboratory certifies that the all-electronic diskette deliverable exactly match the data Summary forms (i.e. Form Is)."

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: NWIRP Bethpage 112G08005-WE13

Project Manager : Ernie Wu

Order ID # Q3248

Test Name: SVOC-SIMGroup1

A. Number of Samples and Date of Receipt:

10 Water samples were received on 10/02/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1. This data package contains results for SVOC-SIMGroup1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_N using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGA. The analysis of SVOC-SIMGroup1 was based on method 8270-Modified and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis except for, BP-VPB-184-GW-600-602 [Terphenyl-d14 - 189%], BP-VPB-184-GW-620-622 [Terphenyl-d14 - 149%] and BP-VPB-184-GW-660-662 [Terphenyl-d14 - 188%]. Failed surrogate is not associated with DOD, Therefore no further corrective action was taken.

The Internal Standards Areas were met for all analysis except for PB169992BSD. Failed internal standard is not associated with DOD, Therefore no further corrective action was taken.

The Retention Times were met for all analysis.

The RPD were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank Spike Duplicate met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration File ID BN038041.D met the requirements except for 2,4,6-Tribromophenol. Failed surrogate is not associated with DOD, Therefore no further corrective action was taken.



The Continuous Calibration File ID BN038066.D met the requirements except for Nitrobenzene-d5 and Phenol-d6. Failed surrogate is not associated with DOD, Therefore no further corrective action was taken.

The Tuning criteria met requirements.

E. Additional Comments:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is)."

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q3248

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 10/15/2025

LAB CHRONICLE

OrderID:	Q3248	OrderDate:	9/30/2025 11:27:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	D31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3248-01	BP-VPB-184-TB-2025 0930	Water			09/30/25			10/02/25
			VOC-TCLVOA-10	8260-Low			10/03/25	
Q3248-02	BP-VPB-184-GW-560- 562	Water			09/30/25			10/02/25
			VOC-TCLVOA-10	8260-Low			10/03/25	
Q3248-03	BP-VPB-184-GW-580- 582	Water			09/30/25			10/02/25
			VOC-TCLVOA-10	8260-Low			10/03/25	
Q3248-04	BP-VPB-184-GW-600- 602	Water			09/30/25			10/02/25
			VOC-TCLVOA-10	8260-Low			10/03/25	
Q3248-05	BP-VPB-184-GW-620- 622	Water			10/01/25			10/02/25
			VOC-TCLVOA-10	8260-Low			10/03/25	
Q3248-06	BP-VPB-184-GW-640- 642	Water			10/01/25			10/02/25
			VOC-TCLVOA-10	8260-Low			10/03/25	
Q3248-07	BP-VPB-184-GW-660- 662	Water			10/01/25			10/02/25
			VOC-TCLVOA-10	8260-Low			10/03/25	
Q3248-08	BP-VPB-184-GW-685- 687	Water			10/02/25			10/02/25
			VOC-TCLVOA-10	8260-Low			10/03/25	

Hit Summary Sheet
SW-846

SDG No.: Q3248
Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID: Q3248-01	BP-VPB-184-TB-20250930 BP-VPB-184-TB-20 Water	Acetone		1.70	J	1.50	3.80	5.00	ug/L
		Total Voc :		1.70					
		Total Concentration:		1.70					
Client ID: Q3248-02	BP-VPB-184-GW-560-562 BP-VPB-184-GW-5 Water	Acetone		9.50		1.50	3.80	5.00	ug/L
Q3248-02	BP-VPB-184-GW-5 Water	Methyl tert-butyl Ether		2.20		0.16	0.50	1.00	ug/L
		Total Voc :		11.7					
		Total Concentration:		11.7					
Client ID: Q3248-03	BP-VPB-184-GW-580-582 BP-VPB-184-GW-5 Water	Acetone		17.7		1.50	3.80	5.00	ug/L
Q3248-03	BP-VPB-184-GW-5 Water	Methyl tert-butyl Ether		5.10		0.16	0.50	1.00	ug/L
Q3248-03	BP-VPB-184-GW-5 Water	2-Butanone		4.10	J	0.98	2.50	5.00	ug/L
		Total Voc :		26.9					
		Total Concentration:		26.9					
Client ID: Q3248-04	BP-VPB-184-GW-600-602 BP-VPB-184-GW-6 Water	Acetone		7.00		1.50	3.80	5.00	ug/L
		Total Voc :		7.00					
		Total Concentration:		7.00					
Client ID: Q3248-05	BP-VPB-184-GW-620-622 BP-VPB-184-GW-6 Water	Acetone		8.80		1.50	3.80	5.00	ug/L
		Total Voc :		8.80					
		Total Concentration:		8.80					
Client ID: Q3248-06	BP-VPB-184-GW-640-642 BP-VPB-184-GW-6 Water	Acetone		10.6		1.50	3.80	5.00	ug/L
		Total Voc :		10.6					
		Total Concentration:		10.6					
Client ID: Q3248-07	BP-VPB-184-GW-660-662 BP-VPB-184-GW-6 Water	Acetone		4.90	J	1.50	3.80	5.00	ug/L
		Total Voc :		4.90					
		Total Concentration:		4.90					
Client ID: Q3248-08	BP-VPB-184-GW-685-687 BP-VPB-184-GW-6 Water	Acetone		5.50		1.50	3.80	5.00	ug/L
		Total Voc :		5.50					
		Total Concentration:		5.50					



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/30/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-TB-20250930		SDG No.:	Q3248	
Lab Sample ID:	Q3248-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047996.D	1	10/03/25 15:55	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	1.70	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/30/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-TB-20250930		SDG No.:	Q3248	
Lab Sample ID:	Q3248-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047996.D	1	10/03/25 15:55	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.1		81 - 118		104%	SPK: 50
1868-53-7	Dibromofluoromethane	49.7		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	48.2		89 - 112		96%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.3		85 - 114		109%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	130000	5.544				
540-36-3	1,4-Difluorobenzene	265000	6.751				
3114-55-4	Chlorobenzene-d5	276000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	139000	12.006				
TENTATIVE IDENTIFIED COMPOUNDS							

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/30/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25
Client Sample ID:	BP-VPB-184-TB-20250930		SDG No.:	Q3248
Lab Sample ID:	Q3248-01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047996.D	1	10/03/25 15:55	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/30/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-GW-560-562		SDG No.:	Q3248	
Lab Sample ID:	Q3248-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047997.D	1	10/03/25 16:15	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	9.50		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	2.20		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/30/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-GW-560-562		SDG No.:	Q3248	
Lab Sample ID:	Q3248-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047997.D	1	10/03/25 16:15	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.1		81 - 118		104%	SPK: 50
1868-53-7	Dibromofluoromethane	48.6		80 - 119		97%	SPK: 50
2037-26-5	Toluene-d8	48.2		89 - 112		96%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.7		85 - 114		109%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	131000	5.544				
540-36-3	1,4-Difluorobenzene	268000	6.751				
3114-55-4	Chlorobenzene-d5	274000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	141000	12.006				
TENTATIVE IDENTIFIED COMPOUNDS							

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/30/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-GW-560-562		SDG No.:	Q3248	
Lab Sample ID:	Q3248-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047997.D	1	10/03/25 16:15	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/30/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-GW-580-582		SDG No.:	Q3248	
Lab Sample ID:	Q3248-03		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047998.D	1	10/03/25 16:36	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	17.7		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	5.10		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	4.10	J	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/30/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-GW-580-582		SDG No.:	Q3248	
Lab Sample ID:	Q3248-03		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047998.D	1	10/03/25 16:36	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.3		81 - 118		107%	SPK: 50
1868-53-7	Dibromofluoromethane	50.2		80 - 119		100%	SPK: 50
2037-26-5	Toluene-d8	49.1		89 - 112		98%	SPK: 50
460-00-4	4-Bromofluorobenzene	56.2		85 - 114		112%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	117000	5.538				
540-36-3	1,4-Difluorobenzene	235000	6.751				
3114-55-4	Chlorobenzene-d5	247000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	125000	12.006				
TENTATIVE IDENTIFIED COMPOUNDS							

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/30/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-GW-580-582		SDG No.:	Q3248	
Lab Sample ID:	Q3248-03		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047998.D	1	10/03/25 16:36	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/30/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-GW-600-602		SDG No.:	Q3248	
Lab Sample ID:	Q3248-04		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047999.D	1	10/03/25 16:57	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	7.00		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/30/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-GW-600-602		SDG No.:	Q3248	
Lab Sample ID:	Q3248-04		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047999.D	1	10/03/25 16:57	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.0		81 - 118		106%	SPK: 50
1868-53-7	Dibromofluoromethane	49.2		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	48.7		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.8		85 - 114		108%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	122000	5.544				
540-36-3	1,4-Difluorobenzene	253000	6.751				
3114-55-4	Chlorobenzene-d5	259000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	128000	12.006				
TENTATIVE IDENTIFIED COMPOUNDS							

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/30/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-GW-600-602		SDG No.:	Q3248	
Lab Sample ID:	Q3248-04		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047999.D	1	10/03/25 16:57	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	10/01/25
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	10/02/25
Client Sample ID:	BP-VPB-184-GW-620-622			SDG No.:	Q3248
Lab Sample ID:	Q3248-05			Matrix:	Water
Analytical Method:	8260D			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID :	0.18	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX048000.D	1	10/03/25 17:18	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	8.80		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/01/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-GW-620-622		SDG No.:	Q3248	
Lab Sample ID:	Q3248-05		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX048000.D	1	10/03/25 17:18	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	51.4		81 - 118		103%	SPK: 50
1868-53-7	Dibromofluoromethane	49.0		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	48.4		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.2		85 - 114		108%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	114000	5.544				
540-36-3	1,4-Difluorobenzene	234000	6.751				
3114-55-4	Chlorobenzene-d5	240000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	120000	12.006				
TENTATIVE IDENTIFIED COMPOUNDS							

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/01/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25
Client Sample ID:	BP-VPB-184-GW-620-622		SDG No.:	Q3248
Lab Sample ID:	Q3248-05		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX048000.D	1	10/03/25 17:18	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/01/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-GW-640-642		SDG No.:	Q3248	
Lab Sample ID:	Q3248-06		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX048001.D	1	10/03/25 17:38	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	10.6		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/01/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-GW-640-642		SDG No.:	Q3248	
Lab Sample ID:	Q3248-06		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX048001.D	1	10/03/25 17:38	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.1		81 - 118		100%	SPK: 50
1868-53-7	Dibromofluoromethane	49.7		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	48.3		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.6		85 - 114		107%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	115000	5.544				
540-36-3	1,4-Difluorobenzene	230000	6.751				
3114-55-4	Chlorobenzene-d5	237000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	118000	12.006				
TENTATIVE IDENTIFIED COMPOUNDS							

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/01/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25
Client Sample ID:	BP-VPB-184-GW-640-642		SDG No.:	Q3248
Lab Sample ID:	Q3248-06		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX048001.D	1	10/03/25 17:38	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	10/01/25	
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-GW-660-662			SDG No.:	Q3248	
Lab Sample ID:	Q3248-07			Matrix:	Water	
Analytical Method:	8260D			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX048002.D	1	10/03/25 17:59	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	4.90	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/01/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-GW-660-662		SDG No.:	Q3248	
Lab Sample ID:	Q3248-07		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX048002.D	1	10/03/25 17:59	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.5		81 - 118		101%	SPK: 50
1868-53-7	Dibromofluoromethane	49.2		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	48.5		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.0		85 - 114		108%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	117000	5.544				
540-36-3	1,4-Difluorobenzene	235000	6.751				
3114-55-4	Chlorobenzene-d5	245000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	118000	12.006				
TENTATIVE IDENTIFIED COMPOUNDS							

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/01/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25
Client Sample ID:	BP-VPB-184-GW-660-662		SDG No.:	Q3248
Lab Sample ID:	Q3248-07		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX048002.D	1	10/03/25 17:59	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/02/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-GW-685-687		SDG No.:	Q3248	
Lab Sample ID:	Q3248-08		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX048004.D	1	10/03/25 18:41	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	5.50		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	10/02/25	
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-GW-685-687			SDG No.:	Q3248	
Lab Sample ID:	Q3248-08			Matrix:	Water	
Analytical Method:	8260D			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX048004.D	1	10/03/25 18:41	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	UM	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	UM	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	UM	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	UM	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	49.1		81 - 118		98%	SPK: 50
1868-53-7	Dibromofluoromethane	47.9		80 - 119		96%	SPK: 50
2037-26-5	Toluene-d8	47.3		89 - 112		95%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.3		85 - 114		105%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	116000	5.544				
540-36-3	1,4-Difluorobenzene	233000	6.751				
3114-55-4	Chlorobenzene-d5	238000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	114000	12.006				
TENTATIVE IDENTIFIED COMPOUNDS							

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/02/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25
Client Sample ID:	BP-VPB-184-GW-685-687		SDG No.:	Q3248
Lab Sample ID:	Q3248-08		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX048004.D	1	10/03/25 18:41	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
75-43-4	Dichlorofluoromethane	N.D					

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: **Q3248**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3248-01	BP-VPB-184-TB-20250930	1,2-Dichloroethane-d4	50	52.1	104		81	118
		Dibromofluoromethane	50	49.7	99		80	119
		Toluene-d8	50	48.2	96		89	112
		4-Bromofluorobenzene	50	54.3	109		85	114
Q3248-02	BP-VPB-184-GW-560-562	1,2-Dichloroethane-d4	50	52.1	104		81	118
		Dibromofluoromethane	50	48.6	97		80	119
		Toluene-d8	50	48.2	96		89	112
		4-Bromofluorobenzene	50	54.7	109		85	114
Q3248-03	BP-VPB-184-GW-580-582	1,2-Dichloroethane-d4	50	53.3	107		81	118
		Dibromofluoromethane	50	50.2	100		80	119
		Toluene-d8	50	49.1	98		89	112
		4-Bromofluorobenzene	50	56.2	112		85	114
Q3248-04	BP-VPB-184-GW-600-602	1,2-Dichloroethane-d4	50	53.0	106		81	118
		Dibromofluoromethane	50	49.2	98		80	119
		Toluene-d8	50	48.7	97		89	112
		4-Bromofluorobenzene	50	53.8	108		85	114
Q3248-05	BP-VPB-184-GW-620-622	1,2-Dichloroethane-d4	50	51.4	103		81	118
		Dibromofluoromethane	50	49.0	98		80	119
		Toluene-d8	50	48.4	97		89	112
		4-Bromofluorobenzene	50	54.2	108		85	114
Q3248-06	BP-VPB-184-GW-640-642	1,2-Dichloroethane-d4	50	50.1	100		81	118
		Dibromofluoromethane	50	49.7	99		80	119
		Toluene-d8	50	48.4	97		89	112
		4-Bromofluorobenzene	50	53.6	107		85	114
Q3248-07	BP-VPB-184-GW-660-662	1,2-Dichloroethane-d4	50	50.5	101		81	118
		Dibromofluoromethane	50	49.2	98		80	119
		Toluene-d8	50	48.5	97		89	112
		4-Bromofluorobenzene	50	54.0	108		85	114
Q3248-08	BP-VPB-184-GW-685-687	1,2-Dichloroethane-d4	50	49.1	98		81	118
		Dibromofluoromethane	50	47.9	96		80	119
		Toluene-d8	50	47.3	95		89	112
		4-Bromofluorobenzene	50	52.3	105		85	114
Q3248-09MS	BP-VPB-184-GW-685-687MS	1,2-Dichloroethane-d4	50	57.1	114		81	118
		Dibromofluoromethane	50	54.6	109		80	119
		Toluene-d8	50	52.9	106		89	112
		4-Bromofluorobenzene	50	56.5	113		85	114
Q3248-10MSD	BP-VPB-184-GW-685-687MSD	1,2-Dichloroethane-d4	50	53.5	107		81	118
		Dibromofluoromethane	50	54.1	108		80	119
		Toluene-d8	50	53.2	106		89	112
		4-Bromofluorobenzene	50	52.8	106		85	114
VX1003WBL01	VX1003WBL01	1,2-Dichloroethane-d4	50	56.6	113		81	118
		Dibromofluoromethane	50	51.0	102		80	119
		Toluene-d8	50	49.4	99		89	112
		4-Bromofluorobenzene	50	57.0	114		85	114
VX1003WBS02	VX1003WBS02	1,2-Dichloroethane-d4	50	53.6	107		81	118
		Dibromofluoromethane	50	55.1	110		80	119
		Toluene-d8	50	53.6	107		89	112
		4-Bromofluorobenzene	50	54.0	108		85	114

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3248

Analytical Method:

SW8260-Low

Client: Tetra Tech NUS, Inc.

Datafile :

VX048005.D

		Sample				Rec	RPD	Limits			
Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID :	Q3248-09MS	Client Sample ID :	BP-VPB-184-GW-685-687MS								
Dichlorodifluoromethane	50	0	43.9	ug/L	88				32	152	
Chloromethane	50	0	41.6	ug/L	83				50	139	
Vinyl chloride	50	0	44.2	ug/L	88				58	137	
Bromomethane	50	0	44.6	ug/L	89				53	141	
Chloroethane	50	0	43.5	ug/L	87				60	138	
Trichlorofluoromethane	50	0	45.3	ug/L	91				65	141	
1,1,2-Trichlorotrifluoroethane	50	0	44.7	ug/L	89				70	136	
1,1-Dichloroethene	50	0	44.5	ug/L	89				71	131	
Acetone	250	5.50	200	ug/L	78				39	160	
Carbon disulfide	50	0	39.0	ug/L	78				64	133	
Methyl tert-butyl Ether	50	0	46.1	ug/L	92				71	124	
Methyl Acetate	50	0	50.7	ug/L	101				56	136	
Methylene Chloride	50	0	45.6	ug/L	91				74	124	
trans-1,2-Dichloroethene	50	0	45.2	ug/L	90				75	124	
1,1-Dichloroethane	50	0	46.6	ug/L	93				77	125	
Cyclohexane	50	0	39.7	ug/L	79				71	130	
2-Butanone	250	0	240	ug/L	96				56	143	
Carbon Tetrachloride	50	0	42.2	ug/L	84				72	136	
cis-1,2-Dichloroethene	50	0	46.6	ug/L	93				78	123	
Bromochloromethane	50	0	51.5	ug/L	103				78	123	
Chloroform	50	0	48.2	ug/L	96				79	124	
1,1,1-Trichloroethane	50	0	46.2	ug/L	92				74	131	
Methylcyclohexane	50	0	37.6	ug/L	75				72	132	
Benzene	50	0	43.5	ug/L	87				79	120	
1,2-Dichloroethane	50	0	44.2	ug/L	88				73	128	
Trichloroethene	50	0	42.9	ug/L	86				79	123	
1,2-Dichloropropane	50	0	45.6	ug/L	91				78	122	
Bromodichloromethane	50	0	45.7	ug/L	91				79	125	
4-Methyl-2-Pentanone	250	0	250	ug/L	100				67	130	
Toluene	50	0	43.6	ug/L	87				80	121	
t-1,3-Dichloropropene	50	0	42.8	ug/L	86				73	127	
cis-1,3-Dichloropropene	50	0	42.4	ug/L	85				75	124	
1,1,2-Trichloroethane	50	0	47.4	ug/L	95				80	119	
2-Hexanone	250	0	240	ug/L	96				57	139	
Dibromochloromethane	50	0	48.1	ug/L	96				74	126	
1,2-Dibromoethane	50	0	46.6	ug/L	93				77	121	
Tetrachloroethene	50	0	37.7	ug/L	75				74	129	
Chlorobenzene	50	0	41.0	ug/L	82				82	118	
Ethyl Benzene	50	0	40.0	ug/L	80				79	121	
m/p-Xylenes	100	0	80.9	ug/L	81				80	121	
o-Xylene	50	0	40.9	ug/L	82				78	122	
Styrene	50	0	40.0	ug/L	80				78	123	
Bromoform	50	0	43.6	ug/L	87				66	130	
Isopropylbenzene	50	0	36.8	ug/L	74				72	131	
1,1,2,2-Tetrachloroethane	50	0	41.7	ug/L	83				71	121	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3248

Analytical Method: SW8260-Low

Client: Tetra Tech NUS, Inc.

Datafile : VX048005.D

Parameter	Spike	Sample	Result	Units	Rec			RPD	Limits		
		Result			Rec	Qual	RPD	Qual	Low	High	RPD
1,3-Dichlorobenzene	50	0	36.6	ug/L	73	*			80	119	
1,4-Dichlorobenzene	50	0	36.4	ug/L	73	*			79	118	
1,2-Dichlorobenzene	50	0	37.6	ug/L	75	*			80	119	
1,2-Dibromo-3-Chloropropane	50	0	39.0	ug/L	78				62	128	
1,2,4-Trichlorobenzene	50	0	34.2	ug/L	68	*			69	130	
1,2,3-Trichlorobenzene	50	0	36.2	ug/L	72				69	129	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3248

Analytical Method: SW8260-Low

Client: Tetra Tech NUS, Inc.

Datafile : VX048006.D

Parameter	Spike	Sample	Result	Units	Rec		RPD		Limits		RPD
		Result			Qual	RPD	Qual	Low	High		
Lab Sample ID :	Q3248-10MSD	Client Sample ID :	BP-VPB-184-GW-685-687MSD								
Dichlorodifluoromethane	50	0	49.1	ug/L	98		11		32	152	20
Chloromethane	50	0	48.9	ug/L	98		16		50	139	20
Vinyl chloride	50	0	50.6	ug/L	101		14		58	137	20
Bromomethane	50	0	52.4	ug/L	105		16		53	141	20
Chloroethane	50	0	51.5	ug/L	103		17		60	138	20
Trichlorofluoromethane	50	0	52.0	ug/L	104		14		65	141	20
1,1,2-Trichlorotrifluoroethane	50	0	50.3	ug/L	101		12		70	136	20
1,1-Dichloroethene	50	0	52.2	ug/L	104		16		71	131	20
Acetone	250	5.50	230	ug/L	90		14		39	160	20
Carbon disulfide	50	0	45.6	ug/L	91		16		64	133	20
Methyl tert-butyl Ether	50	0	55.0	ug/L	110		18		71	124	20
Methyl Acetate	50	0	60.0	ug/L	120		17		56	136	20
Methylene Chloride	50	0	53.0	ug/L	106		15		74	124	20
trans-1,2-Dichloroethene	50	0	52.4	ug/L	105		15		75	124	20
1,1-Dichloroethane	50	0	55.3	ug/L	111		17		77	125	20
Cyclohexane	50	0	46.0	ug/L	92		15		71	130	20
2-Butanone	250	0	280	ug/L	112		15		56	143	20
Carbon Tetrachloride	50	0	51.6	ug/L	103		20		72	136	20
cis-1,2-Dichloroethene	50	0	54.5	ug/L	109		16		78	123	20
Bromochloromethane	50	0	54.0	ug/L	108		5		78	123	20
Chloroform	50	0	56.7	ug/L	113		16		79	124	20
1,1,1-Trichloroethane	50	0	55.1	ug/L	110		18		74	131	20
Methylcyclohexane	50	0	43.8	ug/L	88		15		72	132	20
Benzene	50	0	53.3	ug/L	107		20		79	120	20
1,2-Dichloroethane	50	0	53.7	ug/L	107		19		73	128	20
Trichloroethene	50	0	52.4	ug/L	105		20		79	123	20
1,2-Dichloropropane	50	0	55.8	ug/L	112		20		78	122	20
Bromodichloromethane	50	0	56.2	ug/L	112		21	*	79	125	20
4-Methyl-2-Pentanone	250	0	300	ug/L	120		18		67	130	20
Toluene	50	0	53.7	ug/L	107		21	*	80	121	20
t-1,3-Dichloropropene	50	0	53.4	ug/L	107		22	*	73	127	20
cis-1,3-Dichloropropene	50	0	52.3	ug/L	105		21	*	75	124	20
1,1,2-Trichloroethane	50	0	57.7	ug/L	115		20		80	119	20
2-Hexanone	250	0	290	ug/L	116		19		57	139	20
Dibromochloromethane	50	0	58.8	ug/L	118		20		74	126	20
1,2-Dibromoethane	50	0	58.7	ug/L	117		23	*	77	121	20
Tetrachloroethene	50	0	48.1	ug/L	96		24	*	74	129	20
Chlorobenzene	50	0	53.5	ug/L	107		26	*	82	118	20
Ethyl Benzene	50	0	52.2	ug/L	104		26	*	79	121	20
m/p-Xylenes	100	0	110	ug/L	110		30	*	80	121	20
o-Xylene	50	0	53.4	ug/L	107		27	*	78	122	20
Styrene	50	0	52.4	ug/L	105		27	*	78	123	20
Bromoform	50	0	57.8	ug/L	116		28	*	66	130	20
Isopropylbenzene	50	0	50.0	ug/L	100		30	*	72	131	20
1,1,2,2-Tetrachloroethane	50	0	56.0	ug/L	112		29	*	71	121	20

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3248

Analytical Method: SW8260-Low

Client: Tetra Tech NUS, Inc.

Datafile : VX048006.D

Parameter	Spike	Sample Result	Result	Units	Rec		RPD		Limits		
					Rec	Qual	RPD	Qual	Low	High	RPD
1,3-Dichlorobenzene	50	0	50.3	ug/L	101		32	*	80	119	20
1,4-Dichlorobenzene	50	0	49.3	ug/L	99		30	*	79	118	20
1,2-Dichlorobenzene	50	0	51.6	ug/L	103		31	*	80	119	20
1,2-Dibromo-3-Chloropropane	50	0	53.3	ug/L	107		31	*	62	128	20
1,2,4-Trichlorobenzene	50	0	46.8	ug/L	94		31	*	69	130	20
1,2,3-Trichlorobenzene	50	0	50.2	ug/L	100		32	*	69	129	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3248</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>VX047986.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1003WBS02	Dichlorodifluoromethane	20	18.7	ug/L	94			32	152	
	Chloromethane	20	17.0	ug/L	85			50	139	
	Vinyl chloride	20	18.1	ug/L	91			58	137	
	Bromomethane	20	20.3	ug/L	102			53	141	
	Chloroethane	20	18.2	ug/L	91			60	138	
	Trichlorofluoromethane	20	18.9	ug/L	95			65	141	
	1,1,2-Trichlorotrifluoroethane	20	19.7	ug/L	99			70	136	
	1,1-Dichloroethene	20	18.2	ug/L	91			71	131	
	Acetone	100	75.3	ug/L	75			39	160	
	Carbon disulfide	20	16.4	ug/L	82			64	133	
	Methyl tert-butyl Ether	20	19.1	ug/L	96			71	124	
	Methyl Acetate	20	19.9	ug/L	100			56	136	
	Methylene Chloride	20	19.6	ug/L	98			74	124	
	trans-1,2-Dichloroethene	20	18.5	ug/L	93			75	124	
	1,1-Dichloroethane	20	19.7	ug/L	99			77	125	
	Cyclohexane	20	17.0	ug/L	85			71	130	
	2-Butanone	100	85.5	ug/L	86			56	143	
	Carbon Tetrachloride	20	19.0	ug/L	95			72	136	
	cis-1,2-Dichloroethene	20	19.3	ug/L	97			78	123	
	Bromochloromethane	20	20.3	ug/L	102			78	123	
	Chloroform	20	20.4	ug/L	102			79	124	
	1,1,1-Trichloroethane	20	19.8	ug/L	99			74	131	
	Methylcyclohexane	20	17.7	ug/L	89			72	132	
	Benzene	20	19.6	ug/L	98			79	120	
	1,2-Dichloroethane	20	20.3	ug/L	102			73	128	
	Trichloroethene	20	19.2	ug/L	96			79	123	
	1,2-Dichloropropane	20	20.4	ug/L	102			78	122	
	Bromodichloromethane	20	20.9	ug/L	104			79	125	
	4-Methyl-2-Pentanone	100	96.3	ug/L	96			67	130	
	Toluene	20	19.4	ug/L	97			80	121	
	t-1,3-Dichloropropene	20	19.3	ug/L	97			73	127	
	cis-1,3-Dichloropropene	20	19.6	ug/L	98			75	124	
	1,1,2-Trichloroethane	20	20.9	ug/L	104			80	119	
	2-Hexanone	100	88.5	ug/L	89			57	139	
	Dibromochloromethane	20	21.2	ug/L	106			74	126	
	1,2-Dibromoethane	20	20.7	ug/L	104			77	121	
	Tetrachloroethene	20	18.7	ug/L	94			74	129	
	Chlorobenzene	20	19.4	ug/L	97			82	118	
	Ethyl Benzene	20	18.7	ug/L	94			79	121	
	m/p-Xylenes	40	37.8	ug/L	95			80	121	
	o-Xylene	20	19.1	ug/L	96			78	122	
	Styrene	20	19.0	ug/L	95			78	123	
	Bromoform	20	19.9	ug/L	100			66	130	
	Isopropylbenzene	20	18.3	ug/L	92			72	131	
	1,1,2,2-Tetrachloroethane	20	19.7	ug/L	99			71	121	
	1,3-Dichlorobenzene	20	18.9	ug/L	95			80	119	
	1,4-Dichlorobenzene	20	18.8	ug/L	94			79	118	
	1,2-Dichlorobenzene	20	18.9	ug/L	95			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3248</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>VX047986.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX1003WBS02	1,2-Dibromo-3-Chloropropane	20	16.4	ug/L	82			62	128	
	1,2,4-Trichlorobenzene	20	17.5	ug/L	88			69	130	
	1,2,3-Trichlorobenzene	20	17.7	ug/L	89			69	129	

VOLATILE METHOD BLANK SUMMARY

Client ID

VX1003WBL01

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q3248

Lab File ID: VX047983.D

Lab Sample ID: VX1003WBL01

Date Analyzed: 10/03/2025

Time Analyzed: 10:16

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX1003WBS02	VX1003WBS02	VX047986.D	10/03/2025
BP-VPB-184-TB-20250930	Q3248-01	VX047996.D	10/03/2025
BP-VPB-184-GW-560-562	Q3248-02	VX047997.D	10/03/2025
BP-VPB-184-GW-580-582	Q3248-03	VX047998.D	10/03/2025
BP-VPB-184-GW-600-602	Q3248-04	VX047999.D	10/03/2025
BP-VPB-184-GW-620-622	Q3248-05	VX048000.D	10/03/2025
BP-VPB-184-GW-640-642	Q3248-06	VX048001.D	10/03/2025
BP-VPB-184-GW-660-662	Q3248-07	VX048002.D	10/03/2025
BP-VPB-184-GW-685-687	Q3248-08	VX048004.D	10/03/2025
BP-VPB-184-GW-685-687MS	Q3248-09MS	VX048005.D	10/03/2025
BP-VPB-184-GW-685-687MSD	Q3248-10MSD	VX048006.D	10/03/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: VX047584.D
Instrument ID: MSVOA_X
GC Column: DB-624UI ID: 0.18 (mm)

Contract: TETR06
SDG NO.: Q3248
BFB Injection Date: 09/16/2025
BFB Injection Time: 08:56
Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20
75	30.0 - 60.0% of mass 95	54
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	67.3
175	5.0 - 9.0% of mass 174	5.4 (8.1) 1
176	95.0 - 101.0% of mass 174	66.3 (98.6) 1
177	5.0 - 9.0% of mass 176	4.3 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX047585.D	09/16/2025	09:23
VSTDIC005	VSTDIC005	VX047586.D	09/16/2025	10:16
VSTDIC020	VSTDIC020	VX047587.D	09/16/2025	10:37
VSTDIC050	VSTDIC050	VX047588.D	09/16/2025	10:58
VSTDIC100	VSTDIC100	VX047589.D	09/16/2025	11:40
VSTDIC150	VSTDIC150	VX047590.D	09/16/2025	12:01

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q3248

Lab File ID: VX047980.D

BFB Injection Date: 10/03/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:35

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.8
75	30.0 - 60.0% of mass 95	53.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.9 (1.3) 1
174	50.0 - 100.0% of mass 95	67.2
175	5.0 - 9.0% of mass 174	5.2 (7.7) 1
176	95.0 - 101.0% of mass 174	64.1 (95.4) 1
177	5.0 - 9.0% of mass 176	4.1 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX047981.D	10/03/2025	09:35
VX1003WBL01	VX1003WBL01	VX047983.D	10/03/2025	10:16
VX1003WBS02	VX1003WBS02	VX047986.D	10/03/2025	12:22
BP-VPB-184-TB-20250930	Q3248-01	VX047996.D	10/03/2025	15:55
BP-VPB-184-GW-560-562	Q3248-02	VX047997.D	10/03/2025	16:15
BP-VPB-184-GW-580-582	Q3248-03	VX047998.D	10/03/2025	16:36
BP-VPB-184-GW-600-602	Q3248-04	VX047999.D	10/03/2025	16:57
BP-VPB-184-GW-620-622	Q3248-05	VX048000.D	10/03/2025	17:18
BP-VPB-184-GW-640-642	Q3248-06	VX048001.D	10/03/2025	17:38
BP-VPB-184-GW-660-662	Q3248-07	VX048002.D	10/03/2025	17:59
BP-VPB-184-GW-685-687	Q3248-08	VX048004.D	10/03/2025	18:41
BP-VPB-184-GW-685-687MS	Q3248-09MS	VX048005.D	10/03/2025	19:02
BP-VPB-184-GW-685-687MSD	Q3248-10MSD	VX048006.D	10/03/2025	19:22
VSTDCCC050EC	VSTDCCC050	VX048007.D	10/03/2025	19:43

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06
Lab Code: ACE SDG NO.: Q3248
Lab File ID: VX047981.D Date Analyzed: 10/03/2025
Instrument ID: MSVOA_X Time Analyzed: 09:35
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	131937	5.53	225878	6.74	209083	10.04
UPPER LIMIT	263874	6.031	451756	7.239	418166	10.537
LOWER LIMIT	65968.5	5.031	112939	6.239	104542	9.537
EPA SAMPLE NO.						
BP-VPB-184-TB-20250930	130029	5.54	264754	6.75	276284	10.04
BP-VPB-184-GW-560-562	130796	5.54	267996	6.75	274276	10.04
BP-VPB-184-GW-580-582	116663	5.54	235290	6.75	247287	10.04
BP-VPB-184-GW-600-602	121622	5.54	253461	6.75	259302	10.04
BP-VPB-184-GW-620-622	113992	5.54	234307	6.75	240205	10.04
BP-VPB-184-GW-640-642	114921	5.54	230401	6.75	237241	10.04
BP-VPB-184-GW-660-662	117443	5.54	235229	6.75	244995	10.04
BP-VPB-184-GW-685-687	115930	5.54	233399	6.75	238258	10.04
BP-VPB-184-GW-685-687MS	137909	5.54	257097	6.75	246962	10.04
BP-VPB-184-GW-685-687MSD	118835	5.54	215532	6.75	195286	10.04
VX1003WBL01	132031	5.54	278145	6.75	290715	10.04
VX1003WBS02	126673	5.54	223242	6.75	203303	10.04

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06
Lab Code: ACE SDG NO.: Q3248
Lab File ID: VX047981.D Date Analyzed: 10/03/2025
Instrument ID: MSVOA_X Time Analyzed: 09:35
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	105470	12.00€				
UPPER LIMIT	210940	12.50€				
LOWER LIMIT	52735	11.50€				
EPA SAMPLE NO.						
BP-VPB-184-TB-20250930	139079	12.01				
BP-VPB-184-GW-560-562	141091	12.01				
BP-VPB-184-GW-580-582	124667	12.01				
BP-VPB-184-GW-600-602	127779	12.01				
BP-VPB-184-GW-620-622	119987	12.01				
BP-VPB-184-GW-640-642	118357	12.01				
BP-VPB-184-GW-660-662	118155	12.01				
BP-VPB-184-GW-685-687	114439	12.01				
BP-VPB-184-GW-685-687MS	129827	12.01				
BP-VPB-184-GW-685-687MSD	99502	12.01				
VX1003WBL01	147497	12.01				
VX1003WBS02	101886	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX1003WBL01		SDG No.:	Q3248
Lab Sample ID:	VX1003WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047983.D	1	10/03/25 10:16	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX1003WBL01		SDG No.:	Q3248
Lab Sample ID:	VX1003WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047983.D	1	10/03/25 10:16	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	56.6		81 - 118		113%	SPK: 50
1868-53-7	Dibromofluoromethane	51.0		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	49.4		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	57.0		85 - 114		114%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	132000	5.544				
540-36-3	1,4-Difluorobenzene	278000	6.751				
3114-55-4	Chlorobenzene-d5	291000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	147000	12.006				

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX1003WBL01		SDG No.:	Q3248
Lab Sample ID:	VX1003WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047983.D	1	10/03/25 10:16	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX1003WBS02		SDG No.:	Q3248
Lab Sample ID:	VX1003WBS02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047986.D	1	10/03/25 12:22	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	18.7		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	17.0		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	18.1		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	20.3		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	18.2		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	18.9		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.7		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.2		0.23	0.75	1.00	ug/L
67-64-1	Acetone	75.3		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	16.4		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.1		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	19.9		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	19.6		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.5		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.7		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	17.0		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	85.5		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.0		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.3		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	20.3		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	20.4		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.8		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	17.7		0.16	0.50	1.00	ug/L
71-43-2	Benzene	19.6		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.3		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.2		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	20.4		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	20.9		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	96.3		0.68	2.50	5.00	ug/L
108-88-3	Toluene	19.4		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX1003WBS02		SDG No.:	Q3248
Lab Sample ID:	VX1003WBS02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047986.D	1	10/03/25 12:22	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	19.3		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.6		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.9		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	88.5		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	21.2		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	20.7		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	18.7		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.4		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	18.7		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	37.8		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	19.1		0.12	0.50	1.00	ug/L
100-42-5	Styrene	19.0		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	19.9		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	18.3		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	19.7		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	18.9		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	18.8		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	18.9		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	16.4		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	17.5		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	17.7		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.6		81 - 118		107%	SPK: 50
1868-53-7	Dibromofluoromethane	55.1		80 - 119		110%	SPK: 50
2037-26-5	Toluene-d8	53.6		89 - 112		107%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.0		85 - 114		108%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	127000	5.538				
540-36-3	1,4-Difluorobenzene	223000	6.745				
3114-55-4	Chlorobenzene-d5	203000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	102000	12.006				

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	VX1003WBS02		SDG No.:	Q3248
Lab Sample ID:	VX1003WBS02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX047986.D	1	10/03/25 12:22	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	10/02/25	
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-GW-685-687MS			SDG No.:	Q3248	
Lab Sample ID:	Q3248-09MS			Matrix:	Water	
Analytical Method:	8260D			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX048005.D	1	10/03/25 19:02	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	43.9		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	41.6		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	44.2		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	44.6		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	43.5		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	45.3		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	44.7		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	44.5		0.23	0.75	1.00	ug/L
67-64-1	Acetone	200		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	39.0		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	46.1		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	50.7		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	45.6		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	45.2		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	46.6		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	39.7		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	240		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	42.2		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	46.6		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	51.5		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	48.2		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	46.2		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	37.6		0.16	0.50	1.00	ug/L
71-43-2	Benzene	43.5		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	44.2		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	42.9		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	45.6		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	45.7		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	250		0.68	2.50	5.00	ug/L
108-88-3	Toluene	43.6		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/02/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-GW-685-687MS		SDG No.:	Q3248	
Lab Sample ID:	Q3248-09MS		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX048005.D	1	10/03/25 19:02	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	42.8		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	42.4		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	47.4		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	240		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	48.1		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	46.6		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	37.7		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	41.0		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	40.0		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	80.9		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	40.9		0.12	0.50	1.00	ug/L
100-42-5	Styrene	40.0		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	43.6		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	36.8		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	41.7		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	36.6		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	36.4		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	37.6		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	39.0		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	34.2		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	36.2		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	57.1		81 - 118		114%	SPK: 50
1868-53-7	Dibromofluoromethane	54.6		80 - 119		109%	SPK: 50
2037-26-5	Toluene-d8	52.9		89 - 112		106%	SPK: 50
460-00-4	4-Bromofluorobenzene	56.5		85 - 114		113%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	138000	5.544				
540-36-3	1,4-Difluorobenzene	257000	6.751				
3114-55-4	Chlorobenzene-d5	247000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	130000	12.006				

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/02/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25
Client Sample ID:	BP-VPB-184-GW-685-687MS		SDG No.:	Q3248
Lab Sample ID:	Q3248-09MS		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX048005.D	1	10/03/25 19:02	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.			Date Collected:	10/02/25
Project:	NWIRP Bethpage 112G08005-WE13			Date Received:	10/02/25
Client Sample ID:	BP-VPB-184-GW-685-687MSD			SDG No.:	Q3248
Lab Sample ID:	Q3248-10MSD			Matrix:	Water
Analytical Method:	8260D			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID :	0.18	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX048006.D	1	10/03/25 19:22	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	49.1		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	48.9		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	50.6		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	52.4		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	51.5		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	52.0		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	50.3		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	52.2		0.23	0.75	1.00	ug/L
67-64-1	Acetone	230		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	45.6		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	55.0		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	60.0		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	53.0		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	52.4		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	55.3		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	46.0		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	280		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	51.6		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	54.5		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	54.0		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	56.7		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	55.1		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	43.8		0.16	0.50	1.00	ug/L
71-43-2	Benzene	53.3		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	53.7		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	52.4		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	55.8		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	56.2		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	300		0.68	2.50	5.00	ug/L
108-88-3	Toluene	53.7		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/02/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25	
Client Sample ID:	BP-VPB-184-GW-685-687MSD		SDG No.:	Q3248	
Lab Sample ID:	Q3248-10MSD		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX048006.D	1	10/03/25 19:22	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	53.4		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	52.3		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	57.7		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	290		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	58.8		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	58.7		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	48.1		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	53.5		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	52.2		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	110		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	53.4		0.12	0.50	1.00	ug/L
100-42-5	Styrene	52.4		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	57.8		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	50.0		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	56.0		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	50.3		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	49.3		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	51.6		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	53.3		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	46.8		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	50.2		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.5		81 - 118		107%	SPK: 50
1868-53-7	Dibromofluoromethane	54.1		80 - 119		108%	SPK: 50
2037-26-5	Toluene-d8	53.2		89 - 112		106%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.8		85 - 114		106%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	119000	5.544				
540-36-3	1,4-Difluorobenzene	216000	6.751				
3114-55-4	Chlorobenzene-d5	195000	10.037				
3855-82-1	1,4-Dichlorobenzene-d4	99500	12.006				

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/02/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25
Client Sample ID:	BP-VPB-184-GW-685-687MSD		SDG No.:	Q3248
Lab Sample ID:	Q3248-10MSD		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX048006.D	1	10/03/25 19:22	VX100325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: TETR06
Lab Code: ACE SDG No.: Q3248
Instrument ID: MSVOA_X Calibration Date(s): 09/16/2025 09/16/2025
Heated Purge: (Y/N) N Calibration Time(s): 09:23 12:01
GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:								
RRF001 = VX047585.D			RRF005 = VX047586.D			RRF020 = VX047587.D		
RRF050 = VX047588.D			RRF100 = VX047589.D			RRF150 = VX047590.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Dichlorodifluoromethane	0.447	0.575	0.582	0.501	0.510	0.492	0.518	10
Chloromethane	0.610	0.749	0.739	0.661	0.677	0.647	0.680	7.9
Vinyl Chloride	0.572	0.713	0.716	0.661	0.677	0.658	0.666	7.9
Bromomethane		0.493	0.451	0.424	0.423	0.321	0.422	15
Chloroethane	0.433	0.450	0.469	0.443	0.450	0.426	0.445	3.4
Trichlorofluoromethane	0.865	1.033	1.038	0.995	1.018	0.986	0.989	6.5
1,1,2-Trichlorotrifluoroethane	0.519	0.591	0.603	0.583	0.590	0.583	0.578	5.2
1,1-Dichloroethene	0.526	0.612	0.618	0.604	0.609	0.596	0.594	5.7
Acetone	0.343	0.416	0.400	0.317	0.302	0.299	0.346	14.7
Carbon Disulfide	1.555	1.735	1.758	1.566	1.589	1.555	1.626	5.8
Methyl tert-butyl Ether	1.884	2.238	2.221	2.247	2.212	2.209	2.169	6.5
Methyl Acetate	0.612	0.673	0.714	0.876	0.865	0.881	0.770	15.4
Methylene Chloride	0.718	0.772	0.754	0.730	0.716	0.705	0.732	3.5
trans-1,2-Dichloroethene	0.543	0.703	0.665	0.649	0.643	0.635	0.640	8.3
1,1-Dichloroethane	1.090	1.325	1.299	1.297	1.287	1.278	1.263	6.8
Cyclohexane		1.161	1.092	1.030	0.984	0.993	1.052	7.1
2-Butanone	0.370	0.470	0.473	0.439	0.414	0.424	0.432	9
Carbon Tetrachloride	0.482	0.574	0.555	0.539	0.526	0.518	0.532	6
cis-1,2-Dichloroethene	0.698	0.806	0.803	0.808	0.787	0.797	0.783	5.4
Bromochloromethane	0.457	0.607	0.450	0.577	0.587	0.630	0.551	14.2
Chloroform	1.034	1.364	1.366	1.323	1.282	1.288	1.276	9.7
1,1,1-Trichloroethane	0.901	1.123	1.141	1.127	1.100	1.103	1.082	8.3
Methylcyclohexane	0.490	0.589	0.591	0.565	0.548	0.554	0.556	6.6
Benzene	1.330	1.577	1.566	1.523	1.473	1.459	1.488	6.1
1,2-Dichloroethane	0.513	0.595	0.600	0.582	0.554	0.553	0.566	5.8
Trichloroethene	0.296	0.382	0.382	0.370	0.364	0.363	0.360	8.9
1,2-Dichloropropane	0.326	0.403	0.394	0.397	0.382	0.384	0.381	7.4
Bromodichloromethane	0.461	0.599	0.601	0.607	0.579	0.586	0.572	9.7
4-Methyl-2-Pentanone	0.390	0.493	0.503	0.511	0.479	0.487	0.477	9.3
Toluene	0.841	0.967	0.960	0.940	0.897	0.897	0.917	5.2

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: TETR06
Lab Code: ACE SDG No.: Q3248
Instrument ID: MSVOA_X Calibration Date(s): 09/16/2025 09/16/2025
Heated Purge: (Y/N) N Calibration Time(s): 09:23 12:01
GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:								
RRF001 = VX047585.D			RRF005 = VX047586.D			RRF020 = VX047587.D		
RRF050 = VX047588.D			RRF100 = VX047589.D			RRF150 = VX047590.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
t-1,3-Dichloropropene	0.487	0.588	0.602	0.619	0.595	0.611	0.584	8.3
cis-1,3-Dichloropropene	0.513	0.652	0.644	0.653	0.637	0.644	0.624	8.7
1,1,2-Trichloroethane	0.323	0.365	0.371	0.367	0.344	0.351	0.354	5.1
2-Hexanone	0.274	0.374	0.372	0.360	0.333	0.343	0.343	10.8
Dibromochloromethane	0.321	0.414	0.431	0.435	0.421	0.422	0.407	10.5
1,2-Dibromoethane	0.287	0.379	0.384	0.379	0.364	0.367	0.360	10.1
Tetrachloroethene	0.299	0.354	0.342	0.326	0.320	0.314	0.326	6
Chlorobenzene	0.998	1.182	1.177	1.178	1.139	1.142	1.136	6.2
Ethyl Benzene	1.625	2.077	2.049	2.039	1.996	1.973	1.960	8.6
m/p-Xylenes	0.604	0.774	0.765	0.765	0.739	0.730	0.729	8.8
o-Xylene	0.576	0.734	0.734	0.749	0.721	0.719	0.706	9.1
Styrene	1.019	1.288	1.290	1.299	1.257	1.265	1.236	8.7
Bromoform	0.221	0.293	0.301	0.313	0.312	0.317	0.293	12.4
Isopropylbenzene	3.124	3.907	3.986	4.047	3.907	3.839	3.801	8.9
1,1,2,2-Tetrachloroethane	0.965	1.182	1.203	1.225	1.152	1.173	1.150	8.2
1,3-Dichlorobenzene	1.462	1.800	1.829	1.817	1.768	1.758	1.739	8
1,4-Dichlorobenzene	1.570	1.890	1.868	1.832	1.767	1.783	1.785	6.5
1,2-Dichlorobenzene	1.359	1.691	1.757	1.746	1.696	1.692	1.657	9
1,2-Dibromo-3-Chloropropane	0.173	0.218	0.242	0.254	0.250	0.260	0.233	14.1
1,2,4-Trichlorobenzene	0.831	1.099	1.107	1.156	1.140	1.126	1.077	11.3
1,2,3-Trichlorobenzene	0.753	0.986	1.031	1.088	1.078	1.073	1.002	12.7
1,2-Dichloroethane-d4		0.884	0.647	0.770	0.815	0.863	0.796	11.8
Dibromofluoromethane		0.356	0.266	0.331	0.355	0.368	0.335	12.3
Toluene-d8		1.237	0.943	1.147	1.211	1.263	1.160	11.1
4-Bromofluorobenzene		0.478	0.360	0.427	0.452	0.484	0.440	11.4

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3248</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/03/2025 09:35</u>
Lab File ID:	<u>VX047981.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.577		11.39	20
Chloromethane	0.680	0.659	0.1	-3.09	20
Vinyl Chloride	0.666	0.682		2.4	20
Bromomethane	0.422	0.467		10.66	20
Chloroethane	0.445	0.450		1.12	20
Trichlorofluoromethane	0.989	1.066		7.79	20
1,1,2-Trichlorotrifluoroethane	0.578	0.638		10.38	20
1,1-Dichloroethene	0.594	0.609		2.53	20
Acetone	0.346	0.321		-7.22	20
Carbon Disulfide	1.626	1.511		-7.07	20
Methyl tert-butyl Ether	2.169	2.219		2.31	20
Methyl Acetate	0.770	0.839		8.96	20
Methylene Chloride	0.732	0.762		4.1	20
trans-1,2-Dichloroethene	0.640	0.654		2.19	20
1,1-Dichloroethane	1.263	1.326	0.1	4.99	20
Cyclohexane	1.052	1.024		-2.66	20
2-Butanone	0.432	0.408		-5.56	20
Carbon Tetrachloride	0.532	0.565		6.2	20
cis-1,2-Dichloroethene	0.783	0.827		5.62	20
Bromochloromethane	0.551	0.591		7.26	20
Chloroform	1.276	1.385		8.54	20
1,1,1-Trichloroethane	1.082	1.140		5.36	20
Methylcyclohexane	0.556	0.571		2.7	20
Benzene	1.488	1.585		6.52	20
1,2-Dichloroethane	0.566	0.610		7.77	20
Trichloroethene	0.360	0.385		6.94	20
1,2-Dichloropropane	0.381	0.426		11.81	20
Bromodichloromethane	0.572	0.655		14.51	20
4-Methyl-2-Pentanone	0.477	0.481		0.84	20
Toluene	0.917	0.989		7.85	20
t-1,3-Dichloropropene	0.584	0.657		12.5	20
cis-1,3-Dichloropropene	0.624	0.698		11.86	20
1,1,2-Trichloroethane	0.354	0.398		12.43	20
2-Hexanone	0.343	0.334		-2.62	20
Dibromochloromethane	0.407	0.478		17.44	20
1,2-Dibromoethane	0.360	0.408		13.33	20
Tetrachloroethene	0.326	0.327		0.31	20
Chlorobenzene	1.136	1.205	0.3	6.07	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3248</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/03/2025</u> <u>09:35</u>
Lab File ID:	<u>VX047981.D</u>	Init. Calib. Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	2.036		3.88	20
m/p-Xylenes	0.729	0.772		5.9	20
o-Xylene	0.706	0.741		4.96	20
Styrene	1.236	1.320		6.8	20
Bromoform	0.293	0.314	0.1	7.17	20
Isopropylbenzene	3.801	3.806		0.13	20
1,1,2,2-Tetrachloroethane	1.150	1.146	0.3	-0.35	20
1,3-Dichlorobenzene	1.739	1.775		2.07	20
1,4-Dichlorobenzene	1.785	1.818		1.85	20
1,2-Dichlorobenzene	1.657	1.703		2.78	20
1,2-Dibromo-3-Chloropropane	0.233	0.211		-9.44	20
1,2,4-Trichlorobenzene	1.077	1.037		-3.71	20
1,2,3-Trichlorobenzene	1.002	0.991		-1.1	20
1,2-Dichloroethane-d4	0.796	0.797		0.13	20
Dibromofluoromethane	0.335	0.365		8.95	20
Toluene-d8	1.160	1.215		4.74	20
4-Bromofluorobenzene	0.440	0.468		6.36	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETRO6</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3248</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/03/2025 19:43</u>
Lab File ID:	<u>VX048007.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.527		1.74	50
Chloromethane	0.680	0.643	0.1	-5.44	50
Vinyl Chloride	0.666	0.673		1.05	50
Bromomethane	0.422	0.432		2.37	50
Chloroethane	0.445	0.448		0.67	50
Trichlorofluoromethane	0.989	1.033		4.45	50
1,1,2-Trichlorotrifluoroethane	0.578	0.612		5.88	50
1,1-Dichloroethene	0.594	0.604		1.68	50
Acetone	0.346	0.311		-10.12	50
Carbon Disulfide	1.626	1.465		-9.9	50
Methyl tert-butyl Ether	2.169	2.388		10.1	50
Methyl Acetate	0.770	0.970		25.97	50
Methylene Chloride	0.732	0.762		4.1	50
trans-1,2-Dichloroethene	0.640	0.657		2.66	50
1,1-Dichloroethane	1.263	1.360	0.1	7.6	50
Cyclohexane	1.052	0.988		-6.08	50
2-Butanone	0.432	0.478		10.65	50
Carbon Tetrachloride	0.532	0.561		5.45	50
cis-1,2-Dichloroethene	0.783	0.839		7.15	50
Bromochloromethane	0.551	0.602		9.26	50
Chloroform	1.276	1.435		12.46	50
1,1,1-Trichloroethane	1.082	1.180		9.06	50
Methylcyclohexane	0.556	0.531		-4.5	50
Benzene	1.488	1.592		6.99	50
1,2-Dichloroethane	0.566	0.610		7.77	50
Trichloroethene	0.360	0.385		6.94	50
1,2-Dichloropropane	0.381	0.433		13.65	50
Bromodichloromethane	0.572	0.649		13.46	50
4-Methyl-2-Pentanone	0.477	0.583		22.22	50
Toluene	0.917	0.986		7.53	50
t-1,3-Dichloropropene	0.584	0.639		9.42	50
cis-1,3-Dichloropropene	0.624	0.675		8.17	50
1,1,2-Trichloroethane	0.354	0.416		17.51	50
2-Hexanone	0.343	0.404		17.78	50
Dibromochloromethane	0.407	0.482		18.43	50
1,2-Dibromoethane	0.360	0.426		18.33	50
Tetrachloroethene	0.326	0.336		3.07	50
Chlorobenzene	1.136	1.229	0.3	8.19	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3248</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/03/2025</u> <u>19:43</u>
Lab File ID:	<u>VX048007.D</u>	Init. Calib. Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	2.091		6.68	50
m/p-Xylenes	0.729	0.784		7.55	50
o-Xylene	0.706	0.762		7.93	50
Styrene	1.236	1.348		9.06	50
Bromoform	0.293	0.345	0.1	17.75	50
Isopropylbenzene	3.801	3.923		3.21	50
1,1,2,2-Tetrachloroethane	1.150	1.329	0.3	15.56	50
1,3-Dichlorobenzene	1.739	1.809		4.03	50
1,4-Dichlorobenzene	1.785	1.820		1.96	50
1,2-Dichlorobenzene	1.657	1.770		6.82	50
1,2-Dibromo-3-Chloropropane	0.233	0.257		10.3	50
1,2,4-Trichlorobenzene	1.077	1.098		1.95	50
1,2,3-Trichlorobenzene	1.002	1.073		7.09	50
1,2-Dichloroethane-d4	0.796	0.830		4.27	50
Dibromofluoromethane	0.335	0.360		7.46	50
Toluene-d8	1.160	1.242		7.07	50
4-Bromofluorobenzene	0.440	0.481		9.32	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

LAB CHRONICLE

OrderID:	Q3248	OrderDate:	9/30/2025 11:27:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	D31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3248-04	BP-VPB-184-GW-600-602	Water			09/30/25			10/02/25
			SVOC-SIMGroup1	8270-Modified		10/06/25	10/07/25	
Q3248-05	BP-VPB-184-GW-620-622	Water			10/01/25			10/02/25
			SVOC-SIMGroup1	8270-Modified		10/06/25	10/07/25	
Q3248-07	BP-VPB-184-GW-660-662	Water			10/01/25			10/02/25
			SVOC-SIMGroup1	8270-Modified		10/06/25	10/07/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet SW-846

SDG No.: Q3248

Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID : BP-VPB-184-GW-600-602								
Q3248-04	BP-VPB-184-GW-600-60 WATER	1,4-Dioxane	0.550		0.08	0.24	0.24	ug/L
		Total Svoc :		0.55				
		Total Concentration:		0.55				
Client ID : BP-VPB-184-GW-660-662								
Q3248-07	BP-VPB-184-GW-660-66 WATER	1,4-Dioxane	0.160	J	0.08	0.23	0.23	ug/L
		Total Svoc :		0.16				
		Total Concentration:		0.16				



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/30/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25
Client Sample ID:	BP-VPB-184-GW-600-602		SDG No.:	Q3248
Lab Sample ID:	Q3248-04		Matrix:	Water
Analytical Method:	SW8270ESIM	Level : LOW	% Solid:	0
Sample Wt/Vol:	840 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :	Prep Date: 10/06/25			

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.55		1	0.080	0.24	0.24	ug/L	10/07/25 14:28	PB169992
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.36			30 - 150		91%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.39			30 - 150		98%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.36			55 - 111		91%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.36			53 - 106		90%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.76	*		58 - 132		189%	SPK: 0.4		
INTERNAL STANDARDS										
		Area								
3855-82-1	1,4-Dichlorobenzene-d4	7970								
1146-65-2	Naphthalene-d8	23000								
15067-26-2	Acenaphthene-d10	12500								
1517-22-2	Phenanthrene-d10	25300								
1719-03-5	Chrysene-d12	17200								
1520-96-3	Perylene-d12	12200								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/01/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25
Client Sample ID:	BP-VPB-184-GW-620-622		SDG No.:	Q3248
Lab Sample ID:	Q3248-05		Matrix:	Water
Analytical Method:	SW8270ESIM	Level : LOW	% Solid:	0
Sample Wt/Vol:	870 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :	Prep Date: 10/06/25			

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.23	U	1	0.080	0.23	0.23	ug/L	10/07/25 15:05	PB169992
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.34			30 - 150		84%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.38			30 - 150		94%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.36			55 - 111		89%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.35			53 - 106		88%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.60	*		58 - 132		149%	SPK: 0.4		
INTERNAL STANDARDS										
		Area								
3855-82-1	1,4-Dichlorobenzene-d4	8670								
1146-65-2	Naphthalene-d8	24700								
15067-26-2	Acenaphthene-d10	12200								
1517-22-2	Phenanthrene-d10	21000								
1719-03-5	Chrysene-d12	14000								
1520-96-3	Perylene-d12	12900								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/01/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/02/25
Client Sample ID:	BP-VPB-184-GW-660-662		SDG No.:	Q3248
Lab Sample ID:	Q3248-07		Matrix:	Water
Analytical Method:	SW8270ESIM	Level : LOW	% Solid:	0
Sample Wt/Vol:	860 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :	Prep Date: 10/06/25			

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.16	J	1	0.080	0.23	0.23	ug/L	10/07/25 15:41	PB169992
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.34			30 - 150		86%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.38			30 - 150		95%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.36			55 - 111		89%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.37			53 - 106		93%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.75	*		58 - 132		188%	SPK: 0.4		
INTERNAL STANDARDS										
		Area								
3855-82-1	1,4-Dichlorobenzene-d4	7910								
1146-65-2	Naphthalene-d8	22900								
15067-26-2	Acenaphthene-d10	12400								
1517-22-2	Phenanthrene-d10	24100								
1719-03-5	Chrysene-d12	13900								
1520-96-3	Perylene-d12	10700								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: **Q3248**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **8270-Modified**

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169992BL	PB169992BL	2-Methylnaphthalene-d10	0.4	0.32	81		30	150
		Fluoranthene-d10	0.4	0.33	83		30	150
		Nitrobenzene-d5	0.4	0.33	83		55	111
		2-Fluorobiphenyl	0.4	0.35	88		53	106
		Terphenyl-d14	0.4	0.49	122		58	132
PB169992BS	PB169992BS	2-Methylnaphthalene-d10	0.4	0.35	87		30	150
		Fluoranthene-d10	0.4	0.33	82		30	150
		Nitrobenzene-d5	0.4	0.36	91		55	111
		2-Fluorobiphenyl	0.4	0.42	104		53	106
		Terphenyl-d14	0.4	0.50	125		58	132
PB169992BSD	PB169992BSD	2-Methylnaphthalene-d10	0.4	0.34	85		30	150
		Fluoranthene-d10	0.4	0.32	79		30	150
		Nitrobenzene-d5	0.4	0.35	88		55	111
		2-Fluorobiphenyl	0.4	0.39	97		53	106
		Terphenyl-d14	0.4	0.49	122		58	132
Q3248-04	BP-VPB-184-GW-600-602	2-Methylnaphthalene-d10	0.4	0.36	91		30	150
		Fluoranthene-d10	0.4	0.39	98		30	150
		Nitrobenzene-d5	0.4	0.36	91		55	111
		2-Fluorobiphenyl	0.4	0.36	90		53	106
		Terphenyl-d14	0.4	0.76	189	*	58	132
Q3248-05	BP-VPB-184-GW-620-622	2-Methylnaphthalene-d10	0.4	0.34	84		30	150
		Fluoranthene-d10	0.4	0.38	94		30	150
		Nitrobenzene-d5	0.4	0.36	89		55	111
		2-Fluorobiphenyl	0.4	0.35	88		53	106
		Terphenyl-d14	0.4	0.60	149	*	58	132
Q3248-07	BP-VPB-184-GW-660-662	2-Methylnaphthalene-d10	0.4	0.34	86		30	150
		Fluoranthene-d10	0.4	0.38	95		30	150
		Nitrobenzene-d5	0.4	0.36	89		55	111
		2-Fluorobiphenyl	0.4	0.37	93		53	106
		Terphenyl-d14	0.4	0.75	188	*	58	132

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3248

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN038068.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169992BS	1,4-Dioxane	0.4	0.36	ug/L	90				70	130	

A
B
C
D
E
F
G

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3248

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN038069.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169992BSD	1,4-Dioxane	0.4	0.36	ug/L	90	0			70	130	20

A
B
C
D
E
F
G

4B

SEMIVOLATILE METHOD BLANK SUMMARY

Client ID

PB169992BL

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q3248

Lab File ID: BN038067.D

Lab Sample ID: PB169992BL

Instrument ID: BNA_N

Date Extracted: 10/06/2025

Matrix: (soil/water) Water

Date Analyzed: 10/13/2025

Level: (low/med) LOW

Time Analyzed: 14:23

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB169992BS	PB169992BS	BN038068.D	10/13/2025
PB169992BSD	PB169992BSD	BN038069.D	10/13/2025
BP-VPB-184-GW-600-602	Q3248-04	BN038045.D	10/07/2025
BP-VPB-184-GW-620-622	Q3248-05	BN038046.D	10/07/2025
BP-VPB-184-GW-660-662	Q3248-07	BN038047.D	10/07/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037860.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q3248
DFTPP Injection Date: 09/23/2025
DFTPP Injection Time: 11:33

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	4.2
441	Present, but less than mass 443	72.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	17.7 (20.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN037861.D	09/23/2025	12:13
SSTDICC0.2	SSTDICC0.2	BN037862.D	09/23/2025	12:49
SSTDICCC0.4	SSTDICCC0.4	BN037863.D	09/23/2025	13:25
SSTDICC0.8	SSTDICC0.8	BN037864.D	09/23/2025	14:02
SSTDICC1.6	SSTDICC1.6	BN037865.D	09/23/2025	14:38
SSTDICC3.2	SSTDICC3.2	BN037866.D	09/23/2025	15:15
SSTDICC5.0	SSTDICC5.0	BN037867.D	09/23/2025	15:51

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN038040.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q3248
DFTPP Injection Date: 10/07/2025
DFTPP Injection Time: 10:44

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.7) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	76.8
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.3 (18.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN038041.D	10/07/2025	12:03
BP-VPB-184-GW-600-602	Q3248-04	BN038045.D	10/07/2025	14:28
BP-VPB-184-GW-620-622	Q3248-05	BN038046.D	10/07/2025	15:05
BP-VPB-184-GW-660-662	Q3248-07	BN038047.D	10/07/2025	15:41
SSTDCCC0.4EC	SSTDCCC0.4	BN038058.D	10/07/2025	22:22

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN038065.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q3248
DFTPP Injection Date: 10/13/2025
DFTPP Injection Time: 13:08

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.2) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	70
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.2 (21) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN038066.D	10/13/2025	13:47
PB169992BL	PB169992BL	BN038067.D	10/13/2025	14:23
PB169992BS	PB169992BS	BN038068.D	10/13/2025	14:59
PB169992BSD	PB169992BSD	BN038069.D	10/13/2025	15:35
SSTDCCC0.4EC	SSTDCCC0.4	BN038081.D	10/13/2025	23:24

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3248

Client ID : SSTDCCC0.4

Date Analyzed: 10/07/2025

Lab File ID: BN038041.D

Time Analyzed: 12:03

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	8417	7.804	24486	10.61	13140	14.46
UPPER LIMIT	16834	8.304	48972	11.105	26280	14.962
LOWER LIMIT	4208.5	7.304	12243	10.105	6570	13.962
EPA SAMPLE NO.						
01 BP-VPB-184-GW-600-602	7966	7.80	23022	10.61	12487	14.45
02 BP-VPB-184-GW-620-622	8674	7.80	24704	10.61	12221	14.45
03 BP-VPB-184-GW-660-662	7914	7.80	22866	10.61	12377	14.45

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3248

Client ID: SSTDCCC0.4

Date Analyzed: 10/07/2025

Lab File ID: BN038041.D

Time Analyzed: 12:03

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	24897	17.21	15245	21.393	11156	23.709
UPPER LIMIT	49794	17.71	30490	21.893	22312	24.209
LOWER LIMIT	12448.5	16.71	7622.5	20.893	5578	23.209
EPA SAMPLE NO.						
01 BP-VPB-184-GW-600-602	25250	17.20	17199	21.39	12217	23.71
02 BP-VPB-184-GW-620-622	20976	17.20	13980	21.39	12901	23.71
03 BP-VPB-184-GW-660-662	24052	17.20	13934	21.39	10696	23.71

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3248

Client ID : SSTDCCC0.4

Date Analyzed: 10/13/2025

Lab File ID: BN038066.D

Time Analyzed: 13:47

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	7166	7.797	18651	10.59	10244	14.45
UPPER LIMIT	14332	8.297	37302	11.094	20488	14.952
LOWER LIMIT	3583	7.297	9325.5	10.094	5122	13.952
EPA SAMPLE NO.						
01 PB169992BL	7352	7.80	18582	10.61	9416	14.45
02 PB169992BS	7409	7.80	19910	10.59	9443	14.45
03 PB169992BSD	7306	7.80	19813	10.59	9467	14.45

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3248

Client ID: SSTDCCC0.4

Date Analyzed: 10/13/2025

Lab File ID: BN038066.D

Time Analyzed: 13:47

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	18901	17.198	16827	21.393	17326	23.712
UPPER LIMIT	37802	17.698	33654	21.893	34652	24.212
LOWER LIMIT	9450.5	16.698	8413.5	20.893	8663	23.212
EPA SAMPLE NO.						
01 PB169992BL	15769	17.21	10550	21.39	9532	23.71
02 PB169992BS	15374	17.20	10279	21.39	8827	23.72
03 PB169992BSD	15168	17.20	9943	21.39	8435 *	23.71

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB169992BL		SDG No.:	Q3248
Lab Sample ID:	PB169992BL		Matrix:	Water
Analytical Method:	SW8270ESIM	Level : LOW	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :		Prep Date: 10/06/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.20	U	1	0.070	0.20	0.20	ug/L	10/13/25 14:23	PB169992
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.32			30 - 150		81%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.33			30 - 150		83%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.33			55 - 111		83%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.35			53 - 106		88%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.49			58 - 132		122%	SPK: 0.4		
INTERNAL STANDARDS										
		Area								
3855-82-1	1,4-Dichlorobenzene-d4	7350								
1146-65-2	Naphthalene-d8	18600								
15067-26-2	Acenaphthene-d10	9420								
1517-22-2	Phenanthrene-d10	15800								
1719-03-5	Chrysene-d12	10600								
1520-96-3	Perylene-d12	9530								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB169992BS		SDG No.:	Q3248
Lab Sample ID:	PB169992BS		Matrix:	Water
Analytical Method:	SW8270ESIM	Level : LOW	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :		Prep Date: 10/06/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.36		1	0.070	0.20	0.20	ug/L	10/13/25 14:59	PB169992
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.35			30 - 150		87%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.33			30 - 150		82%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.36			55 - 111		91%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.42			53 - 106		104%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.50			58 - 132		125%	SPK: 0.4		
INTERNAL STANDARDS										
		Area								
3855-82-1	1,4-Dichlorobenzene-d4	7410								
1146-65-2	Naphthalene-d8	19900								
15067-26-2	Acenaphthene-d10	9440								
1517-22-2	Phenanthrene-d10	15400								
1719-03-5	Chrysene-d12	10300								
1520-96-3	Perylene-d12	8830								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	
Client Sample ID:	PB169992BSD	SDG No.:	Q3248
Lab Sample ID:	PB169992BSD	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 mL	Test:	SVOC-SIMGroup1
Prep Method :			
	Level : LOW		
	Final Vol: 1000 uL		
	Prep Date: 10/06/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.36		1	0.070	0.20	0.20	ug/L	10/13/25 15:35	PB169992
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.34			30 - 150		85%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.32			30 - 150		79%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.35			55 - 111		88%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.39			53 - 106		97%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.49			58 - 132		122%	SPK: 0.4		
INTERNAL STANDARDS										
		Area								
3855-82-1	1,4-Dichlorobenzene-d4	7310								
1146-65-2	Naphthalene-d8	19800								
15067-26-2	Acenaphthene-d10	9470								
1517-22-2	Phenanthrene-d10	15200								
1719-03-5	Chrysene-d12	9940								
1520-96-3	Perylene-d12	8440								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN092325.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Sep 24 11:00:01 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN037861.D 0.2 =BN037862.D 0.4 =BN037863.D 0.8 =BN037864.D 1.6 =BN037865.D 3.2 =BN037866.D 5 =BN037867.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.413	0.410	0.403	0.425	0.410	0.376	0.406	4.04	
3)	n-Nitrosodimet...	0.541	0.517	0.526	0.561	0.549	0.547	0.540	3.03	
4) S	2-Fluorophenol	0.944	0.921	0.856	0.839	0.908	0.923	0.998	0.913	5.84
5) S	Phenol-d6	1.190	1.200	1.118	1.088	1.211	1.236	1.353	1.199	7.15
6)	bis(2-Chloroet...	1.133	1.096	1.082	1.066	1.150	1.124	1.143	1.113	2.89
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.299	0.289	0.289	0.290	0.310	0.320	0.338	0.305	6.14
9)	Naphthalene	1.105	1.089	1.048	1.051	1.123	1.137	1.160	1.102	3.84
10)	Hexachlorobuta...	0.198	0.184	0.182	0.183	0.192	0.188	0.188	0.188	2.94
11) SURR2	Methylnaphth...	0.515	0.518	0.499	0.504	0.553	0.587	0.714	0.556	13.74
12)	2-Methylnaphth...	0.678	0.681	0.668	0.682	0.754	0.775	0.802	0.720	7.68
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.148	0.142	0.135	0.139	0.163	0.184	0.152	12.15	
15) S	2-Fluorobiphenyl	1.660	1.631	1.559	1.547	1.654	1.650	1.762	1.638	4.36
16)	Acenaphthylene	1.698	1.721	1.663	1.680	1.907	1.951	2.094	1.816	9.23
17)	Acenaphthene	1.262	1.230	1.198	1.208	1.351	1.371	1.437	1.294	7.16
18)	Fluorene	1.455	1.455	1.452	1.466	1.664	1.717	1.748	1.565	8.77
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.041	0.044	0.046	0.058	0.071	0.084	0.057	29.75	
21)	4-Bromophenyl-...	0.252	0.247	0.244	0.244	0.272	0.275	0.291	0.261	7.07
22)	Hexachlorobenzene	0.317	0.312	0.305	0.300	0.324	0.323	0.332	0.316	3.53
23)	Atrazine	0.180	0.181	0.172	0.178	0.205	0.224	0.249	0.199	14.70
24)	Pentachlorophenol	0.101	0.092	0.097	0.116	0.136	0.108	16.50		
25)	Phenanthrene	1.301	1.271	1.230	1.224	1.358	1.379	1.452	1.316	6.37
26)	Anthracene	1.034	1.035	1.023	1.057	1.209	1.275	1.361	1.142	12.10
27) SURR	Fluoranthene-d10	0.880	0.918	0.887	0.900	1.004	1.085	1.370	1.006	17.59
28)	Fluoranthene	1.278	1.315	1.311	1.358	1.555	1.617	1.713	1.450	12.07
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.576	1.600	1.521	1.476	1.607	1.595	1.679	1.579	4.12
31) S	Terphenyl-d14	0.799	0.791	0.753	0.735	0.798	0.809	0.893	0.797	6.34
32)	Benzo(a)anthra...	1.376	1.359	1.294	1.285	1.423	1.483	1.567	1.398	7.27
33)	Chrysene	1.508	1.526	1.462	1.446	1.559	1.512	1.569	1.512	3.03
34)	Bis(2-ethylhex...	0.592	0.487	0.476	0.518	0.569	0.676	0.553	13.57	
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN092325.M

36)	Indeno(1,2,3-c...	1.715	1.672	1.658	1.707	1.887	1.996	2.159	1.828	10.52
37)	Benzo(b)fluora...	1.599	1.606	1.641	1.590	1.810	1.848	1.981	1.725	8.95
38)	Benzo(k)fluora...	1.620	1.586	1.597	1.628	1.875	1.917	1.926	1.736	9.26
39) C	Benzo(a)pyrene	1.254	1.296	1.231	1.254	1.413	1.492	1.610	1.364	10.61
40)	Dibenzo(a,h)an...	1.285	1.283	1.303	1.333	1.495	1.587	1.705	1.427	11.87
41)	Benzo(g,h,i)pe...	1.407	1.381	1.404	1.397	1.521	1.613	1.773	1.499	9.85

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG No.: Q3248
 Instrument ID: BNA_N Calibration Date/Time: 10/07/2025 12:03
 Lab File ID: BN038041.D Init. Calib. Date(s): 09/23/2025 09/23/2025
 EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 12:13 15:51
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.556	0.513		-7.7	20.0
Fluoranthene-d10	1.006	0.830		-17.5	20.0
2-Fluorophenol	0.913	0.853		-6.6	20.0
Phenol-d6	1.199	1.108		-7.6	20.0
Nitrobenzene-d5	0.305	0.286		-6.2	20.0
2-Fluorobiphenyl	1.638	1.517		-7.4	20.0
2,4,6-Tribromophenol	0.152	0.120		-21.1	20.0
Terphenyl-d14	0.797	0.916		14.9	20.0
1,4-Dioxane	0.406	0.418		3.0	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG No.: Q3248
 Instrument ID: BNA_N Calibration Date/Time: 10/07/2025 22:22
 Lab File ID: BN038058.D Init. Calib. Date(s): 09/23/2025 09/23/2025
 EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 12:13 15:51
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.556	0.505		-9.2	50.0
Fluoranthene-d10	1.006	0.755		-25.0	50.0
2-Fluorophenol	0.913	0.878		-3.8	50.0
Phenol-d6	1.199	1.126		-6.1	50.0
Nitrobenzene-d5	0.305	0.295		-3.3	50.0
2-Fluorobiphenyl	1.638	1.615		-1.4	50.0
2,4,6-Tribromophenol	0.152	0.112		-26.3	50.0
Terphenyl-d14	0.797	0.862		8.2	50.0
1,4-Dioxane	0.406	0.424		4.4	50.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG No.: Q3248
 Instrument ID: BNA_N Calibration Date/Time: 10/13/2025 13:47
 Lab File ID: BN038066.D Init. Calib. Date(s): 09/23/2025 09/23/2025
 EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 12:13 15:51
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.556	0.571		2.7	20.0
Fluoranthene-d10	1.006	1.065		5.9	20.0
2-Fluorophenol	0.913	0.893		-2.2	20.0
Phenol-d6	1.199	0.939		-21.7	20.0
Nitrobenzene-d5	0.305	0.238		-22.0	20.0
2-Fluorobiphenyl	1.638	1.428		-12.8	20.0
2,4,6-Tribromophenol	0.152	0.137		-9.9	20.0
Terphenyl-d14	0.797	0.724		-9.2	20.0
1,4-Dioxane	0.406	0.465		14.5	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG No.: Q3248
 Instrument ID: BNA_N Calibration Date/Time: 10/13/2025 23:24
 Lab File ID: BN038081.D Init. Calib. Date(s): 09/23/2025 09/23/2025
 EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 12:13 15:51
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.556	0.494		-11.2	50.0
Fluoranthene-d10	1.006	0.861		-14.4	50.0
2-Fluorophenol	0.913	0.864		-5.4	50.0
Phenol-d6	1.199	0.927		-22.7	50.0
Nitrobenzene-d5	0.305	0.270		-11.5	50.0
2-Fluorobiphenyl	1.638	1.596		-2.6	50.0
2,4,6-Tribromophenol	0.152	0.107		-29.6	50.0
Terphenyl-d14	0.797	0.757		-5.0	50.0
1,4-Dioxane	0.406	0.485		19.5	50.0

All other compounds must meet a minimum RRF of 0.010.



SHIPPING DOCUMENTS



CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092

(908) 789-8900 Fax: (908) 78-8922

www.chemtech.net

Chemtech Project Number:

Q3248

COC Number:

CLIENT INFORMATION

COMPANY: Tetra Tech

ADDRESS: 4433 Corporation Lane Suite 300

CITY: Virginia Beach

STATE: VA

ZIP: 23462

ATTENTION: Ernie Wu

PHONE: 757-466-4901

FAX: 757-461-4148

PROJECT INFORMATION

PROJECT NAME: NWIRP Bethpage

PROJECT #: 112G08005-WE13 LOCATION: VPB-184

PROJECT MANAGER: Ernie Wu

E-MAIL: ernie.wu@tetratech.com

PHONE: 757-466-4901

FAX: 757-461-4148

BILLING INFORMATION

BILL TO: SEE CONTRACT

PO#

ADDRESS:

CITY:

STATE: ZIP:

ATTENTION:

PHONE:

DATA TURNAROUND INFORMATION

FAX: 2 & 10 DAYS*

HARD COPY: 2 & 10 DAYS*

EDD 2 & 10 DAYS*

* TO BE APPROVED BY CHEMTECH

STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

☐ RESULTS ONLY☐ USEPA CLP☐ RESULTS + QC☐ New York State ASP "B"☐ New Jersey REDUCED☐ New York State ASP "A"☐ New Jersey CLP☐ Other☐ EDD Format

ANALYSIS

VOC(SW846-8260B)

1,4 Dioxane (8270 SIM)

1

2

3

4

5

6

7

8

9

PRESERVATIVES

COMMENTS

<-- Specify Preservatives

A-HCl B-HNO3

C-H2SO4 D-NaOH

E-ICE F-Other

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# of Bottles										
			COMP	GRAB	DATE	TIME		A									
1.	BP-VPB-184-TB-20250930	QA		X	9/30/25	9:00	2	2									Trip Blank
2.	BP-VPB-184-GW-560-562	AQ		X	9/30/25	9:55	3	3									
3.	BP-VPB-184-GW-580-582	AQ		X	9/30/25	12:08	3	3									
4.	BP-VPB-184-GW-600-602	AQ		X	9/30/25	14:11	3	2	1								
5.	BP-VPB-184-GW-620-622	AQ		X	10/1/25	9:50	3	2	1								8260B Duplicate
6.	BP-VPB-184-GW-640-642	AQ		X	10/1/25	12:30	9	9									Extra Due to Sediment
7.	BP-VPB-184-GW-660-662	AQ		X	10/1/25	14:48	3	2	1								
8.	BP-VPB-184-GW-685-687	AQ		X	10/2/25	12:33	9	9									8260B MS/MSD
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER	DATE/TIME	RECEIVED BY	Conditions of bottles or coolers at receipt: q Compliant q Non Compliant q Cooler Temp 2.4°C MeOH extraction requires an additional 4oz. Jar for percent solid Cooler?: Comments: 5 Day TAT - For VOC's see worksheet #15 of SAP 2018 for VPB program VOC list 10-DAY TAT - For 1,4 Dioxane (8270 SIM)
1.	10/2/25	1.	
RELINQUISHED BY	DATE/TIME	RECEIVED BY	
2.	10/2/25	2.	
RELINQUISHED BY	DATE/TIME	RECEIVED FOR LAB BY	
3.	10/2/25	3.	

Page 1 of 1

SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight
CHEMTECH: ☐ Picked Up ☐ Overnight

Shipment Complete
☐ YES ☐ NO

WHITE - CHEMTECH COPY FOR RETURN TO CLIENT

YELLOW - CHEMTECH COPY

PINK - SAMPLER COPY

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q3248 TETR06

Order Date : 9/30/2025 11:27:00 AM

Project Mgr :

Client Name : Tetra Tech NUS, Inc.

Project Name : NWIRP Bethpage 112G080

Report Type : Level 4

Client Contact : Ernie Wu

Receive Date/Time : 10/2/2025 3:35:00 PM

EDD Type : ADAPT

Invoice Name : Tetra Tech NUS, Inc.

Purchase Order : 1813

Hard Copy Date :

Invoice Contact : Ernie Wu

Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q3248-01	BP-VPB-184-GW-20250930	Water	09/30/2025	09:00	VOC-TCLVOA-10		8260-Low	5 Bus. Days	
Q3248-02	BP-VPB-184-GW-560-562	Water	09/30/2025	09:55	VOC-TCLVOA-10		8260-Low	5 Bus. Days	
Q3248-03	BP-VPB-184-GW-580-582	Water	09/30/2025	12:08	VOC-TCLVOA-10		8260-Low	5 Bus. Days	
Q3248-04	BP-VPB-184-GW-600-602	Water	09/30/2025	14:11	VOC-TCLVOA-10		8260-Low	5 Bus. Days	
Q3248-05	BP-VPB-184-GW-620-622	Water	10/01/2025	09:50	VOC-TCLVOA-10		8260-Low	5 Bus. Days	
Q3248-06	BP-VPB-184-GW-640-642	Water	10/01/2025	12:30	VOC-TCLVOA-10		8260-Low	5 Bus. Days	
Q3248-07	BP-VPB-184-GW-660-662	Water	10/01/2025	14:48	VOC-TCLVOA-10		8260-Low	5 Bus. Days	
Q3248-08	BP-VPB-184-GW-685-687	Water	10/02/2025	12:33	VOC-TCLVOA-10		8260-Low	5 Bus. Days	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q3248 TETR06
Client Name : Tetra Tech NUS, Inc.
Client Contact : Ernie Wu
Invoice Name : Tetra Tech NUS, Inc.
Invoice Contact : Ernie Wu

Order Date : 9/30/2025 11:27:00 AM
Project Name : NWIRP Bethpage 112G080
Receive DateTime : 10/2/2025 3:35:00 PM
Purchase Order : 1813

Project Mgr :
Report Type : Level 4
EDD Type : ADAPT
Hard Copy Date :
Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q3248-09	Q3248-08MS	Water	10/02/2025	12:33	VOC-TCLVOA-10		8260-Low 5 Bus. Days		
Q3248-10	Q3248-08MSD	Water	10/02/2025	12:33	VOC-TCLVOA-10		8260-Low 5 Bus. Days		
					VOC-TCLVOA-10		8260-Low 5 Bus. Days		

Relinquished By :

Date / Time :


10-3-25 12:12

Received By :

Date / Time :


10/03/25 12:12 Pg 4 4

Storage Area : VOA Refridgerator Room