

Surrogate Summary

SW-846

SDG No.: **Q3248**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **8270-Modified**

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169992BL	PB169992BL	2-Methylnaphthalene-d10	0.4	0.32	79		30	150
		Fluoranthene-d10	0.4	0.29	72		30	150
		Nitrobenzene-d5	0.4	0.41	101		55	111
		2-Fluorobiphenyl	0.4	0.40	100		53	106
		Terphenyl-d14	0.4	0.51	127		58	132
PB169992BS	PB169992BS	2-Methylnaphthalene-d10	0.4	0.36	90		30	150
		Fluoranthene-d10	0.4	0.28	70		30	150
		Nitrobenzene-d5	0.4	0.43	107		55	111
		2-Fluorobiphenyl	0.4	0.45	112	*	53	106
		Terphenyl-d14	0.4	0.53	132		58	132
PB169992BSD	PB169992BSD	2-Methylnaphthalene-d10	0.4	0.35	88		30	150
		Fluoranthene-d10	0.4	0.30	75		30	150
		Nitrobenzene-d5	0.4	0.42	105		55	111
		2-Fluorobiphenyl	0.4	0.42	104		53	106
		Terphenyl-d14	0.4	0.53	133	*	58	132
Q3248-04	BP-VPB-184-GW-600-602	2-Methylnaphthalene-d10	0.4	0.36	91		30	150
		Fluoranthene-d10	0.4	0.39	98		30	150
		Nitrobenzene-d5	0.4	0.36	91		55	111
		2-Fluorobiphenyl	0.4	0.36	90		53	106
		Terphenyl-d14	0.4	0.76	189	*	58	132
Q3248-05	BP-VPB-184-GW-620-622	2-Methylnaphthalene-d10	0.4	0.34	84		30	150
		Fluoranthene-d10	0.4	0.38	94		30	150
		Nitrobenzene-d5	0.4	0.36	89		55	111
		2-Fluorobiphenyl	0.4	0.35	88		53	106
		Terphenyl-d14	0.4	0.60	149	*	58	132
Q3248-07	BP-VPB-184-GW-660-662	2-Methylnaphthalene-d10	0.4	0.34	86		30	150
		Fluoranthene-d10	0.4	0.38	95		30	150
		Nitrobenzene-d5	0.4	0.36	89		55	111
		2-Fluorobiphenyl	0.4	0.37	93		53	106
		Terphenyl-d14	0.4	0.75	188	*	58	132