

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06
Lab Code: ACE SDG NO.: Q3248
Lab File ID: VX047981.D Date Analyzed: 10/03/2025
Instrument ID: MSVOA_X Time Analyzed: 09:35
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	131937	5.53	225878	6.74	209083	10.04
UPPER LIMIT	263874	6.031	451756	7.239	418166	10.537
LOWER LIMIT	65968.5	5.031	112939	6.239	104542	9.537
EPA SAMPLE NO.						
BP-VPB-184-TB-20250930	130029	5.54	264754	6.75	276284	10.04
BP-VPB-184-GW-560-562	130796	5.54	267996	6.75	274276	10.04
BP-VPB-184-GW-580-582	116663	5.54	235290	6.75	247287	10.04
BP-VPB-184-GW-600-602	121622	5.54	253461	6.75	259302	10.04
BP-VPB-184-GW-620-622	113992	5.54	234307	6.75	240205	10.04
BP-VPB-184-GW-640-642	114921	5.54	230401	6.75	237241	10.04
BP-VPB-184-GW-660-662	117443	5.54	235229	6.75	244995	10.04
BP-VPB-184-GW-685-687	115930	5.54	233399	6.75	238258	10.04
BP-VPB-184-GW-685-687MS	137909	5.54	257097	6.75	246962	10.04
BP-VPB-184-GW-685-687MSD	118835	5.54	215532	6.75	195286	10.04
VX1003WBL01	132031	5.54	278145	6.75	290715	10.04
VX1003WBS02	126673	5.54	223242	6.75	203303	10.04

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.