

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3248</u>
Lab File ID:	<u>VX047981.D</u>	Date Analyzed:	<u>10/03/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>09:35</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	105470	12.00€				
UPPER LIMIT	210940	12.50€				
LOWER LIMIT	52735	11.50€				
EPA SAMPLE NO.						
BP-VPB-184-TB-20250930	139079	12.01				
BP-VPB-184-GW-560-562	141091	12.01				
BP-VPB-184-GW-580-582	124667	12.01				
BP-VPB-184-GW-600-602	127779	12.01				
BP-VPB-184-GW-620-622	119987	12.01				
BP-VPB-184-GW-640-642	118357	12.01				
BP-VPB-184-GW-660-662	118155	12.01				
BP-VPB-184-GW-685-687	114439	12.01				
BP-VPB-184-GW-685-687MS	129827	12.01				
BP-VPB-184-GW-685-687MSD	99502	12.01				
VX1003WBL01	147497	12.01				
VX1003WBS02	101886	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.