

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q3254

Contract: LOCK01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Antimony	4080	4000	102	90 - 110	P	10/02/2025	11:07	LB137406
	Arsenic	4010	4000	100	90 - 110	P	10/02/2025	11:07	LB137406
	Barium	7650	8000	96	90 - 110	P	10/02/2025	11:07	LB137406
	Beryllium	201	200	101	90 - 110	P	10/02/2025	11:07	LB137406
	Cadmium	1970	2000	99	90 - 110	P	10/02/2025	11:07	LB137406
	Chromium	824	800	103	90 - 110	P	10/02/2025	11:07	LB137406
	Cobalt	1970	2000	99	90 - 110	P	10/02/2025	11:07	LB137406
	Copper	1010	1000	101	90 - 110	P	10/02/2025	11:07	LB137406
	Iron	3990	4000	100	90 - 110	P	10/02/2025	11:07	LB137406
	Lead	3870	4000	97	90 - 110	P	10/02/2025	11:07	LB137406
	Manganese	1940	2000	97	90 - 110	P	10/02/2025	11:07	LB137406
	Nickel	1980	2000	99	90 - 110	P	10/02/2025	11:07	LB137406
	Potassium	19600	20000	98	90 - 110	P	10/02/2025	11:07	LB137406
	Selenium	4040	4000	101	90 - 110	P	10/02/2025	11:07	LB137406
	Silver	987	1000	99	90 - 110	P	10/02/2025	11:07	LB137406
	Sodium	19000	20000	95	90 - 110	P	10/02/2025	11:07	LB137406
	Thallium	3980	4000	100	90 - 110	P	10/02/2025	11:07	LB137406
	Vanadium	1960	2000	98	90 - 110	P	10/02/2025	11:07	LB137406
	Zinc	2010	2000	100	90 - 110	P	10/02/2025	11:07	LB137406

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q3254

Contract: LOCK01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Antimony	52.3	50.0	104	80 - 120	P	10/02/2025	11:11	LB137406
	Arsenic	21.7	20.0	108	80 - 120	P	10/02/2025	11:11	LB137406
	Barium	113	100	113	80 - 120	P	10/02/2025	11:11	LB137406
	Beryllium	6.80	6.0	113	80 - 120	P	10/02/2025	11:11	LB137406
	Cadmium	6.03	6.0	100	80 - 120	P	10/02/2025	11:11	LB137406
	Chromium	11.2	10.0	112	80 - 120	P	10/02/2025	11:11	LB137406
	Cobalt	30.1	30.0	100	80 - 120	P	10/02/2025	11:11	LB137406
	Copper	22.2	20.0	111	80 - 120	P	10/02/2025	11:11	LB137406
	Iron	104	100	104	80 - 120	P	10/02/2025	11:11	LB137406
	Lead	13.2	12.0	110	80 - 120	P	10/02/2025	11:11	LB137406
	Manganese	21.5	20.0	108	80 - 120	P	10/02/2025	11:11	LB137406
	Nickel	40.4	40.0	101	80 - 120	P	10/02/2025	11:11	LB137406
	Potassium	1770	2000	88	80 - 120	P	10/02/2025	11:11	LB137406
	Selenium	22.8	20.0	114	80 - 120	P	10/02/2025	11:11	LB137406
	Silver	11.0	10.0	110	80 - 120	P	10/02/2025	11:11	LB137406
	Sodium	1690	2000	84	80 - 120	P	10/02/2025	11:11	LB137406
	Thallium	46.1	40.0	115	80 - 120	P	10/02/2025	11:11	LB137406
	Vanadium	42.2	40.0	106	80 - 120	P	10/02/2025	11:11	LB137406
	Zinc	42.4	40.0	106	80 - 120	P	10/02/2025	11:11	LB137406

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q3254

Contract: LOCK01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Antimony	4950	5000	99	90 - 110	P	10/02/2025	11:58	LB137406
	Arsenic	5100	5000	102	90 - 110	P	10/02/2025	11:58	LB137406
	Barium	9630	10000	96	90 - 110	P	10/02/2025	11:58	LB137406
	Beryllium	267	250	107	90 - 110	P	10/02/2025	11:58	LB137406
	Cadmium	2530	2500	101	90 - 110	P	10/02/2025	11:58	LB137406
	Chromium	1040	1000	104	90 - 110	P	10/02/2025	11:58	LB137406
	Cobalt	2490	2500	100	90 - 110	P	10/02/2025	11:58	LB137406
	Copper	1260	1250	101	90 - 110	P	10/02/2025	11:58	LB137406
	Iron	4740	5000	95	90 - 110	P	10/02/2025	11:58	LB137406
	Lead	5080	5000	102	90 - 110	P	10/02/2025	11:58	LB137406
	Manganese	2440	2500	98	90 - 110	P	10/02/2025	11:58	LB137406
	Nickel	2500	2500	100	90 - 110	P	10/02/2025	11:58	LB137406
	Potassium	23300	25000	93	90 - 110	P	10/02/2025	11:58	LB137406
	Selenium	5120	5000	102	90 - 110	P	10/02/2025	11:58	LB137406
	Silver	1260	1250	101	90 - 110	P	10/02/2025	11:58	LB137406
	Sodium	22800	25000	91	90 - 110	P	10/02/2025	11:58	LB137406
	Thallium	5080	5000	102	90 - 110	P	10/02/2025	11:58	LB137406
	Vanadium	2470	2500	99	90 - 110	P	10/02/2025	11:58	LB137406
	Zinc	2520	2500	101	90 - 110	P	10/02/2025	11:58	LB137406
CCV02	Antimony	4830	5000	97	90 - 110	P	10/02/2025	13:48	LB137406
	Arsenic	5200	5000	104	90 - 110	P	10/02/2025	13:48	LB137406
	Barium	9090	10000	91	90 - 110	P	10/02/2025	13:48	LB137406
	Beryllium	246	250	98	90 - 110	P	10/02/2025	13:48	LB137406
	Cadmium	2570	2500	103	90 - 110	P	10/02/2025	13:48	LB137406
	Chromium	1070	1000	107	90 - 110	P	10/02/2025	13:48	LB137406
	Cobalt	2460	2500	99	90 - 110	P	10/02/2025	13:48	LB137406
	Copper	1260	1250	101	90 - 110	P	10/02/2025	13:48	LB137406
	Iron	5120	5000	102	90 - 110	P	10/02/2025	13:48	LB137406
	Lead	5240	5000	105	90 - 110	P	10/02/2025	13:48	LB137406
	Manganese	2330	2500	93	90 - 110	P	10/02/2025	13:48	LB137406
	Nickel	2490	2500	99	90 - 110	P	10/02/2025	13:48	LB137406
	Potassium	23300	25000	93	90 - 110	P	10/02/2025	13:48	LB137406
	Selenium	5320	5000	106	90 - 110	P	10/02/2025	13:48	LB137406
	Silver	1280	1250	102	90 - 110	P	10/02/2025	13:48	LB137406

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q3254

Contract: LOCK01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Sodium	24800	25000	99	90 - 110	P	10/02/2025	13:48	LB137406
	Thallium	5290	5000	106	90 - 110	P	10/02/2025	13:48	LB137406
	Vanadium	2390	2500	96	90 - 110	P	10/02/2025	13:48	LB137406
	Zinc	2500	2500	100	90 - 110	P	10/02/2025	13:48	LB137406
CCV03	Antimony	4800	5000	96	90 - 110	P	10/02/2025	14:47	LB137406
	Arsenic	5170	5000	104	90 - 110	P	10/02/2025	14:47	LB137406
	Barium	9030	10000	90	90 - 110	P	10/02/2025	14:47	LB137406
	Beryllium	243	250	97	90 - 110	P	10/02/2025	14:47	LB137406
	Cadmium	2560	2500	103	90 - 110	P	10/02/2025	14:47	LB137406
	Chromium	1050	1000	105	90 - 110	P	10/02/2025	14:47	LB137406
	Cobalt	2450	2500	98	90 - 110	P	10/02/2025	14:47	LB137406
	Copper	1250	1250	100	90 - 110	P	10/02/2025	14:47	LB137406
	Iron	5080	5000	102	90 - 110	P	10/02/2025	14:47	LB137406
	Lead	5220	5000	104	90 - 110	P	10/02/2025	14:47	LB137406
	Manganese	2320	2500	93	90 - 110	P	10/02/2025	14:47	LB137406
	Nickel	2470	2500	99	90 - 110	P	10/02/2025	14:47	LB137406
	Potassium	24700	25000	99	90 - 110	P	10/02/2025	14:47	LB137406
	Selenium	5300	5000	106	90 - 110	P	10/02/2025	14:47	LB137406
	Silver	1250	1250	100	90 - 110	P	10/02/2025	14:47	LB137406
	Sodium	24400	25000	98	90 - 110	P	10/02/2025	14:47	LB137406
	Thallium	5300	5000	106	90 - 110	P	10/02/2025	14:47	LB137406
	Vanadium	2360	2500	95	90 - 110	P	10/02/2025	14:47	LB137406
	Zinc	2470	2500	99	90 - 110	P	10/02/2025	14:47	LB137406
CCV04	Antimony	4920	5000	98	90 - 110	P	10/02/2025	15:54	LB137406
	Arsenic	5460	5000	109	90 - 110	P	10/02/2025	15:54	LB137406
	Barium	9070	10000	91	90 - 110	P	10/02/2025	15:54	LB137406
	Beryllium	234	250	93	90 - 110	P	10/02/2025	15:54	LB137406
	Cadmium	2660	2500	106	90 - 110	P	10/02/2025	15:54	LB137406
	Chromium	1070	1000	107	90 - 110	P	10/02/2025	15:54	LB137406
	Cobalt	2480	2500	99	90 - 110	P	10/02/2025	15:54	LB137406
	Copper	1290	1250	103	90 - 110	P	10/02/2025	15:54	LB137406
	Iron	5300	5000	106	90 - 110	P	10/02/2025	15:54	LB137406
	Lead	5270	5000	106	90 - 110	P	10/02/2025	15:54	LB137406
	Manganese	2340	2500	94	90 - 110	P	10/02/2025	15:54	LB137406

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q3254

Contract: LOCK01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Nickel	2520	2500	101	90 - 110	P	10/02/2025	15:54	LB137406
	Potassium	24200	25000	97	90 - 110	P	10/02/2025	15:54	LB137406
	Selenium	4900	5000	98	90 - 110	P	10/02/2025	15:54	LB137406
	Silver	1290	1250	103	90 - 110	P	10/02/2025	15:54	LB137406
	Sodium	23300	25000	93	90 - 110	P	10/02/2025	15:54	LB137406
	Thallium	5170	5000	103	90 - 110	P	10/02/2025	15:54	LB137406
	Vanadium	2400	2500	96	90 - 110	P	10/02/2025	15:54	LB137406
	Zinc	2510	2500	100	90 - 110	P	10/02/2025	15:54	LB137406
CCV05	Antimony	4750	5000	95	90 - 110	P	10/02/2025	17:31	LB137406
	Arsenic	5060	5000	101	90 - 110	P	10/02/2025	17:31	LB137406
	Barium	9160	10000	92	90 - 110	P	10/02/2025	17:31	LB137406
	Beryllium	239	250	96	90 - 110	P	10/02/2025	17:31	LB137406
	Cadmium	2530	2500	101	90 - 110	P	10/02/2025	17:31	LB137406
	Chromium	1020	1000	102	90 - 110	P	10/02/2025	17:31	LB137406
	Cobalt	2430	2500	97	90 - 110	P	10/02/2025	17:31	LB137406
	Copper	1240	1250	99	90 - 110	P	10/02/2025	17:31	LB137406
	Iron	5480	5000	110	90 - 110	P	10/02/2025	17:31	LB137406
	Lead	5140	5000	103	90 - 110	P	10/02/2025	17:31	LB137406
	Manganese	2350	2500	94	90 - 110	P	10/02/2025	17:31	LB137406
	Nickel	2450	2500	98	90 - 110	P	10/02/2025	17:31	LB137406
	Potassium	22500	25000	90	90 - 110	P	10/02/2025	17:31	LB137406
	Selenium	5150	5000	103	90 - 110	P	10/02/2025	17:31	LB137406
	Silver	1230	1250	99	90 - 110	P	10/02/2025	17:31	LB137406
	Sodium	26000	25000	104	90 - 110	P	10/02/2025	17:31	LB137406
	Thallium	4690	5000	94	90 - 110	P	10/02/2025	17:31	LB137406
	Vanadium	2370	2500	95	90 - 110	P	10/02/2025	17:31	LB137406
	Zinc	2430	2500	97	90 - 110	P	10/02/2025	17:31	LB137406
	Antimony	4830	5000	96	90 - 110	P	10/02/2025	18:17	LB137406
	Arsenic	5210	5000	104	90 - 110	P	10/02/2025	18:17	LB137406
	Barium	9140	10000	91	90 - 110	P	10/02/2025	18:17	LB137406
	Beryllium	252	250	101	90 - 110	P	10/02/2025	18:17	LB137406
	Cadmium	2590	2500	104	90 - 110	P	10/02/2025	18:17	LB137406
	Chromium	1040	1000	104	90 - 110	P	10/02/2025	18:17	LB137406
	Cobalt	2460	2500	98	90 - 110	P	10/02/2025	18:17	LB137406

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q3254

Contract: LOCK01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV06	Copper	1270	1250	101	90 - 110	P	10/02/2025	18:17	LB137406
	Iron	5380	5000	108	90 - 110	P	10/02/2025	18:17	LB137406
	Lead	5290	5000	106	90 - 110	P	10/02/2025	18:17	LB137406
	Manganese	2360	2500	94	90 - 110	P	10/02/2025	18:17	LB137406
	Nickel	2490	2500	100	90 - 110	P	10/02/2025	18:17	LB137406
	Potassium	24700	25000	99	90 - 110	P	10/02/2025	18:17	LB137406
	Selenium	5010	5000	100	90 - 110	P	10/02/2025	18:17	LB137406
	Silver	1250	1250	100	90 - 110	P	10/02/2025	18:17	LB137406
	Sodium	26000	25000	104	90 - 110	P	10/02/2025	18:17	LB137406
	Thallium	4730	5000	94	90 - 110	P	10/02/2025	18:17	LB137406
	Vanadium	2370	2500	95	90 - 110	P	10/02/2025	18:17	LB137406
	Zinc	2450	2500	98	90 - 110	P	10/02/2025	18:17	LB137406

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q3254

Contract: LOCK01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Antimony	4030	4000	101	90 - 110	P	10/03/2025	12:18	LB137426
	Arsenic	3940	4000	99	90 - 110	P	10/03/2025	12:18	LB137426
	Barium	7900	8000	99	90 - 110	P	10/03/2025	12:18	LB137426
	Beryllium	199	200	100	90 - 110	P	10/03/2025	12:18	LB137426
	Cadmium	1950	2000	98	90 - 110	P	10/03/2025	12:18	LB137426
	Chromium	799	800	100	90 - 110	P	10/03/2025	12:18	LB137426
	Cobalt	1950	2000	98	90 - 110	P	10/03/2025	12:18	LB137426
	Copper	1010	1000	101	90 - 110	P	10/03/2025	12:18	LB137426
	Iron	3950	4000	99	90 - 110	P	10/03/2025	12:18	LB137426
	Lead	3840	4000	96	90 - 110	P	10/03/2025	12:18	LB137426
	Manganese	1980	2000	99	90 - 110	P	10/03/2025	12:18	LB137426
	Nickel	1960	2000	98	90 - 110	P	10/03/2025	12:18	LB137426
	Potassium	19400	20000	97	90 - 110	P	10/03/2025	12:18	LB137426
	Selenium	3960	4000	99	90 - 110	P	10/03/2025	12:18	LB137426
	Silver	960	1000	96	90 - 110	P	10/03/2025	12:18	LB137426
	Sodium	19300	20000	97	90 - 110	P	10/03/2025	12:18	LB137426
	Thallium	3750	4000	94	90 - 110	P	10/03/2025	12:18	LB137426
	Vanadium	1960	2000	98	90 - 110	P	10/03/2025	12:18	LB137426
	Zinc	1980	2000	99	90 - 110	P	10/03/2025	12:18	LB137426

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q3254

Contract: LOCK01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Antimony	53.8	50.0	108	80 - 120	P	10/03/2025	13:05	LB137426
	Arsenic	22.0	20.0	110	80 - 120	P	10/03/2025	13:05	LB137426
	Barium	99.1	100	99	80 - 120	P	10/03/2025	13:05	LB137426
	Beryllium	6.51	6.0	108	80 - 120	P	10/03/2025	13:05	LB137426
	Cadmium	6.05	6.0	101	80 - 120	P	10/03/2025	13:05	LB137426
	Chromium	10.4	10.0	104	80 - 120	P	10/03/2025	13:05	LB137426
	Cobalt	30.1	30.0	100	80 - 120	P	10/03/2025	13:05	LB137426
	Copper	21.3	20.0	107	80 - 120	P	10/03/2025	13:05	LB137426
	Iron	92.9	100	93	80 - 120	P	10/03/2025	13:05	LB137426
	Lead	12.8	12.0	107	80 - 120	P	10/03/2025	13:05	LB137426
	Manganese	21.2	20.0	106	80 - 120	P	10/03/2025	13:05	LB137426
	Nickel	41.3	40.0	103	80 - 120	P	10/03/2025	13:05	LB137426
	Potassium	1960	2000	98	80 - 120	P	10/03/2025	13:05	LB137426
	Selenium	19.3	20.0	96	80 - 120	P	10/03/2025	13:05	LB137426
	Silver	10.2	10.0	102	80 - 120	P	10/03/2025	13:05	LB137426
	Sodium	1990	2000	100	80 - 120	P	10/03/2025	13:05	LB137426
	Thallium	41.8	40.0	104	80 - 120	P	10/03/2025	13:05	LB137426
	Vanadium	39.2	40.0	98	80 - 120	P	10/03/2025	13:05	LB137426
	Zinc	42.4	40.0	106	80 - 120	P	10/03/2025	13:05	LB137426

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q3254

Contract: LOCK01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Antimony	5040	5000	101	90 - 110	P	10/03/2025	13:37	LB137426
	Arsenic	5030	5000	100	90 - 110	P	10/03/2025	13:37	LB137426
	Barium	10000	10000	100	90 - 110	P	10/03/2025	13:37	LB137426
	Beryllium	247	250	99	90 - 110	P	10/03/2025	13:37	LB137426
	Cadmium	2490	2500	100	90 - 110	P	10/03/2025	13:37	LB137426
	Chromium	989	1000	99	90 - 110	P	10/03/2025	13:37	LB137426
	Cobalt	2480	2500	99	90 - 110	P	10/03/2025	13:37	LB137426
	Copper	1260	1250	101	90 - 110	P	10/03/2025	13:37	LB137426
	Iron	4900	5000	98	90 - 110	P	10/03/2025	13:37	LB137426
	Lead	4940	5000	99	90 - 110	P	10/03/2025	13:37	LB137426
	Manganese	2500	2500	100	90 - 110	P	10/03/2025	13:37	LB137426
	Nickel	2490	2500	100	90 - 110	P	10/03/2025	13:37	LB137426
	Potassium	24200	25000	97	90 - 110	P	10/03/2025	13:37	LB137426
	Selenium	5050	5000	101	90 - 110	P	10/03/2025	13:37	LB137426
	Silver	1230	1250	98	90 - 110	P	10/03/2025	13:37	LB137426
	Sodium	24400	25000	97	90 - 110	P	10/03/2025	13:37	LB137426
	Thallium	4650	5000	93	90 - 110	P	10/03/2025	13:37	LB137426
	Vanadium	2480	2500	99	90 - 110	P	10/03/2025	13:37	LB137426
	Zinc	2480	2500	99	90 - 110	P	10/03/2025	13:37	LB137426
CCV02	Antimony	4990	5000	100	90 - 110	P	10/03/2025	15:00	LB137426
	Arsenic	4990	5000	100	90 - 110	P	10/03/2025	15:00	LB137426
	Barium	9880	10000	99	90 - 110	P	10/03/2025	15:00	LB137426
	Beryllium	244	250	98	90 - 110	P	10/03/2025	15:00	LB137426
	Cadmium	2460	2500	98	90 - 110	P	10/03/2025	15:00	LB137426
	Chromium	984	1000	98	90 - 110	P	10/03/2025	15:00	LB137426
	Cobalt	2450	2500	98	90 - 110	P	10/03/2025	15:00	LB137426
	Copper	1240	1250	99	90 - 110	P	10/03/2025	15:00	LB137426
	Iron	4860	5000	97	90 - 110	P	10/03/2025	15:00	LB137426
	Lead	4890	5000	98	90 - 110	P	10/03/2025	15:00	LB137426
	Manganese	2440	2500	98	90 - 110	P	10/03/2025	15:00	LB137426
	Nickel	2450	2500	98	90 - 110	P	10/03/2025	15:00	LB137426
	Potassium	24400	25000	98	90 - 110	P	10/03/2025	15:00	LB137426
	Selenium	5040	5000	101	90 - 110	P	10/03/2025	15:00	LB137426
	Silver	1230	1250	98	90 - 110	P	10/03/2025	15:00	LB137426

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q3254

Contract: LOCK01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Sodium	24300	25000	97	90 - 110	P	10/03/2025	15:00	LB137426
	Thallium	4580	5000	92	90 - 110	P	10/03/2025	15:00	LB137426
	Vanadium	2470	2500	99	90 - 110	P	10/03/2025	15:00	LB137426
	Zinc	2450	2500	98	90 - 110	P	10/03/2025	15:00	LB137426
CCV03	Antimony	5140	5000	103	90 - 110	P	10/03/2025	15:51	LB137426
	Arsenic	5080	5000	102	90 - 110	P	10/03/2025	15:51	LB137426
	Barium	10000	10000	100	90 - 110	P	10/03/2025	15:51	LB137426
	Beryllium	239	250	96	90 - 110	P	10/03/2025	15:51	LB137426
	Cadmium	2440	2500	98	90 - 110	P	10/03/2025	15:51	LB137426
	Chromium	972	1000	97	90 - 110	P	10/03/2025	15:51	LB137426
	Cobalt	2450	2500	98	90 - 110	P	10/03/2025	15:51	LB137426
	Copper	1250	1250	100	90 - 110	P	10/03/2025	15:51	LB137426
	Iron	5230	5000	105	90 - 110	P	10/03/2025	15:51	LB137426
	Lead	4860	5000	97	90 - 110	P	10/03/2025	15:51	LB137426
	Manganese	2420	2500	97	90 - 110	P	10/03/2025	15:51	LB137426
	Nickel	2450	2500	98	90 - 110	P	10/03/2025	15:51	LB137426
	Potassium	25700	25000	103	90 - 110	P	10/03/2025	15:51	LB137426
	Selenium	5170	5000	103	90 - 110	P	10/03/2025	15:51	LB137426
	Silver	1240	1250	99	90 - 110	P	10/03/2025	15:51	LB137426
	Sodium	27300	25000	109	90 - 110	P	10/03/2025	15:51	LB137426
	Thallium	5250	5000	105	90 - 110	P	10/03/2025	15:51	LB137426
	Vanadium	2450	2500	98	90 - 110	P	10/03/2025	15:51	LB137426
	Zinc	2490	2500	100	90 - 110	P	10/03/2025	15:51	LB137426
CCV04	Antimony	5020	5000	100	90 - 110	P	10/03/2025	16:45	LB137426
	Arsenic	4930	5000	99	90 - 110	P	10/03/2025	16:45	LB137426
	Barium	9780	10000	98	90 - 110	P	10/03/2025	16:45	LB137426
	Beryllium	234	250	93	90 - 110	P	10/03/2025	16:45	LB137426
	Cadmium	2360	2500	94	90 - 110	P	10/03/2025	16:45	LB137426
	Chromium	938	1000	94	90 - 110	P	10/03/2025	16:45	LB137426
	Cobalt	2370	2500	95	90 - 110	P	10/03/2025	16:45	LB137426
	Copper	1210	1250	97	90 - 110	P	10/03/2025	16:45	LB137426
	Iron	5120	5000	102	90 - 110	P	10/03/2025	16:45	LB137426
	Lead	4700	5000	94	90 - 110	P	10/03/2025	16:45	LB137426
	Manganese	2370	2500	95	90 - 110	P	10/03/2025	16:45	LB137426

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q3254

Contract: LOCK01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Nickel	2370	2500	95	90 - 110	P	10/03/2025	16:45	LB137426
	Potassium	24100	25000	96	90 - 110	P	10/03/2025	16:45	LB137426
	Selenium	5060	5000	101	90 - 110	P	10/03/2025	16:45	LB137426
	Silver	1210	1250	97	90 - 110	P	10/03/2025	16:45	LB137426
	Sodium	25400	25000	102	90 - 110	P	10/03/2025	16:45	LB137426
	Thallium	5250	5000	105	90 - 110	P	10/03/2025	16:45	LB137426
	Vanadium	2390	2500	96	90 - 110	P	10/03/2025	16:45	LB137426
	Zinc	2420	2500	97	90 - 110	P	10/03/2025	16:45	LB137426