

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q3254

Contract: LOCK01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169937BS							
Antimony	ug/L	400	393		98	80 - 120	P
Arsenic	ug/L	400	379		95	80 - 120	P
Barium	ug/L	100	93.3		93	80 - 120	P
Beryllium	ug/L	100	96.4		96	80 - 120	P
Cadmium	ug/L	100	93.5		94	80 - 120	P
Chromium	ug/L	200	207		104	80 - 120	P
Cobalt	ug/L	100	94.5		94	80 - 120	P
Copper	ug/L	150	148		99	80 - 120	P
Iron	ug/L	1500	1490		99	80 - 120	P
Lead	ug/L	500	461		92	80 - 120	P
Manganese	ug/L	100	98.0		98	80 - 120	P
Nickel	ug/L	250	247		99	80 - 120	P
Potassium	ug/L	5000	4670		93	80 - 120	P
Selenium	ug/L	1000	961		96	80 - 120	P
Silver	ug/L	37.5	35.5		95	80 - 120	P
Sodium	ug/L	1500	1430		95	80 - 120	P
Thallium	ug/L	1000	926		93	80 - 120	P
Vanadium	ug/L	150	144		96	80 - 120	P
Zinc	ug/L	100	97.7		98	80 - 120	P