

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>LOCK01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3254</u>
Instrument ID:	<u>MSVOA_V</u>	Calibration Date/Time:	<u>10/03/2025 08:53</u>
Lab File ID:	<u>VV039242.D</u>	Init. Calib. Date(s):	<u>09/22/2025 09/22/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:26 16:35</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Chloroethane	50.000	49.76	0.487999	20
Acetone	250.000	261.13	-4.451196	20
Methylene Chloride	50.000	54.17	-8.348	20
2-Hexanone	250.000	270.68	-8.270398	20
Styrene	50.000	57.81	-15.628	20
1,2-Dibromo-3-Chloropropane	50.000	50.27	-0.5360031	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.