

Instrument ID: MSVOA_V

Daily Analysis Runlog For Sequence/QC Batch ID # VV092225

Review By	Mahesh Dadoda	Review On	9/26/2025 1:39:32 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	9/26/2025 1:56:56 AM		
SubDirectory	VV092225	HP Acquire Method	MSVOA_V	HP Processing Method	82v092225w.m

STD. NAME	STD REF.#
Tune/Reschk	VP135580
Initial Calibration Stds	VP135581,VP135582,VP135583,VP135584,VP135585,VP135586
CCC	
Internal Standard/PEM	VP133935
ICV/I.BLK	VP135587
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VV039134.D	22 Sep 2025 10:15		SY/MD	Ok
2	VSTDIC001	VSTDIC001	VV039135.D	22 Sep 2025 10:59	Not use	SY/MD	Not Ok
3	VSTDIC005	VSTDIC005	VV039136.D	22 Sep 2025 11:30	Not use	SY/MD	Not Ok
4	VSTDIC010	VSTDIC010	VV039137.D	22 Sep 2025 12:26		SY/MD	Ok
5	VSTDIC020	VSTDIC020	VV039138.D	22 Sep 2025 12:48	QR- 06,16,20,70,92	SY/MD	Ok
6	VSTDIC050	VSTDIC050	VV039139.D	22 Sep 2025 13:26	LR- 41,58,59	SY/MD	Ok
7	VSTDIC100	VSTDIC100	VV039140.D	22 Sep 2025 13:49		SY/MD	Ok
8	VIBLK	VIBLK	VV039141.D	22 Sep 2025 14:21		SY/MD	Ok
9	VSTDIC005	VSTDIC005	VV039142.D	22 Sep 2025 15:37		SY/MD	Ok,M
10	VSTDIC001	VSTDIC001	VV039143.D	22 Sep 2025 16:35		SY/MD	Ok,M
11	VSTDICV050	ICVVV092225	VV039144.D	22 Sep 2025 16:58	Goof for Non-DOD	SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_V

Daily Analysis Runlog For Sequence/QC Batch ID # VV100325

Review By	Maresh Dadoda	Review On	10/6/2025 4:25:08 AM
Supervise By	Semsettin Yesilyurt	Supervise On	10/6/2025 4:29:07 AM
SubDirectory	VV100325	HP Acquire Method	MSVOA_V
		HP Processing Method	82v092225w.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135740
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135741,VP135742 VP133935

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VV039241.D	03 Oct 2025 08:26		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VV039242.D	03 Oct 2025 08:53	CCC failed High For Com.#51,68,69	SY/MD	Ok
3	VV1003MBL01	VV1003MBL01	VV039243.D	03 Oct 2025 11:04		SY/MD	Ok
4	VV1003WBL01	VV1003WBL01	VV039244.D	03 Oct 2025 11:26		SY/MD	Ok
5	VV1003WBS01	VV1003WBS01	VV039245.D	03 Oct 2025 13:07		SY/MD	Ok
6	VV1003WBSD01	VV1003WBSD01	VV039246.D	03 Oct 2025 13:29		SY/MD	Ok
7	Q3263-01	251818	VV039247.D	03 Oct 2025 14:19	Surrogate fail	SY/MD	ReRun
8	Q3263-02	266380	VV039248.D	03 Oct 2025 14:42	Surrogate fail	SY/MD	ReRun
9	Q3201-07	FIELD-BLANK	VV039249.D	03 Oct 2025 15:04	FB	SY/MD	Ok
10	Q3201-08	TRIP-BLANK	VV039250.D	03 Oct 2025 15:27	TB	SY/MD	Ok
11	Q3254-09	TRIP-BLANK	VV039251.D	03 Oct 2025 15:49	TB	SY/MD	Ok
12	Q3254-01	MW-1	VV039252.D	03 Oct 2025 16:11		SY/MD	Ok
13	Q3254-03	MW-2	VV039253.D	03 Oct 2025 16:34		SY/MD	Ok
14	Q3254-05	MW-3	VV039254.D	03 Oct 2025 16:56		SY/MD	Ok
15	Q3254-07	MW-4	VV039255.D	03 Oct 2025 17:19		SY/MD	Ok
16	Q3201-01	MW6	VV039256.D	03 Oct 2025 17:41	Need 100X	SY/MD	Dilution
17	Q3201-02	MW14	VV039257.D	03 Oct 2025 18:04	Need 100X	SY/MD	Dilution

Instrument ID: MSVOA_V

Daily Analysis Runlog For Sequence/QC Batch ID # VV100325

Review By	Mahesh Dadoda	Review On	10/6/2025 4:25:08 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	10/6/2025 4:29:07 AM		
SubDirectory	VV100325	HP Acquire Method	MSVOA_V	HP Processing Method	82v092225w.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP135740			
CCC		VP135741,VP135742			
Internal Standard/PEM		VP133935			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q3201-03	MW15	VV039258.D	03 Oct 2025 18:26	E flag of previous sample	SY/MD	ReRun
19	Q3201-04	MW10	VV039259.D	03 Oct 2025 18:49	Need Straight Run	SY/MD	Not Ok
20	Q3201-05	MW1	VV039260.D	03 Oct 2025 19:11	CCC Failed High;Need 40X	SY/MD	ReRun
21	Q3201-06	MW3	VV039261.D	03 Oct 2025 19:34	E flag of previous sample; Run @ lower dilution; CCC Failed High	SY/MD	Not Ok
22	VIBLK	VIBLK	VV039262.D	03 Oct 2025 19:56		SY/MD	Ok
23	VSTDCCC050	VSTDCCC050EC	VV039263.D	03 Oct 2025 20:19		SY/MD	Ok

M : Manual Integration