

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: LOCK01
 Lab Code: ACE SDG NO.: Q3254
 Lab File ID: VV039242.D Date Analyzed: 10/03/2025
 Instrument ID: MSVOA_V Time Analyzed: 08:53
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	467640	11.172				
UPPER LIMIT	935280	11.672				
LOWER LIMIT	233820	10.672				
EPA SAMPLE NO.						
MW-1	308059	11.18				
MW-2	261173	11.17				
MW-3	400173	11.17				
MW-4	336486	11.17				
TRIP-BLANK	279370	11.18				
VV1003WBL01	433083	11.17				
VV1003WBS01	506399	11.17				
VV1003WBSD01	481201	11.17				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.