

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137376

Review By	Review On
Supervise By	Supervise On
STD. NAME	STD REF.#
ICAL Standard	MP87388,MP87389,MP87390,MP87391,MP87392,MP87393
ICV Standard	MP87394
CCV Standard	MP87396
ICSA Standard	
CRI Standard	MP87398
LCS Standard	
Chk Standard	MP87395,MP87397,MP87399,MP87403

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/01/25 08:51		mohan	OK
2	S0.2	S0.2	CAL2	10/01/25 08:54		mohan	OK
3	S2.5	S2.5	CAL3	10/01/25 08:56		mohan	OK
4	S5	S5	CAL4	10/01/25 08:58		mohan	OK
5	S7.5	S7.5	CAL5	10/01/25 09:01		mohan	OK
6	S10	S10	CAL6	10/01/25 09:06		mohan	OK
7	ICV53	ICV53	ICV	10/01/25 09:10		mohan	OK
8	ICB53	ICB53	ICB	10/01/25 09:12		mohan	OK
9	CCV67	CCV67	CCV	10/01/25 09:15		mohan	OK
10	CCB67	CCB67	CCB	10/01/25 09:17		mohan	OK
11	CRA	CRA	CRDL	10/01/25 09:20		mohan	OK
12	HighStd	HighStd	HIGH STD	10/01/25 09:22		mohan	OK
13	ChkStd	ChkStd	SAM	10/01/25 09:24		mohan	OK
14	PB169927BL	PB169927BL	MB	10/01/25 09:27		mohan	OK
15	PB169927BS	PB169927BS	LCS	10/01/25 09:30		mohan	OK
16	Q3253-01	ARS20-0009	SAM	10/01/25 09:32		mohan	OK
17	Q3255-01	CONCRETE-1-WC	SAM	10/01/25 09:35		mohan	OK
18	Q3256-01	SOIL-PILE-1-WC	SAM	10/01/25 09:37		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137376

Review By	Review On
Supervise By	Supervise On
STD. NAME	STD REF.#
ICAL Standard	MP87388,MP87389,MP87390,MP87391,MP87392,MP87393
ICV Standard	MP87394
CCV Standard	MP87396
ICSA Standard	
CRI Standard	MP87398
LCS Standard	
Chk Standard	MP87395,MP87397,MP87399,MP87403

19	Q3256-05	SOIL-PILE-2-WC	SAM	10/01/25 09:39	Hg high	mohan	Dilution
20	Q3257-01	ENV1-6FT	SAM	10/01/25 09:42		mohan	OK
21	CCV68	CCV68	CCV	10/01/25 09:45		mohan	OK
22	CCB68	CCB68	CCB	10/01/25 09:47		mohan	OK
23	Q3257-02	ENV2-6FT	SAM	10/01/25 09:50		mohan	OK
24	Q3257-02DUP	ENV2-6FTDUP	DUP	10/01/25 09:52		mohan	OK
25	Q3257-02MS	ENV2-6FTMS	MS	10/01/25 09:54		mohan	OK
26	Q3257-02MSD	ENV2-6FTMSD	MSD	10/01/25 09:56		mohan	OK
27	Q3257-02L	ENV2-6FTL	SD	10/01/25 09:59		mohan	OK
28	Q3257-02A	ENV2-6FTA	PS	10/01/25 10:02		mohan	OK
29	Q3256-05DL	SOIL-PILE-2-WCDL	SAM	10/01/25 10:04	2x for Hg	mohan	Confirms
30	CCV69	CCV69	CCV	10/01/25 10:07		mohan	OK
31	CCB69	CCB69	CCB	10/01/25 10:09		mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137406

Review By	Janvi	Review On	10/3/2025 3:35:00 PM
Supervise By	jaswal	Supervise On	10/7/2025 1:56:16 PM

STD. NAME	STD REF.#
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032
ICV Standard	MP87047,MP87038
CCV Standard	MP87042
ICSA Standard	MP87040,MP87048
CRI Standard	MP87038
LCS Standard	
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/02/25 10:37		Jaswal	OK
2	S1	S1	CAL2	10/02/25 10:42		Jaswal	OK
3	S2	S2	CAL3	10/02/25 10:46		Jaswal	OK
4	S3	S3	CAL4	10/02/25 10:50		Jaswal	OK
5	S4	S4	CAL5	10/02/25 10:54		Jaswal	OK
6	S5	S5	CAL6	10/02/25 10:58		Jaswal	OK
7	ICV01	ICV01	ICV	10/02/25 11:07		Jaswal	OK
8	LLICV01	LLICV01	LLICV	10/02/25 11:11		Jaswal	OK
9	ICB01	ICB01	ICB	10/02/25 11:27		Jaswal	OK
10	CRI01	CRI01	CRDL	10/02/25 11:36		Jaswal	OK
11	ICSA01	ICSA01	ICSA	10/02/25 11:42		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	10/02/25 11:46		Jaswal	OK
13	ICSADL	ICSADL	ICSA	10/02/25 11:50		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	10/02/25 11:54		Jaswal	OK
15	CCV01	CCV01	CCV	10/02/25 11:58		Jaswal	OK
16	CCB01	CCB01	CCB	10/02/25 12:02		Jaswal	OK
17	LR-1	LR-1	HIGH STD	10/02/25 12:06		Jaswal	OK
18	LR-2	LR-2	HIGH STD	10/02/25 12:11		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137406

Review By	Janvi	Review On	10/3/2025 3:35:00 PM
Supervise By	jaswal	Supervise On	10/7/2025 1:56:16 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032		
ICV Standard	MP87047,MP87038		
CCV Standard	MP87042		
ICSA Standard	MP87040,MP87048		
CRI Standard	MP87038		
LCS Standard			
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049		

19	LR-3	LR-3	HIGH STD	10/02/25 12:42		Jaswal	OK
20	Q3255-01	CONCRETE-1-WC	SAM	10/02/25 12:55		Jaswal	OK
21	Q3256-01	SOIL-PILE-1-WC	SAM	10/02/25 13:00		Jaswal	OK
22	Q3256-01DUP	SOIL-PILE-1-WCDUP	DUP	10/02/25 13:04		Jaswal	OK
23	CCV02	CCV02	CCV	10/02/25 13:48		Jaswal	OK
24	CCB02	CCB02	CCB	10/02/25 13:52		Jaswal	OK
25	Q3256-01L	SOIL-PILE-1-WCL	SD	10/02/25 14:07		Jaswal	OK
26	Q3256-01MS	SOIL-PILE-1-WCMS	MS	10/02/25 14:11		Jaswal	OK
27	Q3256-01MSD	SOIL-PILE-1-WCMSD	MSD	10/02/25 14:15		Jaswal	OK
28	Q3256-01A	SOIL-PILE-1-WCA	PS	10/02/25 14:19	0.1ml of M6007 and M6017 in 10ml sample.	Jaswal	OK
29	Q3256-05	SOIL-PILE-2-WC	SAM	10/02/25 14:22		Jaswal	OK
30	Q3257-01	ENV1-6FT	SAM	10/02/25 14:26	Ag is oversaturate	Jaswal	Dilution
31	Q3257-02	ENV2-6FT	SAM	10/02/25 14:30	Ag is oversaturate	Jaswal	Dilution
32	Q3254-01	MW-1	SAM	10/02/25 14:34		Jaswal	OK
33	Q3254-02	MW-1	SAM	10/02/25 14:38		Jaswal	OK
34	Q3254-03	MW-2	SAM	10/02/25 14:43		Jaswal	OK
35	CCV03	CCV03	CCV	10/02/25 14:47		Jaswal	OK
36	CCB03	CCB03	CCB	10/02/25 14:53		Jaswal	OK
37	Q3254-04	MW-2	SAM	10/02/25 14:57		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137406

Review By	Janvi	Review On	10/3/2025 3:35:00 PM
Supervise By	jaswal	Supervise On	10/7/2025 1:56:16 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032		
ICV Standard	MP87047,MP87038		
CCV Standard	MP87042		
ICSA Standard	MP87040,MP87048		
CRI Standard	MP87038		
LCS Standard			
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049		

38	Q3254-05	MW-3	SAM	10/02/25 15:01		Jaswal	OK
39	Q3254-06	MW-3	SAM	10/02/25 15:05		Jaswal	OK
40	Q3254-07	MW-4	SAM	10/02/25 15:09		Jaswal	OK
41	Q3254-07DUP	MW-4DUP	DUP	10/02/25 15:13		Jaswal	OK
42	Q3254-07L	MW-4L	SD	10/02/25 15:18		Jaswal	OK
43	Q3254-07MS	MW-4MS	MS	10/02/25 15:22		Jaswal	OK
44	Q3254-07MSD	MW-4MSD	MSD	10/02/25 15:26		Jaswal	OK
45	Q3254-07A	MW-4A	PS	10/02/25 15:30	0.1ml of M6007 and M6017 in 10ml sample.	Jaswal	OK
46	Q3254-08	MW-4	SAM	10/02/25 15:34		Jaswal	OK
47	CCV04	CCV04	CCV	10/02/25 15:54		Jaswal	OK
48	CCB04	CCB04	CCB	10/02/25 16:05		Jaswal	OK
49	Q3245-24	12-1	SAM	10/02/25 16:47		Jaswal	OK
50	Q3245-16	12-8	SAM	10/02/25 16:51		Jaswal	OK
51	Q3245-08	13-1	SAM	10/02/25 16:55		Jaswal	OK
52	Q3245-24DUP	12-1DUP	DUP	10/02/25 16:59		Jaswal	OK
53	Q3245-24L	12-1L	SD	10/02/25 17:04		Jaswal	OK
54	Q3245-24MS	12-1MS	MS	10/02/25 17:21		Jaswal	OK
55	CCV05	CCV05	CCV	10/02/25 17:31		Jaswal	OK
56	CCB05	CCB05	CCB	10/02/25 17:35		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137406

Review By	Janvi	Review On	10/3/2025 3:35:00 PM
Supervise By	jaswal	Supervise On	10/7/2025 1:56:16 PM

STD. NAME	STD REF.#
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032
ICV Standard	MP87047,MP87038
CCV Standard	MP87042
ICSA Standard	MP87040,MP87048
CRI Standard	MP87038
LCS Standard	
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049

57	Q3245-24MSD	12-1MSD	MSD	10/02/25 17:45		Jaswal	OK
58	Q3245-24A	12-1A	PS	10/02/25 17:49	0.1ml of M6007 and M6017 in 10ml sample.	Jaswal	OK
59	Q2409-05	COP-SOIL-PILE	SAM	10/02/25 17:53		Jaswal	OK
60	Q2409-05DUP	COP-SOIL-PILEDUP	DUP	10/02/25 17:57		Jaswal	OK
61	Q2409-05L	COP-SOIL-PILEL	SD	10/02/25 18:01		Jaswal	OK
62	Q2409-05MS	COP-SOIL-PILEMS	MS	10/02/25 18:05		Jaswal	OK
63	Q2409-05MSD	COP-SOIL-PILEMSD	MSD	10/02/25 18:09		Jaswal	OK
64	Q2409-05A	COP-SOIL-PILEA	PS	10/02/25 18:13	0.1ml of M6007 and M6017 in 10ml sample.	Jaswal	OK
65	CCV06	CCV06	CCV	10/02/25 18:17		Jaswal	OK
66	CCB06	CCB06	CCB	10/02/25 18:21		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137426

Review By	Janvi	Review On	10/7/2025 3:33:56 PM
Supervise By	jaswal	Supervise On	10/7/2025 4:45:27 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032		
ICV Standard	MP87047,MP87038		
CCV Standard	MP87042		
ICSA Standard	MP87040,MP87048		
CRI Standard	MP87038		
LCS Standard			
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/03/25 11:53		Jaswal	OK
2	S1	S1	CAL2	10/03/25 11:58		Jaswal	OK
3	S2	S2	CAL3	10/03/25 12:02		Jaswal	OK
4	S3	S3	CAL4	10/03/25 12:06		Jaswal	OK
5	S4	S4	CAL5	10/03/25 12:10		Jaswal	OK
6	S5	S5	CAL6	10/03/25 12:14		Jaswal	OK
7	ICV01	ICV01	ICV	10/03/25 12:18		Jaswal	OK
8	LLICV01	LLICV01	LLICV	10/03/25 13:05		Jaswal	OK
9	ICB01	ICB01	ICB	10/03/25 13:12		Jaswal	OK
10	CRI01	CRI01	CRDL	10/03/25 13:16		Jaswal	OK
11	ICSA01	ICSA01	ICSA	10/03/25 13:20		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	10/03/25 13:25		Jaswal	OK
13	ICSADL	ICSADL	ICSA	10/03/25 13:29		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	10/03/25 13:33		Jaswal	OK
15	CCV01	CCV01	CCV	10/03/25 13:37		Jaswal	OK
16	CCB01	CCB01	CCB	10/03/25 13:41		Jaswal	OK
17	LR-1	LR-1	HIGH STD	10/03/25 13:45		Jaswal	OK
18	LR-2	LR-2	HIGH STD	10/03/25 13:50		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137426

Review By	Janvi	Review On	10/7/2025 3:33:56 PM
Supervise By	jaswal	Supervise On	10/7/2025 4:45:27 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032		
ICV Standard	MP87047,MP87038		
CCV Standard	MP87042		
ICSA Standard	MP87040,MP87048		
CRI Standard	MP87038		
LCS Standard			
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049		

19	PB169935BS	PB169935BS	LCS	10/03/25 14:03		Jaswal	OK
20	PB169937BL	PB169937BL	MB	10/03/25 14:07		Jaswal	OK
21	PB169937BS	PB169937BS	LCS	10/03/25 14:12		Jaswal	OK
22	PB169958BL	PB169958BL	MB	10/03/25 14:16		Jaswal	OK
23	PB169958BS	PB169958BS	LCS	10/03/25 14:20		Jaswal	OK
24	PB169963BL	PB169963BL	MB	10/03/25 14:24		Jaswal	OK
25	PB169935BL	PB169935BL	MB	10/03/25 14:33		Jaswal	OK
26	LR-3	LR-3	HIGH STD	10/03/25 14:49		Jaswal	OK
27	CCV02	CCV02	CCV	10/03/25 15:00		Jaswal	OK
28	CCB02	CCB02	CCB	10/03/25 15:04		Jaswal	OK
29	PB169963BS	PB169963BS	LCS	10/03/25 15:08		Jaswal	OK
30	PB169960BL	PB169960BL	MB	10/03/25 15:12		Jaswal	OK
31	PB169960BS	PB169960BS	LCS	10/03/25 15:17		Jaswal	OK
32	PB169959BL	PB169959BL	MB	10/03/25 15:21		Jaswal	OK
33	PB169959BS	PB169959BS	LCS	10/03/25 15:25		Jaswal	OK
34	Q3258-06	COMPOSITE	SAM	10/03/25 15:29		Jaswal	OK
35	Q3263-01	251818	SAM	10/03/25 15:34		Jaswal	OK
36	Q3263-02	266380	SAM	10/03/25 15:38		Jaswal	OK
37	Q3263-02DUP	266380DUP	DUP	10/03/25 15:42		Jaswal	OK
38	Q3263-02L	266380L	SD	10/03/25 15:46		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137426

Review By	Janvi	Review On	10/7/2025 3:33:56 PM
Supervise By	jaswal	Supervise On	10/7/2025 4:45:27 PM

STD. NAME	STD REF.#
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032
ICV Standard	MP87047,MP87038
CCV Standard	MP87042
ICSA Standard	MP87040,MP87048
CRI Standard	MP87038
LCS Standard	
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049

39	CCV03	CCV03	CCV	10/03/25 15:51		Jaswal	OK
40	CCB03	CCB03	CCB	10/03/25 16:00		Jaswal	OK
41	Q3263-02MS	266380MS	MS	10/03/25 16:04		Jaswal	OK
42	Q3263-02MSD	266380MSD	MSD	10/03/25 16:08		Jaswal	OK
43	Q3263-02A	266380A	PS	10/03/25 16:12	0.1 ml of each m6008 and m6017 in 10ml sample.	Jaswal	OK
44	Q3269-01	EG1-TP1	SAM	10/03/25 16:16		Jaswal	OK
45	Q3269-04	EG1-TP2	SAM	10/03/25 16:20		Jaswal	OK
46	Q3262-01	PAD-100125	SAM	10/03/25 16:24		Jaswal	OK
47	Q3266-01	ENV-3-6FT	SAM	10/03/25 16:29		Jaswal	OK
48	Q3266-02	ENV-4-6FT	SAM	10/03/25 16:33	NOT USE	Jaswal	Not Ok
49	Q3266-02DUP	ENV-4-6FTDUP	DUP	10/03/25 16:37	NOT USE	Jaswal	Not Ok
50	Q3266-02L	ENV-4-6FTL	SD	10/03/25 16:41	NOT USE	Jaswal	Not Ok
51	CCV04	CCV04	CCV	10/03/25 16:45		Jaswal	OK
52	CCB04	CCB04	CCB	10/03/25 16:49		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137447

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/07/25 10:29		Jaswal	OK
2	S1	S1	CAL2	10/07/25 10:34		Jaswal	OK
3	S2	S2	CAL3	10/07/25 10:38		Jaswal	OK
4	S3	S3	CAL4	10/07/25 10:42		Jaswal	OK
5	S4	S4	CAL5	10/07/25 10:46		Jaswal	OK
6	S5	S5	CAL6	10/07/25 10:50		Jaswal	OK
7	ICV01	ICV01	ICV	10/07/25 10:54		Jaswal	OK
8	LLICV01	LLICV01	LLICV	10/07/25 12:01		Jaswal	OK
9	ICB01	ICB01	ICB	10/07/25 12:06		Jaswal	OK
10	CRI01	CRI01	CRDL	10/07/25 12:10		Jaswal	OK
11	ICSA01	ICSA01	ICSA	10/07/25 12:14		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	10/07/25 12:37		Jaswal	OK
13	ICSADL	ICSADL	ICSA	10/07/25 12:41		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	10/07/25 12:45		Jaswal	OK
15	CCV01	CCV01	CCV	10/07/25 12:50		Jaswal	OK
16	CCB01	CCB01	CCB	10/07/25 12:58		Jaswal	OK
17	Q3283-01	001-WILLETS-PT-BL	SAM	10/07/25 13:21		Jaswal	OK
18	Q3283-02	002-35TH-AVE(SEP)	SAM	10/07/25 13:25		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137447

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

19	Q3283-02DUP	002-35TH-AVE(SEP)	DUP	10/07/25 13:29		Jaswal	OK
20	Q3283-02L	002-35TH-AVE(SEP)	SD	10/07/25 13:33		Jaswal	OK
21	Q3283-02MS	002-35TH-AVE(SEP)	MS	10/07/25 13:38		Jaswal	OK
22	Q3283-02MSD	002-35TH-AVE(SEP)	MSD	10/07/25 13:42		Jaswal	OK
23	Q3283-02A	002-35TH-AVE(SEP)	PS	10/07/25 13:46		Jaswal	OK
24	LR-1	LR-1	HIGH STD	10/07/25 13:50		Jaswal	OK
25	LR-3	LR-3	HIGH STD	10/07/25 13:59		Jaswal	OK
26	LR-2	LR-2	HIGH STD	10/07/25 14:10		Jaswal	OK
27	CCV02	CCV02	CCV	10/07/25 14:19		Jaswal	OK
28	CCB02	CCB02	CCB	10/07/25 14:24		Jaswal	OK
29	Q3270-01	SG-1-TEST-PIT	SAM	10/07/25 14:28		Jaswal	OK
30	Q3271-01	72-11929	SAM	10/07/25 14:32		Jaswal	OK
31	Q3271-01L	72-11929L	SD	10/07/25 14:40		Jaswal	OK
32	Q3271-01MS	72-11929MS	MS	10/07/25 14:44		Jaswal	OK
33	Q3271-01MSD	72-11929MSD	MSD	10/07/25 14:48		Jaswal	OK
34	Q3271-01A	72-11929A	PS	10/07/25 14:52		Jaswal	OK
35	Q3272-01	VNJ-235	SAM	10/07/25 14:56		Jaswal	OK
36	Q3272-03	R7	SAM	10/07/25 15:00	Ca high	Jaswal	Dilution
37	Q3277-01	WC1	SAM	10/07/25 15:04		Jaswal	OK
38	CCV03	CCV03	CCV	10/07/25 15:08		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137447

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

39	CCB03	CCB03	CCB	10/07/25 15:12		Jaswal	OK
40	Q3269-03	EG1-TP1	SAM	10/07/25 15:17		Jaswal	OK
41	Q3269-06	EG1-TP2	SAM	10/07/25 15:21		Jaswal	OK
42	Q3281-01	001-WILLETS-PT-BL	SAM	10/07/25 15:26		Jaswal	OK
43	Q3281-02	002-35TH-AVE(OCT)	SAM	10/07/25 15:30		Jaswal	OK
44	Q3272-03DL	R7DL	SAM	10/07/25 15:34	25x for Ca	Jaswal	Confirms
45	Q3271-01DUP	72-11929DUP	DUP	10/07/25 15:38		Jaswal	OK
46	PB169982BL	PB169982BL	MB	10/07/25 15:43		Jaswal	OK
47	PB169982BS	PB169982BS	LCS	10/07/25 15:48		Jaswal	OK
48	PB169979BL	PB169979BL	MB	10/07/25 15:52		Jaswal	OK
49	PB169979BS	PB169979BS	LCS	10/07/25 15:56		Jaswal	OK
50	CCV04	CCV04	CCV	10/07/25 16:00		Jaswal	OK
51	CCB04	CCB04	CCB	10/07/25 16:04		Jaswal	OK
52	PB169996BS	PB169996BS	LCS	10/07/25 16:08		Jaswal	OK
53	Q3266-02	ENV-4-6FT	SAM	10/07/25 16:13		Jaswal	OK
54	Q3266-02DUP	ENV-4-6FTDUP	DUP	10/07/25 16:17		Jaswal	OK
55	Q3266-02L	ENV-4-6FTL	SD	10/07/25 16:20		Jaswal	OK
56	Q3266-02MS	ENV-4-6FTMS	MS	10/07/25 16:25		Jaswal	OK
57	Q3266-02MSD	ENV-4-6FTMSD	MSD	10/07/25 16:29		Jaswal	OK
58	Q3266-02A	ENV-4-6FTA	PS	10/07/25 16:33		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137447

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

59	Q3258-06DL	COMPOSITEDL	SAM	10/07/25 16:37	NOT USE	Jaswal	Not Ok
60	PB169971TB	PB169971TB	MB	10/07/25 16:41		Jaswal	OK
61	Q3262-03	MOO-25-0280	SAM	10/07/25 16:45		Jaswal	OK
62	CCV05	CCV05	CCV	10/07/25 16:50		Jaswal	OK
63	CCB05	CCB05	CCB	10/07/25 17:01		Jaswal	OK
64	Q3262-06	MOO-25-0281-82	SAM	10/07/25 17:05		Jaswal	OK
65	Q3262-09	279887	SAM	10/07/25 17:10		Jaswal	OK
66	Q3277-02	WC1	SAM	10/07/25 17:14		Jaswal	OK
67	Q3289-02	VNJ-240	SAM	10/07/25 17:19		Jaswal	OK
68	Q3289-04	VNJ-261	SAM	10/07/25 17:23		Jaswal	OK
69	Q3289-06	72-11966	SAM	10/07/25 17:27		Jaswal	OK
70	Q3289-06DUP	72-11966DUP	DUP	10/07/25 17:32		Jaswal	OK
71	Q3289-06L	72-11966L	SD	10/07/25 17:36		Jaswal	OK
72	Q3289-06MS	72-11966MS	MS	10/07/25 17:40		Jaswal	OK
73	Q3289-06MSD	72-11966MSD	MSD	10/07/25 17:44		Jaswal	OK
74	CCV06	CCV06	CCV	10/07/25 17:49		Jaswal	OK
75	CCB06	CCB06	CCB	10/07/25 17:59		Jaswal	OK
76	Q3289-06A	72-11966A	PS	10/07/25 18:03		Jaswal	OK
77	Q3287-02	TSS-25-0001-0002	SAM	10/07/25 18:07	Confirm Wt/vol	Jaswal	OK
78	PB169996BL	PB169996BL	MB	10/07/25 18:11		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137447

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

79	PB169983BL	PB169983BL	MB	10/07/25 18:15		Jaswal	OK
80	PB169983BS	PB169983BS	LCS	10/07/25 18:20		Jaswal	OK
81	Q3288-01	TR-01-100325	SAM	10/07/25 18:24		Jaswal	OK
82	Q3289-01	VNJ-240	SAM	10/07/25 18:28		Jaswal	OK
83	Q3289-03	VNJ-261	SAM	10/07/25 18:32		Jaswal	OK
84	Q3289-05	72-11966	SAM	10/07/25 18:36		Jaswal	OK
85	Q3290-01	CL-02-100325	SAM	10/07/25 18:40	NOT USE	Jaswal	Not Ok
86	CCV07	CCV07	CCV	10/07/25 18:44		Jaswal	OK
87	CCB07	CCB07	CCB	10/07/25 18:48		Jaswal	OK
88	Q3290-01DUP	CL-02-100325DUP	DUP	10/07/25 18:52	NOT USE	Jaswal	Not Ok
89	Q3290-01L	CL-02-100325L	SD	10/07/25 18:56	NOT USE	Jaswal	Not Ok
90	Q3290-01MS	CL-02-100325MS	MS	10/07/25 19:00	NOT USE	Jaswal	Not Ok
91	Q3290-01MSD	CL-02-100325MSD	MSD	10/07/25 19:04	NOT USE	Jaswal	Not Ok
92	Q3290-01A	CL-02-100325A	PS	10/07/25 19:08	NOT USE	Jaswal	Not Ok
93	Q3257-01DL	ENV1-6FTDL	SAM	10/07/25 19:12	5x for Ag	Jaswal	Confirms
94	Q3257-02DL	ENV2-6FTDL	SAM	10/07/25 19:16	5x for Ag	Jaswal	Confirms
95	CCV08	CCV08	CCV	10/07/25 19:20	CCV fail for Na	Jaswal	OK
96	CCB08	CCB08	CCB	10/07/25 19:24		Jaswal	OK