

LAB CHRONICLE

OrderID:	Q3257	OrderDate:	9/30/2025 2:40:01 PM
Client:	CDM Smith	Project:	MWCO CJO WTP Edison 295110
Contact:	Marcie Ann Encinas	Location:	D31,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3257-01	ENV1-6FT	SOIL			09/30/25			09/30/25
			Mercury	7471B		09/30/25	10/01/25	
			Metals ICP-TAL	6010D		10/01/25	10/02/25	
			Metals ICP-TAL	6010D		10/01/25	10/07/25	
Q3257-02	ENV2-6FT	SOIL			09/29/25			09/30/25
			Mercury	7471B		09/30/25	10/01/25	
			Metals ICP-TAL	6010D		10/01/25	10/02/25	
			Metals ICP-TAL	6010D		10/01/25	10/07/25	

Hit Summary Sheet SW-846

SDG No.: Q3257 **Order ID:** Q3257
Client: CDM Smith **Project ID:** MWCO CJO WTP Edison 295110

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : ENV1-6FT								
Q3257-01	ENV1-6FT	SOIL	Aluminum	15700		0.88	5.25	mg/Kg
Q3257-01	ENV1-6FT	SOIL	Arsenic	9.13		0.20	1.05	mg/Kg
Q3257-01	ENV1-6FT	SOIL	Barium	184		0.77	5.25	mg/Kg
Q3257-01	ENV1-6FT	SOIL	Beryllium	2.36		0.026	0.32	mg/Kg
Q3257-01	ENV1-6FT	SOIL	Cadmium	0.37		0.025	0.32	mg/Kg
Q3257-01	ENV1-6FT	SOIL	Calcium	3090		11.7	105	mg/Kg
Q3257-01	ENV1-6FT	SOIL	Chromium	29.1		0.049	0.53	mg/Kg
Q3257-01	ENV1-6FT	SOIL	Cobalt	20.7		0.11	1.57	mg/Kg
Q3257-01	ENV1-6FT	SOIL	Copper	36.9		0.23	1.05	mg/Kg
Q3257-01	ENV1-6FT	SOIL	Iron	29900		4.19	5.25	mg/Kg
Q3257-01	ENV1-6FT	SOIL	Lead	54.3		0.14	0.63	mg/Kg
Q3257-01	ENV1-6FT	SOIL	Magnesium	6490		12.6	105	mg/Kg
Q3257-01	ENV1-6FT	SOIL	Manganese	1770		0.15	1.05	mg/Kg
Q3257-01	ENV1-6FT	SOIL	Mercury	0.048		0.0090	0.015	mg/Kg
Q3257-01	ENV1-6FT	SOIL	Nickel	36.0		0.14	2.10	mg/Kg
Q3257-01	ENV1-6FT	SOIL	Potassium	1370		29.1	105	mg/Kg
Q3257-01	ENV1-6FT	SOIL	Silver	1.40	JD	0.63	2.62	mg/Kg
Q3257-01	ENV1-6FT	SOIL	Sodium	79.8	J	18.7	105	mg/Kg
Q3257-01	ENV1-6FT	SOIL	Vanadium	42.1		0.26	2.10	mg/Kg
Q3257-01	ENV1-6FT	SOIL	Zinc	102		0.24	2.10	mg/Kg
Client ID : ENV2-6FT								
Q3257-02	ENV2-6FT	SOIL	Aluminum	7220		0.83	4.92	mg/Kg
Q3257-02	ENV2-6FT	SOIL	Arsenic	6.03		0.19	0.98	mg/Kg
Q3257-02	ENV2-6FT	SOIL	Barium	211		0.72	4.92	mg/Kg
Q3257-02	ENV2-6FT	SOIL	Beryllium	1.99		0.025	0.30	mg/Kg
Q3257-02	ENV2-6FT	SOIL	Cadmium	1.12		0.024	0.30	mg/Kg
Q3257-02	ENV2-6FT	SOIL	Calcium	3450		10.9	98.4	mg/Kg
Q3257-02	ENV2-6FT	SOIL	Chromium	41.3		0.046	0.49	mg/Kg
Q3257-02	ENV2-6FT	SOIL	Cobalt	22.2		0.098	1.48	mg/Kg
Q3257-02	ENV2-6FT	SOIL	Copper	25.6		0.22	0.98	mg/Kg
Q3257-02	ENV2-6FT	SOIL	Iron	30300		3.93	4.92	mg/Kg
Q3257-02	ENV2-6FT	SOIL	Lead	28.5		0.13	0.59	mg/Kg
Q3257-02	ENV2-6FT	SOIL	Magnesium	6010		11.8	98.4	mg/Kg
Q3257-02	ENV2-6FT	SOIL	Manganese	1570		0.14	0.98	mg/Kg
Q3257-02	ENV2-6FT	SOIL	Nickel	59.1		0.13	1.97	mg/Kg
Q3257-02	ENV2-6FT	SOIL	Potassium	879		27.3	98.4	mg/Kg
Q3257-02	ENV2-6FT	SOIL	Silver	1.04	JD	0.59	2.46	mg/Kg

Hit Summary Sheet
SW-846

SDG No.: Q3257

Order ID: Q3257

Client: CDM Smith

Project ID: MWCO CJO WTP Edison 295110

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q3257-02	ENV2-6FT	SOIL	Sodium	151		17.5	98.4	mg/Kg
Q3257-02	ENV2-6FT	SOIL	Vanadium	32.4		0.25	1.97	mg/Kg
Q3257-02	ENV2-6FT	SOIL	Zinc	57.9		0.23	1.97	mg/Kg



SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	09/30/25
Project:	MWCO CJO WTP Edison 295110	Date Received:	09/30/25
Client Sample ID:	ENV1-6FT	SDG No.:	Q3257
Lab Sample ID:	Q3257-01	Matrix:	SOIL
Level (low/med):	low	% Solid:	84.3

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units(Dry Weight)	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	15700		1	0.88	5.25	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050
7440-36-0	Antimony	0.23	UN	1	0.23	2.62	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050
7440-38-2	Arsenic	9.13		1	0.20	1.05	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050
7440-39-3	Barium	184		1	0.77	5.25	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050
7440-41-7	Beryllium	2.36		1	0.026	0.32	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050
7440-43-9	Cadmium	0.37		1	0.025	0.32	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050
7440-70-2	Calcium	3090		1	11.7	105	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050
7440-47-3	Chromium	29.1	N*	1	0.049	0.53	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050
7440-48-4	Cobalt	20.7		1	0.11	1.57	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050
7440-50-8	Copper	36.9		1	0.23	1.05	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050
7439-89-6	Iron	29900	*	1	4.19	5.25	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050
7439-92-1	Lead	54.3	N	1	0.14	0.63	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050
7439-95-4	Magnesium	6490		1	12.6	105	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050
7439-96-5	Manganese	1770		1	0.15	1.05	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050
7439-97-6	Mercury	0.048		1	0.0090	0.015	mg/Kg	09/30/25 16:30	10/01/25 09:42	7471B	
7440-02-0	Nickel	36.0		1	0.14	2.10	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050
7440-09-7	Potassium	1370		1	29.1	105	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050
7782-49-2	Selenium	0.27	U	1	0.27	1.05	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050
7440-22-4	Silver	1.40	JDN	5	0.63	2.62	mg/Kg	10/01/25 10:40	10/07/25 19:12	6010D	SW3050
7440-23-5	Sodium	79.8	J	1	18.7	105	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050
7440-28-0	Thallium	0.24	U	1	0.24	2.10	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050
7440-62-2	Vanadium	42.1		1	0.26	2.10	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050
7440-66-6	Zinc	102		1	0.24	2.10	mg/Kg	10/01/25 10:40	10/02/25 14:26	6010D	SW3050

Color Before:	Brown	Clarity Before:	Texture:	Medium
Color After:	Yellow	Clarity After:	Artifacts:	No
Comments:	METALS-TAL			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	CDM Smith	Date Collected:	09/29/25
Project:	MWCO CJO WTP Edison 295110	Date Received:	09/30/25
Client Sample ID:	ENV2-6FT	SDG No.:	Q3257
Lab Sample ID:	Q3257-02	Matrix:	SOIL
Level (low/med):	low	% Solid:	84.7

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units(Dry Weight)	Prep Date	Date Ana.	Ana Met.	Prep Met.
7429-90-5	Aluminum	7220		1	0.83	4.92	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050
7440-36-0	Antimony	0.22	UN	1	0.22	2.46	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050
7440-38-2	Arsenic	6.03		1	0.19	0.98	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050
7440-39-3	Barium	211		1	0.72	4.92	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050
7440-41-7	Beryllium	1.99		1	0.025	0.30	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050
7440-43-9	Cadmium	1.12		1	0.024	0.30	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050
7440-70-2	Calcium	3450		1	10.9	98.4	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050
7440-47-3	Chromium	41.3	N*	1	0.046	0.49	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050
7440-48-4	Cobalt	22.2		1	0.098	1.48	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050
7440-50-8	Copper	25.6		1	0.22	0.98	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050
7439-89-6	Iron	30300	*	1	3.93	4.92	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050
7439-92-1	Lead	28.5	N	1	0.13	0.59	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050
7439-95-4	Magnesium	6010		1	11.8	98.4	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050
7439-96-5	Manganese	1570		1	0.14	0.98	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050
7439-97-6	Mercury	0.0090	U	1	0.0090	0.016	mg/Kg	09/30/25 16:30	10/01/25 09:50	7471B	
7440-02-0	Nickel	59.1		1	0.13	1.97	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050
7440-09-7	Potassium	879		1	27.3	98.4	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050
7782-49-2	Selenium	0.26	U	1	0.26	0.98	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050
7440-22-4	Silver	1.04	JDN	5	0.59	2.46	mg/Kg	10/01/25 10:40	10/07/25 19:16	6010D	SW3050
7440-23-5	Sodium	151		1	17.5	98.4	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050
7440-28-0	Thallium	0.23	U	1	0.23	1.97	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050
7440-62-2	Vanadium	32.4		1	0.25	1.97	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050
7440-66-6	Zinc	57.9		1	0.23	1.97	mg/Kg	10/01/25 10:40	10/02/25 14:30	6010D	SW3050

Color Before:	Brown	Clarity Before:	Texture:	Medium
Color After:	Yellow	Clarity After:	Artifacts:	No
Comments:	METALS-TAL			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV53	Mercury	4.17	4.0	104	90 - 110	CV	10/01/2025	09:10	LB137376

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV67	Mercury	5.09	5.0	102	90 - 110	CV	10/01/2025	09:15	LB137376
CCV68	Mercury	5.12	5.0	102	90 - 110	CV	10/01/2025	09:45	LB137376
CCV69	Mercury	5.05	5.0	101	90 - 110	CV	10/01/2025	10:07	LB137376

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7880	8000	99	90 - 110	P	10/02/2025	11:07	LB137406
	Antimony	4080	4000	102	90 - 110	P	10/02/2025	11:07	LB137406
	Arsenic	4010	4000	100	90 - 110	P	10/02/2025	11:07	LB137406
	Barium	7650	8000	96	90 - 110	P	10/02/2025	11:07	LB137406
	Beryllium	201	200	101	90 - 110	P	10/02/2025	11:07	LB137406
	Cadmium	1970	2000	99	90 - 110	P	10/02/2025	11:07	LB137406
	Calcium	19500	20000	98	90 - 110	P	10/02/2025	11:07	LB137406
	Chromium	824	800	103	90 - 110	P	10/02/2025	11:07	LB137406
	Cobalt	1970	2000	99	90 - 110	P	10/02/2025	11:07	LB137406
	Copper	1010	1000	101	90 - 110	P	10/02/2025	11:07	LB137406
	Iron	3990	4000	100	90 - 110	P	10/02/2025	11:07	LB137406
	Lead	3870	4000	97	90 - 110	P	10/02/2025	11:07	LB137406
	Magnesium	19800	20000	99	90 - 110	P	10/02/2025	11:07	LB137406
	Manganese	1940	2000	97	90 - 110	P	10/02/2025	11:07	LB137406
	Nickel	1980	2000	99	90 - 110	P	10/02/2025	11:07	LB137406
	Potassium	19600	20000	98	90 - 110	P	10/02/2025	11:07	LB137406
	Selenium	4040	4000	101	90 - 110	P	10/02/2025	11:07	LB137406
	Silver	987	1000	99	90 - 110	P	10/02/2025	11:07	LB137406
	Sodium	19000	20000	95	90 - 110	P	10/02/2025	11:07	LB137406
	Thallium	3980	4000	100	90 - 110	P	10/02/2025	11:07	LB137406
	Vanadium	1960	2000	98	90 - 110	P	10/02/2025	11:07	LB137406
	Zinc	2010	2000	100	90 - 110	P	10/02/2025	11:07	LB137406

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	120	100	120	80 - 120	P	10/02/2025	11:11	LB137406
	Antimony	52.3	50.0	104	80 - 120	P	10/02/2025	11:11	LB137406
	Arsenic	21.7	20.0	108	80 - 120	P	10/02/2025	11:11	LB137406
	Barium	113	100	113	80 - 120	P	10/02/2025	11:11	LB137406
	Beryllium	6.80	6.0	113	80 - 120	P	10/02/2025	11:11	LB137406
	Cadmium	6.03	6.0	100	80 - 120	P	10/02/2025	11:11	LB137406
	Calcium	2110	2000	106	80 - 120	P	10/02/2025	11:11	LB137406
	Chromium	11.2	10.0	112	80 - 120	P	10/02/2025	11:11	LB137406
	Cobalt	30.1	30.0	100	80 - 120	P	10/02/2025	11:11	LB137406
	Copper	22.2	20.0	111	80 - 120	P	10/02/2025	11:11	LB137406
	Iron	104	100	104	80 - 120	P	10/02/2025	11:11	LB137406
	Lead	13.2	12.0	110	80 - 120	P	10/02/2025	11:11	LB137406
	Magnesium	2220	2000	111	80 - 120	P	10/02/2025	11:11	LB137406
	Manganese	21.5	20.0	108	80 - 120	P	10/02/2025	11:11	LB137406
	Nickel	40.4	40.0	101	80 - 120	P	10/02/2025	11:11	LB137406
	Potassium	1770	2000	88	80 - 120	P	10/02/2025	11:11	LB137406
	Selenium	22.8	20.0	114	80 - 120	P	10/02/2025	11:11	LB137406
	Silver	11.0	10.0	110	80 - 120	P	10/02/2025	11:11	LB137406
	Sodium	1690	2000	84	80 - 120	P	10/02/2025	11:11	LB137406
	Thallium	46.1	40.0	115	80 - 120	P	10/02/2025	11:11	LB137406
	Vanadium	42.2	40.0	106	80 - 120	P	10/02/2025	11:11	LB137406
	Zinc	42.4	40.0	106	80 - 120	P	10/02/2025	11:11	LB137406

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9910	10000	99	90 - 110	P	10/02/2025	11:58	LB137406
	Antimony	4950	5000	99	90 - 110	P	10/02/2025	11:58	LB137406
	Arsenic	5100	5000	102	90 - 110	P	10/02/2025	11:58	LB137406
	Barium	9630	10000	96	90 - 110	P	10/02/2025	11:58	LB137406
	Beryllium	267	250	107	90 - 110	P	10/02/2025	11:58	LB137406
	Cadmium	2530	2500	101	90 - 110	P	10/02/2025	11:58	LB137406
	Calcium	24800	25000	99	90 - 110	P	10/02/2025	11:58	LB137406
	Chromium	1040	1000	104	90 - 110	P	10/02/2025	11:58	LB137406
	Cobalt	2490	2500	100	90 - 110	P	10/02/2025	11:58	LB137406
	Copper	1260	1250	101	90 - 110	P	10/02/2025	11:58	LB137406
	Iron	4740	5000	95	90 - 110	P	10/02/2025	11:58	LB137406
	Lead	5080	5000	102	90 - 110	P	10/02/2025	11:58	LB137406
	Magnesium	25100	25000	100	90 - 110	P	10/02/2025	11:58	LB137406
	Manganese	2440	2500	98	90 - 110	P	10/02/2025	11:58	LB137406
	Nickel	2500	2500	100	90 - 110	P	10/02/2025	11:58	LB137406
	Potassium	23300	25000	93	90 - 110	P	10/02/2025	11:58	LB137406
	Selenium	5120	5000	102	90 - 110	P	10/02/2025	11:58	LB137406
	Silver	1260	1250	101	90 - 110	P	10/02/2025	11:58	LB137406
	Sodium	22800	25000	91	90 - 110	P	10/02/2025	11:58	LB137406
	Thallium	5080	5000	102	90 - 110	P	10/02/2025	11:58	LB137406
	Vanadium	2470	2500	99	90 - 110	P	10/02/2025	11:58	LB137406
	Zinc	2520	2500	101	90 - 110	P	10/02/2025	11:58	LB137406
CCV02	Aluminum	9450	10000	94	90 - 110	P	10/02/2025	13:48	LB137406
	Antimony	4830	5000	97	90 - 110	P	10/02/2025	13:48	LB137406
	Arsenic	5200	5000	104	90 - 110	P	10/02/2025	13:48	LB137406
	Barium	9090	10000	91	90 - 110	P	10/02/2025	13:48	LB137406
	Beryllium	246	250	98	90 - 110	P	10/02/2025	13:48	LB137406
	Cadmium	2570	2500	103	90 - 110	P	10/02/2025	13:48	LB137406
	Calcium	24000	25000	96	90 - 110	P	10/02/2025	13:48	LB137406
	Chromium	1070	1000	107	90 - 110	P	10/02/2025	13:48	LB137406
	Cobalt	2460	2500	99	90 - 110	P	10/02/2025	13:48	LB137406
	Copper	1260	1250	101	90 - 110	P	10/02/2025	13:48	LB137406
	Iron	5120	5000	102	90 - 110	P	10/02/2025	13:48	LB137406
	Lead	5240	5000	105	90 - 110	P	10/02/2025	13:48	LB137406

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24700	25000	99	90 - 110	P	10/02/2025	13:48	LB137406
	Manganese	2330	2500	93	90 - 110	P	10/02/2025	13:48	LB137406
	Nickel	2490	2500	99	90 - 110	P	10/02/2025	13:48	LB137406
	Potassium	23300	25000	93	90 - 110	P	10/02/2025	13:48	LB137406
	Selenium	5320	5000	106	90 - 110	P	10/02/2025	13:48	LB137406
	Silver	1280	1250	102	90 - 110	P	10/02/2025	13:48	LB137406
	Sodium	24800	25000	99	90 - 110	P	10/02/2025	13:48	LB137406
	Thallium	5290	5000	106	90 - 110	P	10/02/2025	13:48	LB137406
	Vanadium	2390	2500	96	90 - 110	P	10/02/2025	13:48	LB137406
	Zinc	2500	2500	100	90 - 110	P	10/02/2025	13:48	LB137406
CCV03	Aluminum	9360	10000	94	90 - 110	P	10/02/2025	14:47	LB137406
	Antimony	4800	5000	96	90 - 110	P	10/02/2025	14:47	LB137406
	Arsenic	5170	5000	104	90 - 110	P	10/02/2025	14:47	LB137406
	Barium	9030	10000	90	90 - 110	P	10/02/2025	14:47	LB137406
	Beryllium	243	250	97	90 - 110	P	10/02/2025	14:47	LB137406
	Cadmium	2560	2500	103	90 - 110	P	10/02/2025	14:47	LB137406
	Calcium	23800	25000	95	90 - 110	P	10/02/2025	14:47	LB137406
	Chromium	1050	1000	105	90 - 110	P	10/02/2025	14:47	LB137406
	Cobalt	2450	2500	98	90 - 110	P	10/02/2025	14:47	LB137406
	Copper	1250	1250	100	90 - 110	P	10/02/2025	14:47	LB137406
	Iron	5080	5000	102	90 - 110	P	10/02/2025	14:47	LB137406
	Lead	5220	5000	104	90 - 110	P	10/02/2025	14:47	LB137406
	Magnesium	24400	25000	98	90 - 110	P	10/02/2025	14:47	LB137406
	Manganese	2320	2500	93	90 - 110	P	10/02/2025	14:47	LB137406
	Nickel	2470	2500	99	90 - 110	P	10/02/2025	14:47	LB137406
	Potassium	24700	25000	99	90 - 110	P	10/02/2025	14:47	LB137406
	Selenium	5300	5000	106	90 - 110	P	10/02/2025	14:47	LB137406
	Silver	1250	1250	100	90 - 110	P	10/02/2025	14:47	LB137406
	Sodium	24400	25000	98	90 - 110	P	10/02/2025	14:47	LB137406
	Thallium	5300	5000	106	90 - 110	P	10/02/2025	14:47	LB137406
	Vanadium	2360	2500	95	90 - 110	P	10/02/2025	14:47	LB137406
	Zinc	2470	2500	99	90 - 110	P	10/02/2025	14:47	LB137406
CCV04	Aluminum	9320	10000	93	90 - 110	P	10/02/2025	15:54	LB137406
	Antimony	4920	5000	98	90 - 110	P	10/02/2025	15:54	LB137406

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith
Contract: CAMP02
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

SDG No.: Q3257
Lab Code: ACE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5460	5000	109	90 - 110	P	10/02/2025	15:54	LB137406
	Barium	9070	10000	91	90 - 110	P	10/02/2025	15:54	LB137406
	Beryllium	234	250	93	90 - 110	P	10/02/2025	15:54	LB137406
	Cadmium	2660	2500	106	90 - 110	P	10/02/2025	15:54	LB137406
	Calcium	23900	25000	96	90 - 110	P	10/02/2025	15:54	LB137406
	Chromium	1070	1000	107	90 - 110	P	10/02/2025	15:54	LB137406
	Cobalt	2480	2500	99	90 - 110	P	10/02/2025	15:54	LB137406
	Copper	1290	1250	103	90 - 110	P	10/02/2025	15:54	LB137406
	Iron	5300	5000	106	90 - 110	P	10/02/2025	15:54	LB137406
	Lead	5270	5000	106	90 - 110	P	10/02/2025	15:54	LB137406
	Magnesium	24100	25000	97	90 - 110	P	10/02/2025	15:54	LB137406
	Manganese	2340	2500	94	90 - 110	P	10/02/2025	15:54	LB137406
	Nickel	2520	2500	101	90 - 110	P	10/02/2025	15:54	LB137406
	Potassium	24200	25000	97	90 - 110	P	10/02/2025	15:54	LB137406
	Selenium	4900	5000	98	90 - 110	P	10/02/2025	15:54	LB137406
	Silver	1290	1250	103	90 - 110	P	10/02/2025	15:54	LB137406
	Sodium	23300	25000	93	90 - 110	P	10/02/2025	15:54	LB137406
	Thallium	5170	5000	103	90 - 110	P	10/02/2025	15:54	LB137406
	Vanadium	2400	2500	96	90 - 110	P	10/02/2025	15:54	LB137406
	Zinc	2510	2500	100	90 - 110	P	10/02/2025	15:54	LB137406
CCV05	Aluminum	9280	10000	93	90 - 110	P	10/02/2025	17:31	LB137406
	Antimony	4750	5000	95	90 - 110	P	10/02/2025	17:31	LB137406
	Arsenic	5060	5000	101	90 - 110	P	10/02/2025	17:31	LB137406
	Barium	9160	10000	92	90 - 110	P	10/02/2025	17:31	LB137406
	Beryllium	239	250	96	90 - 110	P	10/02/2025	17:31	LB137406
	Cadmium	2530	2500	101	90 - 110	P	10/02/2025	17:31	LB137406
	Calcium	23900	25000	96	90 - 110	P	10/02/2025	17:31	LB137406
	Chromium	1020	1000	102	90 - 110	P	10/02/2025	17:31	LB137406
	Cobalt	2430	2500	97	90 - 110	P	10/02/2025	17:31	LB137406
	Copper	1240	1250	99	90 - 110	P	10/02/2025	17:31	LB137406
	Iron	5480	5000	110	90 - 110	P	10/02/2025	17:31	LB137406
	Lead	5140	5000	103	90 - 110	P	10/02/2025	17:31	LB137406
	Magnesium	24100	25000	97	90 - 110	P	10/02/2025	17:31	LB137406
	Manganese	2350	2500	94	90 - 110	P	10/02/2025	17:31	LB137406

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2450	2500	98	90 - 110	P	10/02/2025	17:31	LB137406
	Potassium	22500	25000	90	90 - 110	P	10/02/2025	17:31	LB137406
	Selenium	5150	5000	103	90 - 110	P	10/02/2025	17:31	LB137406
	Silver	1230	1250	99	90 - 110	P	10/02/2025	17:31	LB137406
	Sodium	26000	25000	104	90 - 110	P	10/02/2025	17:31	LB137406
	Thallium	4690	5000	94	90 - 110	P	10/02/2025	17:31	LB137406
	Vanadium	2370	2500	95	90 - 110	P	10/02/2025	17:31	LB137406
	Zinc	2430	2500	97	90 - 110	P	10/02/2025	17:31	LB137406
CCV06	Aluminum	9350	10000	94	90 - 110	P	10/02/2025	18:17	LB137406
	Antimony	4830	5000	96	90 - 110	P	10/02/2025	18:17	LB137406
	Arsenic	5210	5000	104	90 - 110	P	10/02/2025	18:17	LB137406
	Barium	9140	10000	91	90 - 110	P	10/02/2025	18:17	LB137406
	Beryllium	252	250	101	90 - 110	P	10/02/2025	18:17	LB137406
	Cadmium	2590	2500	104	90 - 110	P	10/02/2025	18:17	LB137406
	Calcium	24000	25000	96	90 - 110	P	10/02/2025	18:17	LB137406
	Chromium	1040	1000	104	90 - 110	P	10/02/2025	18:17	LB137406
	Cobalt	2460	2500	98	90 - 110	P	10/02/2025	18:17	LB137406
	Copper	1270	1250	101	90 - 110	P	10/02/2025	18:17	LB137406
	Iron	5380	5000	108	90 - 110	P	10/02/2025	18:17	LB137406
	Lead	5290	5000	106	90 - 110	P	10/02/2025	18:17	LB137406
	Magnesium	24400	25000	98	90 - 110	P	10/02/2025	18:17	LB137406
	Manganese	2360	2500	94	90 - 110	P	10/02/2025	18:17	LB137406
	Nickel	2490	2500	100	90 - 110	P	10/02/2025	18:17	LB137406
	Potassium	24700	25000	99	90 - 110	P	10/02/2025	18:17	LB137406
	Selenium	5010	5000	100	90 - 110	P	10/02/2025	18:17	LB137406
	Silver	1250	1250	100	90 - 110	P	10/02/2025	18:17	LB137406
	Sodium	26000	25000	104	90 - 110	P	10/02/2025	18:17	LB137406
	Thallium	4730	5000	94	90 - 110	P	10/02/2025	18:17	LB137406
	Vanadium	2370	2500	95	90 - 110	P	10/02/2025	18:17	LB137406
	Zinc	2450	2500	98	90 - 110	P	10/02/2025	18:17	LB137406

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7850	8000	98	90 - 110	P	10/03/2025	12:18	LB137426
	Antimony	4030	4000	101	90 - 110	P	10/03/2025	12:18	LB137426
	Arsenic	3940	4000	99	90 - 110	P	10/03/2025	12:18	LB137426
	Barium	7900	8000	99	90 - 110	P	10/03/2025	12:18	LB137426
	Beryllium	199	200	100	90 - 110	P	10/03/2025	12:18	LB137426
	Cadmium	1950	2000	98	90 - 110	P	10/03/2025	12:18	LB137426
	Calcium	19600	20000	98	90 - 110	P	10/03/2025	12:18	LB137426
	Chromium	799	800	100	90 - 110	P	10/03/2025	12:18	LB137426
	Cobalt	1950	2000	98	90 - 110	P	10/03/2025	12:18	LB137426
	Copper	1010	1000	101	90 - 110	P	10/03/2025	12:18	LB137426
	Iron	3950	4000	99	90 - 110	P	10/03/2025	12:18	LB137426
	Lead	3840	4000	96	90 - 110	P	10/03/2025	12:18	LB137426
	Magnesium	19600	20000	98	90 - 110	P	10/03/2025	12:18	LB137426
	Manganese	1980	2000	99	90 - 110	P	10/03/2025	12:18	LB137426
	Nickel	1960	2000	98	90 - 110	P	10/03/2025	12:18	LB137426
	Potassium	19400	20000	97	90 - 110	P	10/03/2025	12:18	LB137426
	Selenium	3960	4000	99	90 - 110	P	10/03/2025	12:18	LB137426
	Silver	960	1000	96	90 - 110	P	10/03/2025	12:18	LB137426
	Sodium	19300	20000	97	90 - 110	P	10/03/2025	12:18	LB137426
	Thallium	3750	4000	94	90 - 110	P	10/03/2025	12:18	LB137426
	Vanadium	1960	2000	98	90 - 110	P	10/03/2025	12:18	LB137426
	Zinc	1980	2000	99	90 - 110	P	10/03/2025	12:18	LB137426

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	114	100	114	80 - 120	P	10/03/2025	13:05	LB137426
	Antimony	53.8	50.0	108	80 - 120	P	10/03/2025	13:05	LB137426
	Arsenic	22.0	20.0	110	80 - 120	P	10/03/2025	13:05	LB137426
	Barium	99.1	100	99	80 - 120	P	10/03/2025	13:05	LB137426
	Beryllium	6.51	6.0	108	80 - 120	P	10/03/2025	13:05	LB137426
	Cadmium	6.05	6.0	101	80 - 120	P	10/03/2025	13:05	LB137426
	Calcium	2060	2000	103	80 - 120	P	10/03/2025	13:05	LB137426
	Chromium	10.4	10.0	104	80 - 120	P	10/03/2025	13:05	LB137426
	Cobalt	30.1	30.0	100	80 - 120	P	10/03/2025	13:05	LB137426
	Copper	21.3	20.0	107	80 - 120	P	10/03/2025	13:05	LB137426
	Iron	92.9	100	93	80 - 120	P	10/03/2025	13:05	LB137426
	Lead	12.8	12.0	107	80 - 120	P	10/03/2025	13:05	LB137426
	Magnesium	2160	2000	108	80 - 120	P	10/03/2025	13:05	LB137426
	Manganese	21.2	20.0	106	80 - 120	P	10/03/2025	13:05	LB137426
	Nickel	41.3	40.0	103	80 - 120	P	10/03/2025	13:05	LB137426
	Potassium	1960	2000	98	80 - 120	P	10/03/2025	13:05	LB137426
	Selenium	19.3	20.0	96	80 - 120	P	10/03/2025	13:05	LB137426
	Silver	10.2	10.0	102	80 - 120	P	10/03/2025	13:05	LB137426
	Sodium	1990	2000	100	80 - 120	P	10/03/2025	13:05	LB137426
	Thallium	41.8	40.0	104	80 - 120	P	10/03/2025	13:05	LB137426
	Vanadium	39.2	40.0	98	80 - 120	P	10/03/2025	13:05	LB137426
	Zinc	42.4	40.0	106	80 - 120	P	10/03/2025	13:05	LB137426

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9960	10000	100	90 - 110	P	10/03/2025	13:37	LB137426
	Antimony	5040	5000	101	90 - 110	P	10/03/2025	13:37	LB137426
	Arsenic	5030	5000	100	90 - 110	P	10/03/2025	13:37	LB137426
	Barium	10000	10000	100	90 - 110	P	10/03/2025	13:37	LB137426
	Beryllium	247	250	99	90 - 110	P	10/03/2025	13:37	LB137426
	Cadmium	2490	2500	100	90 - 110	P	10/03/2025	13:37	LB137426
	Calcium	24800	25000	99	90 - 110	P	10/03/2025	13:37	LB137426
	Chromium	989	1000	99	90 - 110	P	10/03/2025	13:37	LB137426
	Cobalt	2480	2500	99	90 - 110	P	10/03/2025	13:37	LB137426
	Copper	1260	1250	101	90 - 110	P	10/03/2025	13:37	LB137426
	Iron	4900	5000	98	90 - 110	P	10/03/2025	13:37	LB137426
	Lead	4940	5000	99	90 - 110	P	10/03/2025	13:37	LB137426
	Magnesium	24700	25000	99	90 - 110	P	10/03/2025	13:37	LB137426
	Manganese	2500	2500	100	90 - 110	P	10/03/2025	13:37	LB137426
	Nickel	2490	2500	100	90 - 110	P	10/03/2025	13:37	LB137426
	Potassium	24200	25000	97	90 - 110	P	10/03/2025	13:37	LB137426
	Selenium	5050	5000	101	90 - 110	P	10/03/2025	13:37	LB137426
	Silver	1230	1250	98	90 - 110	P	10/03/2025	13:37	LB137426
	Sodium	24400	25000	97	90 - 110	P	10/03/2025	13:37	LB137426
	Thallium	4650	5000	93	90 - 110	P	10/03/2025	13:37	LB137426
	Vanadium	2480	2500	99	90 - 110	P	10/03/2025	13:37	LB137426
	Zinc	2480	2500	99	90 - 110	P	10/03/2025	13:37	LB137426
CCV02	Aluminum	9920	10000	99	90 - 110	P	10/03/2025	15:00	LB137426
	Antimony	4990	5000	100	90 - 110	P	10/03/2025	15:00	LB137426
	Arsenic	4990	5000	100	90 - 110	P	10/03/2025	15:00	LB137426
	Barium	9880	10000	99	90 - 110	P	10/03/2025	15:00	LB137426
	Beryllium	244	250	98	90 - 110	P	10/03/2025	15:00	LB137426
	Cadmium	2460	2500	98	90 - 110	P	10/03/2025	15:00	LB137426
	Calcium	24700	25000	99	90 - 110	P	10/03/2025	15:00	LB137426
	Chromium	984	1000	98	90 - 110	P	10/03/2025	15:00	LB137426
	Cobalt	2450	2500	98	90 - 110	P	10/03/2025	15:00	LB137426
	Copper	1240	1250	99	90 - 110	P	10/03/2025	15:00	LB137426
	Iron	4860	5000	97	90 - 110	P	10/03/2025	15:00	LB137426
	Lead	4890	5000	98	90 - 110	P	10/03/2025	15:00	LB137426

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24400	25000	98	90 - 110	P	10/03/2025	15:00	LB137426
	Manganese	2440	2500	98	90 - 110	P	10/03/2025	15:00	LB137426
	Nickel	2450	2500	98	90 - 110	P	10/03/2025	15:00	LB137426
	Potassium	24400	25000	98	90 - 110	P	10/03/2025	15:00	LB137426
	Selenium	5040	5000	101	90 - 110	P	10/03/2025	15:00	LB137426
	Silver	1230	1250	98	90 - 110	P	10/03/2025	15:00	LB137426
	Sodium	24300	25000	97	90 - 110	P	10/03/2025	15:00	LB137426
	Thallium	4580	5000	92	90 - 110	P	10/03/2025	15:00	LB137426
	Vanadium	2470	2500	99	90 - 110	P	10/03/2025	15:00	LB137426
	Zinc	2450	2500	98	90 - 110	P	10/03/2025	15:00	LB137426
CCV03	Aluminum	9680	10000	97	90 - 110	P	10/03/2025	15:51	LB137426
	Antimony	5140	5000	103	90 - 110	P	10/03/2025	15:51	LB137426
	Arsenic	5080	5000	102	90 - 110	P	10/03/2025	15:51	LB137426
	Barium	10000	10000	100	90 - 110	P	10/03/2025	15:51	LB137426
	Beryllium	239	250	96	90 - 110	P	10/03/2025	15:51	LB137426
	Cadmium	2440	2500	98	90 - 110	P	10/03/2025	15:51	LB137426
	Calcium	24200	25000	97	90 - 110	P	10/03/2025	15:51	LB137426
	Chromium	972	1000	97	90 - 110	P	10/03/2025	15:51	LB137426
	Cobalt	2450	2500	98	90 - 110	P	10/03/2025	15:51	LB137426
	Copper	1250	1250	100	90 - 110	P	10/03/2025	15:51	LB137426
	Iron	5230	5000	105	90 - 110	P	10/03/2025	15:51	LB137426
	Lead	4860	5000	97	90 - 110	P	10/03/2025	15:51	LB137426
	Magnesium	23500	25000	94	90 - 110	P	10/03/2025	15:51	LB137426
	Manganese	2420	2500	97	90 - 110	P	10/03/2025	15:51	LB137426
	Nickel	2450	2500	98	90 - 110	P	10/03/2025	15:51	LB137426
	Potassium	25700	25000	103	90 - 110	P	10/03/2025	15:51	LB137426
	Selenium	5170	5000	103	90 - 110	P	10/03/2025	15:51	LB137426
	Silver	1240	1250	99	90 - 110	P	10/03/2025	15:51	LB137426
	Sodium	27300	25000	109	90 - 110	P	10/03/2025	15:51	LB137426
	Thallium	5250	5000	105	90 - 110	P	10/03/2025	15:51	LB137426
	Vanadium	2450	2500	98	90 - 110	P	10/03/2025	15:51	LB137426
	Zinc	2490	2500	100	90 - 110	P	10/03/2025	15:51	LB137426
CCV04	Aluminum	9420	10000	94	90 - 110	P	10/03/2025	16:45	LB137426
	Antimony	5020	5000	100	90 - 110	P	10/03/2025	16:45	LB137426

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	4930	5000	99	90 - 110	P	10/03/2025	16:45	LB137426
	Barium	9780	10000	98	90 - 110	P	10/03/2025	16:45	LB137426
	Beryllium	234	250	93	90 - 110	P	10/03/2025	16:45	LB137426
	Cadmium	2360	2500	94	90 - 110	P	10/03/2025	16:45	LB137426
	Calcium	23600	25000	94	90 - 110	P	10/03/2025	16:45	LB137426
	Chromium	938	1000	94	90 - 110	P	10/03/2025	16:45	LB137426
	Cobalt	2370	2500	95	90 - 110	P	10/03/2025	16:45	LB137426
	Copper	1210	1250	97	90 - 110	P	10/03/2025	16:45	LB137426
	Iron	5120	5000	102	90 - 110	P	10/03/2025	16:45	LB137426
	Lead	4700	5000	94	90 - 110	P	10/03/2025	16:45	LB137426
	Magnesium	22700	25000	91	90 - 110	P	10/03/2025	16:45	LB137426
	Manganese	2370	2500	95	90 - 110	P	10/03/2025	16:45	LB137426
	Nickel	2370	2500	95	90 - 110	P	10/03/2025	16:45	LB137426
	Potassium	24100	25000	96	90 - 110	P	10/03/2025	16:45	LB137426
	Selenium	5060	5000	101	90 - 110	P	10/03/2025	16:45	LB137426
	Silver	1210	1250	97	90 - 110	P	10/03/2025	16:45	LB137426
	Sodium	25400	25000	102	90 - 110	P	10/03/2025	16:45	LB137426
	Thallium	5250	5000	105	90 - 110	P	10/03/2025	16:45	LB137426
	Vanadium	2390	2500	96	90 - 110	P	10/03/2025	16:45	LB137426
	Zinc	2420	2500	97	90 - 110	P	10/03/2025	16:45	LB137426

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7920	8000	99	90 - 110	P	10/07/2025	10:54	LB137447
	Antimony	4110	4000	103	90 - 110	P	10/07/2025	10:54	LB137447
	Arsenic	4000	4000	100	90 - 110	P	10/07/2025	10:54	LB137447
	Barium	7940	8000	99	90 - 110	P	10/07/2025	10:54	LB137447
	Beryllium	201	200	100	90 - 110	P	10/07/2025	10:54	LB137447
	Cadmium	1980	2000	99	90 - 110	P	10/07/2025	10:54	LB137447
	Calcium	19900	20000	99	90 - 110	P	10/07/2025	10:54	LB137447
	Chromium	817	800	102	90 - 110	P	10/07/2025	10:54	LB137447
	Cobalt	1980	2000	99	90 - 110	P	10/07/2025	10:54	LB137447
	Copper	1020	1000	102	90 - 110	P	10/07/2025	10:54	LB137447
	Iron	4130	4000	103	90 - 110	P	10/07/2025	10:54	LB137447
	Lead	3890	4000	97	90 - 110	P	10/07/2025	10:54	LB137447
	Magnesium	19900	20000	100	90 - 110	P	10/07/2025	10:54	LB137447
	Manganese	1990	2000	100	90 - 110	P	10/07/2025	10:54	LB137447
	Nickel	1980	2000	99	90 - 110	P	10/07/2025	10:54	LB137447
	Potassium	20200	20000	101	90 - 110	P	10/07/2025	10:54	LB137447
	Selenium	4000	4000	100	90 - 110	P	10/07/2025	10:54	LB137447
	Silver	1030	1000	103	90 - 110	P	10/07/2025	10:54	LB137447
	Sodium	20100	20000	100	90 - 110	P	10/07/2025	10:54	LB137447
	Thallium	3720	4000	93	90 - 110	P	10/07/2025	10:54	LB137447
	Vanadium	1980	2000	99	90 - 110	P	10/07/2025	10:54	LB137447
	Zinc	2020	2000	101	90 - 110	P	10/07/2025	10:54	LB137447

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	92.2	100	92	80 - 120	P	10/07/2025	12:01	LB137447
	Antimony	49.4	50.0	99	80 - 120	P	10/07/2025	12:01	LB137447
	Arsenic	18.4	20.0	92	80 - 120	P	10/07/2025	12:01	LB137447
	Barium	95.7	100	96	80 - 120	P	10/07/2025	12:01	LB137447
	Beryllium	6.20	6.0	103	80 - 120	P	10/07/2025	12:01	LB137447
	Cadmium	5.85	6.0	98	80 - 120	P	10/07/2025	12:01	LB137447
	Calcium	2020	2000	101	80 - 120	P	10/07/2025	12:01	LB137447
	Chromium	9.65	10.0	96	80 - 120	P	10/07/2025	12:01	LB137447
	Cobalt	29.0	30.0	97	80 - 120	P	10/07/2025	12:01	LB137447
	Copper	21.9	20.0	110	80 - 120	P	10/07/2025	12:01	LB137447
	Iron	119	100	119	80 - 120	P	10/07/2025	12:01	LB137447
	Lead	11.1	12.0	93	80 - 120	P	10/07/2025	12:01	LB137447
	Magnesium	2070	2000	104	80 - 120	P	10/07/2025	12:01	LB137447
	Manganese	20.8	20.0	104	80 - 120	P	10/07/2025	12:01	LB137447
	Nickel	39.6	40.0	99	80 - 120	P	10/07/2025	12:01	LB137447
	Potassium	1910	2000	95	80 - 120	P	10/07/2025	12:01	LB137447
	Selenium	23.6	20.0	118	80 - 120	P	10/07/2025	12:01	LB137447
	Silver	9.88	10.0	99	80 - 120	P	10/07/2025	12:01	LB137447
	Sodium	1890	2000	94	80 - 120	P	10/07/2025	12:01	LB137447
	Thallium	43.0	40.0	107	80 - 120	P	10/07/2025	12:01	LB137447
	Vanadium	37.3	40.0	93	80 - 120	P	10/07/2025	12:01	LB137447
	Zinc	36.9	40.0	92	80 - 120	P	10/07/2025	12:01	LB137447

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	10100	10000	101	90 - 110	P	10/07/2025	12:50	LB137447
	Antimony	5050	5000	101	90 - 110	P	10/07/2025	12:50	LB137447
	Arsenic	5080	5000	102	90 - 110	P	10/07/2025	12:50	LB137447
	Barium	9980	10000	100	90 - 110	P	10/07/2025	12:50	LB137447
	Beryllium	252	250	101	90 - 110	P	10/07/2025	12:50	LB137447
	Cadmium	2510	2500	100	90 - 110	P	10/07/2025	12:50	LB137447
	Calcium	25200	25000	101	90 - 110	P	10/07/2025	12:50	LB137447
	Chromium	1030	1000	103	90 - 110	P	10/07/2025	12:50	LB137447
	Cobalt	2490	2500	100	90 - 110	P	10/07/2025	12:50	LB137447
	Copper	1270	1250	101	90 - 110	P	10/07/2025	12:50	LB137447
	Iron	5200	5000	104	90 - 110	P	10/07/2025	12:50	LB137447
	Lead	4980	5000	100	90 - 110	P	10/07/2025	12:50	LB137447
	Magnesium	25200	25000	101	90 - 110	P	10/07/2025	12:50	LB137447
	Manganese	2490	2500	100	90 - 110	P	10/07/2025	12:50	LB137447
	Nickel	2490	2500	100	90 - 110	P	10/07/2025	12:50	LB137447
	Potassium	25500	25000	102	90 - 110	P	10/07/2025	12:50	LB137447
	Selenium	5150	5000	103	90 - 110	P	10/07/2025	12:50	LB137447
	Silver	1280	1250	103	90 - 110	P	10/07/2025	12:50	LB137447
	Sodium	25600	25000	102	90 - 110	P	10/07/2025	12:50	LB137447
	Thallium	4870	5000	97	90 - 110	P	10/07/2025	12:50	LB137447
	Vanadium	2520	2500	101	90 - 110	P	10/07/2025	12:50	LB137447
	Zinc	2600	2500	104	90 - 110	P	10/07/2025	12:50	LB137447
CCV02	Aluminum	9930	10000	99	90 - 110	P	10/07/2025	14:19	LB137447
	Antimony	5050	5000	101	90 - 110	P	10/07/2025	14:19	LB137447
	Arsenic	5080	5000	102	90 - 110	P	10/07/2025	14:19	LB137447
	Barium	9770	10000	98	90 - 110	P	10/07/2025	14:19	LB137447
	Beryllium	250	250	100	90 - 110	P	10/07/2025	14:19	LB137447
	Cadmium	2510	2500	100	90 - 110	P	10/07/2025	14:19	LB137447
	Calcium	24900	25000	100	90 - 110	P	10/07/2025	14:19	LB137447
	Chromium	1030	1000	103	90 - 110	P	10/07/2025	14:19	LB137447
	Cobalt	2480	2500	99	90 - 110	P	10/07/2025	14:19	LB137447
	Copper	1260	1250	101	90 - 110	P	10/07/2025	14:19	LB137447
	Iron	5340	5000	107	90 - 110	P	10/07/2025	14:19	LB137447
	Lead	4960	5000	99	90 - 110	P	10/07/2025	14:19	LB137447

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24900	25000	100	90 - 110	P	10/07/2025	14:19	LB137447
	Manganese	2480	2500	99	90 - 110	P	10/07/2025	14:19	LB137447
	Nickel	2490	2500	100	90 - 110	P	10/07/2025	14:19	LB137447
	Potassium	26100	25000	104	90 - 110	P	10/07/2025	14:19	LB137447
	Selenium	5190	5000	104	90 - 110	P	10/07/2025	14:19	LB137447
	Silver	1290	1250	103	90 - 110	P	10/07/2025	14:19	LB137447
	Sodium	25000	25000	100	90 - 110	P	10/07/2025	14:19	LB137447
	Thallium	5200	5000	104	90 - 110	P	10/07/2025	14:19	LB137447
	Vanadium	2500	2500	100	90 - 110	P	10/07/2025	14:19	LB137447
	Zinc	2600	2500	104	90 - 110	P	10/07/2025	14:19	LB137447
CCV03	Aluminum	10000	10000	100	90 - 110	P	10/07/2025	15:08	LB137447
	Antimony	4960	5000	99	90 - 110	P	10/07/2025	15:08	LB137447
	Arsenic	4970	5000	99	90 - 110	P	10/07/2025	15:08	LB137447
	Barium	9860	10000	99	90 - 110	P	10/07/2025	15:08	LB137447
	Beryllium	253	250	101	90 - 110	P	10/07/2025	15:08	LB137447
	Cadmium	2440	2500	98	90 - 110	P	10/07/2025	15:08	LB137447
	Calcium	25100	25000	100	90 - 110	P	10/07/2025	15:08	LB137447
	Chromium	1010	1000	101	90 - 110	P	10/07/2025	15:08	LB137447
	Cobalt	2420	2500	97	90 - 110	P	10/07/2025	15:08	LB137447
	Copper	1240	1250	99	90 - 110	P	10/07/2025	15:08	LB137447
	Iron	5120	5000	102	90 - 110	P	10/07/2025	15:08	LB137447
	Lead	4840	5000	97	90 - 110	P	10/07/2025	15:08	LB137447
	Magnesium	25100	25000	101	90 - 110	P	10/07/2025	15:08	LB137447
	Manganese	2510	2500	100	90 - 110	P	10/07/2025	15:08	LB137447
	Nickel	2430	2500	97	90 - 110	P	10/07/2025	15:08	LB137447
	Potassium	24200	25000	97	90 - 110	P	10/07/2025	15:08	LB137447
	Selenium	5070	5000	101	90 - 110	P	10/07/2025	15:08	LB137447
	Silver	1250	1250	100	90 - 110	P	10/07/2025	15:08	LB137447
	Sodium	23700	25000	95	90 - 110	P	10/07/2025	15:08	LB137447
	Thallium	4850	5000	97	90 - 110	P	10/07/2025	15:08	LB137447
CCV04	Vanadium	2540	2500	102	90 - 110	P	10/07/2025	15:08	LB137447
	Zinc	2520	2500	101	90 - 110	P	10/07/2025	15:08	LB137447
	Aluminum	9980	10000	100	90 - 110	P	10/07/2025	16:00	LB137447
	Antimony	5080	5000	102	90 - 110	P	10/07/2025	16:00	LB137447

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith

Contract: CAMP02

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

SDG No.: Q3257

Lab Code: ACE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5090	5000	102	90 - 110	P	10/07/2025	16:00	LB137447
	Barium	9920	10000	99	90 - 110	P	10/07/2025	16:00	LB137447
	Beryllium	254	250	102	90 - 110	P	10/07/2025	16:00	LB137447
	Cadmium	2490	2500	100	90 - 110	P	10/07/2025	16:00	LB137447
	Calcium	25100	25000	100	90 - 110	P	10/07/2025	16:00	LB137447
	Chromium	1020	1000	102	90 - 110	P	10/07/2025	16:00	LB137447
	Cobalt	2480	2500	99	90 - 110	P	10/07/2025	16:00	LB137447
	Copper	1260	1250	101	90 - 110	P	10/07/2025	16:00	LB137447
	Iron	5080	5000	102	90 - 110	P	10/07/2025	16:00	LB137447
	Lead	4950	5000	99	90 - 110	P	10/07/2025	16:00	LB137447
	Magnesium	25200	25000	101	90 - 110	P	10/07/2025	16:00	LB137447
	Manganese	2530	2500	101	90 - 110	P	10/07/2025	16:00	LB137447
	Nickel	2480	2500	99	90 - 110	P	10/07/2025	16:00	LB137447
	Potassium	24300	25000	97	90 - 110	P	10/07/2025	16:00	LB137447
	Selenium	5180	5000	104	90 - 110	P	10/07/2025	16:00	LB137447
	Silver	1260	1250	101	90 - 110	P	10/07/2025	16:00	LB137447
	Sodium	23700	25000	95	90 - 110	P	10/07/2025	16:00	LB137447
	Thallium	5000	5000	100	90 - 110	P	10/07/2025	16:00	LB137447
	Vanadium	2540	2500	102	90 - 110	P	10/07/2025	16:00	LB137447
	Zinc	2550	2500	102	90 - 110	P	10/07/2025	16:00	LB137447
CCV05	Aluminum	9820	10000	98	90 - 110	P	10/07/2025	16:50	LB137447
	Antimony	4960	5000	99	90 - 110	P	10/07/2025	16:50	LB137447
	Arsenic	5000	5000	100	90 - 110	P	10/07/2025	16:50	LB137447
	Barium	9540	10000	95	90 - 110	P	10/07/2025	16:50	LB137447
	Beryllium	245	250	98	90 - 110	P	10/07/2025	16:50	LB137447
	Cadmium	2480	2500	99	90 - 110	P	10/07/2025	16:50	LB137447
	Calcium	24500	25000	98	90 - 110	P	10/07/2025	16:50	LB137447
	Chromium	1030	1000	103	90 - 110	P	10/07/2025	16:50	LB137447
	Cobalt	2450	2500	98	90 - 110	P	10/07/2025	16:50	LB137447
	Copper	1250	1250	100	90 - 110	P	10/07/2025	16:50	LB137447
	Iron	5320	5000	106	90 - 110	P	10/07/2025	16:50	LB137447
	Lead	4890	5000	98	90 - 110	P	10/07/2025	16:50	LB137447
	Magnesium	24700	25000	99	90 - 110	P	10/07/2025	16:50	LB137447
	Manganese	2430	2500	97	90 - 110	P	10/07/2025	16:50	LB137447

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2460	2500	98	90 - 110	P	10/07/2025	16:50	LB137447
	Potassium	26200	25000	105	90 - 110	P	10/07/2025	16:50	LB137447
	Selenium	5090	5000	102	90 - 110	P	10/07/2025	16:50	LB137447
	Silver	1280	1250	103	90 - 110	P	10/07/2025	16:50	LB137447
	Sodium	26000	25000	104	90 - 110	P	10/07/2025	16:50	LB137447
	Thallium	5150	5000	103	90 - 110	P	10/07/2025	16:50	LB137447
	Vanadium	2490	2500	99	90 - 110	P	10/07/2025	16:50	LB137447
	Zinc	2620	2500	105	90 - 110	P	10/07/2025	16:50	LB137447
CCV06	Aluminum	9870	10000	99	90 - 110	P	10/07/2025	17:49	LB137447
	Antimony	5030	5000	100	90 - 110	P	10/07/2025	17:49	LB137447
	Arsenic	5040	5000	101	90 - 110	P	10/07/2025	17:49	LB137447
	Barium	9460	10000	95	90 - 110	P	10/07/2025	17:49	LB137447
	Beryllium	244	250	98	90 - 110	P	10/07/2025	17:49	LB137447
	Cadmium	2460	2500	98	90 - 110	P	10/07/2025	17:49	LB137447
	Calcium	24400	25000	98	90 - 110	P	10/07/2025	17:49	LB137447
	Chromium	1000	1000	100	90 - 110	P	10/07/2025	17:49	LB137447
	Cobalt	2440	2500	98	90 - 110	P	10/07/2025	17:49	LB137447
	Copper	1250	1250	100	90 - 110	P	10/07/2025	17:49	LB137447
	Iron	4840	5000	97	90 - 110	P	10/07/2025	17:49	LB137447
	Lead	4880	5000	98	90 - 110	P	10/07/2025	17:49	LB137447
	Magnesium	24700	25000	99	90 - 110	P	10/07/2025	17:49	LB137447
	Manganese	2420	2500	97	90 - 110	P	10/07/2025	17:49	LB137447
	Nickel	2450	2500	98	90 - 110	P	10/07/2025	17:49	LB137447
	Potassium	25200	25000	101	90 - 110	P	10/07/2025	17:49	LB137447
	Selenium	5150	5000	103	90 - 110	P	10/07/2025	17:49	LB137447
	Silver	1230	1250	98	90 - 110	P	10/07/2025	17:49	LB137447
	Sodium	23800	25000	95	90 - 110	P	10/07/2025	17:49	LB137447
	Thallium	4750	5000	95	90 - 110	P	10/07/2025	17:49	LB137447
	Vanadium	2460	2500	98	90 - 110	P	10/07/2025	17:49	LB137447
	Zinc	2480	2500	99	90 - 110	P	10/07/2025	17:49	LB137447
CCV07	Aluminum	9840	10000	98	90 - 110	P	10/07/2025	18:44	LB137447
	Antimony	5100	5000	102	90 - 110	P	10/07/2025	18:44	LB137447
	Arsenic	5120	5000	102	90 - 110	P	10/07/2025	18:44	LB137447
	Barium	9570	10000	96	90 - 110	P	10/07/2025	18:44	LB137447

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	244	250	98	90 - 110	P	10/07/2025	18:44	LB137447
	Cadmium	2470	2500	99	90 - 110	P	10/07/2025	18:44	LB137447
	Calcium	24400	25000	98	90 - 110	P	10/07/2025	18:44	LB137447
	Chromium	997	1000	100	90 - 110	P	10/07/2025	18:44	LB137447
	Cobalt	2460	2500	98	90 - 110	P	10/07/2025	18:44	LB137447
	Copper	1260	1250	101	90 - 110	P	10/07/2025	18:44	LB137447
	Iron	4950	5000	99	90 - 110	P	10/07/2025	18:44	LB137447
	Lead	4920	5000	98	90 - 110	P	10/07/2025	18:44	LB137447
	Magnesium	24600	25000	98	90 - 110	P	10/07/2025	18:44	LB137447
	Manganese	2440	2500	98	90 - 110	P	10/07/2025	18:44	LB137447
	Nickel	2470	2500	99	90 - 110	P	10/07/2025	18:44	LB137447
	Potassium	24000	25000	96	90 - 110	P	10/07/2025	18:44	LB137447
	Selenium	5280	5000	106	90 - 110	P	10/07/2025	18:44	LB137447
	Silver	1230	1250	98	90 - 110	P	10/07/2025	18:44	LB137447
	Sodium	23100	25000	92	90 - 110	P	10/07/2025	18:44	LB137447
	Thallium	5050	5000	101	90 - 110	P	10/07/2025	18:44	LB137447
	Vanadium	2480	2500	99	90 - 110	P	10/07/2025	18:44	LB137447
	Zinc	2480	2500	99	90 - 110	P	10/07/2025	18:44	LB137447
CCV08	Aluminum	9510	10000	95	90 - 110	P	10/07/2025	19:20	LB137447
	Antimony	4850	5000	97	90 - 110	P	10/07/2025	19:20	LB137447
	Arsenic	4850	5000	97	90 - 110	P	10/07/2025	19:20	LB137447
	Barium	9160	10000	92	90 - 110	P	10/07/2025	19:20	LB137447
	Beryllium	238	250	95	90 - 110	P	10/07/2025	19:20	LB137447
	Cadmium	2370	2500	95	90 - 110	P	10/07/2025	19:20	LB137447
	Calcium	23500	25000	94	90 - 110	P	10/07/2025	19:20	LB137447
	Chromium	967	1000	97	90 - 110	P	10/07/2025	19:20	LB137447
	Cobalt	2360	2500	94	90 - 110	P	10/07/2025	19:20	LB137447
	Copper	1210	1250	96	90 - 110	P	10/07/2025	19:20	LB137447
	Iron	4810	5000	96	90 - 110	P	10/07/2025	19:20	LB137447
	Lead	4720	5000	94	90 - 110	P	10/07/2025	19:20	LB137447
	Magnesium	23900	25000	96	90 - 110	P	10/07/2025	19:20	LB137447
	Manganese	2350	2500	94	90 - 110	P	10/07/2025	19:20	LB137447
	Nickel	2360	2500	94	90 - 110	P	10/07/2025	19:20	LB137447
	Potassium	23100	25000	92	90 - 110	P	10/07/2025	19:20	LB137447

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	4950	5000	99	90 - 110	P	10/07/2025	19:20	LB137447
	Silver	1200	1250	96	90 - 110	P	10/07/2025	19:20	LB137447
	Sodium	22400	25000	90	90 - 110	P	10/07/2025	19:20	LB137447
	Thallium	4540	5000	91	90 - 110	P	10/07/2025	19:20	LB137447
	Vanadium	2380	2500	95	90 - 110	P	10/07/2025	19:20	LB137447
	Zinc	2420	2500	97	90 - 110	P	10/07/2025	19:20	LB137447

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: CDM Smith **SDG No.:** Q3257
Contract: CAMP02 **Lab Code:** ACE
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.22	0.2	110	70 - 130	CV	10/01/2025	09:20	LB137376
CRI01	Aluminum	116	100	116	65 - 135	P	10/02/2025	11:36	LB137406
	Antimony	53.4	50.0	107	65 - 135	P	10/02/2025	11:36	LB137406
	Arsenic	23.1	20.0	116	65 - 135	P	10/02/2025	11:36	LB137406
	Barium	113	100	113	65 - 135	P	10/02/2025	11:36	LB137406
	Beryllium	6.95	6.0	116	65 - 135	P	10/02/2025	11:36	LB137406
	Cadmium	6.02	6.0	100	65 - 135	P	10/02/2025	11:36	LB137406
	Calcium	2080	2000	104	65 - 135	P	10/02/2025	11:36	LB137406
	Chromium	10.5	10.0	105	65 - 135	P	10/02/2025	11:36	LB137406
	Cobalt	30.2	30.0	101	65 - 135	P	10/02/2025	11:36	LB137406
	Copper	22.5	20.0	112	65 - 135	P	10/02/2025	11:36	LB137406
	Iron	88.5	100	88	65 - 135	P	10/02/2025	11:36	LB137406
	Lead	13.3	12.0	111	65 - 135	P	10/02/2025	11:36	LB137406
	Magnesium	2250	2000	112	65 - 135	P	10/02/2025	11:36	LB137406
	Manganese	21.0	20.0	105	65 - 135	P	10/02/2025	11:36	LB137406
	Nickel	40.8	40.0	102	65 - 135	P	10/02/2025	11:36	LB137406
	Potassium	1530	2000	77	65 - 135	P	10/02/2025	11:36	LB137406
	Selenium	23.8	20.0	119	65 - 135	P	10/02/2025	11:36	LB137406
	Silver	10.3	10.0	103	65 - 135	P	10/02/2025	11:36	LB137406
	Sodium	1490	2000	74	65 - 135	P	10/02/2025	11:36	LB137406
	Thallium	46.6	40.0	117	65 - 135	P	10/02/2025	11:36	LB137406
	Vanadium	37.2	40.0	93	65 - 135	P	10/02/2025	11:36	LB137406
	Zinc	41.5	40.0	104	65 - 135	P	10/02/2025	11:36	LB137406
CRI01	Aluminum	118	100	118	65 - 135	P	10/03/2025	13:16	LB137426
	Antimony	50.6	50.0	101	65 - 135	P	10/03/2025	13:16	LB137426
	Arsenic	21.4	20.0	107	65 - 135	P	10/03/2025	13:16	LB137426
	Barium	100	100	100	65 - 135	P	10/03/2025	13:16	LB137426
	Beryllium	6.49	6.0	108	65 - 135	P	10/03/2025	13:16	LB137426
	Cadmium	6.04	6.0	101	65 - 135	P	10/03/2025	13:16	LB137426
	Calcium	2060	2000	103	65 - 135	P	10/03/2025	13:16	LB137426
	Chromium	9.90	10.0	99	65 - 135	P	10/03/2025	13:16	LB137426
	Cobalt	29.7	30.0	99	65 - 135	P	10/03/2025	13:16	LB137426
	Copper	21.2	20.0	106	65 - 135	P	10/03/2025	13:16	LB137426
	Iron	91.2	100	91	65 - 135	P	10/03/2025	13:16	LB137426
	Lead	10.2	12.0	85	65 - 135	P	10/03/2025	13:16	LB137426

Metals
- 2b -
CRDL STANDARD FOR AA & ICP

Client: CDM Smith **SDG No.:** Q3257
Contract: CAMP02 **Lab Code:** ACE
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Magnesium	2140	2000	107	65 - 135	P	10/03/2025	13:16	LB137426
	Manganese	21.0	20.0	105	65 - 135	P	10/03/2025	13:16	LB137426
	Nickel	40.9	40.0	102	65 - 135	P	10/03/2025	13:16	LB137426
	Potassium	1890	2000	94	65 - 135	P	10/03/2025	13:16	LB137426
	Selenium	21.2	20.0	106	65 - 135	P	10/03/2025	13:16	LB137426
	Silver	10.5	10.0	105	65 - 135	P	10/03/2025	13:16	LB137426
	Sodium	1910	2000	95	65 - 135	P	10/03/2025	13:16	LB137426
	Thallium	39.4	40.0	98	65 - 135	P	10/03/2025	13:16	LB137426
	Vanadium	39.8	40.0	100	65 - 135	P	10/03/2025	13:16	LB137426
	Zinc	42.0	40.0	105	65 - 135	P	10/03/2025	13:16	LB137426
CRI01	Aluminum	90.0	100	90	65 - 135	P	10/07/2025	12:10	LB137447
	Antimony	49.9	50.0	100	65 - 135	P	10/07/2025	12:10	LB137447
	Arsenic	17.7	20.0	88	65 - 135	P	10/07/2025	12:10	LB137447
	Barium	97.4	100	97	65 - 135	P	10/07/2025	12:10	LB137447
	Beryllium	6.27	6.0	104	65 - 135	P	10/07/2025	12:10	LB137447
	Cadmium	5.96	6.0	99	65 - 135	P	10/07/2025	12:10	LB137447
	Calcium	2040	2000	102	65 - 135	P	10/07/2025	12:10	LB137447
	Chromium	9.90	10.0	99	65 - 135	P	10/07/2025	12:10	LB137447
	Cobalt	29.6	30.0	99	65 - 135	P	10/07/2025	12:10	LB137447
	Copper	22.2	20.0	111	65 - 135	P	10/07/2025	12:10	LB137447
	Iron	123	100	123	65 - 135	P	10/07/2025	12:10	LB137447
	Lead	12.1	12.0	101	65 - 135	P	10/07/2025	12:10	LB137447
	Magnesium	2090	2000	104	65 - 135	P	10/07/2025	12:10	LB137447
	Manganese	20.9	20.0	105	65 - 135	P	10/07/2025	12:10	LB137447
	Nickel	40.1	40.0	100	65 - 135	P	10/07/2025	12:10	LB137447
	Potassium	1880	2000	94	65 - 135	P	10/07/2025	12:10	LB137447
	Selenium	22.6	20.0	113	65 - 135	P	10/07/2025	12:10	LB137447
	Silver	9.85	10.0	98	65 - 135	P	10/07/2025	12:10	LB137447
	Sodium	1910	2000	95	65 - 135	P	10/07/2025	12:10	LB137447
	Thallium	45.0	40.0	113	65 - 135	P	10/07/2025	12:10	LB137447
	Vanadium	37.6	40.0	94	65 - 135	P	10/07/2025	12:10	LB137447
	Zinc	37.9	40.0	95	65 - 135	P	10/07/2025	12:10	LB137447



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:

CDM Smith

SDG No.:

Q3257

Contract:

CAMP02

Lab Code:

ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB53	Mercury	0.076	+/-0.2	U	0.20	CV	10/01/2025	09:12	LB137376

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith **SDG No.:** Q3257
Contract: CAMP02 **Lab Code:** ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB67	Mercury	0.076	+/-0.2	U	0.20	CV	10/01/2025	09:17	LB137376
CCB68	Mercury	0.076	+/-0.2	U	0.20	CV	10/01/2025	09:47	LB137376
CCB69	Mercury	0.076	+/-0.2	U	0.20	CV	10/01/2025	10:09	LB137376

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	17.8	+/-50	J	100	P	10/02/2025	11:27	LB137406
	Antimony	6.76	+/-25	U	50.0	P	10/02/2025	11:27	LB137406
	Arsenic	5.12	+/-10	U	20.0	P	10/02/2025	11:27	LB137406
	Barium	15.3	+/-50	J	100	P	10/02/2025	11:27	LB137406
	Beryllium	0.56	+/-3	U	6.00	P	10/02/2025	11:27	LB137406
	Cadmium	0.50	+/-3	U	6.00	P	10/02/2025	11:27	LB137406
	Calcium	234	+/-1000	U	2000	P	10/02/2025	11:27	LB137406
	Chromium	2.12	+/-5	U	10.0	P	10/02/2025	11:27	LB137406
	Cobalt	2.26	+/-15	U	30.0	P	10/02/2025	11:27	LB137406
	Copper	4.60	+/-10	U	20.0	P	10/02/2025	11:27	LB137406
	Iron	23.4	+/-50	U	100	P	10/02/2025	11:27	LB137406
	Lead	3.04	+/-6	J	12.0	P	10/02/2025	11:27	LB137406
	Magnesium	244	+/-1000	U	2000	P	10/02/2025	11:27	LB137406
	Manganese	5.94	+/-10	U	20.0	P	10/02/2025	11:27	LB137406
	Nickel	3.06	+/-20	U	40.0	P	10/02/2025	11:27	LB137406
	Potassium	918	+/-1000	U	2000	P	10/02/2025	11:27	LB137406
	Selenium	9.64	+/-10	U	20.0	P	10/02/2025	11:27	LB137406
	Silver	1.62	+/-5	U	10.0	P	10/02/2025	11:27	LB137406
	Sodium	868	+/-1000	U	2000	P	10/02/2025	11:27	LB137406
	Thallium	4.38	+/-20	U	40.0	P	10/02/2025	11:27	LB137406
	Vanadium	6.26	+/-20	U	40.0	P	10/02/2025	11:27	LB137406
	Zinc	16.7	+/-20	U	40.0	P	10/02/2025	11:27	LB137406

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	10/02/2025	12:02	LB137406
	Antimony	6.76	+/-25	U	50.0	P	10/02/2025	12:02	LB137406
	Arsenic	5.12	+/-10	U	20.0	P	10/02/2025	12:02	LB137406
	Barium	14.6	+/-50	U	100	P	10/02/2025	12:02	LB137406
	Beryllium	0.56	+/-3	U	6.00	P	10/02/2025	12:02	LB137406
	Cadmium	0.50	+/-3	U	6.00	P	10/02/2025	12:02	LB137406
	Calcium	234	+/-1000	U	2000	P	10/02/2025	12:02	LB137406
	Chromium	2.12	+/-5	U	10.0	P	10/02/2025	12:02	LB137406
	Cobalt	2.26	+/-15	U	30.0	P	10/02/2025	12:02	LB137406
	Copper	4.60	+/-10	U	20.0	P	10/02/2025	12:02	LB137406
	Iron	23.4	+/-50	U	100	P	10/02/2025	12:02	LB137406
	Lead	2.30	+/-6	U	12.0	P	10/02/2025	12:02	LB137406
	Magnesium	244	+/-1000	U	2000	P	10/02/2025	12:02	LB137406
	Manganese	5.94	+/-10	U	20.0	P	10/02/2025	12:02	LB137406
	Nickel	3.06	+/-20	U	40.0	P	10/02/2025	12:02	LB137406
	Potassium	918	+/-1000	U	2000	P	10/02/2025	12:02	LB137406
	Selenium	9.64	+/-10	U	20.0	P	10/02/2025	12:02	LB137406
	Silver	1.62	+/-5	U	10.0	P	10/02/2025	12:02	LB137406
	Sodium	868	+/-1000	U	2000	P	10/02/2025	12:02	LB137406
	Thallium	4.38	+/-20	U	40.0	P	10/02/2025	12:02	LB137406
	Vanadium	6.26	+/-20	U	40.0	P	10/02/2025	12:02	LB137406
	Zinc	16.7	+/-20	U	40.0	P	10/02/2025	12:02	LB137406
CCB02	Aluminum	11.3	+/-50	U	100	P	10/02/2025	13:52	LB137406
	Antimony	7.17	+/-25	J	50.0	P	10/02/2025	13:52	LB137406
	Arsenic	5.12	+/-10	U	20.0	P	10/02/2025	13:52	LB137406
	Barium	36.1	+/-50	J	100	P	10/02/2025	13:52	LB137406
	Beryllium	0.56	+/-3	U	6.00	P	10/02/2025	13:52	LB137406
	Cadmium	0.50	+/-3	U	6.00	P	10/02/2025	13:52	LB137406
	Calcium	234	+/-1000	U	2000	P	10/02/2025	13:52	LB137406
	Chromium	2.12	+/-5	U	10.0	P	10/02/2025	13:52	LB137406
	Cobalt	2.26	+/-15	U	30.0	P	10/02/2025	13:52	LB137406
	Copper	4.60	+/-10	U	20.0	P	10/02/2025	13:52	LB137406
	Iron	23.4	+/-50	U	100	P	10/02/2025	13:52	LB137406
	Lead	2.30	+/-6	U	12.0	P	10/02/2025	13:52	LB137406
	Magnesium	244	+/-1000	U	2000	P	10/02/2025	13:52	LB137406
	Manganese	5.94	+/-10	U	20.0	P	10/02/2025	13:52	LB137406
	Nickel	3.06	+/-20	U	40.0	P	10/02/2025	13:52	LB137406
	Potassium	918	+/-1000	U	2000	P	10/02/2025	13:52	LB137406
	Selenium	9.64	+/-10	U	20.0	P	10/02/2025	13:52	LB137406

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	10/02/2025	13:52	LB137406
	Sodium	868	+/-1000	U	2000	P	10/02/2025	13:52	LB137406
	Thallium	9.07	+/-20	J	40.0	P	10/02/2025	13:52	LB137406
	Vanadium	6.26	+/-20	U	40.0	P	10/02/2025	13:52	LB137406
	Zinc	16.7	+/-20	U	40.0	P	10/02/2025	13:52	LB137406
CCB03	Aluminum	11.3	+/-50	U	100	P	10/02/2025	14:53	LB137406
	Antimony	9.80	+/-25	J	50.0	P	10/02/2025	14:53	LB137406
	Arsenic	5.12	+/-10	U	20.0	P	10/02/2025	14:53	LB137406
	Barium	35.3	+/-50	J	100	P	10/02/2025	14:53	LB137406
	Beryllium	0.56	+/-3	U	6.00	P	10/02/2025	14:53	LB137406
	Cadmium	0.50	+/-3	U	6.00	P	10/02/2025	14:53	LB137406
	Calcium	234	+/-1000	U	2000	P	10/02/2025	14:53	LB137406
	Chromium	2.12	+/-5	U	10.0	P	10/02/2025	14:53	LB137406
	Cobalt	2.26	+/-15	U	30.0	P	10/02/2025	14:53	LB137406
	Copper	4.60	+/-10	U	20.0	P	10/02/2025	14:53	LB137406
	Iron	23.4	+/-50	U	100	P	10/02/2025	14:53	LB137406
	Lead	4.60	+/-6	J	12.0	P	10/02/2025	14:53	LB137406
	Magnesium	244	+/-1000	U	2000	P	10/02/2025	14:53	LB137406
	Manganese	5.94	+/-10	U	20.0	P	10/02/2025	14:53	LB137406
	Nickel	3.06	+/-20	U	40.0	P	10/02/2025	14:53	LB137406
	Potassium	918	+/-1000	U	2000	P	10/02/2025	14:53	LB137406
	Selenium	9.64	+/-10	U	20.0	P	10/02/2025	14:53	LB137406
	Silver	1.62	+/-5	U	10.0	P	10/02/2025	14:53	LB137406
	Sodium	868	+/-1000	U	2000	P	10/02/2025	14:53	LB137406
	Thallium	8.77	+/-20	J	40.0	P	10/02/2025	14:53	LB137406
	Vanadium	6.26	+/-20	U	40.0	P	10/02/2025	14:53	LB137406
	Zinc	16.7	+/-20	U	40.0	P	10/02/2025	14:53	LB137406
CCB04	Aluminum	11.3	+/-50	U	100	P	10/02/2025	16:05	LB137406
	Antimony	13.8	+/-25	J	50.0	P	10/02/2025	16:05	LB137406
	Arsenic	5.72	+/-10	J	20.0	P	10/02/2025	16:05	LB137406
	Barium	44.2	+/-50	J	100	P	10/02/2025	16:05	LB137406
	Beryllium	0.56	+/-3	U	6.00	P	10/02/2025	16:05	LB137406
	Cadmium	0.50	+/-3	U	6.00	P	10/02/2025	16:05	LB137406
	Calcium	234	+/-1000	U	2000	P	10/02/2025	16:05	LB137406
	Chromium	2.12	+/-5	U	10.0	P	10/02/2025	16:05	LB137406
	Cobalt	2.26	+/-15	U	30.0	P	10/02/2025	16:05	LB137406
	Copper	4.60	+/-10	U	20.0	P	10/02/2025	16:05	LB137406
	Iron	23.4	+/-50	U	100	P	10/02/2025	16:05	LB137406
	Lead	2.67	+/-6	J	12.0	P	10/02/2025	16:05	LB137406

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	10/02/2025	16:05	LB137406
	Manganese	5.94	+/-10	U	20.0	P	10/02/2025	16:05	LB137406
	Nickel	3.06	+/-20	U	40.0	P	10/02/2025	16:05	LB137406
	Potassium	918	+/-1000	U	2000	P	10/02/2025	16:05	LB137406
	Selenium	9.64	+/-10	U	20.0	P	10/02/2025	16:05	LB137406
	Silver	1.62	+/-5	U	10.0	P	10/02/2025	16:05	LB137406
	Sodium	868	+/-1000	U	2000	P	10/02/2025	16:05	LB137406
	Thallium	17.5	+/-20	J	40.0	P	10/02/2025	16:05	LB137406
	Vanadium	6.26	+/-20	U	40.0	P	10/02/2025	16:05	LB137406
	Zinc	16.7	+/-20	U	40.0	P	10/02/2025	16:05	LB137406
CCB05	Aluminum	20.7	+/-50	J	100	P	10/02/2025	17:35	LB137406
	Antimony	6.76	+/-25	U	50.0	P	10/02/2025	17:35	LB137406
	Arsenic	5.12	+/-10	U	20.0	P	10/02/2025	17:35	LB137406
	Barium	32.0	+/-50	J	100	P	10/02/2025	17:35	LB137406
	Beryllium	0.56	+/-3	U	6.00	P	10/02/2025	17:35	LB137406
	Cadmium	0.50	+/-3	U	6.00	P	10/02/2025	17:35	LB137406
	Calcium	234	+/-1000	U	2000	P	10/02/2025	17:35	LB137406
	Chromium	2.12	+/-5	U	10.0	P	10/02/2025	17:35	LB137406
	Cobalt	2.26	+/-15	U	30.0	P	10/02/2025	17:35	LB137406
	Copper	4.60	+/-10	U	20.0	P	10/02/2025	17:35	LB137406
	Iron	23.4	+/-50	U	100	P	10/02/2025	17:35	LB137406
	Lead	2.81	+/-6	J	12.0	P	10/02/2025	17:35	LB137406
	Magnesium	244	+/-1000	U	2000	P	10/02/2025	17:35	LB137406
	Manganese	5.94	+/-10	U	20.0	P	10/02/2025	17:35	LB137406
	Nickel	3.06	+/-20	U	40.0	P	10/02/2025	17:35	LB137406
	Potassium	918	+/-1000	U	2000	P	10/02/2025	17:35	LB137406
	Selenium	9.64	+/-10	U	20.0	P	10/02/2025	17:35	LB137406
	Silver	1.62	+/-5	U	10.0	P	10/02/2025	17:35	LB137406
	Sodium	868	+/-1000	U	2000	P	10/02/2025	17:35	LB137406
	Thallium	6.83	+/-20	J	40.0	P	10/02/2025	17:35	LB137406
	Vanadium	6.26	+/-20	U	40.0	P	10/02/2025	17:35	LB137406
	Zinc	16.7	+/-20	U	40.0	P	10/02/2025	17:35	LB137406
CCB06	Aluminum	11.3	+/-50	U	100	P	10/02/2025	18:21	LB137406
	Antimony	9.78	+/-25	J	50.0	P	10/02/2025	18:21	LB137406
	Arsenic	5.12	+/-10	U	20.0	P	10/02/2025	18:21	LB137406
	Barium	37.5	+/-50	J	100	P	10/02/2025	18:21	LB137406
	Beryllium	0.56	+/-3	U	6.00	P	10/02/2025	18:21	LB137406
	Cadmium	0.50	+/-3	U	6.00	P	10/02/2025	18:21	LB137406
	Calcium	234	+/-1000	U	2000	P	10/02/2025	18:21	LB137406

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	10/02/2025	18:21	LB137406
	Cobalt	2.26	+/-15	U	30.0	P	10/02/2025	18:21	LB137406
	Copper	4.60	+/-10	U	20.0	P	10/02/2025	18:21	LB137406
	Iron	23.4	+/-50	U	100	P	10/02/2025	18:21	LB137406
	Lead	3.53	+/-6	J	12.0	P	10/02/2025	18:21	LB137406
	Magnesium	244	+/-1000	U	2000	P	10/02/2025	18:21	LB137406
	Manganese	5.94	+/-10	U	20.0	P	10/02/2025	18:21	LB137406
	Nickel	3.06	+/-20	U	40.0	P	10/02/2025	18:21	LB137406
	Potassium	918	+/-1000	U	2000	P	10/02/2025	18:21	LB137406
	Selenium	9.64	+/-10	U	20.0	P	10/02/2025	18:21	LB137406
	Silver	1.62	+/-5	U	10.0	P	10/02/2025	18:21	LB137406
	Sodium	868	+/-1000	U	2000	P	10/02/2025	18:21	LB137406
	Thallium	12.0	+/-20	J	40.0	P	10/02/2025	18:21	LB137406
	Vanadium	6.26	+/-20	U	40.0	P	10/02/2025	18:21	LB137406
	Zinc	16.7	+/-20	U	40.0	P	10/02/2025	18:21	LB137406

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.6	+/-50	J	100	P	10/03/2025	13:12	LB137426
	Antimony	6.76	+/-25	U	50.0	P	10/03/2025	13:12	LB137426
	Arsenic	5.12	+/-10	U	20.0	P	10/03/2025	13:12	LB137426
	Barium	14.6	+/-50	U	100	P	10/03/2025	13:12	LB137426
	Beryllium	0.56	+/-3	U	6.00	P	10/03/2025	13:12	LB137426
	Cadmium	0.50	+/-3	U	6.00	P	10/03/2025	13:12	LB137426
	Calcium	234	+/-1000	U	2000	P	10/03/2025	13:12	LB137426
	Chromium	2.12	+/-5	U	10.0	P	10/03/2025	13:12	LB137426
	Cobalt	2.26	+/-15	U	30.0	P	10/03/2025	13:12	LB137426
	Copper	4.60	+/-10	U	20.0	P	10/03/2025	13:12	LB137426
	Iron	23.4	+/-50	U	100	P	10/03/2025	13:12	LB137426
	Lead	2.30	+/-6	U	12.0	P	10/03/2025	13:12	LB137426
	Magnesium	244	+/-1000	U	2000	P	10/03/2025	13:12	LB137426
	Manganese	5.94	+/-10	U	20.0	P	10/03/2025	13:12	LB137426
	Nickel	3.06	+/-20	U	40.0	P	10/03/2025	13:12	LB137426
	Potassium	918	+/-1000	U	2000	P	10/03/2025	13:12	LB137426
	Selenium	9.64	+/-10	U	20.0	P	10/03/2025	13:12	LB137426
	Silver	1.62	+/-5	U	10.0	P	10/03/2025	13:12	LB137426
	Sodium	868	+/-1000	U	2000	P	10/03/2025	13:12	LB137426
	Thallium	4.38	+/-20	U	40.0	P	10/03/2025	13:12	LB137426
	Vanadium	6.26	+/-20	U	40.0	P	10/03/2025	13:12	LB137426
	Zinc	16.7	+/-20	U	40.0	P	10/03/2025	13:12	LB137426

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	10/03/2025	13:41	LB137426
	Antimony	6.76	+/-25	U	50.0	P	10/03/2025	13:41	LB137426
	Arsenic	5.12	+/-10	U	20.0	P	10/03/2025	13:41	LB137426
	Barium	14.6	+/-50	U	100	P	10/03/2025	13:41	LB137426
	Beryllium	0.56	+/-3	U	6.00	P	10/03/2025	13:41	LB137426
	Cadmium	0.50	+/-3	U	6.00	P	10/03/2025	13:41	LB137426
	Calcium	234	+/-1000	U	2000	P	10/03/2025	13:41	LB137426
	Chromium	2.12	+/-5	U	10.0	P	10/03/2025	13:41	LB137426
	Cobalt	2.26	+/-15	U	30.0	P	10/03/2025	13:41	LB137426
	Copper	4.60	+/-10	U	20.0	P	10/03/2025	13:41	LB137426
	Iron	23.4	+/-50	U	100	P	10/03/2025	13:41	LB137426
	Lead	2.30	+/-6	U	12.0	P	10/03/2025	13:41	LB137426
	Magnesium	244	+/-1000	U	2000	P	10/03/2025	13:41	LB137426
	Manganese	5.94	+/-10	U	20.0	P	10/03/2025	13:41	LB137426
	Nickel	3.06	+/-20	U	40.0	P	10/03/2025	13:41	LB137426
	Potassium	918	+/-1000	U	2000	P	10/03/2025	13:41	LB137426
	Selenium	9.64	+/-10	U	20.0	P	10/03/2025	13:41	LB137426
	Silver	1.62	+/-5	U	10.0	P	10/03/2025	13:41	LB137426
	Sodium	868	+/-1000	U	2000	P	10/03/2025	13:41	LB137426
	Thallium	4.38	+/-20	U	40.0	P	10/03/2025	13:41	LB137426
	Vanadium	6.26	+/-20	U	40.0	P	10/03/2025	13:41	LB137426
	Zinc	16.7	+/-20	U	40.0	P	10/03/2025	13:41	LB137426
CCB02	Aluminum	11.3	+/-50	U	100	P	10/03/2025	15:04	LB137426
	Antimony	6.76	+/-25	U	50.0	P	10/03/2025	15:04	LB137426
	Arsenic	5.12	+/-10	U	20.0	P	10/03/2025	15:04	LB137426
	Barium	14.6	+/-50	U	100	P	10/03/2025	15:04	LB137426
	Beryllium	0.56	+/-3	U	6.00	P	10/03/2025	15:04	LB137426
	Cadmium	0.50	+/-3	U	6.00	P	10/03/2025	15:04	LB137426
	Calcium	234	+/-1000	U	2000	P	10/03/2025	15:04	LB137426
	Chromium	2.12	+/-5	U	10.0	P	10/03/2025	15:04	LB137426
	Cobalt	2.26	+/-15	U	30.0	P	10/03/2025	15:04	LB137426
	Copper	4.60	+/-10	U	20.0	P	10/03/2025	15:04	LB137426
	Iron	23.4	+/-50	U	100	P	10/03/2025	15:04	LB137426
	Lead	2.30	+/-6	U	12.0	P	10/03/2025	15:04	LB137426
	Magnesium	244	+/-1000	U	2000	P	10/03/2025	15:04	LB137426
	Manganese	5.94	+/-10	U	20.0	P	10/03/2025	15:04	LB137426
	Nickel	3.06	+/-20	U	40.0	P	10/03/2025	15:04	LB137426
	Potassium	918	+/-1000	U	2000	P	10/03/2025	15:04	LB137426
	Selenium	9.64	+/-10	U	20.0	P	10/03/2025	15:04	LB137426

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith **SDG No.:** Q3257
Contract: CAMP02 **Lab Code:** ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	10/03/2025	15:04	LB137426
	Sodium	868	+/-1000	U	2000	P	10/03/2025	15:04	LB137426
	Thallium	10.1	+/-20	J	40.0	P	10/03/2025	15:04	LB137426
	Vanadium	6.26	+/-20	U	40.0	P	10/03/2025	15:04	LB137426
	Zinc	16.7	+/-20	U	40.0	P	10/03/2025	15:04	LB137426
CCB03	Aluminum	11.3	+/-50	U	100	P	10/03/2025	16:00	LB137426
	Antimony	6.76	+/-25	U	50.0	P	10/03/2025	16:00	LB137426
	Arsenic	5.12	+/-10	U	20.0	P	10/03/2025	16:00	LB137426
	Barium	14.6	+/-50	U	100	P	10/03/2025	16:00	LB137426
	Beryllium	0.56	+/-3	U	6.00	P	10/03/2025	16:00	LB137426
	Cadmium	0.50	+/-3	U	6.00	P	10/03/2025	16:00	LB137426
	Calcium	234	+/-1000	U	2000	P	10/03/2025	16:00	LB137426
	Chromium	2.12	+/-5	U	10.0	P	10/03/2025	16:00	LB137426
	Cobalt	2.26	+/-15	U	30.0	P	10/03/2025	16:00	LB137426
	Copper	4.60	+/-10	U	20.0	P	10/03/2025	16:00	LB137426
	Iron	23.4	+/-50	U	100	P	10/03/2025	16:00	LB137426
	Lead	2.30	+/-6	U	12.0	P	10/03/2025	16:00	LB137426
	Magnesium	244	+/-1000	U	2000	P	10/03/2025	16:00	LB137426
	Manganese	5.94	+/-10	U	20.0	P	10/03/2025	16:00	LB137426
	Nickel	3.06	+/-20	U	40.0	P	10/03/2025	16:00	LB137426
	Potassium	918	+/-1000	U	2000	P	10/03/2025	16:00	LB137426
	Selenium	9.64	+/-10	U	20.0	P	10/03/2025	16:00	LB137426
	Silver	1.62	+/-5	U	10.0	P	10/03/2025	16:00	LB137426
	Sodium	868	+/-1000	U	2000	P	10/03/2025	16:00	LB137426
	Thallium	4.38	+/-20	U	40.0	P	10/03/2025	16:00	LB137426
	Vanadium	6.26	+/-20	U	40.0	P	10/03/2025	16:00	LB137426
	Zinc	16.7	+/-20	U	40.0	P	10/03/2025	16:00	LB137426
CCB04	Aluminum	11.3	+/-50	U	100	P	10/03/2025	16:49	LB137426
	Antimony	6.76	+/-25	U	50.0	P	10/03/2025	16:49	LB137426
	Arsenic	5.12	+/-10	U	20.0	P	10/03/2025	16:49	LB137426
	Barium	14.6	+/-50	U	100	P	10/03/2025	16:49	LB137426
	Beryllium	0.56	+/-3	U	6.00	P	10/03/2025	16:49	LB137426
	Cadmium	0.50	+/-3	U	6.00	P	10/03/2025	16:49	LB137426
	Calcium	234	+/-1000	U	2000	P	10/03/2025	16:49	LB137426
	Chromium	2.12	+/-5	U	10.0	P	10/03/2025	16:49	LB137426
	Cobalt	2.26	+/-15	U	30.0	P	10/03/2025	16:49	LB137426
	Copper	4.60	+/-10	U	20.0	P	10/03/2025	16:49	LB137426
	Iron	23.4	+/-50	U	100	P	10/03/2025	16:49	LB137426
	Lead	2.30	+/-6	U	12.0	P	10/03/2025	16:49	LB137426

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	10/03/2025	16:49	LB137426
	Manganese	5.94	+/-10	U	20.0	P	10/03/2025	16:49	LB137426
	Nickel	3.06	+/-20	U	40.0	P	10/03/2025	16:49	LB137426
	Potassium	918	+/-1000	U	2000	P	10/03/2025	16:49	LB137426
	Selenium	9.64	+/-10	U	20.0	P	10/03/2025	16:49	LB137426
	Silver	1.62	+/-5	U	10.0	P	10/03/2025	16:49	LB137426
	Sodium	868	+/-1000	U	2000	P	10/03/2025	16:49	LB137426
	Thallium	4.38	+/-20	U	40.0	P	10/03/2025	16:49	LB137426
	Vanadium	6.26	+/-20	U	40.0	P	10/03/2025	16:49	LB137426
	Zinc	16.7	+/-20	U	40.0	P	10/03/2025	16:49	LB137426

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	10/07/2025	12:06	LB137447
	Antimony	6.76	+/-25	U	50.0	P	10/07/2025	12:06	LB137447
	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	12:06	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	12:06	LB137447
	Beryllium	0.56	+/-3	U	6.00	P	10/07/2025	12:06	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	12:06	LB137447
	Calcium	234	+/-1000	U	2000	P	10/07/2025	12:06	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	12:06	LB137447
	Cobalt	2.26	+/-15	U	30.0	P	10/07/2025	12:06	LB137447
	Copper	4.60	+/-10	U	20.0	P	10/07/2025	12:06	LB137447
	Iron	23.4	+/-50	U	100	P	10/07/2025	12:06	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	12:06	LB137447
	Magnesium	244	+/-1000	U	2000	P	10/07/2025	12:06	LB137447
	Manganese	5.94	+/-10	U	20.0	P	10/07/2025	12:06	LB137447
	Nickel	3.06	+/-20	U	40.0	P	10/07/2025	12:06	LB137447
	Potassium	918	+/-1000	U	2000	P	10/07/2025	12:06	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	12:06	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	12:06	LB137447
	Sodium	868	+/-1000	U	2000	P	10/07/2025	12:06	LB137447
	Thallium	5.24	+/-20	J	40.0	P	10/07/2025	12:06	LB137447
	Vanadium	6.26	+/-20	U	40.0	P	10/07/2025	12:06	LB137447
	Zinc	16.7	+/-20	U	40.0	P	10/07/2025	12:06	LB137447

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	10/07/2025	12:58	LB137447
	Antimony	6.76	+/-25	U	50.0	P	10/07/2025	12:58	LB137447
	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	12:58	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	12:58	LB137447
	Beryllium	0.56	+/-3	U	6.00	P	10/07/2025	12:58	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	12:58	LB137447
	Calcium	234	+/-1000	U	2000	P	10/07/2025	12:58	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	12:58	LB137447
	Cobalt	2.26	+/-15	U	30.0	P	10/07/2025	12:58	LB137447
	Copper	4.60	+/-10	U	20.0	P	10/07/2025	12:58	LB137447
	Iron	23.4	+/-50	U	100	P	10/07/2025	12:58	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	12:58	LB137447
	Magnesium	244	+/-1000	U	2000	P	10/07/2025	12:58	LB137447
	Manganese	5.94	+/-10	U	20.0	P	10/07/2025	12:58	LB137447
	Nickel	3.06	+/-20	U	40.0	P	10/07/2025	12:58	LB137447
	Potassium	918	+/-1000	U	2000	P	10/07/2025	12:58	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	12:58	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	12:58	LB137447
	Sodium	868	+/-1000	U	2000	P	10/07/2025	12:58	LB137447
	Thallium	4.38	+/-20	U	40.0	P	10/07/2025	12:58	LB137447
	Vanadium	6.26	+/-20	U	40.0	P	10/07/2025	12:58	LB137447
	Zinc	16.7	+/-20	U	40.0	P	10/07/2025	12:58	LB137447
CCB02	Aluminum	14.4	+/-50	J	100	P	10/07/2025	14:24	LB137447
	Antimony	6.76	+/-25	U	50.0	P	10/07/2025	14:24	LB137447
	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	14:24	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	14:24	LB137447
	Beryllium	0.56	+/-3	U	6.00	P	10/07/2025	14:24	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	14:24	LB137447
	Calcium	234	+/-1000	U	2000	P	10/07/2025	14:24	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	14:24	LB137447
	Cobalt	2.26	+/-15	U	30.0	P	10/07/2025	14:24	LB137447
	Copper	4.60	+/-10	U	20.0	P	10/07/2025	14:24	LB137447
	Iron	28.4	+/-50	J	100	P	10/07/2025	14:24	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	14:24	LB137447
	Magnesium	244	+/-1000	U	2000	P	10/07/2025	14:24	LB137447
	Manganese	5.94	+/-10	U	20.0	P	10/07/2025	14:24	LB137447
	Nickel	3.06	+/-20	U	40.0	P	10/07/2025	14:24	LB137447
	Potassium	918	+/-1000	U	2000	P	10/07/2025	14:24	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	14:24	LB137447

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith **SDG No.:** Q3257
Contract: CAMP02 **Lab Code:** ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	10/07/2025	14:24	LB137447
	Sodium	868	+/-1000	U	2000	P	10/07/2025	14:24	LB137447
	Thallium	5.15	+/-20	J	40.0	P	10/07/2025	14:24	LB137447
	Vanadium	6.26	+/-20	U	40.0	P	10/07/2025	14:24	LB137447
	Zinc	16.7	+/-20	U	40.0	P	10/07/2025	14:24	LB137447
CCB03	Aluminum	11.3	+/-50	U	100	P	10/07/2025	15:12	LB137447
	Antimony	6.76	+/-25	U	50.0	P	10/07/2025	15:12	LB137447
	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	15:12	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	15:12	LB137447
	Beryllium	0.56	+/-3	U	6.00	P	10/07/2025	15:12	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	15:12	LB137447
	Calcium	234	+/-1000	U	2000	P	10/07/2025	15:12	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	15:12	LB137447
	Cobalt	2.26	+/-15	U	30.0	P	10/07/2025	15:12	LB137447
	Copper	4.60	+/-10	U	20.0	P	10/07/2025	15:12	LB137447
	Iron	23.4	+/-50	U	100	P	10/07/2025	15:12	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	15:12	LB137447
	Magnesium	244	+/-1000	U	2000	P	10/07/2025	15:12	LB137447
	Manganese	5.94	+/-10	U	20.0	P	10/07/2025	15:12	LB137447
	Nickel	3.06	+/-20	U	40.0	P	10/07/2025	15:12	LB137447
	Potassium	918	+/-1000	U	2000	P	10/07/2025	15:12	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	15:12	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	15:12	LB137447
	Sodium	868	+/-1000	U	2000	P	10/07/2025	15:12	LB137447
	Thallium	6.22	+/-20	J	40.0	P	10/07/2025	15:12	LB137447
	Vanadium	6.26	+/-20	U	40.0	P	10/07/2025	15:12	LB137447
	Zinc	16.7	+/-20	U	40.0	P	10/07/2025	15:12	LB137447
CCB04	Aluminum	11.3	+/-50	U	100	P	10/07/2025	16:04	LB137447
	Antimony	6.76	+/-25	U	50.0	P	10/07/2025	16:04	LB137447
	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	16:04	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	16:04	LB137447
	Beryllium	0.56	+/-3	U	6.00	P	10/07/2025	16:04	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	16:04	LB137447
	Calcium	234	+/-1000	U	2000	P	10/07/2025	16:04	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	16:04	LB137447
	Cobalt	2.26	+/-15	U	30.0	P	10/07/2025	16:04	LB137447
	Copper	4.60	+/-10	U	20.0	P	10/07/2025	16:04	LB137447
	Iron	23.4	+/-50	U	100	P	10/07/2025	16:04	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	16:04	LB137447

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	10/07/2025	16:04	LB137447
	Manganese	5.94	+/-10	U	20.0	P	10/07/2025	16:04	LB137447
	Nickel	3.06	+/-20	U	40.0	P	10/07/2025	16:04	LB137447
	Potassium	918	+/-1000	U	2000	P	10/07/2025	16:04	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	16:04	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	16:04	LB137447
	Sodium	868	+/-1000	U	2000	P	10/07/2025	16:04	LB137447
	Thallium	5.51	+/-20	J	40.0	P	10/07/2025	16:04	LB137447
	Vanadium	6.26	+/-20	U	40.0	P	10/07/2025	16:04	LB137447
	Zinc	16.7	+/-20	U	40.0	P	10/07/2025	16:04	LB137447
CCB05	Aluminum	11.3	+/-50	U	100	P	10/07/2025	17:01	LB137447
	Antimony	6.76	+/-25	U	50.0	P	10/07/2025	17:01	LB137447
	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	17:01	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	17:01	LB137447
	Beryllium	0.56	+/-3	U	6.00	P	10/07/2025	17:01	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	17:01	LB137447
	Calcium	234	+/-1000	U	2000	P	10/07/2025	17:01	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	17:01	LB137447
	Cobalt	2.26	+/-15	U	30.0	P	10/07/2025	17:01	LB137447
	Copper	4.60	+/-10	U	20.0	P	10/07/2025	17:01	LB137447
	Iron	23.4	+/-50	U	100	P	10/07/2025	17:01	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	17:01	LB137447
	Magnesium	244	+/-1000	U	2000	P	10/07/2025	17:01	LB137447
	Manganese	5.94	+/-10	U	20.0	P	10/07/2025	17:01	LB137447
	Nickel	3.06	+/-20	U	40.0	P	10/07/2025	17:01	LB137447
	Potassium	918	+/-1000	U	2000	P	10/07/2025	17:01	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	17:01	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	17:01	LB137447
	Sodium	868	+/-1000	U	2000	P	10/07/2025	17:01	LB137447
	Thallium	5.07	+/-20	J	40.0	P	10/07/2025	17:01	LB137447
	Vanadium	6.26	+/-20	U	40.0	P	10/07/2025	17:01	LB137447
	Zinc	16.7	+/-20	U	40.0	P	10/07/2025	17:01	LB137447
CCB06	Aluminum	11.3	+/-50	U	100	P	10/07/2025	17:59	LB137447
	Antimony	6.76	+/-25	U	50.0	P	10/07/2025	17:59	LB137447
	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	17:59	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	17:59	LB137447
	Beryllium	0.56	+/-3	U	6.00	P	10/07/2025	17:59	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	17:59	LB137447
	Calcium	234	+/-1000	U	2000	P	10/07/2025	17:59	LB137447

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	17:59	LB137447
	Cobalt	2.26	+/-15	U	30.0	P	10/07/2025	17:59	LB137447
	Copper	4.60	+/-10	U	20.0	P	10/07/2025	17:59	LB137447
	Iron	23.4	+/-50	U	100	P	10/07/2025	17:59	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	17:59	LB137447
	Magnesium	244	+/-1000	U	2000	P	10/07/2025	17:59	LB137447
	Manganese	5.94	+/-10	U	20.0	P	10/07/2025	17:59	LB137447
	Nickel	3.06	+/-20	U	40.0	P	10/07/2025	17:59	LB137447
	Potassium	918	+/-1000	U	2000	P	10/07/2025	17:59	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	17:59	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	17:59	LB137447
	Sodium	868	+/-1000	U	2000	P	10/07/2025	17:59	LB137447
	Thallium	4.38	+/-20	U	40.0	P	10/07/2025	17:59	LB137447
	Vanadium	6.26	+/-20	U	40.0	P	10/07/2025	17:59	LB137447
	Zinc	16.7	+/-20	U	40.0	P	10/07/2025	17:59	LB137447
CCB07	Aluminum	11.3	+/-50	U	100	P	10/07/2025	18:48	LB137447
	Antimony	6.76	+/-25	U	50.0	P	10/07/2025	18:48	LB137447
	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	18:48	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	18:48	LB137447
	Beryllium	0.56	+/-3	U	6.00	P	10/07/2025	18:48	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	18:48	LB137447
	Calcium	234	+/-1000	U	2000	P	10/07/2025	18:48	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	18:48	LB137447
	Cobalt	2.26	+/-15	U	30.0	P	10/07/2025	18:48	LB137447
	Copper	4.60	+/-10	U	20.0	P	10/07/2025	18:48	LB137447
	Iron	23.4	+/-50	U	100	P	10/07/2025	18:48	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	18:48	LB137447
	Magnesium	244	+/-1000	U	2000	P	10/07/2025	18:48	LB137447
	Manganese	5.94	+/-10	U	20.0	P	10/07/2025	18:48	LB137447
	Nickel	3.06	+/-20	U	40.0	P	10/07/2025	18:48	LB137447
	Potassium	918	+/-1000	U	2000	P	10/07/2025	18:48	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	18:48	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	18:48	LB137447
	Sodium	868	+/-1000	U	2000	P	10/07/2025	18:48	LB137447
	Thallium	4.38	+/-20	U	40.0	P	10/07/2025	18:48	LB137447
	Vanadium	6.26	+/-20	U	40.0	P	10/07/2025	18:48	LB137447
	Zinc	16.7	+/-20	U	40.0	P	10/07/2025	18:48	LB137447
CCB08	Aluminum	11.3	+/-50	U	100	P	10/07/2025	19:24	LB137447
	Antimony	6.76	+/-25	U	50.0	P	10/07/2025	19:24	LB137447

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	19:24	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	19:24	LB137447
	Beryllium	0.56	+/-3	U	6.00	P	10/07/2025	19:24	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	19:24	LB137447
	Calcium	234	+/-1000	U	2000	P	10/07/2025	19:24	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	19:24	LB137447
	Cobalt	2.26	+/-15	U	30.0	P	10/07/2025	19:24	LB137447
	Copper	4.60	+/-10	U	20.0	P	10/07/2025	19:24	LB137447
	Iron	23.4	+/-50	U	100	P	10/07/2025	19:24	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	19:24	LB137447
	Magnesium	244	+/-1000	U	2000	P	10/07/2025	19:24	LB137447
	Manganese	5.94	+/-10	U	20.0	P	10/07/2025	19:24	LB137447
	Nickel	3.06	+/-20	U	40.0	P	10/07/2025	19:24	LB137447
	Potassium	918	+/-1000	U	2000	P	10/07/2025	19:24	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	19:24	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	19:24	LB137447
	Sodium	868	+/-1000	U	2000	P	10/07/2025	19:24	LB137447
	Thallium	4.38	+/-20	U	40.0	P	10/07/2025	19:24	LB137447
	Vanadium	6.26	+/-20	U	40.0	P	10/07/2025	19:24	LB137447
	Zinc	16.7	+/-20	U	40.0	P	10/07/2025	19:24	LB137447

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3257

Instrument: CV1

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB169927BL		SOLID		Batch Number:	PB169927		Prep Date:	09/30/2025	
	Mercury	0.0070	<0.012	U	0.012	CV	10/01/2025	09:27	LB137376

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q3257

Instrument: P4

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB169935BL	SOLID			Batch Number:	PB169935		Prep Date:	10/01/2025	
	Aluminum	0.84	<2.5	U	5.00	P	10/03/2025	14:33	LB137426
	Antimony	0.22	<1.25	U	2.50	P	10/03/2025	14:33	LB137426
	Arsenic	0.19	<0.5	U	1.00	P	10/03/2025	14:33	LB137426
	Barium	0.73	<2.5	U	5.00	P	10/03/2025	14:33	LB137426
	Beryllium	0.025	<0.15	U	0.30	P	10/03/2025	14:33	LB137426
	Cadmium	0.024	<0.15	U	0.30	P	10/03/2025	14:33	LB137426
	Calcium	11.1	<50	U	100	P	10/03/2025	14:33	LB137426
	Chromium	0.078	<0.25	J	0.50	P	10/03/2025	14:33	LB137426
	Cobalt	0.10	<0.75	U	1.50	P	10/03/2025	14:33	LB137426
	Copper	0.22	<0.5	U	1.00	P	10/03/2025	14:33	LB137426
	Iron	3.99	<2.5	U	5.00	P	10/03/2025	14:33	LB137426
	Lead	0.26	<0.3	J	0.60	P	10/03/2025	14:33	LB137426
	Magnesium	12.0	<50	U	100	P	10/03/2025	14:33	LB137426
	Manganese	0.25	<0.5	J	1.00	P	10/03/2025	14:33	LB137426
	Nickel	0.13	<1	U	2.00	P	10/03/2025	14:33	LB137426
	Potassium	27.7	<50	U	100	P	10/03/2025	14:33	LB137426
	Selenium	0.26	<0.5	U	1.00	P	10/03/2025	14:33	LB137426
	Silver	0.12	<0.25	U	0.50	P	10/03/2025	14:33	LB137426
	Sodium	17.8	<50	U	100	P	10/03/2025	14:33	LB137426
	Thallium	0.23	<1	U	2.00	P	10/03/2025	14:33	LB137426
	Vanadium	0.25	<1	U	2.00	P	10/03/2025	14:33	LB137426
	Zinc	0.23	<1	U	2.00	P	10/03/2025	14:33	LB137426

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>CDM Smith</u>	SDG No.: <u>Q3257</u>
Contract: <u>CAMP02</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	245000	255000	96	216000	294000	10/02/2025	11:42	LB137406
	Antimony	1.99			-50	50	10/02/2025	11:42	LB137406
	Arsenic	5.85			-20	20	10/02/2025	11:42	LB137406
	Barium	18.7	6.0	312	-94	106	10/02/2025	11:42	LB137406
	Beryllium	1.98			-6	6	10/02/2025	11:42	LB137406
	Cadmium	-1.23	1.0	123	-5	7	10/02/2025	11:42	LB137406
	Calcium	236000	245000	96	208000	282000	10/02/2025	11:42	LB137406
	Chromium	50.8	52.0	98	42	62	10/02/2025	11:42	LB137406
	Cobalt	0.27			-30	30	10/02/2025	11:42	LB137406
	Copper	3.53	2.0	176	-18	22	10/02/2025	11:42	LB137406
	Iron	96000	101000	95	85600	116500	10/02/2025	11:42	LB137406
	Lead	-7.87			-12	12	10/02/2025	11:42	LB137406
	Magnesium	251000	255000	98	216000	294000	10/02/2025	11:42	LB137406
	Manganese	0.23	7.0	3	-13	27	10/02/2025	11:42	LB137406
	Nickel	3.41	2.0	170	-38	42	10/02/2025	11:42	LB137406
	Potassium	-102			0	0	10/02/2025	11:42	LB137406
	Selenium	-8.13			-20	20	10/02/2025	11:42	LB137406
	Silver	-3.77			-10	10	10/02/2025	11:42	LB137406
	Sodium	-181			0	0	10/02/2025	11:42	LB137406
	Thallium	14.0			-40	40	10/02/2025	11:42	LB137406
	Vanadium	2.54			-40	40	10/02/2025	11:42	LB137406
	Zinc	-0.077			-40	40	10/02/2025	11:42	LB137406
ICSAB01	Aluminum	249000	247000	101	209000	285000	10/02/2025	11:46	LB137406
	Antimony	622	618	101	525	711	10/02/2025	11:46	LB137406
	Arsenic	110	104	106	88.4	120	10/02/2025	11:46	LB137406
	Barium	515	537	96	437	637	10/02/2025	11:46	LB137406
	Beryllium	548	495	111	420	570	10/02/2025	11:46	LB137406
	Cadmium	1040	972	107	826	1120	10/02/2025	11:46	LB137406
	Calcium	240000	235000	102	199000	271000	10/02/2025	11:46	LB137406
	Chromium	594	542	110	460	624	10/02/2025	11:46	LB137406
	Cobalt	515	476	108	404	548	10/02/2025	11:46	LB137406
	Copper	510	511	100	434	588	10/02/2025	11:46	LB137406
	Iron	98300	99300	99	84400	114500	10/02/2025	11:46	LB137406
	Lead	42.6	49.0	87	37	61	10/02/2025	11:46	LB137406
	Magnesium	255000	248000	103	210000	286000	10/02/2025	11:46	LB137406
	Manganese	497	507	98	430	584	10/02/2025	11:46	LB137406
	Nickel	1030	954	108	810	1100	10/02/2025	11:46	LB137406
	Potassium	-265			0	0	10/02/2025	11:46	LB137406
	Selenium	46.7	46.0	102	26	66	10/02/2025	11:46	LB137406
	Silver	212	201	106	170	232	10/02/2025	11:46	LB137406
	Sodium	-229			0	0	10/02/2025	11:46	LB137406
	Thallium	115	108	106	68	148	10/02/2025	11:46	LB137406

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>CDM Smith</u>	SDG No.: <u>Q3257</u>
Contract: <u>CAMP02</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	499	491	102	417	565	10/02/2025	11:46	LB137406
	Zinc	1060	952	111	809	1095	10/02/2025	11:46	LB137406
ICSA01	Aluminum	245000	255000	96	216000	294000	10/03/2025	13:20	LB137426
	Antimony	3.52			-50	50	10/03/2025	13:20	LB137426
	Arsenic	8.41			-20	20	10/03/2025	13:20	LB137426
	Barium	1.66	6.0	28	-94	106	10/03/2025	13:20	LB137426
	Beryllium	1.42			-6	6	10/03/2025	13:20	LB137426
	Cadmium	0.91	1.0	91	-5	7	10/03/2025	13:20	LB137426
	Calcium	236000	245000	96	208000	282000	10/03/2025	13:20	LB137426
	Chromium	46.6	52.0	90	42	62	10/03/2025	13:20	LB137426
	Cobalt	0.68			-30	30	10/03/2025	13:20	LB137426
	Copper	3.06	2.0	153	-18	22	10/03/2025	13:20	LB137426
	Iron	99600	101000	99	85600	116500	10/03/2025	13:20	LB137426
	Lead	0.97			-12	12	10/03/2025	13:20	LB137426
	Magnesium	247000	255000	97	216000	294000	10/03/2025	13:20	LB137426
	Manganese	-1.76	7.0	25	-13	27	10/03/2025	13:20	LB137426
	Nickel	3.93	2.0	196	-38	42	10/03/2025	13:20	LB137426
	Potassium	187			0	0	10/03/2025	13:20	LB137426
	Selenium	9.27			-20	20	10/03/2025	13:20	LB137426
	Silver	1.41			-10	10	10/03/2025	13:20	LB137426
	Sodium	127			0	0	10/03/2025	13:20	LB137426
	Thallium	11.3			-40	40	10/03/2025	13:20	LB137426
	Vanadium	4.84			-40	40	10/03/2025	13:20	LB137426
	Zinc	1.66			-40	40	10/03/2025	13:20	LB137426
ICSAB01	Aluminum	247000	247000	100	209000	285000	10/03/2025	13:25	LB137426
	Antimony	609	618	98	525	711	10/03/2025	13:25	LB137426
	Arsenic	110	104	106	88.4	120	10/03/2025	13:25	LB137426
	Barium	496	537	92	437	637	10/03/2025	13:25	LB137426
	Beryllium	516	495	104	420	570	10/03/2025	13:25	LB137426
	Cadmium	980	972	101	826	1120	10/03/2025	13:25	LB137426
	Calcium	235000	235000	100	199000	271000	10/03/2025	13:25	LB137426
	Chromium	553	542	102	460	624	10/03/2025	13:25	LB137426
	Cobalt	491	476	103	404	548	10/03/2025	13:25	LB137426
	Copper	484	511	95	434	588	10/03/2025	13:25	LB137426
	Iron	97500	99300	98	84400	114500	10/03/2025	13:25	LB137426
	Lead	49.1	49.0	100	37	61	10/03/2025	13:25	LB137426
	Magnesium	248000	248000	100	210000	286000	10/03/2025	13:25	LB137426
	Manganese	492	507	97	430	584	10/03/2025	13:25	LB137426
	Nickel	981	954	103	810	1100	10/03/2025	13:25	LB137426
	Potassium	15.2			0	0	10/03/2025	13:25	LB137426
	Selenium	56.4	46.0	123	26	66	10/03/2025	13:25	LB137426
	Silver	208	201	104	170	232	10/03/2025	13:25	LB137426

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>CDM Smith</u>	SDG No.: <u>Q3257</u>
Contract: <u>CAMP02</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Sodium	92.2			0	0	10/03/2025	13:25	LB137426
	Thallium	105	108	97	68	148	10/03/2025	13:25	LB137426
	Vanadium	485	491	99	417	565	10/03/2025	13:25	LB137426
	Zinc	1030	952	108	809	1095	10/03/2025	13:25	LB137426
ICSA01	Aluminum	243000	255000	95	216000	294000	10/07/2025	12:14	LB137447
	Antimony	-5.05			-50	50	10/07/2025	12:14	LB137447
	Arsenic	3.53			-20	20	10/07/2025	12:14	LB137447
	Barium	-8.30	6.0	138	-94	106	10/07/2025	12:14	LB137447
	Beryllium	1.74			-6	6	10/07/2025	12:14	LB137447
	Cadmium	2.40	1.0	240	-5	7	10/07/2025	12:14	LB137447
	Calcium	234000	245000	96	208000	282000	10/07/2025	12:14	LB137447
	Chromium	47.2	52.0	91	42	62	10/07/2025	12:14	LB137447
	Cobalt	1.06			-30	30	10/07/2025	12:14	LB137447
	Copper	13.8	2.0	690	-18	22	10/07/2025	12:14	LB137447
	Iron	94400	101000	94	85600	116500	10/07/2025	12:14	LB137447
	Lead	-1.11			-12	12	10/07/2025	12:14	LB137447
	Magnesium	246000	255000	96	216000	294000	10/07/2025	12:14	LB137447
	Manganese	6.71	7.0	96	-13	27	10/07/2025	12:14	LB137447
	Nickel	2.60	2.0	130	-38	42	10/07/2025	12:14	LB137447
	Potassium	-102			0	0	10/07/2025	12:14	LB137447
	Selenium	8.06			-20	20	10/07/2025	12:14	LB137447
	Silver	-0.34			-10	10	10/07/2025	12:14	LB137447
	Sodium	-109			0	0	10/07/2025	12:14	LB137447
	Thallium	33.6			-40	40	10/07/2025	12:14	LB137447
	Vanadium	-0.97			-40	40	10/07/2025	12:14	LB137447
	Zinc	-0.92			-40	40	10/07/2025	12:14	LB137447
ICSAB01	Aluminum	247000	247000	100	209000	285000	10/07/2025	12:37	LB137447
	Antimony	604	618	98	525	711	10/07/2025	12:37	LB137447
	Arsenic	107	104	103	88.4	120	10/07/2025	12:37	LB137447
	Barium	477	537	89	437	637	10/07/2025	12:37	LB137447
	Beryllium	496	495	100	420	570	10/07/2025	12:37	LB137447
	Cadmium	990	972	102	826	1120	10/07/2025	12:37	LB137447
	Calcium	231000	235000	98	199000	271000	10/07/2025	12:37	LB137447
	Chromium	540	542	100	460	624	10/07/2025	12:37	LB137447
	Cobalt	493	476	104	404	548	10/07/2025	12:37	LB137447
	Copper	489	511	96	434	588	10/07/2025	12:37	LB137447
	Iron	97100	99300	98	84400	114500	10/07/2025	12:37	LB137447
	Lead	50.4	49.0	103	37	61	10/07/2025	12:37	LB137447
	Magnesium	246000	248000	99	210000	286000	10/07/2025	12:37	LB137447
	Manganese	477	507	94	430	584	10/07/2025	12:37	LB137447
	Nickel	982	954	103	810	1100	10/07/2025	12:37	LB137447
	Potassium	-317			0	0	10/07/2025	12:37	LB137447

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>CDM Smith</u>	SDG No.: <u>Q3257</u>
Contract: <u>CAMP02</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Selenium	60.1	46.0	131	26	66	10/07/2025	12:37	LB137447
	Silver	205	201	102	170	232	10/07/2025	12:37	LB137447
	Sodium	-119			0	0	10/07/2025	12:37	LB137447
	Thallium	140	108	130	68	148	10/07/2025	12:37	LB137447
	Vanadium	475	491	97	417	565	10/07/2025	12:37	LB137447
	Zinc	1020	952	107	809	1095	10/07/2025	12:37	LB137447



METAL QC DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client: CDM Smith **level:** low **sdg no.:** Q3257
contract: CAMP02 **lab code:** ACE
matrix: Solid **sample id:** Q3256-01 **client id:** SOIL-PILE-1-WCMS

Percent Solids for Sample: 87.7 **Spiked ID:** Q3256-01MS **Percent Solids for Spike Sample:** 87.7

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	mg/Kg	75 - 125	7730		6980		95.0	790		P
Antimony	mg/Kg	75 - 125	17.8		2.42	U	38.0	47	N	P
Arsenic	mg/Kg	75 - 125	43.8		6.80		38.0	97		P
Barium	mg/Kg	75 - 125	68.3		66.9		9.5	15		P
Beryllium	mg/Kg	75 - 125	9.50		0.61		9.5	94		P
Cadmium	mg/Kg	75 - 125	10.8		0.28	J	9.5	110		P
Calcium	mg/Kg	75 - 125	8930		11400		47.5	-5176		P
Chromium	mg/Kg	75 - 125	32.6		13.0		19.0	103		P
Cobalt	mg/Kg	75 - 125	14.4		4.93		9.5	100		P
Copper	mg/Kg	75 - 125	48.8		31.6		14.3	120		P
Iron	mg/Kg	75 - 125	9340		10000		140	-504		P
Lead	mg/Kg	75 - 125	206		158		47.5	102		P
Magnesium	mg/Kg	75 - 125	1510		1400		95.0	112		P
Manganese	mg/Kg	75 - 125	132		157		9.5	-262		P
Nickel	mg/Kg	75 - 125	35.0		10.0		23.8	105		P
Potassium	mg/Kg	75 - 125	939		548		480	81		P
Selenium	mg/Kg	75 - 125	91.5		0.97	U	95.0	96		P
Silver	mg/Kg	75 - 125	2.47		0.48	U	3.6	69	N	P
Sodium	mg/Kg	75 - 125	1310		1370		140	-45		P
Thallium	mg/Kg	75 - 125	99.3		1.93	U	95.0	104		P
Vanadium	mg/Kg	75 - 125	30.8		17.6		14.3	92		P
Zinc	mg/Kg	75 - 125	117		109		9.5	84		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: CDM Smith **level:** low **sdg no.:** Q3257
contract: CAMP02 **lab code:** ACE
matrix: Solid **sample id:** Q3256-01 **client id:** SOIL-PILE-1-WCMSD

Percent Solids for Sample: 87.7 **Spiked ID:** Q3256-01MSD **Percent Solids for Spike Sample:** 87.7

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Aluminum	mg/Kg	75 - 125	7310		6980		95.0	347		P
Antimony	mg/Kg	75 - 125	17.0		2.42	U	38.0	45	N	P
Arsenic	mg/Kg	75 - 125	43.2		6.80		38.0	96		P
Barium	mg/Kg	75 - 125	68.1		66.9		9.5	13		P
Beryllium	mg/Kg	75 - 125	9.08		0.61		9.5	89		P
Cadmium	mg/Kg	75 - 125	10.2		0.28	J	9.5	104		P
Calcium	mg/Kg	75 - 125	9940		11400		47.5	-3033		P
Chromium	mg/Kg	75 - 125	43.3		13.0		19.0	160	N	P
Cobalt	mg/Kg	75 - 125	13.8		4.93		9.5	94		P
Copper	mg/Kg	75 - 125	48.1		31.6		14.3	115		P
Iron	mg/Kg	75 - 125	12600		10000		140	1829		P
Lead	mg/Kg	75 - 125	224		158		47.5	139	N	P
Magnesium	mg/Kg	75 - 125	1470		1400		95.0	73		P
Manganese	mg/Kg	75 - 125	138		157		9.5	-209		P
Nickel	mg/Kg	75 - 125	33.6		10.0		23.8	99		P
Potassium	mg/Kg	75 - 125	940		548		480	82		P
Selenium	mg/Kg	75 - 125	87.2		0.97	U	95.0	92		P
Silver	mg/Kg	75 - 125	2.06		0.48	U	3.6	57	N	P
Sodium	mg/Kg	75 - 125	1330		1370		140	-30		P
Thallium	mg/Kg	75 - 125	95.6		1.93	U	95.0	101		P
Vanadium	mg/Kg	75 - 125	31.5		17.6		14.3	98		P
Zinc	mg/Kg	75 - 125	119		109		9.5	100		P

metals
- 5a -
MATRIX SPIKE SUMMARY

client: CDM Smith level: low sdg no.: Q3257
contract: CAMP02 lab code: ACE
matrix: Solid sample id: Q3257-02 client id: ENV2-6FTMS
Percent Solids for Sample: 84.7 Spiked ID: Q3257-02MS Percent Solids for Spike Sample: 84.7

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	mg/Kg	80 - 120	0.31		0.016	U	0.3	103		CV

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client:	CDM Smith	level:	low	sdg no.:	Q3257
contract:	CAMP02			lab code:	ACE
matrix:	Solid	sample id:	Q3257-02	client id:	ENV2-6FTMSD

Percent Solids for Sample:	84.7	Spiked ID:	Q3257-02MSD	Percent Solids for Spike Sample:	84.7
----------------------------	------	------------	-------------	----------------------------------	------

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	mg/Kg	80 - 120	0.32		0.016	U	0.3	106		CV

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: <u>CDM Smith</u>	SDG No.: <u>Q3257</u>	
Contract: <u>CAMP02</u>	Lab Code: <u>ACE</u>	
Matrix: <u>Solid</u>	Level: <u>LOW</u>	Client ID: <u>SOIL-PILE-1-WCA</u>
Sample ID: <u>Q3256-01</u>	Spiked ID: <u>Q3256-01A</u>	

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Antimony	mg/Kg	75 - 125	33.4		2.42	U	38.7	86		P
Chromium	mg/Kg	75 - 125	31.2		13.0		19.3	94		P
Lead	mg/Kg	75 - 125	206		158		48.3	100		P
Silver	mg/Kg	75 - 125	2.26		0.48	U	3.60	63	N	P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>CDM Smith</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3257</u>
Contract:	<u>CAMP02</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Solid</u>	Sample ID:	<u>Q3256-01</u>	Client ID:	<u>SOIL-PILE-1-WCDUP</u>
Percent Solids for Sample:	87.7	Duplicate ID	Q3256-01DUP	Percent Solids for Spike Sample:	87.7

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	mg/Kg	20	6980		7060		1		P
Antimony	mg/Kg	20	2.42	U	2.48	U			P
Arsenic	mg/Kg	20	6.80		5.91		14		P
Barium	mg/Kg	20	66.9		67.6		1		P
Beryllium	mg/Kg	20	0.61		0.63		3		P
Cadmium	mg/Kg	20	0.28	J	0.30	J	5		P
Calcium	mg/Kg	20	11400		9620		17		P
Chromium	mg/Kg	20	13.0		12.6		3		P
Cobalt	mg/Kg	20	4.93		4.69		5		P
Copper	mg/Kg	20	31.6		28.2		11		P
Iron	mg/Kg	20	10000		9500		5		P
Lead	mg/Kg	20	158		159		1		P
Magnesium	mg/Kg	20	1400		1480		6		P
Manganese	mg/Kg	20	157		134		16		P
Nickel	mg/Kg	20	10.0		10.3		3		P
Potassium	mg/Kg	20	548		522		5		P
Selenium	mg/Kg	20	0.97	U	0.99	U			P
Silver	mg/Kg	20	0.48	U	0.50	U			P
Sodium	mg/Kg	20	1370		1210		12		P
Thallium	mg/Kg	20	1.93	U	1.98	U			P
Vanadium	mg/Kg	20	17.6		16.7		5		P
Zinc	mg/Kg	20	109		128		16		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>CDM Smith</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3257</u>
Contract:	<u>CAMP02</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Solid</u>	Sample ID:	<u>Q3256-01MS</u>	Client ID:	<u>SOIL-PILE-1-WCMSD</u>
Percent Solids for Sample:	87.7	Duplicate ID	Q3256-01MSD	Percent Solids for Spike Sample:	87.7

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	mg/Kg	20	7730		7310		6		P
Antimony	mg/Kg	20	17.8		17.0		5		P
Arsenic	mg/Kg	20	43.8		43.2		1		P
Barium	mg/Kg	20	68.3		68.1		0		P
Beryllium	mg/Kg	20	9.50		9.08		5		P
Cadmium	mg/Kg	20	10.8		10.2		6		P
Calcium	mg/Kg	20	8930		9940		11		P
Chromium	mg/Kg	20	32.6		43.3		28	*	P
Cobalt	mg/Kg	20	14.4		13.8		4		P
Copper	mg/Kg	20	48.8		48.1		1		P
Iron	mg/Kg	20	9340		12600		30	*	P
Lead	mg/Kg	20	206		224		8		P
Magnesium	mg/Kg	20	1510		1470		3		P
Manganese	mg/Kg	20	132		138		4		P
Nickel	mg/Kg	20	35.0		33.6		4		P
Potassium	mg/Kg	20	939		940		0		P
Selenium	mg/Kg	20	91.5		87.2		5		P
Silver	mg/Kg	20	2.47		2.06		18		P
Sodium	mg/Kg	20	1310		1330		2		P
Thallium	mg/Kg	20	99.3		95.6		4		P
Vanadium	mg/Kg	20	30.8		31.5		2		P
Zinc	mg/Kg	20	117		119		2		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>CDM Smith</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3257</u>
Contract:	<u>CAMP02</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Solid</u>	Sample ID:	<u>Q3257-02</u>	Client ID:	<u>ENV2-6FTDUP</u>
Percent Solids for Sample:	84.7	Duplicate ID	Q3257-02DUP	Percent Solids for Spike Sample:	84.7

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	mg/Kg	20	0.016	U	0.016	U			CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>CDM Smith</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3257</u>
Contract:	<u>CAMP02</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Solid</u>	Sample ID:	<u>Q3257-02MS</u>	Client ID:	<u>ENV2-6FTMSD</u>
Percent Solids for Sample:	84.7	Duplicate ID	Q3257-02MSD	Percent Solids for Spike Sample:	84.7

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	mg/Kg	20	0.31		0.32		4		CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169927BS Mercury	mg/Kg	0.25	0.27		106	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client:	<u>CDM Smith</u>	SDG No.:	<u>Q3257</u>
Contract:	<u>CAMP02</u>	Lab Code:	<u>ACE</u>

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169935BS							
Aluminum	mg/Kg	100	98.6		99	80 - 120	P
Antimony	mg/Kg	40.0	39.9		100	80 - 120	P
Arsenic	mg/Kg	40.0	38.3		96	80 - 120	P
Barium	mg/Kg	10.0	9.54		95	80 - 120	P
Beryllium	mg/Kg	10.0	9.92		99	80 - 120	P
Cadmium	mg/Kg	10.0	9.48		95	80 - 120	P
Calcium	mg/Kg	50.0	48.5	J	97	80 - 120	P
Chromium	mg/Kg	20.0	20.8		104	80 - 120	P
Cobalt	mg/Kg	10.0	9.57		96	80 - 120	P
Copper	mg/Kg	15.0	14.9		99	80 - 120	P
Iron	mg/Kg	150	151		101	80 - 120	P
Lead	mg/Kg	50.0	46.7		93	80 - 120	P
Magnesium	mg/Kg	100	99.5	J	100	80 - 120	P
Manganese	mg/Kg	10.0	9.99		100	80 - 120	P
Nickel	mg/Kg	25.0	24.8		99	80 - 120	P
Potassium	mg/Kg	500	491		98	80 - 120	P
Selenium	mg/Kg	100	97.3		97	80 - 120	P
Silver	mg/Kg	3.8	3.67		97	80 - 120	P
Sodium	mg/Kg	150	151		101	80 - 120	P
Thallium	mg/Kg	100	94.9		95	80 - 120	P
Vanadium	mg/Kg	15.0	14.5		97	80 - 120	P
Zinc	mg/Kg	10.0	10.0		100	80 - 120	P

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

SOIL-PILE-1-WCL

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **Lb No.:** lb137406 **Lab Sample ID :** Q3256-01L **SDG No.:** Q3257
Matrix (soil/water): Solid **Level (low/med):** LOW
Concentration Units: mg/Kg

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Differ-ence	Q	M
Aluminum	6980		7720		11		P
Antimony	2.42	U	12.1	U			P
Arsenic	6.80		6.75		1		P
Barium	66.9		79.4		19		P
Beryllium	0.61		0.92	J	50		P
Cadmium	0.28	J	0.17	J	39		P
Calcium	11400		12800		12		P
Chromium	13.0		14.5		12		P
Cobalt	4.93		4.49	J	9		P
Copper	31.6		34.5		9		P
Iron	10000		10300		3		P
Lead	158		152		4		P
Magnesium	1400		1620		16		P
Manganese	157		176		12		P
Nickel	10.0		9.20	J	8		P
Potassium	548		475	J	13		P
Selenium	0.97	U	4.83	U			P
Silver	0.48	U	2.42	U			P
Sodium	1370		1270		8		P
Thallium	1.93	U	1.43	J	100.0		P
Vanadium	17.6		20.6		17		P
Zinc	109		118		8		P

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

ENV2-6FTL

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **Lb No.:** lb137376 **Lab Sample ID :** Q3257-02L **SDG No.:** Q3257
Matrix (soil/water): Solid **Level (low/med):** LOW
Concentration Units: mg/Kg

Analyte	Initial Sample Result (I)	Serial Dilution Result (S)	% Differ- ence	Q	M
Mercury	0.016 U	0.079 U			CV



METAL PREPARATION & INSTRUMENT DATA

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Al	Ca	Fe	Mg	Ag
Aluminum	396.100	0.0000000	-0.0002060	0.0000000	0.0000000	0.0000000
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	-0.0000440	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000930	0.0000000	0.0000000
Calcium	373.690	0.0000000	0.0000000	-0.0075970	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0007850	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	-0.0000920	0.0000000	0.0000380	0.0000000	0.0000000
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	-0.0001440	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	-0.0001490	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0001050	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMEN CORRECTION FACTORS

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Instrument ID:

Date:

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		As	Ba	Be	Cd	Co
Aluminum	396.100	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0002870
Calcium	373.690	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0000000	0.0009530
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	-0.0039600
Lead	220.353	0.0000000	0.0003170	0.0000000	0.0000000	0.0000000
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	-0.0003570
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0054900
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Cr	Cu	K	Mn	Mo
Aluminum	396.100	0.0000000	0.0000000	0.0000590	0.0000000	0.0396900
Antimony	206.833	0.0122000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	-0.0029000	0.0000000	0.0000000	0.0000000	0.0004900
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	-0.0000710	-0.0003400
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Calcium	373.690	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000070	0.0002200	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	-0.0007860
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0006510	0.0020500
Iron	240.488	0.0000000	0.0000000	0.0000730	0.0000000	-0.0015250
Lead	220.353	0.0000000	0.0000000	0.0000000	0.0001400	-0.0008600
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0007460	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	-0.0000120
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0017400	-0.0100400
Vanadium	292.402	-0.0025100	0.0000000	0.0000000	0.0000000	-0.0072000
Zinc	213.800	0.0000000	0.0009010	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Na	Ni	Pb	Sb	Se
Aluminum	396.100	0.0000000	0.0000000	0.0012800	0.0000000	0.0000000
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Calcium	373.690	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	-0.0047000	0.0036100	0.0000000	0.0000000
Iron	240.488	0.0000000	-0.0017000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	0.0006580	0.0000000	0.0000000	0.0001290
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0003330	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0067600	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMEN CORRECTION FACTORS

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Instrument ID:

Date:

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Sn	Ti	Tl	V	Zn
Aluminum	396.100	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Antimony	206.833	-0.0035600	-0.0007970	0.0000000	-0.0018900	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000630	0.0001280	0.0000000	0.0000000
Calcium	373.690	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0001110	0.0000000
Cobalt	228.616	0.0000000	0.0018800	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0003840	0.0000000	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	-0.0003610	0.0000000	0.0000000	0.0000000
Magnesium	279.079	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Silver	328.068	0.0000000	-0.0007420	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	-0.0039700	0.0000000	-0.0115600	0.0000000
Vanadium	292.402	0.0000000	0.0005320	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

METAL PREPARATION & ANALYICAL SUMMARY

Metals
- 13 -

SAMPLE PREPARATION SUMMARY

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Method: _____

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(g)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB169927							
PB169927BL	PB169927BL	MB	SOLID	09/30/2025	0.57	35.0	100.00
PB169927BS	PB169927BS	LCS	SOLID	09/30/2025	0.55	35.0	100.00
Q3257-01	ENV1-6FT	SAM	SOLID	09/30/2025	0.54	35.0	84.30
Q3257-02	ENV2-6FT	SAM	SOLID	09/30/2025	0.52	35.0	84.70
Q3257-02DUP	ENV2-6FTDUP	DUP	SOLID	09/30/2025	0.53	35.0	84.70
Q3257-02MS	ENV2-6FTMS	MS	SOLID	09/30/2025	0.56	35.0	84.70
Q3257-02MSD	ENV2-6FTMSD	MSD	SOLID	09/30/2025	0.55	35.0	84.70

Metals
- 13 -

SAMPLE PREPARATION SUMMARY

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Method: _____

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(g)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB169935							
PB169935BL	PB169935BL	MB	SOLID	10/01/2025	2.00	100.0	100.00
PB169935BS	PB169935BS	LCS	SOLID	10/01/2025	2.00	100.0	100.00
Q3256-01DUP	SOIL-PILE-1-WCDUP	DUP	SOLID	10/01/2025	2.30	100.0	87.70
Q3256-01MS	SOIL-PILE-1-WCMS	MS	SOLID	10/01/2025	2.40	100.0	87.70
Q3256-01MSD	SOIL-PILE-1-WCMSD	MSD	SOLID	10/01/2025	2.40	100.0	87.70
Q3257-01	ENV1-6FT	SAM	SOLID	10/01/2025	2.26	100.0	84.30
Q3257-02	ENV2-6FT	SAM	SOLID	10/01/2025	2.40	100.0	84.70

metals
- 14 -
ANALYSIS RUN LOG

Client: CDM Smith

Contract: CAMP02

Lab code: ACE

Sdg no.: Q3257

Instrument id number: _____

Method: _____

Run number: LB137376

Start date: 10/01/2025

End date: 10/01/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	0851	HG
S0.2	S0.2	1	0854	HG
S2.5	S2.5	1	0856	HG
S5	S5	1	0858	HG
S7.5	S7.5	1	0901	HG
S10	S10	1	0906	HG
ICV53	ICV53	1	0910	HG
ICB53	ICB53	1	0912	HG
CCV67	CCV67	1	0915	HG
CCB67	CCB67	1	0917	HG
CRA	CRA	1	0920	HG
PB169927BL	PB169927BL	1	0927	HG
PB169927BS	PB169927BS	1	0930	HG
Q3257-01	ENV1-6FT	1	0942	HG
CCV68	CCV68	1	0945	HG
CCB68	CCB68	1	0947	HG
Q3257-02	ENV2-6FT	1	0950	HG
Q3257-02DUP	ENV2-6FTDUP	1	0952	HG
Q3257-02MS	ENV2-6FTMS	1	0954	HG
Q3257-02MSD	ENV2-6FTMSD	1	0956	HG
Q3257-02L	ENV2-6FTL	5	0959	HG
CCV69	CCV69	1	1007	HG
CCB69	CCB69	1	1009	HG

metals
- 14 -
ANALYSIS RUN LOG

Client: CDM Smith

Contract: CAMP02

Lab code: ACE

Sdg no.: Q3257

Instrument id number:

Method:

Run number: LB137406

Start date: 10/02/2025

End date: 10/02/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1037	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S1	S1	1	1042	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S2	S2	1	1046	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S3	S3	1	1050	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S4	S4	1	1054	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S5	S5	1	1058	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICV01	ICV01	1	1107	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
LLICV01	LLICV01	1	1111	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICB01	ICB01	1	1127	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CRI01	CRI01	1	1136	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSA01	ICSA01	1	1142	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSAB01	ICSAB01	1	1146	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV01	CCV01	1	1158	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB01	CCB01	1	1202	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3256-01DUP	SOIL-PILE-1-WCDUP	1	1304	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV02	CCV02	1	1348	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB02	CCB02	1	1352	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3256-01L	SOIL-PILE-1-WCL	5	1407	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3256-01MS	SOIL-PILE-1-WCMS	1	1411	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3256-01MSD	SOIL-PILE-1-WCMSD	1	1415	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3256-01A	SOIL-PILE-1-WCA	1	1419	Ag,Cr,Pb,Sb
Q3257-01	ENV1-6FT	1	1426	Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3257-02	ENV2-6FT	1	1430	Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV03	CCV03	1	1447	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB03	CCB03	1	1453	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV04	CCV04	1	1554	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB04	CCB04	1	1605	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV05	CCV05	1	1731	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB05	CCB05	1	1735	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV06	CCV06	1	1817	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB06	CCB06	1	1821	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn

metals
- 14 -
ANALYSIS RUN LOG

Client: CDM Smith

Contract: CAMP02

Lab code: ACE

Sdg no.: Q3257

Instrument id number:

Method:

Run number: LB137426

Start date: 10/03/2025

End date: 10/03/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1153	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S1	S1	1	1158	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S2	S2	1	1202	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S3	S3	1	1206	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S4	S4	1	1210	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S5	S5	1	1214	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICV01	ICV01	1	1218	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
LLICV01	LLICV01	1	1305	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICB01	ICB01	1	1312	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CRI01	CRI01	1	1316	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSA01	ICSA01	1	1320	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSAB01	ICSAB01	1	1325	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV01	CCV01	1	1337	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB01	CCB01	1	1341	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
PB169935BS	PB169935BS	1	1403	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
PB169935BL	PB169935BL	1	1433	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV02	CCV02	1	1500	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB02	CCB02	1	1504	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV03	CCV03	1	1551	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB03	CCB03	1	1600	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV04	CCV04	1	1645	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB04	CCB04	1	1649	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn

metals
- 14 -
ANALYSIS RUN LOG

Client: CDM Smith

Contract: CAMP02

Lab code: ACE

Sdg no.: Q3257

Instrument id number: _____

Method: _____

Run number: LB137447

Start date: 10/07/2025

End date: 10/07/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1029	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S1	S1	1	1034	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S2	S2	1	1038	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S3	S3	1	1042	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S4	S4	1	1046	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S5	S5	1	1050	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICV01	ICV01	1	1054	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
LLICV01	LLICV01	1	1201	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICB01	ICB01	1	1206	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CRI01	CRI01	1	1210	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSA01	ICSA01	1	1214	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSAB01	ICSAB01	1	1237	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV01	CCV01	1	1250	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB01	CCB01	1	1258	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV02	CCV02	1	1419	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB02	CCB02	1	1424	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV03	CCV03	1	1508	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB03	CCB03	1	1512	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV04	CCV04	1	1600	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB04	CCB04	1	1604	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV05	CCV05	1	1650	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB05	CCB05	1	1701	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV06	CCV06	1	1749	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB06	CCB06	1	1759	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV07	CCV07	1	1844	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB07	CCB07	1	1848	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q3257-01	ENV1-6FT	5	1912	Ag
Q3257-02	ENV2-6FT	5	1916	Ag
CCV08	CCV08	1	1920	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB08	CCB08	1	1924	Ag,Al,As,Ba,Be,Ca,Cd,Co,Cr,Cu,Fe,K,Mg,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn