

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169927BS Mercury	mg/Kg	0.25	0.27		106	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith

SDG No.: Q3257

Contract: CAMP02

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169935BS							
Aluminum	mg/Kg	100	98.6		99	80 - 120	P
Antimony	mg/Kg	40.0	39.9		100	80 - 120	P
Arsenic	mg/Kg	40.0	38.3		96	80 - 120	P
Barium	mg/Kg	10.0	9.54		95	80 - 120	P
Beryllium	mg/Kg	10.0	9.92		99	80 - 120	P
Cadmium	mg/Kg	10.0	9.48		95	80 - 120	P
Calcium	mg/Kg	50.0	48.5	J	97	80 - 120	P
Chromium	mg/Kg	20.0	20.8		104	80 - 120	P
Cobalt	mg/Kg	10.0	9.57		96	80 - 120	P
Copper	mg/Kg	15.0	14.9		99	80 - 120	P
Iron	mg/Kg	150	151		101	80 - 120	P
Lead	mg/Kg	50.0	46.7		93	80 - 120	P
Magnesium	mg/Kg	100	99.5	J	100	80 - 120	P
Manganese	mg/Kg	10.0	9.99		100	80 - 120	P
Nickel	mg/Kg	25.0	24.8		99	80 - 120	P
Potassium	mg/Kg	500	491		98	80 - 120	P
Selenium	mg/Kg	100	97.3		97	80 - 120	P
Silver	mg/Kg	3.8	3.67		97	80 - 120	P
Sodium	mg/Kg	150	151		101	80 - 120	P
Thallium	mg/Kg	100	94.9		95	80 - 120	P
Vanadium	mg/Kg	15.0	14.5		97	80 - 120	P
Zinc	mg/Kg	10.0	10.0		100	80 - 120	P