

Continuing Calibration Report for SequenceID : FE100225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE056147.D

Aliphatic C9-C12	7965010.000	60.000	3.311	6.945	132750.167	140027.957	5.197
Aliphatic C12-C16	6502829.000	40.000	6.946	10.395	162570.725	153376.491	-5.995
Aliphatic C16-C21	10562347.000	60.000	10.396	13.772	176039.117	165879.737	-6.125
Aliphatic C21-C28	12151939.000	80.000	13.773	17.442	151899.238	148527.151	-2.270
Aliphatic C28-C40	16261900.000	120.000	17.443	22.448	135515.833	136251.923	0.540
Aliphatic EPH	53444025.000	360.000	3.311	22.448	148455.625	146449.789	-1.370