

Continuing Calibration Report for SequenceID : FE100225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE056160.D

Aliphatic C9-C12	8211859.000	60.000	3.311	6.945	136864.317	140027.957	2.259
Aliphatic C12-C16	6753799.000	40.000	6.946	10.395	168844.975	153376.491	-10.085
Aliphatic C16-C21	11048212.000	60.000	10.396	13.772	184136.867	165879.737	-11.006
Aliphatic C21-C28	12784235.000	80.000	13.773	17.442	159802.938	148527.151	-7.592
Aliphatic C28-C40	17078354.000	120.000	17.443	22.448	142319.617	136251.923	-4.453
Aliphatic EPH	55876459.000	360.000	3.311	22.448	155212.386	146449.789	-5.983