

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

ENV2-6FTMS

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q3257

Lab Sample ID: Q3257-02MS

Date(s) Analyzed: 10/03/2025 10/03/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP

ID: 0.32 (mm)

GC Column:(2): RTX-CLP2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
DICHLORPROP	1	8.17	8.12	8.22	88.2	12.1
	2	8.65	8.60	8.70	78.1	
2,4-D	1	8.42	8.37	8.47	56.3	12.8
	2	8.99	8.94	9.04	64.0	
2,4,5-TP(Silvex)	1	9.28	9.23	9.33	69.6	28.9
	2	9.89	9.84	9.94	93.1	
2,4,5-T	1	9.58	9.53	9.63	62.2	15.4
	2	10.32	10.27	10.37	72.6	
2,4-DB	1	10.16	10.11	10.21	67.3	21
	2	10.89	10.84	10.94	54.5	
Dinoseb	1	11.35	11.30	11.40	45.4	29.5
	2	11.26	11.21	11.31	61.1	
DICAMBA	1	7.46	7.41	7.51	61.6	20.3
	2	7.93	7.88	7.98	75.5	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

ENV2-6FTMSD

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q3257

Lab Sample ID: Q3257-02MSD

Date(s) Analyzed: 10/03/2025 10/03/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP

ID: 0.32 (mm)

GC Column:(2): RTX-CLP2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
DICAMBA	1	7.46	7.41	7.51	61.4	19.9
	2	7.93	7.88	7.98	75.0	
DICHLORPROP	1	8.17	8.12	8.22	84.8	8.2
	2	8.65	8.60	8.70	78.1	
2,4-D	1	8.42	8.37	8.47	62.1	15.2
	2	9.00	8.95	9.05	72.3	
2,4,5-TP(Silvex)	1	9.28	9.23	9.33	69.6	29.1
	2	9.89	9.84	9.94	93.3	
2,4,5-T	1	9.58	9.53	9.63	63.8	12.6
	2	10.32	10.27	10.37	72.4	
2,4-DB	1	10.16	10.11	10.21	61.4	14.9
	2	10.89	10.84	10.94	52.9	
Dinoseb	1	11.35	11.30	11.40	44.1	25.9
	2	11.26	11.21	11.31	57.2	

COMPOUND DETECTION SUMMARY

CLIENT SAMPLE NO.

PB169931BS

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q3257

Lab Sample ID: PB169931BS

Date(s) Analyzed: 10/01/2025 10/01/2025

Instrument ID (1): ECD_S

Instrument ID (2): ECD_S

GC Column: (1): RTX-CLP

ID: 0.32 (mm)

GC Column:(2): RTX-CLP2

ID: 0.32 (mm)

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	%RPD
			FROM	TO		
2,4,5-T	1	9.58	9.53	9.63	186	10.8
	2	10.32	10.27	10.37	167	
2,4,5-TP(Silvex)	1	9.28	9.23	9.33	180	4
	2	9.89	9.84	9.94	173	
2,4-D	1	8.41	8.36	8.46	197	16.5
	2	9.00	8.95	9.05	167	
2,4-DB	1	10.16	10.11	10.21	179	18.3
	2	10.89	10.84	10.94	149	
DICHLORPROP	1	8.17	8.12	8.22	175	1.7
	2	8.66	8.61	8.71	178	
Dinoseb	1	11.35	11.30	11.40	189	1.1
	2	11.26	11.21	11.31	187	
DICAMBA	1	7.47	7.42	7.52	172	2.4
	2	7.94	7.89	7.99	168	