

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q3257

Continuing Calib Date: 10/01/2025

Initial Calibration Date(s): 09/09/2025

09/09/2025

Continuing Calib Time: 18:09

Initial Calibration Time(s): 11:32

13:08

GC Column: RTX-CLP ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|------------------|------------|-----------|------------------------|-------|------------|
| DICAMBA          | 7.46       | 7.47      | 7.37                   | 7.57  | 0.01       |
| 2,4-DCAA         | 7.28       | 7.29      | 7.19                   | 7.39  | 0.01       |
| DICHLORPROP      | 8.17       | 8.18      | 8.08                   | 8.28  | 0.01       |
| 2,4-D            | 8.41       | 8.41      | 8.31                   | 8.51  | 0.00       |
| 2,4,5-TP(Silvex) | 9.28       | 9.29      | 9.19                   | 9.39  | 0.01       |
| 2,4,5-T          | 9.58       | 9.58      | 9.48                   | 9.68  | 0.00       |
| 2,4-DB           | 10.15      | 10.16     | 10.06                  | 10.26 | 0.01       |
| Dinoseb          | 11.35      | 11.35     | 11.25                  | 11.45 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q3257

Continuing Calib Date: 10/01/2025

Initial Calibration Date(s): 09/09/2025

09/09/2025

Continuing Calib Time: 18:09

Initial Calibration Time(s): 11:32

13:08

GC Column: RTX-CLP2 ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|------------------|------------|-----------|------------------------|-------|------------|
| DICAMBA          | 7.94       | 7.94      | 7.84                   | 8.04  | 0.00       |
| 2,4-DCAA         | 7.74       | 7.74      | 7.64                   | 7.84  | 0.00       |
| DICHLORPROP      | 8.66       | 8.66      | 8.56                   | 8.76  | 0.00       |
| 2,4-D            | 9.00       | 9.00      | 8.90                   | 9.10  | 0.00       |
| 2,4,5-TP(Silvex) | 9.89       | 9.89      | 9.79                   | 9.99  | 0.00       |
| 2,4,5-T          | 10.32      | 10.32     | 10.22                  | 10.42 | 0.00       |
| 2,4-DB           | 10.89      | 10.89     | 10.79                  | 10.99 | 0.00       |
| Dinoseb          | 11.26      | 11.27     | 11.17                  | 11.37 | 0.01       |