284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02

Lab Code: ACE **SDG NO.:** Q3257

Continuing Calib Date: 10/02/2025 **Initial Calibration Date(s):** 09/09/2025 09/09/2025

Continuing Calib Time: 14:05 **Initial Calibration Time(s):** 11:32 13:08

GC Column: RTX-CLP **ID:** 0.32 **(mm)**

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
DICAMBA	7.46	7.47	7.37	7.57	0.01
2,4-DCAA	7.28	7.29	7.19	7.39	0.01
DICHLORPROP	8.17	8.18	8.08	8.28	0.01
2,4-D	8.41	8.41	8.31	8.51	0.00
2,4,5-TP(Silvex)	9.28	9.29	9.19	9.39	0.01
2,4,5-T	9.58	9.58	9.48	9.68	0.01
2,4-DB	10.15	10.16	10.06	10.26	0.01
Dinoseb	11.34	11.35	11.25	11.45	0.01

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02

Lab Code: ACE **SDG NO.:** Q3257

Continuing Calib Date: 10/02/2025 **Initial Calibration Date(s):** 09/09/2025 09/09/2025

Continuing Calib Time: 14:05 **Initial Calibration Time(s):** 11:32 13:08

GC Column: RTX-CLP2 **ID:** 0.32 **(mm)**

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
DICAMBA	7.94	7.94	7.84	8.04	0.01
2,4-DCAA	7.74	7.74	7.64	7.84	0.00
DICHLORPROP	8.65	8.66	8.56	8.76	0.01
2,4-D	9.00	9.00	8.90	9.10	0.00
2,4,5-TP(Silvex)	9.89	9.89	9.79	9.99	0.00
2,4,5-T	10.32	10.32	10.22	10.42	0.00
2,4-DB	10.89	10.89	10.79	10.99	0.00
Dinoseb	11.26	11.27	11.17	11.37	0.01