

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3257
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/01/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031913.D **Time Analyzed:** 13:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.577	9.483	9.683	728.780	712.500	2.3
2,4,5-TP(Silvex)	9.280	9.185	9.385	695.950	712.500	-2.3
2,4-D	8.409	8.314	8.514	740.720	705.000	5.1
2,4-DB	10.154	10.058	10.258	723.760	712.500	1.6
2,4-DCAA	7.281	7.185	7.385	720.030	750.000	-4.0
DICAMBA	7.464	7.369	7.569	674.850	705.000	-4.3
DICHLORPROP	8.172	8.077	8.277	675.310	705.000	-4.2
Dinoseb	11.348	11.254	11.454	728.050	705.000	3.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3257
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/01/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031913.D **Time Analyzed:** 13:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.318	10.221	10.421	673.380	712.500	-5.5
2,4,5-TP(Silvex)	9.888	9.793	9.993	682.090	712.500	-4.3
2,4-D	8.998	8.900	9.100	660.740	705.000	-6.3
2,4-DB	10.885	10.790	10.990	620.090	712.500	-13.0
2,4-DCAA	7.740	7.643	7.843	704.240	750.000	-6.1
DICAMBA	7.935	7.839	8.039	669.390	705.000	-5.1
DICHLORPROP	8.654	8.559	8.759	693.190	705.000	-1.7
Dinoseb	11.261	11.166	11.366	736.440	705.000	4.5