

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3257
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/01/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031918.D **Time Analyzed:** 18:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.576	9.483	9.683	732.220	712.500	2.8
2,4,5-TP(Silvex)	9.280	9.185	9.385	699.540	712.500	-1.8
2,4-D	8.410	8.314	8.514	756.810	705.000	7.3
2,4-DB	10.154	10.058	10.258	705.490	712.500	-1.0
2,4-DCAA	7.281	7.185	7.385	719.030	750.000	-4.1
DICAMBA	7.464	7.369	7.569	674.040	705.000	-4.4
DICHLORPROP	8.172	8.077	8.277	677.680	705.000	-3.9
Dinoseb	11.348	11.254	11.454	732.710	705.000	3.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3257
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/01/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031918.D **Time Analyzed:** 18:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.319	10.221	10.421	676.120	712.500	-5.1
2,4,5-TP(Silvex)	9.890	9.793	9.993	682.000	712.500	-4.3
2,4-D	8.999	8.900	9.100	662.770	705.000	-6.0
2,4-DB	10.888	10.790	10.990	614.570	712.500	-13.7
2,4-DCAA	7.742	7.643	7.843	709.240	750.000	-5.4
DICAMBA	7.937	7.839	8.039	669.700	705.000	-5.0
DICHLORPROP	8.656	8.559	8.759	665.770	705.000	-5.6
Dinoseb	11.263	11.166	11.366	735.800	705.000	4.4