

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3257
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/02/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031921.D **Time Analyzed:** 10:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.576	9.483	9.683	704.360	712.500	-1.1
2,4,5-TP(Silvex)	9.278	9.185	9.385	688.610	712.500	-3.4
2,4-D	8.409	8.314	8.514	722.160	705.000	2.4
2,4-DB	10.152	10.058	10.258	657.830	712.500	-7.7
2,4-DCAA	7.280	7.185	7.385	683.620	750.000	-8.9
DICAMBA	7.462	7.369	7.569	666.500	705.000	-5.5
DICHLORPROP	8.171	8.077	8.277	664.430	705.000	-5.8
Dinoseb	11.344	11.254	11.454	738.320	705.000	4.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3257
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/02/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031921.D **Time Analyzed:** 10:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.317	10.221	10.421	653.360	712.500	-8.3
2,4,5-TP(Silvex)	9.887	9.793	9.993	676.510	712.500	-5.1
2,4-D	8.998	8.900	9.100	642.250	705.000	-8.9
2,4-DB	10.886	10.790	10.990	581.470	712.500	-18.4
2,4-DCAA	7.740	7.643	7.843	689.290	750.000	-8.1
DICAMBA	7.934	7.839	8.039	665.950	705.000	-5.5
DICHLORPROP	8.654	8.559	8.759	685.790	705.000	-2.7
Dinoseb	11.259	11.166	11.366	753.790	705.000	6.9