

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3257
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/02/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031929.D **Time Analyzed:** 14:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.575	9.483	9.683	730.650	712.500	2.5
2,4,5-TP(Silvex)	9.278	9.185	9.385	694.880	712.500	-2.5
2,4-D	8.407	8.314	8.514	757.520	705.000	7.4
2,4-DB	10.150	10.058	10.258	748.890	712.500	5.1
2,4-DCAA	7.280	7.185	7.385	719.200	750.000	-4.1
DICAMBA	7.463	7.369	7.569	669.340	705.000	-5.1
DICHLORPROP	8.170	8.077	8.277	673.650	705.000	-4.4
Dinoseb	11.344	11.254	11.454	735.350	705.000	4.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3257
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL04 **Date Analyzed:** 10/02/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031929.D **Time Analyzed:** 14:05

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.317	10.221	10.421	692.950	712.500	-2.7
2,4,5-TP(Silvex)	9.888	9.793	9.993	693.210	712.500	-2.7
2,4-D	8.996	8.900	9.100	678.040	705.000	-3.8
2,4-DB	10.885	10.790	10.990	590.960	712.500	-17.1
2,4-DCAA	7.739	7.643	7.843	726.000	750.000	-3.2
DICAMBA	7.935	7.839	8.039	682.080	705.000	-3.3
DICHLORPROP	8.654	8.559	8.759	679.370	705.000	-3.6
Dinoseb	11.259	11.166	11.366	759.620	705.000	7.7