

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3257
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/03/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031941.D **Time Analyzed:** 09:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.578	9.483	9.683	734.040	712.500	3.0
2,4,5-TP(Silvex)	9.279	9.185	9.385	711.950	712.500	-0.1
2,4-D	8.410	8.314	8.514	674.060	705.000	-4.4
2,4-DB	10.153	10.058	10.258	573.360	712.500	-19.5
2,4-DCAA	7.281	7.185	7.385	733.770	750.000	-2.2
DICAMBA	7.463	7.369	7.569	692.490	705.000	-1.8
DICHLORPROP	8.171	8.077	8.277	689.420	705.000	-2.2
Dinoseb	11.346	11.254	11.454	757.370	705.000	7.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3257
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL05 **Date Analyzed:** 10/03/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031941.D **Time Analyzed:** 09:58

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.318	10.221	10.421	680.880	712.500	-4.4
2,4,5-TP(Silvex)	9.888	9.793	9.993	706.290	712.500	-0.9
2,4-D	8.999	8.900	9.100	666.300	705.000	-5.5
2,4-DB	10.887	10.790	10.990	580.670	712.500	-18.5
2,4-DCAA	7.741	7.643	7.843	719.490	750.000	-4.1
DICAMBA	7.936	7.839	8.039	696.580	705.000	-1.2
DICHLORPROP	8.655	8.559	8.759	713.700	705.000	1.2
Dinoseb	11.260	11.166	11.366	774.510	705.000	9.9