

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3257
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL06 **Date Analyzed:** 10/03/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031945.D **Time Analyzed:** 14:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.577	9.483	9.683	760.220	712.500	6.7
2,4,5-TP(Silvex)	9.279	9.185	9.385	723.390	712.500	1.5
2,4-D	8.409	8.314	8.514	692.220	705.000	-1.8
2,4-DB	10.151	10.058	10.258	679.770	712.500	-4.6
2,4-DCAA	7.281	7.185	7.385	748.450	750.000	-0.2
DICAMBA	7.464	7.369	7.569	702.070	705.000	-0.4
DICHLORPROP	8.172	8.077	8.277	701.870	705.000	-0.4
Dinoseb	11.346	11.254	11.454	782.960	705.000	11.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance
Contract: CAMP02
Lab Code: ACE
SDG NO.: Q3257
GC Column: RTX-CLP2
ID: 0.32 (mm)
 Initi. Calib. Date(s): 09/09/2025 09/09/2025

Client Sample No.: CCAL06
Date Analyzed: 10/03/2025
Lab Sample No.: HSTDCCC750
Data File : PS031945.D
Time Analyzed: 14:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.318	10.221	10.421	706.990	712.500	-0.8
2,4,5-TP(Silvex)	9.889	9.793	9.993	719.650	712.500	1.0
2,4-D	8.999	8.900	9.100	694.570	705.000	-1.5
2,4-DB	10.886	10.790	10.990	634.370	712.500	-11.0
2,4-DCAA	7.741	7.643	7.843	742.930	750.000	-0.9
DICAMBA	7.936	7.839	8.039	709.290	705.000	0.6
DICHLORPROP	8.655	8.559	8.759	729.060	705.000	3.4
Dinoseb	11.260	11.166	11.366	804.060	705.000	14.1