

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3257</u>	Analytical Method:	<u>8151A</u>
Client:	<u>CDM Smith</u>	DataFile :	<u>PS031943.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD		Low	Limits	
			Result	Result				RPD	Qual		High	RPD
Lab Sample ID:	Q3257-02MSD (Column 1)	Client Sample ID:		ENV2-6FTMSD								
	DICAMBA	196.5	0	61.4	ug/Kg	31		0		10	112	20
	DICHLORPROP	196.5	0	84.8	ug/Kg	43		5		10	113	20
	2,4-D	196.5	0	62.1	ug/Kg	32		10		10	144	20
	2,4,5-TP(Silvex)	196.5	0	69.6	ug/Kg	35		0		10	114	20
	2,4,5-T	196.5	0	63.8	ug/Kg	32		0		10	115	20
	2,4-DB	196.5	0	61.4	ug/Kg	31		9		10	140	20
	Dinoseb	196.5	0	44.1	ug/Kg	22		4		10	118	20
Lab Sample ID:	Q3257-02MSD (Column 2)	Client Sample ID:		ENV2-6FTMSD								
	DICAMBA	196.5	0	75.0	ug/Kg	38		0		10	112	20
	DICHLORPROP	196.5	0	78.1	ug/Kg	40		0		10	113	20
	2,4-D	196.5	0	72.3	ug/Kg	37		11		10	144	20
	2,4,5-TP(Silvex)	196.5	0	93.3	ug/Kg	47		0		10	114	20
	2,4,5-T	196.5	0	72.4	ug/Kg	37		0		10	115	20
	2,4-DB	196.5	0	52.9	ug/Kg	27		4		10	140	20
	Dinoseb	196.5	0	57.2	ug/Kg	29		7		10	118	20