

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3257
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/01/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097469.D **Time Analyzed:** 10:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.663	6.566	6.766	55.030	50.000	10.1
4,4'-DDE	6.155	6.056	6.256	52.670	50.000	5.3
4,4'-DDT	6.978	6.880	7.080	51.120	50.000	2.2
Aldrin	5.230	5.132	5.332	49.480	50.000	-1.0
alpha-BHC	3.972	3.875	4.075	50.010	50.000	0.0
alpha-Chlordane	5.983	5.886	6.086	50.410	50.000	0.8
beta-BHC	4.488	4.390	4.590	49.380	50.000	-1.2
Decachlorobiphenyl	9.000	8.902	9.102	47.370	50.000	-5.3
delta-BHC	4.732	4.635	4.835	51.310	50.000	2.6
Dieldrin	6.304	6.206	6.406	51.740	50.000	3.5
Endosulfan I	6.030	5.934	6.134	50.370	50.000	0.7
Endosulfan II	6.743	6.646	6.846	47.750	50.000	-4.5
Endosulfan sulfate	7.107	7.007	7.207	52.870	50.000	5.7
Endrin	6.531	6.432	6.632	53.250	50.000	6.5
Endrin aldehyde	6.872	6.774	6.974	55.820	50.000	11.6
Endrin ketone	7.586	7.487	7.687	54.630	50.000	9.3
gamma-BHC (Lindane)	4.300	4.202	4.402	49.130	50.000	-1.7
gamma-Chlordane	5.903	5.806	6.006	51.340	50.000	2.7
Heptachlor	4.891	4.793	4.993	48.080	50.000	-3.8
Heptachlor epoxide	5.648	5.552	5.752	52.150	50.000	4.3
Methoxychlor	7.451	7.352	7.552	48.420	50.000	-3.2
Tetrachloro-m-xylene	3.526	3.428	3.628	45.630	50.000	-8.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3257
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/01/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097469.D **Time Analyzed:** 10:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.847	5.750	5.950	48.060	50.000	-3.9
4,4'-DDE	5.295	5.197	5.397	46.430	50.000	-7.1
4,4'-DDT	6.100	6.002	6.202	41.890	50.000	-16.2
Aldrin	4.288	4.191	4.391	47.560	50.000	-4.9
alpha-BHC	3.327	3.229	3.429	48.060	50.000	-3.9
alpha-Chlordane	5.106	5.009	5.209	46.580	50.000	-6.8
beta-BHC	3.955	3.857	4.057	44.490	50.000	-11.0
Decachlorobiphenyl	7.982	7.884	8.084	40.080	50.000	-19.8
delta-BHC	4.187	4.090	4.290	46.420	50.000	-7.2
Dieldrin	5.424	5.328	5.528	48.470	50.000	-3.1
Endosulfan I	5.161	5.064	5.264	43.050	50.000	-13.9
Endosulfan II	5.992	5.895	6.095	47.060	50.000	-5.9
Endosulfan sulfate	6.394	6.296	6.496	46.440	50.000	-7.1
Endrin	5.701	5.603	5.803	54.780	50.000	9.6
Endrin aldehyde	6.170	6.073	6.273	46.110	50.000	-7.8
Endrin ketone	6.898	6.800	7.000	52.710	50.000	5.4
gamma-BHC (Lindane)	3.659	3.561	3.761	47.190	50.000	-5.6
gamma-Chlordane	5.042	4.945	5.145	44.960	50.000	-10.1
Heptachlor	4.006	3.909	4.109	45.790	50.000	-8.4
Heptachlor epoxide	4.790	4.693	4.893	48.310	50.000	-3.4
Methoxychlor	6.672	6.574	6.774	41.340	50.000	-17.3
Tetrachloro-m-xylene	2.822	2.724	2.924	42.810	50.000	-14.4