

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3257
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/01/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097491.D **Time Analyzed:** 16:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.661	6.566	6.766	54.360	50.000	8.7
4,4'-DDE	6.154	6.056	6.256	50.790	50.000	1.6
4,4'-DDT	6.976	6.880	7.080	49.220	50.000	-1.6
Aldrin	5.230	5.132	5.332	47.470	50.000	-5.1
alpha-BHC	3.972	3.875	4.075	48.280	50.000	-3.4
alpha-Chlordane	5.982	5.886	6.086	47.850	50.000	-4.3
beta-BHC	4.488	4.390	4.590	47.360	50.000	-5.3
Decachlorobiphenyl	8.996	8.902	9.102	46.270	50.000	-7.5
delta-BHC	4.731	4.635	4.835	49.690	50.000	-0.6
Dieldrin	6.302	6.206	6.406	49.830	50.000	-0.3
Endosulfan I	6.030	5.934	6.134	50.060	50.000	0.1
Endosulfan II	6.741	6.646	6.846	46.840	50.000	-6.3
Endosulfan sulfate	7.105	7.007	7.207	51.640	50.000	3.3
Endrin	6.529	6.432	6.632	50.880	50.000	1.8
Endrin aldehyde	6.870	6.774	6.974	54.330	50.000	8.7
Endrin ketone	7.583	7.487	7.687	53.420	50.000	6.8
gamma-BHC (Lindane)	4.300	4.202	4.402	47.860	50.000	-4.3
gamma-Chlordane	5.901	5.806	6.006	48.850	50.000	-2.3
Heptachlor	4.891	4.793	4.993	45.860	50.000	-8.3
Heptachlor epoxide	5.648	5.552	5.752	50.160	50.000	0.3
Methoxychlor	7.449	7.352	7.552	46.640	50.000	-6.7
Tetrachloro-m-xylene	3.526	3.428	3.628	44.150	50.000	-11.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q3257
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/01/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097491.D **Time Analyzed:** 16:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.847	5.750	5.950	47.860	50.000	-4.3
4,4'-DDE	5.295	5.197	5.397	45.680	50.000	-8.6
4,4'-DDT	6.099	6.002	6.202	40.830	50.000	-18.3
Aldrin	4.289	4.191	4.391	46.770	50.000	-6.5
alpha-BHC	3.328	3.229	3.429	47.460	50.000	-5.1
alpha-Chlordane	5.106	5.009	5.209	47.100	50.000	-5.8
beta-BHC	3.955	3.857	4.057	44.020	50.000	-12.0
Decachlorobiphenyl	7.981	7.884	8.084	40.360	50.000	-19.3
delta-BHC	4.188	4.090	4.290	45.800	50.000	-8.4
Dieldrin	5.424	5.328	5.528	46.970	50.000	-6.1
Endosulfan I	5.161	5.064	5.264	42.690	50.000	-14.6
Endosulfan II	5.992	5.895	6.095	46.420	50.000	-7.2
Endosulfan sulfate	6.393	6.296	6.496	46.360	50.000	-7.3
Endrin	5.701	5.603	5.803	54.390	50.000	8.8
Endrin aldehyde	6.170	6.073	6.273	45.540	50.000	-8.9
Endrin ketone	6.897	6.800	7.000	53.200	50.000	6.4
gamma-BHC (Lindane)	3.659	3.561	3.761	46.530	50.000	-6.9
gamma-Chlordane	5.042	4.945	5.145	46.550	50.000	-6.9
Heptachlor	4.007	3.909	4.109	45.040	50.000	-9.9
Heptachlor epoxide	4.790	4.693	4.893	46.970	50.000	-6.1
Methoxychlor	6.671	6.574	6.774	40.990	50.000	-18.0
Tetrachloro-m-xylene	2.823	2.724	2.924	42.030	50.000	-15.9