

Instrument ID: MSVOA_W

Daily Analysis Runlog For Sequence/QC Batch ID # VW100125

Review By	Maresh Dadoda	Review On	10/3/2025 2:27:29 AM
Supervise By	Semsettin Yesilyurt	Supervise On	10/3/2025 2:31:29 AM
SubDirectory	VW100125	HP Acquire Method	MSVOA_W
		HP Processing Method	82w100125s.m

STD. NAME	STD REF.#
Tune/Reschk	VP135712
Initial Calibration Stds	VP135713,VP135714,VP135715,VP135716,VP135717,VP135718
CCC	
Internal Standard/PEM	VP133934
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VW032312.D	01 Oct 2025 08:12		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VW032313.D	01 Oct 2025 13:12	Pass for DOD	SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VW032314.D	01 Oct 2025 13:34	LR-20	SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VW032315.D	01 Oct 2025 14:15		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VW032316.D	01 Oct 2025 14:37		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VW032317.D	01 Oct 2025 14:59		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VW032318.D	01 Oct 2025 15:20		SY/MD	Ok,M
8	VIBLK	VIBLK	VW032319.D	01 Oct 2025 15:42		SY/MD	Ok
9	VSTDICV050	ICVVW100125	VW032320.D	01 Oct 2025 16:04		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_W

Daily Analysis Runlog For Sequence/QC Batch ID # VW100225

Review By	Maresh Dadoda	Review On	10/3/2025 4:06:22 PM
Supervise By		Supervise On	
SubDirectory	VW100225	HP Acquire Method	MSVOA_W HP Processing Method 82w100125s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135728
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135729,VP135730 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VW032321.D	02 Oct 2025 09:19		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VW032322.D	02 Oct 2025 10:53		SY/MD	Ok,M
3	VW1002SBL01	VW1002SBL01	VW032323.D	02 Oct 2025 11:20		SY/MD	Ok
4	VW1002SBS01	VW1002SBS01	VW032324.D	02 Oct 2025 11:48		SY/MD	Ok,M
5	VW1002SBSD01	VW1002SBSD01	VW032325.D	02 Oct 2025 12:10		SY/MD	Ok,M
6	Q3150-04RE	Mid Site Grid 5-7RE	VW032326.D	02 Oct 2025 12:51	vial-B ISTD Fail	SY/MD	Confirms
7	Q3257-01	ENV1-6FT	VW032327.D	02 Oct 2025 13:13	vial-A	SY/MD	Ok
8	Q3257-02	ENV2-6FT	VW032328.D	02 Oct 2025 13:35	vial-A	SY/MD	Ok
9	Q3266-01	ENV-3-6FT	VW032329.D	02 Oct 2025 13:57	vial-A	SY/MD	Ok
10	Q3266-02	ENV-4-6FT	VW032330.D	02 Oct 2025 14:19	vial-A	SY/MD	Ok
11	Q3262-02	MOO-25-0280	VW032331.D	02 Oct 2025 14:41	vial-A	SY/MD	Ok
12	Q3262-04	MOO-25-0282	VW032332.D	02 Oct 2025 15:03	vial-A ISTD Fail	SY/MD	ReRun
13	Q3262-07	279887	VW032333.D	02 Oct 2025 15:25	vial-A	SY/MD	Ok
14	Q3269-01	EG1-TP1	VW032334.D	02 Oct 2025 15:47	vial-A Surrogate Fail	SY/MD	ReRun
15	Q3269-04	EG1-TP2	VW032335.D	02 Oct 2025 16:09	vial-A	SY/MD	Ok
16	VIBLK	VIBLK	VW032336.D	02 Oct 2025 16:31		SY/MD	Ok
17	VSTDCCC050	VSTDCCC050EC	VW032337.D	02 Oct 2025 16:53		SY/MD	Ok,M

M : Manual Integration