

Surrogate Summary

SDG No.: **Q3257**

Client: **CDM Smith**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3257-01	ENV1-6FT	1,2-Dichloroethane-d4	50	67.0	134		63	155
		Dibromofluoromethane	50	54.8	110		70	134
		Toluene-d8	50	56.8	114		74	123
		4-Bromofluorobenzene	50	57.5	115		17	146
Q3257-02	ENV2-6FT	1,2-Dichloroethane-d4	50	71.2	142		63	155
		Dibromofluoromethane	50	57.0	114		70	134
		Toluene-d8	50	57.1	114		74	123
		4-Bromofluorobenzene	50	58.1	116		17	146
VW1002SBL01	VW1002SBL01	1,2-Dichloroethane-d4	50	59.9	120		63	155
		Dibromofluoromethane	50	52.7	105		70	134
		Toluene-d8	50	53.3	107		74	123
		4-Bromofluorobenzene	50	50.8	102		17	146
VW1002SBS01	VW1002SBS01	1,2-Dichloroethane-d4	50	52.2	104		63	155
		Dibromofluoromethane	50	50.1	100		70	134
		Toluene-d8	50	51.8	104		74	123
		4-Bromofluorobenzene	50	51.5	103		17	146
VW1002SBSD0	VW1002SBSD01	1,2-Dichloroethane-d4	50	49.3	99		63	155
		Dibromofluoromethane	50	49.7	99		70	134
		Toluene-d8	50	50.2	100		74	123
		4-Bromofluorobenzene	50	49.3	99		17	146