

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CAMP02
 Lab Code: ACE SDG NO.: Q3257
 Lab File ID: VW032322.D Date Analyzed: 10/02/2025
 Instrument ID: MSVOA_W Time Analyzed: 10:53
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	241629	13.556				
UPPER LIMIT	483258	14.056				
LOWER LIMIT	120815	13.056				
EPA SAMPLE NO.						
ENV1-6FT	191597	13.56				
ENV2-6FT	232780	13.56				
VW1002SBL01	204423	13.56				
VW1002SBS01	242228	13.56				
VW1002SBSD01	239286	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.