



284 Sheffield Street, Mountainside, NJ 7092, Phone: 908 789 8900, Fax: 908 789 8922

CASE NARRATIVE

Dal-Tile - Dickson Plant

Project Name: Waste Water

Project # N/A

Order ID # Q3258

Test Name: Mercury,Metal ICP-Group1Ammonia,BOD5,Cyanide,Hexavalent Chromium,Oil and Grease,Phosphorus-Total,TSS

A. Number of Samples and Date of Receipt:

6 Water samples were received on 10/01/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Mercury,Metal ICP-Group1Ammonia,BOD5,Cyanide,Hexavalent Chromium,Oil and Grease,Phosphorus-Total,TSS. This data package contains results for Mercury(7470A), Metal ICP-Group1(6010D)Ammonia(SM4500-NH3),BOD5(SM5210 B), Cyanide(SM4500-CN C,E),Hexavalent Chromium(7196A),Oil and Grease(1664A), Phosphorus-Total(365.3),TSS(SM2540 D).

C. Analytical Techniques:

Mercury,Metals ICP-Group1 : The analysis of Metals ICP-Group1 was based on method 6010D, digestion based on method 3010 (waters). The analysis and digestion of Mercury was based on method 7470A.

Wetchem : The analysis of Oil and Grease was based on method 1664A, The analysis of Phosphorus-Total was based on method 365.3, The analysis of Hexavalent Chromium was based on method 7196A, The analysis of TSS was based on method SM2540 D, The analysis of Cyanide was based on method SM4500-CN C,E, The analysis of Ammonia was based on method SM4500-NH3 and The analysis of BOD5 was based on method SM5210 B.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The MS recoveries met the requirements for all compounds.

The MSD recoveries met the acceptable requirements.

The RPD recoveries met criteria.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.



284 Sheffield Street, Mountainside, NJ 7092, Phone: 908 789 8900, Fax: 908 789 8922

The Duplicate analysis met criteria for all compounds.
The Serial Dilution met the acceptable requirements.

E. Additional Comments:

Ammonia, BOD5, Cyanide, Hexavalent Chromium, Oil and Grease, Phosphorus-Total, TSS :
As per method, aqueous sample for Hexavalent Chromium analysis should be filtered within 15 minutes of collection time. However, samples were not filtered as per requirement therefore Lab has filtered the samples in-house and analyzed within 24 hours from collection.

Due to limited volume DUP, MS and MSD were not performed for Hexavalent Chromium.

As per method 1664A, MS/MSD is required to be performed with the sample analysis. However, Lab did not receive sufficient volume to perform the MS/MSD for Q3258 therefore Lab reported MS-MSD from Q3229

Due to low volume MS and MSD were not performed for PB170085 of Cyanide.

This package has been revised because Sample#08 and Sample#09 added for Cyanide.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____