

Initial and Continuing Calibration Verification

Client: Dal-Tile - Dickson Plant

SDG No.: Q3258

Project: Waste Water

RunNo.: LB137386

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.502	0.5	100	90-110	10/01/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.497	0.5	99	90-110	10/01/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.504	0.5	101	90-110	10/01/2025

Initial and Continuing Calibration Verification

Client: Dal-Tile - Dickson Plant

SDG No.: Q3258

Project: Waste Water

RunNo.: LB137417

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.98	1	98	90-110	10/03/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.97	1	97	90-110	10/03/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.97	1	97	90-110	10/03/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	10/03/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.98	1	98	90-110	10/03/2025

Initial and Continuing Calibration Verification

Client: Dal-Tile - Dickson Plant

SDG No.: Q3258

Project: Waste Water

RunNo.: LB137435

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.095	0.099	96	85-115	10/06/2025
Sample ID: CCV1 Cyanide	mg/L	0.24	0.25	96	90-110	10/06/2025
Sample ID: CCV2 Cyanide	mg/L	0.24	0.25	96	90-110	10/06/2025

Initial and Continuing Calibration Verification

Client: Dal-Tile - Dickson Plant

SDG No.: Q3258

Project: Waste Water

RunNo.: LB137481

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.489	0.50	98	90-110	10/09/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.506	0.50	101	90-110	10/09/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.502	0.50	100	90-110	10/09/2025

Initial and Continuing Calibration Verification

Client: Dal-Tile - Dickson Plant

SDG No.: Q3258

Project: Waste Water

RunNo.: LB137517

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.096	0.099	97	85-115	10/14/2025
Sample ID: CCV1 Cyanide	mg/L	0.24	0.25	96	90-110	10/14/2025
Sample ID: CCV2 Cyanide	mg/L	0.24	0.25	96	90-110	10/14/2025