

Laboratory Control Sample Summary

Client: Dal-Tile - Dickson Plant

SDG No.: Q3258

Project: Waste Water

Run No.: LB137386

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137386BS							
Hexavalent Chromium	mg/L	0.5	0.51		103	1	90-111	10/01/2025

Laboratory Control Sample Summary

Client: Dal-Tile - Dickson Plant

SDG No.: Q3258

Project: Waste Water

Run No.: LB137388

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137388BS							
BOD5	mg/L	198	181		92	1	84.6-115.4	10/01/2025

Laboratory Control Sample Summary

Client: Dal-Tile - Dickson Plant

SDG No.: Q3258

Project: Waste Water

Run No.: LB137413

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137413BS							
Oil and Grease	mg/L	20.0	16.8		84	1	78-114	10/03/2025

Laboratory Control Sample Summary

Client: Dal-Tile - Dickson Plant

SDG No.: Q3258

Project: Waste Water

Run No.: LB137419

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137419BS							
TSS	mg/L	550	532		97	1	90-110	10/03/2025

Laboratory Control Sample Summary

Client: Dal-Tile - Dickson Plant

SDG No.: Q3258

Project: Waste Water

Run No.: LB137417

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169946BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	10/03/2025

Laboratory Control Sample Summary

Client: Dal-Tile - Dickson Plant

SDG No.: Q3258

Project: Waste Water

Run No.: LB137435

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169962BS							
Cyanide	mg/L	0.1	0.096		96	1	85-115	10/06/2025

Laboratory Control Sample Summary

Client: Dal-Tile - Dickson Plant

SDG No.: Q3258

Project: Waste Water

Run No.: LB137481

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170046BS							
Phosphorus, Total	mg/L	0.50	0.48		97	1	90-110	10/09/2025

Laboratory Control Sample Summary

Client: Dal-Tile - Dickson Plant

SDG No.: Q3258

Project: Waste Water

Run No.: LB137517

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170085BS							
Cyanide	mg/L	0.1	0.097		97	1	85-115	10/14/2025