



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Dal-Tile - Dickson Plant

SDG No.: Q3258

Contract: DALT01

Lab Code: ACE

Initial Calibration Source:

Continuing Calibration Source:

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.18	0.2	90	70 - 130	CV	10/03/2025	11:21	LB137412
CRI01	Arsenic	21.4	20.0	107	65 - 135	P	10/03/2025	13:16	LB137426
	Cadmium	6.04	6.0	101	65 - 135	P	10/03/2025	13:16	LB137426
	Chromium	9.90	10.0	99	65 - 135	P	10/03/2025	13:16	LB137426
	Copper	21.2	20.0	106	65 - 135	P	10/03/2025	13:16	LB137426
	Lead	10.2	12.0	85	65 - 135	P	10/03/2025	13:16	LB137426
	Nickel	40.9	40.0	102	65 - 135	P	10/03/2025	13:16	LB137426
	Selenium	21.2	20.0	106	65 - 135	P	10/03/2025	13:16	LB137426
	Zinc	42.0	40.0	105	65 - 135	P	10/03/2025	13:16	LB137426