



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal-Tile - Dickson Plant

SDG No.: Q3258

Contract: DALT01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB58	Mercury	0.076	+/-0.2	U	0.20	CV	10/03/2025	11:14	LB137412

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal-Tile - Dickson Plant

SDG No.: Q3258

Contract: DALT01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB80	Mercury	0.076	+/-0.2	U	0.20	CV	10/03/2025	11:18	LB137412
CCB81	Mercury	0.076	+/-0.2	U	0.20	CV	10/03/2025	11:53	LB137412
CCB82	Mercury	0.076	+/-0.2	U	0.20	CV	10/03/2025	12:40	LB137412

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal-Tile - Dickson Plant

SDG No.: Q3258

Contract: DALT01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	10/03/2025	13:12	LB137426
	Cadmium	0.50	+/-3	U	6.00	P	10/03/2025	13:12	LB137426
	Chromium	2.12	+/-5	U	10.0	P	10/03/2025	13:12	LB137426
	Copper	4.60	+/-10	U	20.0	P	10/03/2025	13:12	LB137426
	Lead	2.30	+/-6	U	12.0	P	10/03/2025	13:12	LB137426
	Nickel	3.06	+/-20	U	40.0	P	10/03/2025	13:12	LB137426
	Selenium	9.64	+/-10	U	20.0	P	10/03/2025	13:12	LB137426
	Zinc	16.7	+/-20	U	40.0	P	10/03/2025	13:12	LB137426

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Dal-Tile - Dickson Plant

SDG No.: Q3258

Contract: DALT01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	10/03/2025	13:41	LB137426
	Cadmium	0.50	+/-3	U	6.00	P	10/03/2025	13:41	LB137426
	Chromium	2.12	+/-5	U	10.0	P	10/03/2025	13:41	LB137426
	Copper	4.60	+/-10	U	20.0	P	10/03/2025	13:41	LB137426
	Lead	2.30	+/-6	U	12.0	P	10/03/2025	13:41	LB137426
	Nickel	3.06	+/-20	U	40.0	P	10/03/2025	13:41	LB137426
	Selenium	9.64	+/-10	U	20.0	P	10/03/2025	13:41	LB137426
	Zinc	16.7	+/-20	U	40.0	P	10/03/2025	13:41	LB137426
CCB02	Arsenic	5.12	+/-10	U	20.0	P	10/03/2025	15:04	LB137426
	Cadmium	0.50	+/-3	U	6.00	P	10/03/2025	15:04	LB137426
	Chromium	2.12	+/-5	U	10.0	P	10/03/2025	15:04	LB137426
	Copper	4.60	+/-10	U	20.0	P	10/03/2025	15:04	LB137426
	Lead	2.30	+/-6	U	12.0	P	10/03/2025	15:04	LB137426
	Nickel	3.06	+/-20	U	40.0	P	10/03/2025	15:04	LB137426
	Selenium	9.64	+/-10	U	20.0	P	10/03/2025	15:04	LB137426
	Zinc	16.7	+/-20	U	40.0	P	10/03/2025	15:04	LB137426
CCB03	Arsenic	5.12	+/-10	U	20.0	P	10/03/2025	16:00	LB137426
	Cadmium	0.50	+/-3	U	6.00	P	10/03/2025	16:00	LB137426
	Chromium	2.12	+/-5	U	10.0	P	10/03/2025	16:00	LB137426
	Copper	4.60	+/-10	U	20.0	P	10/03/2025	16:00	LB137426
	Lead	2.30	+/-6	U	12.0	P	10/03/2025	16:00	LB137426
	Nickel	3.06	+/-20	U	40.0	P	10/03/2025	16:00	LB137426
	Selenium	9.64	+/-10	U	20.0	P	10/03/2025	16:00	LB137426
	Zinc	16.7	+/-20	U	40.0	P	10/03/2025	16:00	LB137426
CCB04	Arsenic	5.12	+/-10	U	20.0	P	10/03/2025	16:49	LB137426
	Cadmium	0.50	+/-3	U	6.00	P	10/03/2025	16:49	LB137426
	Chromium	2.12	+/-5	U	10.0	P	10/03/2025	16:49	LB137426
	Copper	4.60	+/-10	U	20.0	P	10/03/2025	16:49	LB137426
	Lead	2.30	+/-6	U	12.0	P	10/03/2025	16:49	LB137426
	Nickel	3.06	+/-20	U	40.0	P	10/03/2025	16:49	LB137426
	Selenium	9.64	+/-10	U	20.0	P	10/03/2025	16:49	LB137426
	Zinc	16.7	+/-20	U	40.0	P	10/03/2025	16:49	LB137426