

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client:	<u>Dal-Tile - Dickson Plant</u>	SDG No.:	<u>Q3258</u>
Contract:	<u>DALT01</u>	Lab Code:	<u>ACE</u>
ICS Source:	<u>EPA</u>	Instrument ID:	<u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	8.41			-20	20	10/03/2025	13:20	LB137426
	Cadmium	0.91	1.0	91	-5	7	10/03/2025	13:20	LB137426
	Chromium	46.6	52.0	90	42	62	10/03/2025	13:20	LB137426
	Copper	3.06	2.0	153	-18	22	10/03/2025	13:20	LB137426
	Lead	0.97			-12	12	10/03/2025	13:20	LB137426
	Nickel	3.93	2.0	196	-38	42	10/03/2025	13:20	LB137426
	Selenium	9.27			-20	20	10/03/2025	13:20	LB137426
	Zinc	1.66			-40	40	10/03/2025	13:20	LB137426
ICSAB01	Arsenic	110	104	106	88.4	120	10/03/2025	13:25	LB137426
	Cadmium	980	972	101	826	1120	10/03/2025	13:25	LB137426
	Chromium	553	542	102	460	624	10/03/2025	13:25	LB137426
	Copper	484	511	95	434	588	10/03/2025	13:25	LB137426
	Lead	49.1	49.0	100	37	61	10/03/2025	13:25	LB137426
	Nickel	981	954	103	810	1100	10/03/2025	13:25	LB137426
	Selenium	56.4	46.0	123	26	66	10/03/2025	13:25	LB137426
	Zinc	1030	952	108	809	1095	10/03/2025	13:25	LB137426