

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Dal-Tile - Dickson Plant</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3258</u>
Contract:	<u>DALT01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3252-05</u>	Client ID:	<u>AUD-25-0161DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3252-05DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.20	U	0.20	U			CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Dal-Tile - Dickson Plant</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3258</u>
Contract:	<u>DALT01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3252-05MS</u>	Client ID:	<u>AUD-25-0161MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3252-05MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	3.66		3.85		5		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Dal-Tile - Dickson Plant</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3258</u>
Contract:	<u>DALT01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3263-02</u>	Client ID:	<u>266380DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3263-02DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	10.0	U	10.0	U			P
Cadmium	ug/L	20	3.00	U	3.00	U			P
Chromium	ug/L	20	5.00	U	5.00	U			P
Copper	ug/L	20	10.0	U	10.0	U			P
Lead	ug/L	20	6.00	U	6.00	U			P
Nickel	ug/L	20	6.71	J	7.09	J	6		P
Selenium	ug/L	20	10.0	U	10.0	U			P
Zinc	ug/L	20	75.6		80.4		6		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Dal-Tile - Dickson Plant</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3258</u>
Contract:	<u>DALT01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3263-02MS</u>	Client ID:	<u>266380MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3263-02MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	382		381		0		P
Cadmium	ug/L	20	88.3		88.0		0		P
Chromium	ug/L	20	188		185		2		P
Copper	ug/L	20	141		140		1		P
Lead	ug/L	20	430		430		0		P
Nickel	ug/L	20	233		232		0		P
Selenium	ug/L	20	961		956		1		P
Zinc	ug/L	20	173		173		0		P