

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Dal-Tile - Dickson Plant

SDG No.: Q3258

Contract: DALT01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169958BS							
Arsenic	ug/L	400	384		96	80 - 120	P
Cadmium	ug/L	100	94.8		95	80 - 120	P
Chromium	ug/L	200	208		104	80 - 120	P
Copper	ug/L	150	150		100	80 - 120	P
Lead	ug/L	500	466		93	80 - 120	P
Nickel	ug/L	250	247		99	80 - 120	P
Selenium	ug/L	1000	971		97	80 - 120	P
Zinc	ug/L	100	99.5		100	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Dal-Tile - Dickson Plant

SDG No.: Q3258

Contract: DALT01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169961BS Mercury	ug/L	4.0	4.10		102	80 - 120	CV