

Duplicate Sample Summary

Client:	ATG - AKRON LAB	SDG No.:	Q3259
Project:	PO 25061955	Sample ID:	Q3259-01
Client ID:	25092166-001DUP	Percent Solids for Spike Sample:	82.1

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/Kg	+/-20	0.051	U	0.051	U	1	0		10/02/2025

Duplicate Sample Summary

Client: ATG - AKRON LAB	SDG No.: Q3259
Project: PO 25061955	Sample ID: Q3259-01
Client ID: 25092166-001MSD	Percent Solids for Spike Sample: 82.1

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/Kg	+/-20	2.20		2.20		1	0		10/02/2025

Duplicate Sample Summary

Client:	ATG - AKRON LAB	SDG No.:	Q3259
Project:	PO 25061955	Sample ID:	Q3259-21
Client ID:	25092166-023DUP	Percent Solids for Spike Sample:	84.3

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/Kg	+/-20	0.11	J	0.10	J	1	10		10/02/2025

Duplicate Sample Summary

Client:	ATG - AKRON LAB	SDG No.:	Q3259
Project:	PO 25061955	Sample ID:	Q3259-21
Client ID:	25092166-023MSD	Percent Solids for Spike Sample:	84.3

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/Kg	+/-20	2.20		2.20		1	0		10/02/2025

Duplicate Sample Summary

Client:	ATG - AKRON LAB	SDG No.:	Q3259
Project:	PO 25061955	Sample ID:	Q3259-41
Client ID:	25092166-049DUP	Percent Solids for Spike Sample:	86.1

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/Kg	+/-20	0.091	J	0.088	J	1	3		10/02/2025

Duplicate Sample Summary

Client:	ATG - AKRON LAB	SDG No.:	Q3259
Project:	PO 25061955	Sample ID:	Q3259-41
Client ID:	25092166-049MSD	Percent Solids for Spike Sample:	86.1

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/Kg	+/-20	2.20		2.20		1	0		10/02/2025