

Laboratory Control Sample Summary

Client: ATG - AKRON LAB

SDG No.: Q3259

Project: PO 25061955

Run No.: LB137396

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169942BS							
Cyanide	mg/Kg	5	4.80		96	1	85-115	10/02/2025

Laboratory Control Sample Summary

Client: ATG - AKRON LAB

SDG No.: Q3259

Project: PO 25061955

Run No.: LB137396

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169943BS							
Cyanide	mg/Kg	5	4.80		96	1	85-115	10/02/2025

Laboratory Control Sample Summary

Client: ATG - AKRON LAB

SDG No.: Q3259

Project: PO 25061955

Run No.: LB137405

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169944BS							
Cyanide	mg/Kg	5	4.80		96	1	85-115	10/02/2025