



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Aramark Uniforms

SDG No.: Q3260

Contract: ARAM01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	5.73	6.0	96	65 - 135	P	10/08/2025	12:41	LB137466
	Copper	21.6	20.0	108	65 - 135	P	10/08/2025	12:41	LB137466
	Lead	12.2	12.0	102	65 - 135	P	10/08/2025	12:41	LB137466
	Nickel	39.3	40.0	98	65 - 135	P	10/08/2025	12:41	LB137466
	Zinc	43.1	40.0	108	65 - 135	P	10/08/2025	12:41	LB137466
CRA	Mercury	0.041	0.05	82	60 - 140	CV	10/15/2025	13:03	LB137542