

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Aramark Uniforms

SDG No.: Q3260

Contract: ARAM01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169959BS							
Cadmium	ug/L	100	92.2		92	85 - 115	P
Copper	ug/L	150	149		99	85 - 115	P
Lead	ug/L	500	459		92	85 - 115	P
Nickel	ug/L	250	235		94	85 - 115	P
Zinc	ug/L	100	100		100	85 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Aramark Uniforms

SDG No.: Q3260

Contract: ARAM01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170118BS Mercury	ug/L	4.0	3.89		97	85 - 115	CV