

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q3261

Project: Linden Generating Station - 402

Run No.: LB137414

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137414BS							
TPH	mg/L	20.0	16.7		84	1	78-114	10/03/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q3261

Project: Linden Generating Station - 402

Run No.: LB137414

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137414BSD							
TPH	mg/L	20.0	17.0		85	1	78-114	10/03/2025